



Requirements for Xenon International: Revision 2

August 2018

James C. Hayes
Mathew A. Covert
Michael P. Foxe
Warren W. Harper
Charles W. Hubbard
Paul H. Humble
Scott J. Morris
Amanda M. Prinke
Timothy L. Stewart

Matt W. Cooper
James H. Ely
Derek A. Haas
Tom R. Heimbigner
Daniel T. Keller
Jill C. Madison
Mark E. Panisko
Michael D. Ripplinger

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor Battelle Memorial Institute, nor any of their employees, makes **any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights.** Reference herein to any specific commercial product, process or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or Battelle Memorial Institute. The views and opinions of the authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

PACIFIC NORTHWEST NATIONAL LABORATORY
operated by
BATTELLE
for the
UNITED STATES DEPARTMENT OF ENERGY
Under Contract DE-AC05-76RL01830

Printed in the United States of America

Online Access: U.S. Department of Energy (DOE) reports produced after 1991 and a growing number of pre-1991 documents are available free via DOE's SciTech Connect (<http://www.osti.gov/scitech>)

Reports not in digital format may be purchased by the public from the National Technical Information Service (NTIS):

U.S. Department of Commerce
National Technical Information Service
5301 Shawnee Rd
Alexandria, VA 22312
www.ntis.gov

Phone: (800) 553-NTIS (6847) or (703) 605-6000
Fax: (703) 605-6900
Email: orders@ntis.gov

Reports not in digital format are available to DOE and DOE contractors from the Office of Scientific and Technical Information (OSTI):

U.S. Department of Energy
Office of Scientific and Technical Information
P.O. Box 62
Oak Ridge, TN 37831-0062
www.osti.gov
Phone: (865) 576-8401
Fax: (865) 576-5728
Email: reports@osti.gov

This document has been assigned the OSTI Identifier 1095446



This document was printed on recycled paper.
(8/2010)

Contents

Acronyms and Abbreviations	v
1.0 Document Overview	1
1.1 Mission Scenario and Needs Statement	1
1.2 Operational Environment	1
2.0 Capabilities Required	2
2.1 Requirements Definition	2
2.1.1 Threshold.....	2
2.1.2 Objective	2
2.1.3 Maintenance Concept.....	2
2.2 Key Performance Parameters (KPP) and Key System Attributes (KSA)	3
2.3 Operational Modes	4
3.0 System Requirements	6
F 1 Gas Processing	6
F 1.1 Collection	6
F 1.2 Processing.....	6
F 1.3 Quantification.....	7
F 2 Nuclear Detection.....	7
F 2.2 Count	7
F 2.3 Nuclear Detector Calibration.....	8
F 2.4 Nuclear Detector Quality Control (QC)	9
F 3 Process Control (PNNL-developed GFE software)	10
F 3.1 Software Control	10
F 3.2 Operational Modes	10
F 4 Data	11
F 4.1 Data Collection and Format	11
F 4.2 Data Support Tools.....	12
F 5 Archive System	12
F 6 Design Constraints	13
F 6.1 Size and Weight	13
F 6.2 Environmental	13
F 6.3 Electrical.....	14
F 6.4 Materials.....	15
F 6.5 System Expendables.....	15
F 6.6 Security	15
F 6.7 Safety.....	15
F 6.8 Shutdown.....	16

F 6.9 Usability	16
F 7 Documentation	16
F 8 Sustainment	17
F 8.1 Availability	17
F 8.2 Maintainability	17
F 8.3 Mean Time between Critical Failures (MTBCF)	17
F 8.4 Mean Time between Failures (MTBF)	17
F 8.5 Mean Time to Repair (MTTR)	18
F 8.6 Maximum Corrective Maintenance Time (C_{MxtMX})	18
F 8.7 Support Equipment	18
F 8.8 Transportation and Facilities	18
Appendix A	19
Appendix B	20

Tables

Table 2-1 Summary of Key Operations Requirements	3
Table 2-2 Key System Attributes	3
Table 2-3 Operational Modes	5

Acronyms and Abbreviations

$C_{MXt_{MX}}$	maximum corrective maintenance time
Co	cobalt
CO ₂	carbon dioxide
Cs	cesium
FMC	fully mission capable
GFE	government-furnished equipment
H ₂ O	water
INGE	International Noble Gas Experiment
MDC	minimal detectable concentrations
MTBCF	mean time between critical failures
MTBF	mean time between failures
MTTR	mean time to repair
N ₂ O	nitrous oxide
NDC	National Data Center
Pb	lead
PMC	partially mission capable
PNNL	Pacific Northwest National Laboratory
QC	quality control
Rn	radon
ROI	region of interest (the region within a beta-gamma spectrum that contains the sample counts used in calculating the activity)
SOH	state of health
STP	standard temperature and pressure – a temperature of 273.15 K (0 °C, 32 °F) and an absolute pressure of 100 kPa (14.504 psi, 0.986 atm, 1 bar).
UPS	uninterruptable power supply
Xe	xenon

1.0 Document Overview

This document defines the requirements for the Xenon International radioxenon system to measure radioactive xenon concentrations in the atmosphere. This document contains requirements for the Xenon International system for development and for production. Each requirement will be verified during test and evaluation of the system, and lessons learned from testing will be incorporated into the design of the production system. **(Updated 2018)**

The Xenon International system has a lifetime goal of 10-15 years. This takes into account that some of the parts will need to be maintained and replaced during that time. This is a design and planning goal rather than a manufacturing requirement, as years of field experience are not yet available to verify the expected lifetime of the system.

NOTE TO READER: Requirements in **Bold** are the Key Performance Requirements and Key System Attributes. Several requirement have changed since version 1 of this document; change dates are listed after the changed paragraph or requirement highlighted in **Bold**.

This document supersedes the previous two versions of the Xenon International Requirements document. The changes are shown in the change log in Appendix A.

1.1 Mission Scenario and Needs Statement

The Xenon International system will be installed in a fixed location and will operate 24 hours per day, 7 days per week to measure radioactive xenon concentrations in the atmosphere. The hardware and software will be sharable for their intended use of treaty monitoring.

Detection of xenon isotopes is a proven and important method for distinguishing nuclear explosions from earthquakes, and is particularly well suited to detecting undeclared underground testing. The radioxenon isotopes ^{131m}Xe , ^{133}Xe , ^{133m}Xe , and ^{135}Xe are of particularly high value in identifying nuclear explosions and thus are the focus of current noble gas detection systems. The Xenon International system will incorporate the important lessons learned during the International Noble Gas Experiment (INGE). Redesigned subsystems will increase sensitivity to minor xenon isotopes, increase sampling frequency, and improve system reliability and performance, all of which are the main drivers of this new system. The presence of an improved xenon monitoring system will increase ability to detect and distinguish nuclear explosion activity.

1.2 Operational Environment

The system will be installed and operated indoors in a building or facility that will have access to outdoor air for sampling. The system will obtain electrical power from the building. It is assumed the system will be operated in a lockable room that will allow easy access for operation and maintenance. Storage shall be available for the consumables, tools, and spare parts deemed necessary to operate and maintain the system. There will be no radioactive sources that can interfere with the system measurement capability and that can lead to a false spectrum interpretation. If sources are stored near the system, they shall be properly shielded.

The indoor (ambient) temperature of the room housing the system shall remain within the operational temperature range (+50°F and +95°F) of the system components with humidity between 50% and 80% (non-condensing). **(Updated 2018)**

This document contains the requirements necessary to manufacture a Xenon International system. Requirements that pertain exclusively to external systems such as uninterruptable power supply (UPS), or authentication hardware are outside the scope of this document. Similarly, calibration or certification requirements are outside the scope of the manufacturing requirements.

1.3 Deleted (Updated 2018)

2.0 Capabilities Required

2.1 Requirements Definition

This section defines terms used within some of the system requirements found in Section 3.

2.1.1 Threshold

A threshold is a minimum acceptable capability or value for a system characteristic that is necessary to provide the operational capability that satisfies the mission need. Unless otherwise noted, all requirements in this document are thresholds.

2.1.2 Objective

An objective is a capability or value beyond the threshold that could have a measurable and beneficial impact on the system performance, supportability, or operational concept of employment. It is that value or capability desired by the user, which the system is ultimately attempting to obtain. An objective is a cost-effective increment above the threshold for each program capability, and represents a desire for performance.

2.1.3 Maintenance Concept

The Xenon International maintenance concept consists of the following levels of maintenance:

- on-site maintenance - consists of those maintenance tasks normally performed on-site.
- depot maintenance - consists of those tasks performed using highly specialized skills, sophisticated shop equipment, or special facilities of a supporting directorate; commercial activity (sending part[s] or component[s] to the supplier or manufacturer); or interservice agency, at specialized repair centers, centralized repair facilities, or in some cases, at an operating location.

2.2 Key Performance Parameters (KPP) and Key System Attributes (KSA)

Xenon International will have increased capability (increased sensitivity to minor xenon isotopes, increased sampling frequency, and improved system reliability and performance), compared to currently fielded systems. Table 2-1 presents a summary of some of the key requirements of the system that will be detailed further in the Section 3. Table 2-2 details the Key System Attributes. **(Updated 2018)**

Table 2-1 Summary of Key Performance Parameters

Characteristics	Threshold KPP	Objective KPP	Requirement ID
Collection Time	The system shall collect xenon samples every 8 hrs. ± 0.10 hrs.	The system shall collect xenon samples every 6 hrs. ± 0.1 hrs.	R 1.1.3
Process and Analyze	The system shall complete each individual xenon samples analysis within 24 hours ± 0.1 hrs. of the start of sample collection ($c+p+a \leq 24$ hours) collection=c, processing=p, analysis=a	Threshold = Objective	R 2.2.9
Processing Efficiency	The system shall have a xenon collection efficiency of $\geq 65\%$ and vary no more than 10% during a sample collection	The system shall have a xenon collection efficiency of $\geq 75\%$ and vary no more than 10% during a sample collection	R 1.1.7 R 1.1.8 (Updated 2018)
Isotopes Determined	The system shall measure four isotopes of Xe: Xe-131m, Xe-133m, Xe-133, Xe-135	Threshold = Objective	R 1.2.5
Communication	The system shall meet DoD Net Ready requirements and have 2-way data connectivity	Threshold = Objective	R 3.1.1

Table 2-2 Key System Attributes

Characteristics	Threshold KSA	Objective KSA	Requirement ID
Data Transmittal	≤ 1.0 hrs. post analysis	Threshold = Objective	R 4.1.10
Minimum Detectable Concentration	0.3 mBq/m ³ for Xe-131m, Xe-133m, Xe-133 1.0 mBq/m ³ for Xe-135 based on defined detector background	0.15 mBq/m ³ for Xe-131m, Xe-133m, Xe-133 0.5 mBq/m ³ for Xe-135 based on defined detector background	R 2.2.2

Characteristics	Threshold KSA	Objective KSA	Requirement ID
Minimum Quantifiable Concentration	1.0 mBq/m ³ ± 20% RSD for Xe131m, Xe133m, Xe133 3.0 mBq/m ³ ± 20% RSD for Xe135 based on defined detector background	0.5 mBq/m ³ ± 10% RSD for Xe131m, Xe133m, Xe133 1.5 mBq/m ³ ± 10% RSD for Xe135 based on defined detector background	R 2.2.2
Data Availability	≥ 90%	≥95 %	R 4.1.1
State of Health	Daily transmittal of minimum data set with retention of full set for 400 days. Data call up available	Threshold = Objective	R 4.1.1.4 R 4.1.1.8
Size	≤ 45 in. x 32 in. x 80 in.	≤ 45 in. x 30 in. x 76 in.	R 6.1.1
Weight	≤ 3000 lbs.	≤ 2500 lbs.	R 6.1.2
Power	Single phase 110-140 V 50/60 Hz	Threshold = Objective	R 6.3.1
UPS Operational Time	The UPS shall provide sufficient power to permit the system to remain operational during short outages (≤ 10 minutes) and have sufficient power to permit the safe shutdown of the unit	The UPS shall provide sufficient power to permit the system to remain operational during short outages (≤ 15 minutes) and have sufficient power to permit the safe shutdown of the unit (Updated 2018)	R 6.3.6
Archive Capability	The Xenon International system shall provide an interface for an archiving system	The Xenon International system shall have a built-in archive system	R 5.1.1

2.3 Operational Modes

The system shall allow for the modes of operation detailed in Table 2-3.

Table 2-3 Operational Modes

Mode	Required Functionality
Fully mission capable (FMC)	Full function operation
Partially mission capable (PMC)	Less than full capability, but able to produce results at least intermittently (for example, operating with a failed sensor or nuclear detector)
Regeneration	The system must be fully regenerated and ready to start collection within 8 hours.
Shutdown mode	Emergency and controlled shutdowns shall be operational when necessary
Maintenance	Maintenance is performed on the system
Failure	The system is unable to collect or record data

3.0 System Requirements

The Xenon International system requirements are as follows. Functions are indicated with F, and requirements are indicated by R. Note: requirements in **bold** are the Key Operations Requirements listed in Table 2.1.

F 1 Gas Processing

F 1.1 Collection

- R 1.1.1 The system shall be an automated, unattended system that continuously collects whole air from the atmosphere 24 hours a day, 7 days a week.
- R 1.1.2 The system shall automatically collect, process, transfer, and analyze xenon gas without human intervention.
- R 1.1.3 The system shall support continuous collection of xenon from the atmosphere with cycle times of 8 hours (threshold), 6 hours (objective).**
- R 1.1.4 The system shall dry the air to a dew point of at least -85 °C standard temperature and pressure (STP).
- R 1.1.5 The system shall contain a certified precision flow meter to monitor the stability of the air flow. The precision of the measured flow rate shall be equal or better than 10%. The certification requirement may be satisfied through calibration at the manufacturer and verified with the certification documentation that comes with the sensor.
- R 1.1.6 The system shall remove water vapor (H₂O), carbon dioxide (CO₂) and nitrous oxide (N₂O) to concentration levels below 1.0 ppm each as measured between the driers and the collection trap.
- R 1.1.7 The system shall have a collection efficiency of greater than or equal to 65% (threshold), 75% (objective).
- R 1.1.8 The system shall have a stable xenon collection efficiency that varies no more than 10% during a sampling cycle.
- R 1.1.9 The sampling cycles for the Xenon International system shall be of the same duration within $\pm 10\%$.

F 1.2 Processing

- R 1.2.1 The system shall perform on-site gas processing of xenon from the whole air collection.
- R 1.2.2 Removed (incorporated into R 1.1.6).
- R 1.2.3 The system shall use a nitrogen generator. **(Updated 2018)**
- R 1.2.4 The system shall remove radon (Rn) greater than a factor of 10^7 from the processing stream as Rn causes significant interference with the nuclear measurement of the target xenon isotopes.

R 1.2.5 The system shall separate xenon gas from the whole air and measure the following isotopes: ^{133}Xe , $^{131\text{m}}\text{Xe}$, $^{133\text{m}}\text{Xe}$, and ^{135}Xe .

R 1.2.6 The system shall deliver a minimum total volume of 1 cc of stable xenon at STP conditions over a period of 8 hours in the measuring cell.

R 1.2.7 The cross-contamination between samples shall be less than 1% (i.e., less than 1% of a sample shall be carried over to a subsequent sample). This requirement does not address the memory effect in the measurement cells. The memory effect is addressed in section F 2.2 Nuclear Detection Count.

F 1.3 Quantification

R 1.3.1 The system shall include a gas analysis system that is capable of determining the volume of xenon gas (volume at STP) in the measurement cell with an uncertainty of 10 % or better.

R 1.3.2 The calculations for the Xe volume uncertainty measurement shall be documented.

R 1.3.3 The system shall report a measurement value at the end of acquisition and within 48 hours after the start of collection.

F 2 Nuclear Detection

R 2.1.1 The Xenon International system shall use a beta-gamma coincidence detector.

F 2.2 Count

R 2.2.1 The system shall carry out a gas background measurement prior to each sample cycle measurement.

R 2.2.2 The system shall be capable of performing measurements of radioxenon activity concentrations in a minimum background environment of 10 $\mu\text{R/hr}$ due to ambient room radiation (such as radon and ^{40}K) with a minimal detectable concentration (MDC) calculated per standard methodology shown in the appendix:

- 0.3 mBq/m^3 (threshold), 0.15 mBq/m^3 (objective) for ^{133}Xe , $^{133\text{m}}\text{Xe}$, and $^{131\text{m}}\text{Xe}$
- 1.0 mBq/m^3 (threshold), 0.5 mBq/m^3 (objective) for ^{135}Xe

The system shall demonstrate a minimum quantifiable concentration (MQC) calculated per standard methodology:

- 1.0 $\text{mBq/m}^3 \pm 20\% \text{ RSD}$ (threshold), 0.5 $\text{mBq/m}^3 \pm 10\% \text{ RSD}$ (objective) for ^{133}Xe , $^{133\text{m}}\text{Xe}$, and $^{131\text{m}}\text{Xe}$
- 3.0 $\text{mBq/m}^3 \pm 20\% \text{ RSD}$ (threshold), 1.5 $\text{mBq/m}^3 \pm 10\% \text{ RSD}$ (objective) for ^{135}Xe (Updated 2018)

R 2.2.3 The memory effect on the system shall have an upper limit of 5% (e.g., less than 5% of the measured xenon activity shall remain in the cell at the start of the next measurement).

- R 2.2.4 The loss of sample during counting shall not be larger than 5% for each measurement cycle.
- R 2.2.5 The gamma-resolution shall be better than 15% full width at half maximum (FWHM) at 80 keV and the beta resolution shall be better than 40 keV at 129 keV.
- R 2.2.6 The shielding design shall minimize the direct external radiation which reaches the detector, such that the detector is capable of meeting the minimal detection limit requirements specified in R 2.2.2 without additional measures.
- R 2.2.7 The system shielding shall attenuate the radiation of a ^{137}Cs source at 1 m distance from the detector by at least a factor of 200. **Note:** This would reduce the count rate of a calibration source with typical source strength (40 kBq) to approximately 0.1 counts per second with a detector that has a cross section of 50 cm².
- R 2.2.8 The nuclear gamma and beta detectors gains shall be stable within 2% during the sample count time.
- R 2.2.9 The acquisition of the sample measurement shall be less than or equal to 24 hours.**

F 2.3 Nuclear Detector Calibration

- R 2.3.1 The system shall provide a calibration verification method for the nuclear detector, and shall be capable of reporting background and spectrum data when requested by the user.
- R 2.3.2 The system shall be capable of having the nuclear detectors replaced and maintaining calibration through update of the calibration files. **(Updated 2018)**
- R 2.3.3 The system shall be capable of collecting background spectra periodically when commanded by on-site or remote users.
- R 2.3.4 The system shall be capable of collecting energy spectra calibration data periodically.
- R 2.3.5 The system shall limit the flux of radiation (cross-talk) between detector cells to less than 0.1% of the counts observed in an adjacent detector during calibration (excludes electrical crosstalk and gas migration between cells).
- R 2.3.6 The system shall use a radioactive Xe gas calibration standard to calibrate the nuclear detector (typically loaded manually into the cell).
- R 2.3.7 The system shall have a statistical uncertainty of the calibration sample activity measurement of 1% or less.
- R 2.3.8 The gamma detectors and nuclear measurement acquisition electronics shall be configured such that the operating range for gamma energy detection spans from at least 25 keV to 730 keV at approximately 5 keV per channel. Gamma calibration points are to be observed at approximately 81, 250, 295, 352 and 609 keV.

- R 2.3.9 The beta detectors and nuclear measurement acquisition electronics shall be configured such that the operating range for the beta energy spans at least 25 keV to 900 keV, at approximately 5 keV per channel. Beta calibration points are to be observed at approximately 129 to 199, 346 to 391, and up to 905 keV.
- R 2.3.10 The detector shall be calibrated for efficiency for each of the four xenon isotopes ($^{131\text{m}}\text{Xe}$, ^{133}Xe , $^{133\text{m}}\text{Xe}$ and ^{135}Xe).
- R 2.3.11 The detector shall be calibrated for interference factors for all operationally relevant interfering nuclides: ^{133}Xe , ^{135}Xe , ^{222}Rn and its decay products at minimum.
- R 2.3.12 The calculations for the nuclear detector energy, resolution, and efficiency calibration shall be documented.

F 2.4 Nuclear Detector Quality Control (QC)

- R 2.4.1 The system shall be able to perform spectral acquisition using a QC source.
- R 2.4.2 The system shall take a QC source measurement before each (cycle) sample measurement.
- R 2.4.3 The system shall automatically retract and store the QC source before sample acquisition.
- R 2.4.4 The system shall limit (i.e., shield) the signal of the QC sources to minimize false identification of any xenon isotope and shall not significantly raise the spectrum baseline.
- R 2.4.5 The QC source shall have a peak at gamma energy of at least 662 keV in order to detect small energy drifts.
- R 2.4.6 The system shall use a QC source that is below the limit for exempt packaging as specified in U.S. Department of Transportation shipping standards.
- R 2.4.7 The system shall provide a QC source certificate on the composition and activity of the source. If the source is used for calibration, a certificate by an accredited institution shall be provided on composition, activity, uncertainties of activity, and dimensions of the active source.
- R 2.4.8 The QC source shall be a sealed source per ANSI/HPS N43.6 or ISO 2919 classification designation.
- R 2.4.9 The QC source to detector configuration shall be reproducible for the QC data collections.
- R 2.4.10 The counting statistics of the QC measurement shall be better than 5% (over the validity period of the QC source) for the 662 keV peak of the ^{137}Cs source in the gamma detector.
- R 2.4.11 The quality assurance/quality control (QA/QC) source when counted, shall be placed in the same location ± 2 mm.

F 3 Process Control (PNNL-developed GFE software)

F 3.1 Software Control

- R 3.1.1 The system shall meet DoD Net Ready requirements and allow a two-way communication with the user. The user shall be able to send commands to the stations (i.e., for requesting data) and perform unscheduled tasks (Updated 2018)**
- R 3.1.2 The system shall provide a means to verify, remotely and on-site, the status of the Xenon International system.
- R 3.1.3 The system shall use software meeting industry recognized quality assurance standards, e.g., software grading, development, and verification and validation requirements. **(Updated 2018)**
- R 3.1.4 The system shall be equipped and configured to allow an operator (whether remote or on-site) to control system functions. These functions include:
 - (1) view sensor parameters/system status
 - (2) initiate calibration validation and self-diagnostic functions
 - (3) retrieve archived data.
- R 3.1.5 The system shall allow an operator (either remotely or on-site) to switch operation modes and update the system configuration.
- R 3.1.6 The system shall not require user interaction to retrieve data after a shutdown and restart.
- R 3.1.7 The system shall be configured to support as a minimum the following protocols and/or services:
 - (1) secure shell (SSH)
 - (2) secure copy (SCP) or secure shell file transfer protocol (SFTP)
 - (3) simple mail transfer protocol (SMTP)
 - (4) network time protocol (NTP).
- R 3.1.8 The system shall include a text-based menu set for the command and control functions.
- R 3.1.9 The system shall include a graphical user interface (GUI) for the command and control functions.
- R 3.1.10 The system shall use spectral acquisition software that allows saving the partial spectra every 2 hours, so that in case of a system failure, the spectrum acquired until that event is not completely lost.
- R 3.1.11 The system shall use a category-5 unshielded twisted pair (UTP) cable with a RJ-45 connector (IEEE 802.3 standard) for physical connections.

F 3.2 Operational Modes

- R 3.2.1 The system shall have inherent capability for the following modes:

- (1) fully mission capable (FMC) - full functional operation
- (2) partially mission capable (PMC) - less than full capability, but able to produce results at least intermittently (for example, operating with a failed sensor or nuclear detector)
- (3) regeneration - the system must be fully regenerated and ready to process within 8 hours
- (4) shutdown mode - emergency and controlled shutdowns shall be operational when necessary
- (5) maintenance - maintenance is being performed on the system
- (6) failure - the system is unable to collect or record data.

F 4 Data

F 4.1 Data Collection and Format

- R 4.1.1 The system shall have a data availability rate of 95% or greater. Data availability refers to availability of the system and does not include downtime to receive new parts, delivery, availability of maintenance workers, etc. The algorithm documented in the appendix shall be used to calculate data availability.**
- R 4.1.2 The system shall produce, at a minimum, the following five categories of data: collection, spectral, state-of-health (SOH), event logger, and stable gas and activity concentrations.
- R 4.1.3 The system shall archive data of all types generated or collected internally. These data shall be readily accessible locally or remotely.
- R 4.1.4 The system shall produce, at a minimum, the following six categories of data: collection, spectral, state-of-health (SOH), alert, event logger, and stable gas and activity concentrations. **(Updated 2018)**
- R 4.1.5 The system shall record collection data that include all information necessary to uniquely identify a sample and the particulars of its collection (e.g., collection start/stop).
- R 4.1.6 The system shall automatically generate concentration data that contain the activity and the stable gas quantification and reported in activity concentration (mBq/m³).
- R 4.1.7 The system shall report and archive the results in the format of current systems (pbg file).
- R 4.1.8 The system shall allow all recorded data to be downloaded from the system to an external device.
- R 4.1.9 Full sample spectra shall be sent to the data center after each processing and counting cycle.

R 4.1.10 The system shall, at least daily, transmit the resulting data to the National Data Center (NDC) either automatically or when requested, such as prior to a planned system shutdown.

R 4.1.11 The system shall locally archive all data generated in the station after being sent to the NDC.

F 4.1.1 State of Health Data

R 4.1.1.1 The system shall be equipped with sensors and instrumentation to provide SOH data.

R 4.1.1.2 The SOH data shall be written in non-binary format.

R 4.1.1.3 The SOH data shall be saved every 1-30 minutes (based on a user-defined interval).

R 4.1.1.4 The system shall be able to transmit the SOH file to the NDC.

R 4.1.1.5 The SOH data transmission interval to send data to the data center shall be able to be set at 5 minute intervals between 5 minutes and 120 minutes. **(Updated 2018)**

R 4.1.1.6 The xenon collection efficiency shall be documented in the SOH file.

R 4.1.1.7 The SOH data shall be sufficient to track the flow of the sample through the subsystems including: gas compression, dryer, collection, separation, quantification, and nuclear detection.

R 4.1.1.8 The system shall retain a full set of SOH data for 400 days. **(Updated 2018)**

F 4.1.2 Event Logger Data

R 4.1.2.1 The Xenon International system shall generate logger data (event data), that shall include a record of all commands issued by the software and by an operator.

R 4.1.2.2 The event logger data shall be written in non-binary format.

R 4.1.2.3 When an alert occurs, the system shall transmit an alarm to the NDC.

F 4.2 Data Support Tools

R 4.2.1 The system shall include a detector histogram viewer to facilitate accurate analysis of detector performance and xenon isotope samples.

R 4.2.2 The system shall include a SOH viewer and an event logger file viewer that include graphical user representation.

F 5 Archive System **(Updated 2018)**

R 5.1.1 The Xenon International system shall include a sample archiving system.

R 5.1.2 The archive system shall hold samples for a minimum period of three/four days (objective/threshold)

R 5.1.3 The archive containers shall be leak tight and shall be easily detachable from the system for shipment to a laboratory for further analysis.

- R 5.1.4 Sample loss from the archive container: after a storage and transport period of 10 days at least 0.32 cm³ of xenon shall be stored in the archive container.
- R 5.1.5 Relevant collection, processing, and analysis parameters shall be captured electronically with the archive container for use at the confirmatory analysis laboratory.
- R 5.1.6 After the holding time of the xenon gas in the archive, the sample shall be removed from the archive, e.g., by evacuation or by flushing with a carrier gas. After the regeneration of the archive container, less than 1% of the pervious xenon gas sample shall be leftover. The system shall overwrite electronic data after the holding time of the xenon gas in the archive.
- R 5.1.7 The exhaust gas of the archive system shall not interfere with the air intake of the sampling system and shall not be blown into the inside of the station if this violates local radiation protection regulations. In this context, in particular the possibly higher activities of calibration gases and spikes have to be considered.

F 6 Design Constraints

F 6.1 Size and Weight

- R 6.1.1 The system shall not exceed a maximum size of 45-in. wide x 32-in. deep x 80-in. high (threshold) 45-in. x 30-in. x 76-in. (objective). The footprint will not include the uninterruptable power supply (UPS), archive system, and nitrogen generator. (Updated 2018)**
- R 6.1.2 The system shall not exceed a weight of 3000 pounds (threshold) 2500 pounds (objective). The weight requirements will not include the uninterruptable power supply (UPS) and nitrogen generator. (Updated 2018)**

F 6.2 Environmental

- R 6.2.1 The system shall be equipped with a filter that will remove or filter out any particle greater than 25 µm, preventing these particles from entering the processing stream at the air inlet.
- R 6.2.2 The system shall be designed so operators can easily replace and/or clean the filter
- R 6.2.3 Operating conditions:
 - R 6.2.3.1 The system shall be able to operate in a temperature of ambient air: 50 °F (lower limit), to 95 °F (upper limit).
 - R 6.2.3.2 The temperature of intake air shall be: -40 °F (lower limit), 100 °F (upper limit).
 - R 6.2.3.3 The relative humidity (non-condensed) of ambient air shall be: 50% (lower limit), 80% (upper limit).

R 6.2.3.4 The relative humidity (non-condensed) of intake air shall be:
5% (lower limit), 100% (upper limit).

R 6.2.3.5 The storage temperature shall be (induced):
-50 °F (lower limit), 140 °F (upper limit).

Note: Extreme temperatures below 0 and above 100 are not expected to persist for greater than 12 hours. **(Updated 2018)**

R 6.2.4 Transport conditions:

R 6.2.4.1 The relative humidity (non-condensed) during transport shall be:
5% (lower limit), 100% (upper limit).

R 6.2.4.2 The temperature during transport shall be (induced):
-50 °F (lower limit), 140 °F (upper limit).

Note: Extreme temperatures below 0 °F and above 100 °F are not expected to persist for greater than 12 hours. **(Updated 2018)**

R 6.2.4.3 The system shall be designed for the following shock: short trips (less than 150 miles) over rough roads (i.e., repeated impulses at higher frequencies) and normal warehouse handling.

R 6.2.4.4 The system shall be designed to withstand the following vibration: long trips (over 1000 miles) over smooth roads (i.e., long-duration exposure to lower frequencies).

R 6.2.4.5 The packaging shall be robust enough to prevent damage of the equipment because of thermal shock.

F 6.3 Electrical

R 6.3.1 The voltage of the system shall function on single phase domestic power (threshold) and single phase international power (objective).

R 6.3.2 The system shall provide a visible indicator that it is receiving input power. An example of this may be the emergency power shutoff switch has a light indicating power going to the system.

R 6.3.3 The system's power interface shall include line conditioning and surge protection.

R 6.3.4 The system's grounding, bonding, and shielding interface and installation requirements shall be in accordance with acceptable industry standards and U.S. (threshold) and international (objective) electrical codes.

R 6.3.5 The system shall include an interface to attach to a UPS to facilitate a controlled shut-down upon a sudden, unexpected power loss. The UPS is not required to be a part of the footprint.

- R 6.3.6 This UPS shall ensure operation of all essential equipment for at least 10 minutes (Threshold)/15 minute (Objective). Essential equipment will include the control computer and the electronics enclosure. **(Updated 2018)**
- R 6.3.7 The system shall provide current overload protection and easily visible indications that a circuit breaker has been tripped.
- R 6.3.8 Details of the electrical circuitry shall be included in system documentation.

F 6.4 Materials

- R 6.4.1 All support structure and fabricated components (i.e., frame and housing of the system), shall be designed to minimize corrosion while promoting system grounding. **(Updated 2018)**
- R 6.4.2 The system shall not require liquid nitrogen or use of any Class I or Class II ozone-depleting refrigerants in development, operation, or maintenance.
- R 6.4.3 **(Deleted 2018)**

F 6.5 System Expendables

- R 6.5.1 The system shall be designed to limit the number/amount of fielded expendables (e.g., no excess carrier gas).

F 6.6 Security

- R 6.6.1 The internal hardware and software shall not be accessible to the operator, only to the authorized maintenance technician.
- R 6.6.2 The computer system(s) shall be protected by password(s) and by other similar cyber-security features.

F 6.7 Safety

- R 6.7.1 The system shall allow maintenance personnel to override the interlock and other safety features as necessary for maintenance with proper warnings noted in the technical manuals.
- R 6.7.2 The system shall properly mark any system element that is hazardous such as high voltages/currents, high/low temperatures, radioactive, or chemically hazardous materials to meet U.S. safety requirement codes.
- R 6.7.3 The system shall be designed to shut down without damaging personnel, anything on the outside of the system, and with minimal damage to the system in the case of a power outage or a sensor failure.
- R 6.7.4 The system shall be equipped with a manual emergency shutdown button.
- R 6.7.5 The system shall be equipped with software safety controls that shall manage overheating and overpressure. Pending software failure, redundant safety features shall be installed such as pressure relief valves, or rupture disks. Electrical safety features are identified in section F 6.3.

F 6.8 Shutdown

- R 6.8.1 The system shall not require recalibration after a shutdown.
- R 6.8.2 The system shall, once the power has been restored, be able to automatically regenerate itself after a power shutdown or power outage (among others: evacuating the partial process samples).
- R 6.8.3 The system shall, after an uncontrolled shut down, be able to sequence and retrieve the 2-hour spectra being acquired until shutdown.
- R 6.8.4 The system shall have a controlled shutdown procedure.
- R 6.8.5 The system shall be regenerated and start a collection within 8 hours after power resumed (after a power outage of less than 24 hours).
- R 6.8.6 The system shall send a spectrum before any controlled shutdown.

F 6.9 Usability

- R 6.9.1 The system's control panels shall be visible, easily readable, and accessible by the operator under all possible ambient lighting conditions. "Easily readable" is defined such that a government operator (with 20/20 corrected vision) can read displayed system information at a distance of 3 feet.
- R 6.9.2 The user shall be able to shut down the system either on-site or remotely. The system shall allow immediate shutdown in case of emergency and a controlled shutdown that will preserve data.

F 7 Documentation

- R 7.1.1 The system shall include a mechanical drawing package including assemblies.
- R 7.1.2 The system shall include an electrical drawing package.
- R 7.1.3 The system shall include an electrical connection scheme.
- R 7.1.4 The system documentation shall include the layout (equipment, connections, ducting, etc.) of a generic station site and equipment.
- R 7.1.5 The system's documentation package shall include an operations/user manual that will include a troubleshooting guide for the system.
- R 7.1.6 The system's documentation package shall include test documentation.
- R 7.1.7 The system's documentation package shall include intellectual property releases.
- R 7.1.8 The system's documentation package shall include software design and test documentation.

- R 7.1.9 The system's documentation package shall include a training package that includes training material on the installation, theory of operation, operation, preventive maintenance, and demand maintenance of the system for on-site and depot maintenance for operator and maintenance personnel. The training shall be at the level targeting personnel with a high school diploma or a journeyman/apprentice-type education.

F 8 Sustainment

F 8.1 Availability

- R 8.1.1 The system shall be designed to not exceed 7 consecutive days or 15 days total annually of downtime. This includes both scheduled and unscheduled downtime.**

F 8.2 Maintainability

- R 8.2.1 The system shall be designed in such a way that critical components can be easily accessed for troubleshooting, repair, and system evolution.
- R 8.2.2 The system shall be designed to minimize maintenance because of environmental conditions such as dust, condensation, humidity, etc.

F 8.3 Mean Time between Critical Failures (MTBCF)

- R 8.3.1 The Xenon International shall be designed to meet a minimum MTBCF of 8760 hours. Critical failure is defined as any failure that would cause the Xenon International to lose its collection, processing, and/or measurement capabilities and requires immediate maintenance action.

F 8.4 Mean Time between Failures (MTBF)

- R 8.4.1 The Xenon International system shall be designed to meet a minimum MTBF of 4800 hours. Failure is defined as any failure that does not cause the Xenon International to lose its collection, processing, and/or measurement capabilities (i.e., any capability less than FMC mode).

F 8.5 Mean Time to Repair (MTTR)

- R 8.5.1 The Xenon International system shall be designed to meet on-site maintenance MTTR of 4 hours. This applies to “active” corrective maintenance only, and does not include time for scheduled maintenance, warm-up or cool-down cycles, or administrative and logistics delays. MTTR shall not include normal regeneration time; however, any time in excess of the required maximum regeneration times shall be included in the MTTR. Corrective maintenance shall begin upon verification of a fault by technician and includes time to:
- (1) localize the fault
 - (2) isolate the fault
 - (3) remove and replace the faulted item
 - (4) calibrate and/or align the Xenon International (if necessary)
 - (5) test and check out the Xenon International (excluding the regeneration of gas system filters)
 - (6) initiate system start sequence

F 8.6 Maximum Corrective Maintenance Time ($C_{MXt_{MX}}$)

- R 8.6.1 The Xenon International system shall be designed to meet an on-site maintenance $C_{MXt_{MX}}$ of 8 hours. $C_{MXt_{MX}}$ shall not include normal regeneration time; however, any time in excess of the required maximum regeneration times shall be included in the $C_{MXt_{MX}}$.

F 8.7 Support Equipment

- R 8.7.1 The system shall, to the maximum extent possible, be capable of being maintained with common hand tools, and support and test equipment.
- R 8.7.2 The system shall require no special test equipment to operate (threshold), and no special test equipment will be required to maintain the system for on-site maintenance (objective). Calibration sources are an exception to this requirement.

F 8.8 Transportation and Facilities

- R 8.8.1 The packaging, handling, storage, and transportation (PHS&T) considerations shall be identified and documented (including proper markings) to ensure the system and/or components of the system shall withstand all modes of transportation used. The packaging must protect sensitive system components from temperature extremes below 0 °F and above 100 °F for up to 12 hours. **(Updated 2018)**
- R 8.8.2 The system shall use reusable containers for critical spare parts and assemblies.

Appendix A

Calculations

A.1 Minimal Detectable Concentrations Calculation

To calculate the minimal detectable concentration (MDC), the following equation is used:

$$MDC\left(\frac{mBq}{m^3_{air}}\right) = \left(\frac{2.71 + 4.65\sigma_0}{\varepsilon_\gamma \varepsilon_\beta \gamma_{BR} \beta_{BR}} \right) \frac{\lambda^2 * T_C}{(1 - \exp(-\lambda T_C)) \exp(-\lambda T_P) (1 - \exp(-\lambda T_A))} \left(\frac{1000}{V_{Air}} \right)$$

where the first term takes into account the typical MDC expression using the standard deviation of the measured blank sample (σ_0) with the gamma and beta detector efficiencies and branching fractions respectively (ε_γ , ε_β , γ_{BR} , β_{BR}); the second term accounts for the decay of the xenon isotope during the collection time (T_C), the processing time (T_P) and the nuclear acquisition time (T_A) using the decay constant for the specific isotope (λ); and the final third term accounts for the volume of air measured (V_{air}) and conversion to millibecquerel.

A.2 Data Availability Calculation

Data availability for the Xenon International system is the ratio of data received by the National Data Center (NDC) that can be categorized relative to the data expected to be received from that station, expressed as a percentage.

$$\text{Data Availability}(\%) = \frac{\text{Data Received by the NDC}}{\text{Expected Data}} \cdot 100$$

Appendix B

Change Log of “Requirements for Xenon International”

Date	Change Description	Pages Changed
09/2013	PNNL-22227, DOI: https://doi.org/10.2172/1095446	Original Requirements Document
12/2015	Revision 1: PNNL-22227 Rev.1, DOI: https://doi.org/10.2172/1122330	
12/30/2015	Revised R 1.1.6 to incorporate R 1.2.2 and specify maximum acceptable impurity concentration levels.	6
12/30/2015	Removed R 1.2.2 (incorporated into R 1.1.6).	6
12/30/2015	Revised R 2.2.2 to refine expected ambient background radiation.	7
12/30/2015	Revised R 2.2.6 to improve manufacturability.	8
12/30/2015	Revised R 2.3.5 to clarify acceptable interference levels (crosstalk) from calibration samples in adjacent detector cells.	8
12/30/2015	Revised R 2.3.8 to clarify acceptable detector acquisition gamma energy configurations.	8
12/30/2015	Revised R 2.3.9 to clarify acceptable detector acquisition beta energy configurations.	9
12/30/2015	Revised R 2.3.11 to clarify minimum interference factor calibrations.	9
12/30/2015	Revised R 2.4.3 to refine QC source storage sequence.	9
12/30/2015	Revised R 4.1.1.4 to clarify state of health transmission requirements.	12
08/2018	Revision 2: PNNL-22227 Rev.2: DOI: TBD	
04/2018	Revised Key Operations Parameters to Key Performance Parameters	3
05/2018	Revised Processing efficiency requirements IDs	3
05/2018	Added Objective to UPS requirements in KPP requirements	4
05/2018	Added Key System Attributes	4
04/2018	Revised R 1.2.3 to clarify use of liquid nitrogen generator	6
04/2018	Revised R 2.2.2 to add minimum quantifiable concentration (MQC)	7
04/2018	Revised R 2.3.2 clarifying calibration method for nuclear detectors	8
04/2018	Revised R 3.1.1 to include DoD Net ready requirements	10
04/2018	Revised R 3.1.3 to clarify software quality standards	10
05/2018	Revised 4.1.4 to update data categories	11
04/2018	Revised R 4.1.1.5 to specify SOH transmission rate	12
04/2018	Added R 4.1.1.8 to specify how long SOH is retained on system	12
04/2018	Revised R 5.1.1 to include sample archiving	12
04/2018	Added R 5.1.2 to specify how long archive samples are kept	12

Date	Change Description	Pages Changed
04/2018	Added R 5.1.3 to specify archive container integrity	13
04/2018	Added 5.1.4 to specify how much product is retained in archive container	13
04/2018	Added 5.1.5 to specify archive container electronic data storage	13
04/2018	Added R 5.1.6 to specify carryover from sample to sample in archive	13
04/2018	Added R 5.1.7 to specify disposal of archive samples	13
05/2018	Revised R 6.1.1 to specify objective size requirement	13
04/2018	Revised R 6.1.2 to specify components included in weight limit	13
05/2018	Revised R 6.2.3.5 temperature storage requirements	14
05/2018	Revised 6.2.4.2 Revised transportation temperature	14
06/2018	Revised 6.3.6 to include objective for UPS	15
04/2018	Revised 6.4.1 to clarify corrosion resistance	15
04/2018	Removed R 6.4.3	15
05/2018	Revised R 8.8.1 for packaging, shipping, storage, and transportation	18