

Statistical Analysis of the Critical Mass Laboratory (209-E) Filter Data

T. L. Welsh

June 2005



Prepared for the U.S. Department of Energy
under Contract DE-AC05-76RL01830

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor Battelle Memorial Institute, nor any of their employees, makes **any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights.** Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or Battelle Memorial Institute. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

PACIFIC NORTHWEST NATIONAL LABORATORY

operated by

BATTELLE

for the

UNITED STATES DEPARTMENT OF ENERGY

under Contract DE-AC05-76RL01830

Printed in the United States of America

Available to DOE and DOE contractors from the
Office of Scientific and Technical Information,
P.O. Box 62, Oak Ridge, TN 37831-0062;
ph: (865) 576-8401
fax: (865) 576-5728
email: reports@adonis.osti.gov

Available to the public from the National Technical Information Service,
U.S. Department of Commerce, 5285 Port Royal Rd., Springfield, VA 22161
ph: (800) 553-6847
fax: (703) 605-6900
email: orders@ntis.fedworld.gov
online ordering: <http://www.ntis.gov/ordering.htm>



This document was printed on recycled paper.

(9/2003)

Statistical Analysis of the Critical Mass Laboratory (209-E) Filter Data

T. L. Welsh

June 2005

Prepared for
the U.S. Department of Energy
under Contract DE-AC05-76RL01830

Pacific Northwest National Laboratory
Richland, Washington 99352

Summary

The potential-to-emit (PTE) calculation for the Critical Mass Laboratory (209-E) exhaust stack, which was developed around 1992, was questioned by the Washington Department of Health (WDOH). WDOH issued a notice of assurance of discontinuance and a commitment was made in the subsequent reply to develop a new PTE. Samples were obtained from the pre-filters and the first bank of HEPA filters in the 209-E exhaust system and analyzed per the sampling and analysis plan (LeBaron 2005). The sample information from the exhaust pre-filters and the first bank of HEPA filters provided in Table 5 will be used to develop a PTE. If the PTE shows that the stack emissions are greater than 0.1 mrem/year, the stack will be designated as a major stack. If the PTE shows that the stack emissions are less than 0.1 mrem/year, the stack will be designated as a minor stack.

Contents

Report	1
References	4

Tables

1	Pre-Filter Sample Results	5
2	HEPA Filter Sample Results	6
3	Mean Concentrations and the Associated Variability	13
4	Pre-filter versus HEPA filter Mean Concentrations	13
5	Final Mean Concentrations and the Associated Variability	14
6	Systematic and Total Measurement Uncertainty Estimates	15

Figures

1	Total Alpha Analytical Results	7
2	Americium-241 Analytical Results	7
3	Total Beta Analytical Results	8
4	Plutonium-239/240 Analytical Results	8
5	Plutonium/Americium-241 Analytical Results	9
6	Plutonium/Americium-242 Analytical Results	9
7	Uranium-234 Analytical Results	10
8	Uranium-235 Analytical Results	10
9	Uranium-236 Analytical Results	11
10	Uranium-238 Analytical Results	12

Statistical Analysis of the Critical Mass Laboratory (209-E) Filter Data

Summary

The potential-to-emit (PTE) calculation for the Critical Mass Laboratory (209-E) exhaust stack, which was developed around 1992, was questioned by the Washington Department of Health (WDOH). WDOH issued a notice of assurance of discontinuance and a commitment was made in the subsequent reply to develop a new PTE. Samples were obtained from the pre-filters and the first bank of HEPA filters in the 209-E exhaust system and analyzed per the sampling and analysis plan (LeBaron 2005). The sample information from the exhaust pre-filters and the first bank of HEPA filters provided in Table 5 will be used to develop a PTE. If the PTE shows that the stack emissions are greater than 0.1 mrem/year, the stack will be designated as a major stack. If the PTE shows that the stack emissions are less than 0.1 mrem/year, the stack will be designated as a minor stack.

Purpose

The purpose of the statistical analysis of the data is to determine the mean concentration for each radionuclide used in determining the PTE.

Sampling Design

In order to obtain samples from the two banks of filters (the pre-filters and the first bank of HEPA filters), a sampling design was developed using the methodology outlined in SW-846 (EPA, 2005). The number of samples was determined using Equation 8 from Table 9-1 varying the mean, variability, and the difference from the threshold (see Appendix B in LeBaron 2005). Equation 8 assumes that the data are from a normal distribution. It was determined that ten samples would be obtained from each bank of filters.

Each of the two banks of filters is a two-by-two array consisting of four filters each. The filter media of all four filters is considered the filter face (filter bank) through which the exhaust air flows. An imaginary grid was superimposed over each filter face and a series of consecutive numbers was assigned to each cell of the grid. Ten random numbers for each filter face (each number representing a cell) were selected using a random number generator from SPLUS (Statistical Sciences, 1993). A sample core (a full penetration sample; from front to back) was obtained for each selected cell. All samples were sent to 222-S Analytical Laboratory for dissolution and analysis.

Analytical Results

The twenty samples were analyzed according to the sampling and analysis plan (LeBaron 2005) at 222-S Analytical Laboratory. The data were reported in Bushaw 2005. The data are listed in Table 1 for the pre-filter sample results and Table 2 for the HEPA filter sample results. The data are illustrated in Figures 1- 10.

To compare the results for analytes measured using two different analytical methods some of the data were converted from $\mu\text{g}/\text{sample}$ to $\mu\text{Ci}/\text{sample}$ (or vice versus) using the specific activities reported in DOE-STD-1027-92 (Chow et.al., 1995). As can be seen in Figure 2, the two analytical methods (GEA versus TRU SPEC) for americium-241 provide equivalent results within the total measurement uncertainty. In addition, after summing the plutonium-239 and plutonium-240 results obtained using ICP/MS, the results are equivalent (within the total

measurement uncertainty of the two methods) to the plutonium-239/240 results obtained by the TRU SPEC analytical method (see Figure 4).

Statistical Analysis

If the reported analytical results were a mixture of “real” values and “less than” values, then three methodologies were used to determine a mean concentration. The first method did not use the “less than” values in the statistical analysis. The second method uses the upper bound of the “less than” value (e.g., use 6.51E-04 for <6.51E-04). The third method uses half of the upper bound of the “less than” value. All three methods produce (1) a bias (typically positive) of unknown magnitude in the mean concentration and (2) a bias (direction not known) of unknown magnitude in the variance associated with the mean. The method which produced the largest mean concentration (so as not to underestimate the PTE) is reported.

Normal Distribution

Two independent tests were performed to test the hypothesis that the random sample comes from a normal distribution with unspecified mean and variance versus the alternative that the distribution is not normal. The two tests are called the Lilliefors Test for Normality and the Shapiro-Wilk Test for Normality (Bowen 1988, chapter 9). The normality tests were performed for each set of filter face analytical results.

The hypothesis that the random sample comes from a normal distribution was not rejected using either test at the 0.05 level of significance for all analytes except for the Total Beta results from the HEPA filter samples. If one result (S05051-20) was deleted from the HEPA filter data set for Total Beta, the hypothesis that the random sample comes from a normal distribution would not be rejected. However, the following two statements (Ostle, 1988) indicate that using the mean concentration for the HEPA filter Total Beta result is reasonable. “That is, unless the population distribution is extremely different from a normal distribution and the sample size is extremely small, the sampling distribution of \bar{X} will be approximately normal. Therefore we can obtain reasonably good approximation by using normal theory.” “However, for many reasonable distributions, (see chapter 4), the Central Limit Theorem will allow us to use a normal approximation with a sample size as small as $n=5$.”

Mean Concentration and the Variability Associated with the Mean

For each analyte for each data set (filter bank), the data were statistically evaluated using one-way analysis of variance (ANOVA): the data are identified by one variable (the sample). The one-way ANOVA statistical model used to describe the structure of the data is

$$Y_{ij} = \mu + S_i + A_{ij},$$

$$i = 1, 2, \dots, a, \quad j = 1, 2, \dots, n_i,$$

where

Y_{ij}	=	concentration from the j^{th} analytical result from the i^{th} sample
μ	=	the grand mean
S_i	=	the effect of the i^{th} sample
A_{ij}	=	the effect of the j^{th} analytical result from the i^{th} sample
a	=	the number of samples (10)
n_i	=	the number of analytical results from the i^{th} sample (1 or 2).

The variable S_i is assumed to be a random effect (the samples were randomly chosen from the filter face). This variable, as well as A_{ij} , are assumed to be normally distributed with means zero and variances $\sigma^2(S)$ and $\sigma^2(A)$, respectively. Estimates of $\sigma^2(S)$ and $\sigma^2(A)$ were obtained using Restricted Maximum Likelihood Estimation (REML) techniques. This method applied to variance component estimation is described in Harville, 1977. The results using the REML techniques were obtained using the statistical analysis package S-PLUS (Statistical Sciences, 1993).

The estimator, $\hat{\mu}$, is the maximum likelihood estimate of the mean. This estimator was determined by the structure of the data reflected by the statistical model. The estimate of $\hat{\mu}$ was obtained using REML techniques in S-PLUS. The estimated standard deviation of the mean (where the mean is the maximum likelihood estimate), $\hat{\sigma}_{\hat{\mu}}$, is the square root of a function of the variance estimates $\sigma^2(S)$ and $\sigma^2(A)$. Since the data are unbalanced, $\hat{\sigma}_{\hat{\mu}}$, is a complicated function of these two variances.

The mean analyte concentration ($\hat{\mu}$) and the relative standard deviation (RSD) associated with the mean ($\hat{\sigma}_{\hat{\mu}} / \hat{\mu} \times 100$) are listed in Table 3 for both the Pre-filter sample results and the HEPA filter sample results.

The comparisons of the Pre-filter results to the HEPA filter results indicate a fairly constant ratio. The data, the ratio, and the summary statistics for the ratio data are provided in Table 4. Based on these comparisons, the HEPA filter mean concentration for analytes with all data reported as “less than” values, was determined by multiplying the Pre-filter mean concentration by 0.5. This replaces using an upper value from a “less than” value for the mean concentration which produces a positive bias; the mean concentration is overstated.

The final mean concentrations for the Pre-filter sample results and the HEPA filter sample results are listed in Table 5. If the HEPA filter mean concentration was determined from the Pre-filter mean concentration then the uncertainty for this HEPA filter result was calculated using both the variability in the Pre-filter result and the variability in the mean ratio result.

Reference standards and/or spike analyses were performed as specified in the analytical procedures. All reference standards were within statistical process control limits. The reference standards and the spike recoveries provide assurances that the analytical procedures are producing valid results. The reference standard or spike results provide estimates of the “systematic” uncertainty associated with the analytical procedures for the sample matrix. The systematic uncertainties ranged from 3% (1 RSD) to 14%. The total measurement uncertainty for a mean concentration was determined by adding in quadrature the variance associated with the mean (1 RSD²) and the systematic variance estimate (1 RSD²). The systematic uncertainty along with the total measurement uncertainty are provided for each analyte per filter type in Table 6. Except for cesium-137 and strontium-89/90 the total measurement uncertainty ranges from approximately 5% to approximately 20%.

Conclusion

The mean concentrations from which the PTE will be developed are listed in Table 5.

References

- Bowen, W. Michael and Carl A. Bennett (ed.). 1988. *Statistical Methods for Nuclear Material Management*, NUREG/CR-4604.
- Bushaw, R.A. 2005. *Final Report for the Critical Mass Laboratory (209-E) Filter Samples Collected in March 2005*, CH2M-0501330. CH2M Hill Hanford Group, Inc, Richland, WA.
- Chow, Jim, et.al.. 1995. *Table of DOE-STD-1027-92 Hazard Category 3 Threshold Quantities for the ICRP-30 List of 757 Raionuclides*. Los Alamos National Laboratory, Los Alamos, New Mexico.
- Environmental Protection Agency. 2005. *Test Methods for Evaluating Solid Waste, Physical/Chemical Methods* (SW-846).
- Harville, D.A. 1977. "Maximum Likelihood Approaches to Variance Component Estimation and to Related Problems," *Journal of the American Statistical Association*, pp. 320-340.
- LeBaron, G.J. 2005. *Sample Analysis Plan for the Critical Mass Laboratory (209-E)*, HNF-23373 Revision 0. Fluor Hanford, Richland, WA.
- Ostle, Bernard and Linda C. Malone. 1988. *Statistics in Research, Fourth Edition*. , Iowa State University Press, Ames, Iowa.
- S-PLUS Reference Manual*, Version 3.3. 1993. Statistical Sciences, Inc., StatSci, a division of MathSoft, Inc. Seattle, Washington.

Table 1. Pre-Filter Sample Results

Customer ID	Analysis	Alpha Env	Am-241	Am-241	Beta in Env	Pu239
		Solid/Misc	TRU-SPEC	GEA	Samples	ICP/MS
		µCi/sample	µCi/sample	µCi/sample	µCi/sample	µg/sample
S05051-01	1	3.82E-02	1.28E-02	1.20E-02	2.59E-03	2.73E-01
S05051-01	2	3.61E-02	1.29E-02	1.16E-02	2.45E-03	2.70E-01
S05051-02	1	5.86E-02	1.98E-02	1.89E-02	4.17E-03	4.29E-01
S05051-03	1	3.18E-02	1.11E-02	1.05E-02	2.09E-03	2.51E-01
S05051-04	1	4.49E-02	1.55E-02	1.49E-02	2.93E-03	3.28E-01
S05051-05	1	4.87E-02	1.59E-02	1.55E-02	3.43E-03	3.60E-01
S05051-06	1	4.89E-02	1.60E-02	1.54E-02	3.16E-03	3.60E-01
S05051-06	2	4.65E-02	n/a	n/a	3.20E-03	n/a
S05051-07	1	1.97E-02	6.45E-03	6.03E-03	1.33E-03	1.41E-01
S05051-08	1	2.81E-02	9.22E-03	8.49E-03	1.96E-03	1.91E-01
S05051-09	1	6.40E-02	2.12E-02	2.08E-02	4.50E-03	4.77E-01
S05051-10	1	6.34E-02	2.02E-02	1.95E-02	4.38E-03	4.55E-01
Customer ID	Analysis	Pu240	Pu244/Cm244	Pu/Am242	Pu/Am241	Pu238
		ICP/MS	ICP/MS	ICP/MS	ICP/MS	TRU-SPEC
		µg/sample	µg/sample	µg/sample	µg/sample	µCi/sample
S05051-01	1	2.97E-02	1.96E-05	9.70E-04	4.37E-03	2.46E-03
S05051-01	2	2.91E-02	1.90E-05	9.52E-04	4.33E-03	2.62E-03
S05051-02	1	4.57E-02	2.27E-05	1.47E-03	6.92E-03	3.65E-03
S05051-03	1	2.63E-02	1.56E-05	8.27E-04	3.88E-03	<2.39E-03
S05051-04	1	3.48E-02	2.00E-05	1.12E-03	5.35E-03	3.44E-03
S05051-05	1	3.91E-02	3.82E-05	1.30E-03	5.80E-03	4.06E-03
S05051-06	1	3.91E-02	2.32E-05	1.25E-03	5.76E-03	3.84E-03
S05051-06	2	n/a	n/a	n/a	n/a	n/a
S05051-07	1	1.49E-02	2.79E-05	4.78E-04	2.30E-03	<1.70E-03
S05051-08	1	2.06E-02	3.46E-05	6.69E-04	3.20E-03	<2.13E-03
S05051-09	1	5.10E-02	3.50E-05	1.61E-03	7.54E-03	<5.26E-03
S05051-10	1	4.90E-02	1.72E-05	1.56E-03	7.41E-03	4.93E-03
Customer ID	Analysis	Pu239/240	U233	U234	U235	U236
		TRU-SPEC	ICP/MS Acid Add	ICP/MS Acid Add	ICP/MS Acid Add	ICP/MS Acid Add
		µCi/sample	µg/sample	µg/sample	µg/sample	µg/sample
S05051-01	1	2.07E-02	1.91E-05	5.97E-05	1.70E-03	1.35E-04
S05051-01	2	2.13E-02	2.18E-05	6.35E-05	1.76E-03	1.20E-04
S05051-02	1	3.23E-02	2.83E-05	1.03E-04	2.95E-03	2.29E-04
S05051-03	1	2.23E-02	1.74E-05	5.84E-05	1.54E-03	1.11E-04
S05051-04	1	3.04E-02	2.39E-05	7.96E-05	2.31E-03	1.62E-04
S05051-05	1	3.41E-02	2.06E-05	8.46E-05	2.08E-03	1.56E-04
S05051-06	1	3.21E-02	2.46E-05	7.34E-05	2.05E-03	1.60E-04
S05051-06	2	n/a	n/a	n/a	n/a	n/a
S05051-07	1	1.41E-02	1.03E-05	4.15E-05	1.35E-03	6.45E-05
S05051-08	1	1.83E-02	1.39E-05	5.02E-05	1.35E-03	9.93E-05
S05051-09	1	4.70E-02	3.04E-05	1.21E-04	3.30E-03	2.26E-04
S05051-10	1	4.25E-02	3.36E-05	1.15E-04	3.58E-03	2.15E-04

Table 1. Pre-Filter Sample Results (continued)

Customer ID	Analysis	U238	Cs137	Cs137		
		ICP/MS Acid Add	GEA	GEA		
		µg/sample	µCi/sample	µCi/sample		
S05051-01	1	1.50E-01	<6.14E-05	<6.14E-05		
S05051-01	2	1.57E-01	<6.24E-05	<6.24E-05		
S05051-02	1	2.46E-01	<6.39E-05	<6.39E-05		
S05051-03	1	1.36E-01	<6.39E-05	<6.39E-05		
S05051-04	1	1.92E-01	<6.51E-05	<6.51E-05		
S05051-05	1	1.71E-01	<6.35E-05	<6.35E-05		
S05051-06	1	1.76E-01	<6.35E-05	<6.35E-05		
S05051-06	2	n/a	n/a	n/a		
S05051-07	1	9.38E-02	<6.39E-05	<6.39E-05		
S05051-08	1	1.20E-01	<6.42E-05	<6.42E-05		
S05051-09	1	2.30E-01	<6.38E-05	<6.38E-05		
S05051-10	1	2.49E-01	<6.47E-05	<6.47E-05		

Table 2. HEPA Filter Sample Results

Customer ID	Analysis	Alpha Env	Am-241	Am-241	Beta in Env	Pu239
		Solid/Misc	TRU-SPEC	GEA	Samples	ICP/MS
		µCi/sample	µCi/sample	µCi/sample	µCi/sample	µg/sample
S05051-11	1	2.54E-02	7.30E-03	7.99E-03	2.15E-03	1.80E-01
S05051-11	2	2.53E-02	7.98E-03	6.36E-03	2.51E-03	1.86E-01
S05051-12	1	1.68E-02	5.41E-03	5.96E-03	1.49E-03	1.16E-01
S05051-13	1	1.71E-02	5.55E-03	5.44E-03	1.49E-03	1.14E-01
S05051-14	1	2.20E-02	8.07E-03	6.28E-03	1.82E-03	1.80E-01
S05051-15	1	2.46E-02	7.89E-03	8.74E-03	2.06E-03	1.73E-01
S05051-16	1	2.28E-02	7.88E-03	7.86E-03	2.14E-03	1.75E-01
S05051-16	2	2.33E-02	n/a	n/a	2.19E-03	n/a
S05051-17	1	2.16E-02	7.04E-03	7.36E-03	2.19E-03	1.57E-01
S05051-18	1	1.74E-02	5.91E-03	5.52E-03	1.53E-03	1.28E-01
S05051-19	1	2.21E-02	7.90E-03	8.40E-03	1.93E-03	1.70E-01
S05051-20	1	1.85E-02	6.83E-03	6.58E-03	4.48E-03	1.42E-01
Customer ID	Analysis	Pu240	Pu244/Cm244	Pu/Am242	Pu/Am241	Pu238
		ICP/MS	ICP/MS	ICP/MS	ICP/MS	TRU-SPEC
		µg/sample	µg/sample	µg/sample	µg/sample	µCi/sample
S05051-11	1	1.92E-02	<4.50E-05	5.54E-04	2.36E-03	<4.65E-03
S05051-11	2	1.96E-02	<2.50E-05	5.63E-04	2.47E-03	<3.56E-03
S05051-12	1	1.30E-02	<4.50E-05	4.56E-04	1.71E-03	<2.52E-03
S05051-13	1	1.28E-02	<4.50E-05	4.74E-04	1.58E-03	<2.68E-03
S05051-14	1	1.92E-02	<4.50E-05	6.20E-04	2.53E-03	<2.68E-03
S05051-15	1	1.85E-02	<4.50E-05	6.00E-04	2.41E-03	<3.15E-03
S05051-16	1	1.90E-02	<4.50E-05	6.50E-04	2.49E-03	<3.02E-03
S05051-16	2	n/a	n/a	n/a	n/a	n/a
S05051-17	1	1.73E-02	<4.50E-05	5.71E-04	2.26E-03	<4.49E-03
S05051-18	1	1.40E-02	<4.50E-05	4.74E-04	1.74E-03	<2.17E-03
S05051-19	1	1.84E-02	<4.50E-05	6.12E-04	2.40E-03	<2.15E-03
S05051-20	1	1.58E-02	<4.50E-05	5.26E-04	1.96E-03	<3.71E-03

Table 2. HEPA Filter Sample Results (continued)

Customer ID	Analysis	Pu239/240	U233	U234	U235	U236
		TRU-SPEC	ICP/MS Acid Add	ICP/MS Acid Add	ICP/MS Acid Add	ICP/MS Acid Add
		µCi/sample	µg/sample	µg/sample	µg/sample	µg/sample
S05051-11	1	1.62E-02	<4.50e-05	3.70E-05	1.24E-03	6.50E-05
S05051-11	2	1.74E-02	<4.50e-05	3.70E-05	1.30E-03	8.55E-05
S05051-12	1	1.06E-02	<4.50e-05	3.80E-05	9.95E-04	5.90E-05
S05051-13	1	1.07E-02	<4.50e-05	2.20E-05	1.15E-03	5.50E-05
S05051-14	1	1.59E-02	<4.50e-05	4.25E-05	1.36E-03	6.55E-05
S05051-15	1	1.50E-02	<4.50e-05	3.15E-05	1.76E-03	6.90E-05
S05051-16	1	1.54E-02	<4.50e-05	5.45E-05	1.29E-03	7.10E-05
S05051-16	2	n/a	n/a	n/a	n/a	n/a
S05051-17	1	1.40E-02	<4.50e-05	4.95E-05	1.60E-03	6.20E-05
S05051-18	1	1.14E-02	<4.50e-05	2.70E-05	1.16E-03	6.15E-05
S05051-19	1	1.46E-02	<4.50e-05	3.70E-05	1.44E-03	7.45E-05
S05051-20	1	1.30E-02	<4.50e-05	5.30E-05	1.39E-03	6.75E-05
Customer ID	Analysis	U238	Cs137	Sr89/90		
		ICP/MS Acid Add	GEA	Env. Misc		
		µg/sample	µCi/sample	µCi/sample		
S05051-11	1	1.04E-01	<3.95E-04	<4.93E-05		
S05051-11	2	1.07E-01	<3.56E-04	<3.18E-05		
S05051-12	1	8.68E-02	<3.71E-04	<3.20E-05		
S05051-13	1	8.30E-02	<3.87E-04	<5.07E-05		
S05051-14	1	1.30E-01	<3.72E-04	<5.50E-05		
S05051-15	1	1.22E-01	<3.87E-04	<3.77E-05		
S05051-16	1	1.08E-01	<3.44E-04	<3.32E-05		
S05051-16	2	n/a	n/a	n/a		
S05051-17	1	1.52E-01	<3.60E-04	<3.11E-05		
S05051-18	1	1.03E-01	<3.40E-04	<3.29E-05		
S05051-19	1	1.14E-01	<3.87E-04	<4.67E-05		
S05051-20	1	1.52E-01	<3.44E-04	<3.23E-05		




Figure 1. Total Alpha Analytical Results




Figure 2. Americium 241 Analytical Results




Figure 3. Total Beta Analytical Results




Figure 4. Plutonium-239/240 Analytical Results




Figure 5. Plutonium/Americium-241 Analytical Results




Figure 6. Plutonium/Americium-242 Analytical Results




Figure 7. Uranium-234 Analytical Results




Figure 8. Uranium-235 Analytical Results




Figure 9. Uranium-236 Analytical Results




Figure 10. Uranium-238 Analytical Results

Table 3. Mean Concentrations and the Associated Variability

Analyte	Units	Mean Concentration		Random Uncertainty RSD associated with the mean (%)	
		Pre-Filter	HEPA Filter	Pre-Filter	HEPA Filter
Alpha Env: Solid/Misc (Each)	μCi/sample	4.44E-02	2.09E-02	10.8	4.8
Am-241 by TRU-SPEC Resin IonEx	μCi/sample	1.48E-02	7.01E-03	10.5	4.7
Americium-241 by GEA	μCi/sample	1.42E-02	6.93E-03	10.9	5.3
Beta in Env. Samples (Each)	μCi/sample	3.05E-03	2.15E-03	11.3	12.9
Cesium-137 by GEA	μCi/sample	<6.51E-05 ^a	<3.95E-04 ^a	NA	NA
Plutonium-239 by ICP/MS	μg/sample	3.26E-01	1.54E-01	10.9	5.5
Plutonium-240 by ICP/MS	μg/sample	3.50E-02	1.67E-02	10.9	5.0
Pu 244/Cm 244 by ICP/MS	μg/sample	2.54E-05	<4.50E-05 ^a	10.1	NA
Pu/Am 242 ICP/MS	μg/sample	1.12E-03	5.54E-04	10.8	3.9
Pu/Am-241 by ICP/MS	μg/sample	5.25E-03	2.15E-03	10.8	5.4
Pu-238 by TRU-SPEC Resin IonEx	μCi/sample	3.74E-03 ^b	<4.65E-03	8.5	NA
Pu-239/240 by TRU-SPEC Resin	μCi/sample	2.94E-02	1.37E-02	11.3	5.1
Pu239+Pu240 by ICP/MS by addition	μCi/sample	2.83E-02	1.34E-02	10.9	5.3
Sr-89/90 Env. Misc. (ea.)	μCi/sample	1.32E-04 ^c	<5.50E-05 ^a	45.0	NA
Uranium-233 by ICP/MS Acid Add	μg/sample	2.24E-05	<4.50E-05 ^a	10.4	NA
Uranium-234 by ICP/MS Acid Add	μg/sample	7.88E-05	3.92E-05	10.9	8.7
Uranium-235 by ICP/MS Acid Add	μg/sample	2.22E-03	1.34E-03	11.4	5.3
Uranium-236 by ICP/MS Acid Add	μg/sample	1.55E-04	6.60E-05	11.4	3.2
Uranium-238 by ICP/MS Acid Add	μg/sample	1.77E-01	1.16E-01	9.5	6.5

a the largest “less than” value

b “less than” values deleted from the statistical analysis

c only result that was not a “less than” value; RSD is from the counting statistics

NA not available with the stated data

Table 4. Pre-filter versus HEPA filter Mean Concentrations

Analyte	Units	Mean Concentration		HEPA mean/ Pre-filter mean	
		Pre-Filter	HEPA Filter	All Data	Exclude Total Beta
Alpha Env: Solid/Misc (Each)	μCi/sample	4.44E-02	2.09E-02	0.470	0.470
Am-241 by TRU-SPEC Resin IonEx	μCi/sample	1.48E-02	7.01E-03	0.473	0.473
Americium-241 by GEA	μCi/sample	1.42E-02	6.93E-03	0.489	0.489
Beta in Env. Samples (Each)	μCi/sample	3.05E-03	2.15E-03	0.705	---
Plutonium-239 by ICP/MS	μg/sample	3.26E-01	1.54E-01	0.471	0.471
Plutonium-240 by ICP/MS	μg/sample	3.50E-02	1.67E-02	0.478	0.478
Pu/Am 242 ICP/MS	μg/sample	1.12E-03	5.54E-04	0.493	0.493
Pu/Am-241 by ICP/MS	μg/sample	5.25E-03	2.15E-03	0.409	0.409
Pu-239/240 by TRU-SPEC Resin	μCi/sample	2.94E-02	1.37E-02	0.467	0.467
Pu239+Pu240 by ICP/MS by addition	μCi/sample	2.83E-02	1.34E-02	0.473	0.473
Uranium-234 by ICP/MS Acid Add	ug/sample	7.88E-05	3.92E-05	0.497	0.497
Uranium-235 by ICP/MS Acid Add	ug/sample	2.22E-03	1.34E-03	0.603	0.603
Uranium-236 by ICP/MS Acid Add	ug/sample	1.55E-04	6.60E-05	0.426	0.426
Uranium-238 by ICP/MS Acid Add	ug/sample	1.77E-01	1.16E-01	0.654	0.654
Mean				0.508	0.493
Standard Deviation				0.085	0.066
Standard Deviation of the Mean				0.023	0.018

Table 5. Final Mean Concentrations and the Associated Variability

Analyte	Units	Mean Concentration		Random Uncertainty RSD associated with the mean (%)	
		Pre-Filter	HEPA Filter	Pre-Filter	HEPA Filter
Alpha Env: Solid/Misc (Each)	μCi/sample	4.44E-02	2.09E-02	10.8	4.8
Am-241 by TRU-SPEC Resin IonEx	μCi/sample	1.48E-02	7.01E-03	10.5	4.7
Americium-241 by GEA	μCi/sample	1.42E-02	6.93E-03	10.9	5.3
Beta in Env. Samples (Each)	μCi/sample	3.05E-03	2.15E-03	11.3	12.9
Cesium-137 by GEA	μCi/sample	6.51E-05 ^a	3.26E-05 ^b	50.0 ^c	50.2
Plutonium-239 by ICP/MS	μg/sample	3.26E-01	1.54E-01	10.9	5.5
Plutonium-240 by ICP/MS	μg/sample	3.50E-02	1.67E-02	10.9	5.0
Pu 244/Cm 244 by ICP/MS	μg/sample	2.54E-05	1.27E-05 ^b	10.1	11.0
Pu/Am 242 ICP/MS	μg/sample	1.12E-03	5.54E-04	10.8	3.9
Pu/Am-241 by ICP/MS	μg/sample	5.25E-03	2.15E-03	10.8	5.4
Pu-238 by TRU-SPEC Resin IonEx	μCi/sample	3.74E-03 ^d	1.87E-03 ^b	8.5	9.6
Pu-239/240 by TRU-SPEC Resin	μCi/sample	2.94E-02	1.37E-02	11.3	5.1
Pu239+Pu240 by ICP/MS by addition	μCi/sample	2.83E-02	1.34E-02	10.9	5.3
Sr-89/90 Env. Misc. (ea.)	μCi/sample	1.32E-04 ^e	6.60E-05 ^b	45.0	45.2
Uranium-233 by ICP/MS Acid Add	μg/sample	2.24E-05	1.12E-05 ^b	10.4	11.3
Uranium-234 by ICP/MS Acid Add	μg/sample	7.88E-05	3.92E-05	10.9	8.7
Uranium-235 by ICP/MS Acid Add	μg/sample	2.22E-03	1.34E-03	11.4	5.3
Uranium-236 by ICP/MS Acid Add	μg/sample	1.55E-04	6.60E-05	11.4	3.2
Uranium-238 by ICP/MS Acid Add	μg/sample	1.77E-01	1.16E-01	9.5	6.5

a the upper value of the largest “less than” (e.g., 6.51E-05 for <6.51E-05)

b assumes 50% of the Pre-filter result

c assumption; the less than implies that the result could be anywhere below the stated number. using 50% as 1 RSD and 100% as 2 RSD allows the result to be as low as zero

d “less than” values were deleted from the statistical analysis (6 samples instead of 10)

e only result that was not a “less than” value; RSD is from the counting statistics

Table 6. Systematic and Total Measurement Uncertainty Estimates

Analyte	Systematic Uncertainty	Total Measurement Uncertainty associated with the mean (%)	
		Pre-Filter	HEPA Filter
Alpha Env: Solid/Misc (Each)	8	13.4	9.3
Am-241 by TRU-SPEC Resin IonEx	9	13.8	10.2
Americium-241 by GEA	8	13.5	9.6
Beta in Env. Samples (Each)	9	14.4	15.7
Cesium-137 by GEA	8	50.6 ^a	50.8
Plutonium-239 by ICP/MS	3	11.3	6.3
Plutonium-240 by ICP/MS	3	11.3	5.8
Pu 244/Cm 244 by ICP/MS	3	10.5	11.4
Pu/Am 242 ICP/MS	3	11.2	4.9
Pu/Am-241 by ICP/MS	9	14.0	10.5
Pu-238 by TRU-SPEC Resin IonEx	7	11.0	11.9
Pu-239/240 by TRU-SPEC Resin	7	13.3	8.7
Pu239+Pu240 by ICP/MS by addition	3	11.3	6.1
Sr-89/90 Env. Misc. (ea.)	5	45.3	45.5
Uranium-233 by ICP/MS Acid Add	14	17.4	18.0
Uranium-234 by ICP/MS Acid Add	14	17.8	16.5
Uranium-235 by ICP/MS Acid Add	14	18.1	15.0
Uranium-236 by ICP/MS Acid Add	14	18.1	14.4
Uranium-238 by ICP/MS Acid Add	8	12.4	10.3

a assumes a random uncertainty of 50%

b “less than” values were deleted from the statistical analysis (6 samples instead of 10)

c only result that was not a “less than” value; RSD is from the counting statistics

Distribution

No. of Copies		No. of Copies	
OFFSITE		ONSITE	
3 Fluor Hanford		2 Pacific Northwest National Laboratory	
D.L. Dyekman	H8-13	B.A. Pulsipher	K6-08
G. LeBaron	S2-42	TL Welsh	K6-08
M. Wright	S2-42		