PNNL-27432 RPT-DFTP-006, Rev. 0



Proudly Operated by Battelle Since 1965

Multi-Cycle Cesium Ion Exchange Testing Using Spherical Resorcinol-Formaldehyde Resin with Diluted Hanford Tank Waste 241-AP-105

April 2018

SK Fiskum JR Allred HA Colburn AM Rovira MR Smoot RA Peterson



Prepared for the U.S. Department of Energy under Contract DE-AC05-76RL01830

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor Battelle Memorial Institute, nor any of their employees, makes **any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or Battelle Memorial Institute. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.**

PACIFIC NORTHWEST NATIONAL LABORATORY operated by BATTELLE for the UNITED STATES DEPARTMENT OF ENERGY under Contract DE-AC05-76RL01830

Printed in the United States of America

Available to DOE and DOE contractors from the Office of Scientific and Technical Information, P.O. Box 62, Oak Ridge, TN 37831-0062; ph: (865) 576-8401 fax: (865) 576-5728 email: reports@adonis.osti.gov

Available to the public from the National Technical Information Service 5301 Shawnee Rd., Alexandria, VA 22312 ph: (800) 553-NTIS (6847) email: <u>orders@ntis.gov</u> <http://www.ntis.gov/about/form.aspx> Online ordering: http://www.ntis.gov



Multi-Cycle Cesium Ion Exchange Testing Using Spherical Resorcinol-Formaldehyde Resin with Diluted Hanford Tank Waste 241-AP-105

SK Fiskum JR Allred HA Colburn AM Rovira MR Smoot RA Peterson

April 2018

Prepared for the U.S. Department of Energy under Contract DE-AC05-76RL01830

Pacific Northwest National Laboratory Richland, Washington 99352

Executive Summary

Washington River Protection Solutions requested that at least 4 L of Hanford tank waste collected from tank 241-AP-105 and diluted to approximately 5.6 M Na (AP-105DF) be prepared for use in vitrification studies. Cesium removal was required to meet this objective and the waste pretreatment platform, established at Pacific Northwest National Laboratory, was used to create the vitrification feed. The Test Platform mimicked the Low-Activity Waste Pretreatment System (LAWPS) facility planned to pretreat Hanford tank waste supernate by removing solids in a cross flow filter apparatus and processing the supernate through ion exchange columns to remove cesium. At the time of testing, the ion exchange media was targeted to be spherical resorcinol-formaldehyde (SRF) resin.¹ The SRF resin has been tested with a wide array of simulants and process scales, but column performance testing with actual tank waste had been somewhat limited to two Hanford tank wastes (AP-101 and AN-102) in up to two process cycles. This report describes testing conducted in a total of six load elute cycles with the AP-105DF tank waste. The column system was first tested with simulant and was previously described^{2,3}; this report describes the six process cycle results with the AP-105DF tank waste.

Column testing was conducted on SRF resin provided by Microbeads AS (Skedsmokorset, Norway, batch number 1F-370/1392), which was manufactured in August 2011. The column testing was prototypic to the intended LAWPS operations in a lead-lag column format, albeit on a small-scale basis with 10-mL resin beds. In this configuration, neither the length-to-diameter ratio nor the superficial flow velocity matched the full-scale design. In this process, the feed was directed downflow through the lead column and then through the lag column. Loading continued until the lag column reached 10% of the Hanford Tank Waste Treatment and Immobilization Plant contract limit for receiving supernatant waste for vitrification (a function of the Na and ¹³⁷Cs concentrations). As a result of this process condition, the lead column was nearly saturated with Cs. After loading, the feed was displaced with 0.1 M NaOH, and then the columns were rinsed with water in a downflow lead-to-lag configuration. Elution was conducted downflow from the lag to the lead column with 0.45 M HNO₃ followed by a water rinse. The resin was returned to the Na-form by processing 1 M NaOH downflow from the lag to the lead column, which deviated from the intended LAWPS process operation where the regeneration would occur in upflow to fluidize the resin beds. Variations to the feed flowrates, elution volumes, and elution flowrates were implemented to evaluate effects on the Cs load behavior and Cs leakage to the next process cycle.

Cs load and elution profiles were generated. From the load profile, the number of bed volumes (BVs) processed to reach 50% breakthrough on the lead column was determined along with the number of BVs processed before reaching 10% of the contract limit on the lag column. For AP-105DF, the 10% contract limit is set to ~0.015% of the influent ¹³⁷Cs concentration; this requires a Cs decontamination factor of 6620. Table ES.1 summarizes the observed column performance for the six process cycles.

¹ The LAWPS system approach had been modified subsequent to the testing described in this report; test parameters were current at the testing time.

² Fiskum SK, HA Colburn, MR Smoot, JR Allred, RA Peterson. 2017. *Cesium Ion Exchange Using Spherical Resorcinol-Formaldehyde Resin in Support of Waste Qualification Testing for LAWPS*. PNNL-26837, RPT-DFTP-003, Rev. 0.0. Pacific Northwest National Laboratory, Richland, Washington.

³ The first cycle AP-105 actual waste testing was also described by Fiskum et al. (2017).

Test Number	Flowrate, BV/h	Elution Volume, BVs	50% Cs Breakthrough, BVs	Mass Transfer Zone, BVs ^(a)	Cs Load Capacity, mg Cs/g H-form Resin	Contract Limit Breakthrough, BVs
1	1.80	15.9	206	95	6.64	275
2	3.05	29.9	205	124	6.38	245
3	4.53	23.0	189	149	6.09	189
4 (Take 1)	3.04	24.9	NA	NA	NA	NA
4 (Take 2)	2.95	29.5	182	112	6.06	193
5	3.17	29.1	181	128	5.53	202

 Table ES.1.
 Column Performance Summary

(a) The mass transfer zone was the range between 1% and 90% C/C $_0$ breakthrough.

NA = not applicable; Test 4 Take 1 processing was short, with only 52.9 BVs.

One BV equals 9.9 mL.

Cs leakage from the lag column occurred from one process run to the next. Increased elution beyond 16 BVs was apparently required to keep the product effluent below the 10% contract limit. The leakage appeared to reduce as more feed was processed until the expected Cs breakthrough profile was established.

The repeated cycling resulted in significant resin swelling. The initial Na-form resin beds were 9.9 mL; after the final process cycle, the Na-form resin BVs were 10.9 mL (lead column) and 12.1 mL (lag column), expansions of 10% and 20%, respectively. Volume expansions of resin in the H-form were 106% and 112%, respectively.

The effluent and eluate were characterized to assess distribution of cations, anions, and radionuclides during the ion exchange process. In addition to Cs, small fractions of Fe, Ca, Cu, Pb, and Zn were measured in the eluate; all other cations and anions were quantitatively recovered in the Cs-decontaminated effluent. Within 10% analytical uncertainty, all ⁹⁰Sr, Pu, and Am reported to the Cs-decontaminated effluent. However, 66% of the Cm and 78% of the ⁹⁹Tc were found in the effluent; less than 4% was found in the eluate, indicating the difference may be held by the resin.

Acknowledgments

The authors thank the Shielded Facilities Operations staff LaWanda Grow, Jarrod Turner, Michael Rojas, Jeff Chenault, Robert Cox, and Jordan Condray for in-cell sample handling and systems manipulations. We thank the Analytical Support Operations (ASO) count room staff Truc Trang-Le, Bruce Pierson, and Mike Cantaloub for rapid ¹³⁷Cs analysis for the column load/elution sample analysis. Special thanks to Bruce Pierson for support in calculating the dose to which the SRF was exposed. We thank the ASO staff Karl Pool, Lori Darnell, Crystal Rutherford, Garrett Brown, Chuck Soderquist, and Deborah Coffey for the sample analysis processing and reporting. We thank Renee Russell and Bruce Pierson for reviewing the many ion exchange calculation files and test data packages. We thank Renee Russell for reviewing this technical report. We also thank Matt Wilburn for technical editing of this report.

Acronyms and Abbreviations

AP-105DF	Hanford tank waste from 241-AP-105, diluted to 5.7 M sodium					
ASO	Analytical Support Operations					
ASR	Analytical Service Request					
AV	apparatus volume					
BV	bed volume					
DF	decontamination factor					
DI	deionized (water)					
FD	feed displacement					
GEA	gamma energy analysis					
IC	ion chromatography					
ICP-MS	inductively coupled plasma mass spectrometry					
ICP-OES	inductively coupled plasma optical emission spectrometry					
ID	identification (number) or internal diameter, see context					
KPA	kinetic phosphorescence analysis					
LAW	low-activity waste					
LAWPS	Low-Activity Waste Pretreatment System					
MDL	method detection limit					
NA	not applicable					
PNNL	Pacific Northwest National Laboratory					
QA	quality assurance					
R&D	research and development					
SRF	spherical resorcinol-formaldehyde					
TIC	total inorganic carbon					
TOC	total organic carbon					
WRPS	Washington River Protection Solutions					
WTP	Hanford Tank Waste Treatment and Immobilization Plant					
WWFTP	WRPS Waste Form Testing Program					

Contents

Exec	cutive	e Summary	iii
Ack	nowle	edgments	v
Acro	onym	s and Abbreviations	vi
1.0	Intro	oduction	. 1.1
	1.1	Quality Assurance	. 1.2
2.0	Test	Conditions	. 2.1
	2.1	SRF Resin	. 2.1
	2.2	AP-105 Tank Waste	. 2.1
	2.3	Ion Exchange Process Testing	. 2.3
		2.3.1 Test 1 Process Notes	. 2.8
		2.3.2 Test 2 Process Notes	. 2.8
		2.3.3 Test 3 Process Notes	. 2.8
		2.3.4 Test 4 Process Notes	. 2.8
		2.3.5 Test 5 Process Notes	. 2.9
	2.4	Sample Analysis	2.10
		2.4.1 Cesium Load and Elution Sample Analysis	2.10
		2.4.2 Composite Feed, Effluent, and Eluate Sample Analysis	2.10
3.0	Colu	umn Test Results with AP-105DF	. 3.1
	3.1	Cs Load, Feed Displacement, and Water Rinse Results	. 3.1
	3.2	Cesium Elution Results	. 3.5
	3.3	Activity Balance	. 3.6
	3.4	Ion Exchange Column Absorbed Dose Calculation	. 3.9
	3.5	Chemical and Radiochemical Composition	3.19
4.0	Disc	cussion	. 4.1
	4.1	Cs Load Capacity	. 4.1
	4.2	Column Performance	. 4.2
		4.2.1 50% Breakthrough	. 4.2
		4.2.2 Mass Transfer Zone	. 4.2
		4.2.3 Cs Leakage	. 4.4
5.0	Resi	in Bed Physical Properties	. 5.1
6.0	Mul	ti-Cycle Testing with AP-105DF Conclusions	. 6.1
7.0	Refe	erences	. 7.1
App	endix	A Column Load and Elution Data	A.1
App	endix	B Analytical Reports	. B .1

Figures

2.1. Cesium Ion Exchange Process Schematic Showing Downflow Lead-to-Lag Processing	2.3
2.2. Column Assembly in the Hot Cell	2.4
2.3. Gaps in Lag Column during Test 4 Take 1 Elution Processing	2.9
3.1. Test 1 Cesium Load Curve for AP-101DF, 1.80 BV/h	3.1
3.2. Test 2 Cesium Load Curve for AP-101DF, 3.05 BV/h	3.2
3.3. Test 3 Cesium Load Curve for AP-101DF, 4.53 BV/h	3.2
3.4. Test 4 Take 1 Cesium Load Curve for AP-101DF, 3.04 BV/h	3.3
3.5. Test 4 Take 2 Cesium Load Curve for AP-101DF, 2.95 BV/h to 162 BVs, then 3.60 BV/h	3.3
3.6. Test 5 Cesium Load Curve for AP-101DF, 3.17 BV/h	3.4
3.7. Elution Profiles for Tests 1-3	3.6
3.8. Test 1 Column Absorbed Gamma Dose as a Function of Time	3.12
3.9. Test 1 Column Absorbed Beta Dose as a Function of Time	3.12
3.10. Test 2 Column Absorbed Gamma Dose as a Function of Time	3.13
3.11. Test 2 Column Absorbed Beta Dose as a Function of Time	3.13
3.12. Test 3 Column Absorbed Gamma Dose as a Function of Time	3.14
3.13. Test 3 Column Absorbed Beta Dose as a Function of Time	3.14
3.14. Test 4 Take 1 Column Absorbed Gamma Dose as a Function of Time	3.15
3.15. Test 4 Take 1 Column Absorbed Beta Dose as a Function of Time	3.15
3.16. Test 4 Take 2 Column Absorbed Gamma Dose as a Function of Time	3.16
3.17. Test 4 Take 2 Column Absorbed Beta Dose as a Function of Time	3.16
3.18. Test 5 Column Absorbed Gamma Dose as a Function of Time	3.17
3.19. Test 5 Column Absorbed Beta Dose as a Function of Time	3.17
3.20. Column Absorbed Dose as a Function of Time during Tests 1-5 (including background from hot cell)	3 18
4.1 Cesium Breakthrough Curves as a Function of Flowrate on SRF Resin	43
4.2 Lead Column Cesium Breakthrough Profiles at Constant Flowrate	4 4
4.3 Comparison of Lag Column Cs Breakthrough	4 5
4.4 Final Water Rinse Sample Cs Concentration as a Function of the Elution BVs Processed	113
Tests 0 to 5	4.6
4.5. Initial Cs Leakage versus Eluent BVs Processed in Previous Cycle	4.7
5.1. Resin Bed Expansion and Contraction History	5.2

Tables

2.1.	Characterization of Sample 5AP-16-01 Collected from Hanford Tank AP-105 (ASR 0272)	2.2
2.2.	AP-105DF Cs Isotopic Composition (ASR 0329)	2.3
2.3.	Test 1 Experimental Conditions for AP-105DF Column Processing, July 10-21, 2017	2.5
2.4.	Test 2 Experimental Conditions for AP-105DF Column Processing, July 31-August 5, 2017	2.5
2.5.	Test 3 Experimental Conditions for AP-105DF Column Processing, August 21-25, 2017	2.6
2.6.	Test 4 (Take 1) Experimental Conditions for AP-105DF Column Processing, September 12- 15, 2017	2.6
2.7.	Test 4 (Take 2) Experimental Conditions for AP-105DF Column Processing, September 18- 22, 2017	2.7
2.8.	Test 5 Experimental Conditions for AP-105DF Column Processing, October 9-13, 2017	2.7
2.9.	Analytical Scope	. 2.11
3.1.	AP-105DF Effluent Composites ¹³⁷ Cs Content and Decontamination Factor	3.5
3.2.	¹³⁷ Cs Activity Balance for AP-105DF Test 1	3.7
3.3.	¹³⁷ Cs Activity Balance for AP-105DF Test 2	3.7
3.4.	¹³⁷ Cs Activity Balance for AP-105DF Test 3	3.8
3.5.	¹³⁷ Cs Activity Balance for AP-105DF Test 4, Take 1	3.8
3.6.	¹³⁷ Cs Activity Balance for AP-105DF Test 4, Take 2	3.9
3.7.	¹³⁷ Cs Activity Balance for AP-105DF Test 5	3.9
3.8.	Absorbed Dose Summary for Tests1-5	. 3.19
3.9.	AP-105DF Feed, Effluent, and Eluate Compositions (Test 1) ASR 0335, Radionuclides	. 3.20
3.10	. AP-105DF Feed, Effluent, and Eluate Compositions (Test 1) ASR 0335, Inorganic and	
	Carbon Analytes	. 3.21
3.11	. Test 2 and Test 3 Analytical Results Summary (ASR 372)	. 3.23
3.12	. Test 4 Take 1 and Test 4 Take 2 Analytical Results Summary (ASR 372)	. 3.24
3.13	. Test 5 Analytical Results Summary (ASR 372)	. 3.25
4.1.	Cesium Capacity as a Function of Process Cycle	4.1
4.2.	50% Breakthrough as a Function of Process Cycle	4.2
4.3.	Mass Transfer Zone (1% to 90% C/C ₀)	4.3
4.4.	Lag Column Initial Cs Leakage	4.6

1.0 Introduction

Decanted tank waste supernatant will be pretreated in the Low-Activity Waste Pretreatment System (LAWPS) to meet the Hanford Tank Waste Treatment and Immobilization Plant (WTP) Low-Activity Waste (LAW) facility waste acceptance criteria in preparation for vitrification.¹ Washington River Protection Solutions (WRPS) is designing the LAWPS facility. The key process operations for treating the tank waste supernatants include solids filtration and cesium removal. At the time of this work (July 2017), the LAWPS design specified solids removal by cross flow filtration and cesium removal with spherical resorcinol-formaldehyde (SRF) ion exchange resin. Testing of actual tank waste processing on SRF resin has been conducted (Fiskum et al. 2006b, c; Duignan and Nash 2009) where up to two process cycles were employed. Duignan and Nash (2009) processed Savannah River National Laboratory tank waste up to 100% breakthrough in the lead column, whereas Fiskum et al. (2006a, b) processed the lead column to 50% breakthrough in the mode in which WTP plans to process tank waste.

Supernate waste delivered to the WTP LAW facility is required to be <3.18E-5 Ci ¹³⁷Cs/mole of Na (contract limit). In the LAWPS design at the time of this work, the SRF resin was to be loaded in two columns, ~300-gallon resin beds each, arranged in a lead-lag format.² Feed was to be processed until the lag column effluent reached 10% of the contract limit (i.e., 3.18E-6 Ci ¹³⁷Cs/mole of Na). The 10% contract limit was implemented by WRPS to manage lag column Cs loading and concomitant Cs leakage to the next process run. The Cs-decontaminated effluent was then to be forwarded to the WTP for vitrification.

The key functional parameters of the LAWPS and WTP LAW facilities were adapted for small-scale testing at Pacific Northwest National Laboratory (PNNL) and the designated Test Platform. The cesium ion exchange component of the Test Platform was constructed using small-scale ion exchange columns (each column contained 9.9 mL Na-form SRF resin in a 1.43-cm-diameter column). Other attributes of the ion exchange system were scaled to be generally prototypic of the LAWPS system at the time of this work. The ion exchange system was previously described by Fiskum et al. (2017).

Approximately 8 L of Hanford tank waste were collected from tank 241-AP-105 (hereafter called AP-105) to process through the Test Platform. WRPS requested that the tank waste be diluted to a target of 5.6 M Na and then filtered, and that cesium be removed to support vitrification testing. To complete this effort, 10 L of diluted AP-105 (AP-105DF) were processed through the ion exchange system, requiring a total of six load-elute cycles.

This report discusses results of multi-cycle testing using SRF resin with AP-105DF. Cesium capacity as a function of process cycle; cesium leakage into subsequent runs; and radionuclide, cation, and anion mass balances (with special emphasis on Cs, Pu, Na, K, Tc, transition metals) were determined.

¹ 24590-WTP-ICD-MG-01-030, Rev. 0. 2015. *ICD 30 – Interface Control Document for Direct LAW Feed*. Bechtel National, Inc., Richland, Washington.

² After processing described herein, WRPS staff are considering alternative processing and ion exchange media.

1.1 Quality Assurance

The work described in this report was conducted with funding from WRPS contract 36437/212, *DFLAW Radioactive Waste Test Platform*. This contract was managed under PNNL Project 69832. All research and development (R&D) work at PNNL is performed in accordance with PNNL's Laboratory-Level Quality Management Program to R&D activities, which is based on a graded application of NQA-1-2000, *Quality Assurance Requirements for Nuclear Facility Applications* (ASME 2000). To ensure that all client quality assurance (QA) expectations were addressed, the QA controls of the PNNL's WRPS Waste Form Testing Program (WWFTP) QA program were also implemented for this work. The WWFTP QA program implements the requirements of NQA-1-2008, *Quality Assurance Requirements for Nuclear Facility Applications* (ASME 2008) and NQA-1a-2009, *Addenda to ASME NQA-1-2008* (ASME 2009), and consists of the *WWFTP Quality Assurance Plan* (QA-WWFTP-001) and associated QA-NSLWnumbered procedures that provide detailed instructions for implementing NQA-1 requirements for R&D work.

The work described in this report was assigned the technology level "Applied Research" and was planned, performed, documented, and reported in accordance with procedure QA-NSLW-1102, *Scientific Investigation for Applied Research*. All staff members contributing to the work received proper technical and QA training prior to performing quality-affecting work.

2.0 Test Conditions

This section describes the SRF resin, diluted AP-105 tank waste, and column ion exchange conditions. All testing was conducted in accordance with a test plan prepared by PNNL and approved by WRPS.¹

2.1 SRF Resin

The SRF ion exchange resin (Spheromers[®] RF 380) was provided by Microbeads AS, Skedsmokorset, Norway, batch number 1F-370/1392, and was manufactured in August 2011. A full description of the resin retrieval and pretreatment was previously provided by Fiskum et al. (2017).

The SRF resin was pretreated according to established protocol.² The SRF resin loaded in the column assembly was exposed to a process cycle before use with the AP-105DF as previously described (Fiskum et al. 2017). This process cycle used a simple simulant (Russell et al. 2017) spiked with 60 μ g/mL Cs and tracer 0.1 μ Ci/mL¹³⁷Cs.

2.2 AP-105 Tank Waste

As previously reported (Fiskum et al. 2017), multiple samples (32 each at nominally 250 mL) were collected at four different depths (91 in., 172 in., 253 in., and 334 in.) from the AP-105 Hanford tank. The first sample collected, 5AP-16-01, was subsampled for a limited analysis suite. The density was measured in cell using a 10-mL volumetric flask. All other measurements were conducted by the Analytical Support Operations (ASO) according to Analytical Service Request (ASR) 0272; results are provided in Table 2.1.

¹ TP-DFTP-001, Rev.0.2. *DFLAW Test Platform Cesium Ion Exchange Testing with AP-105 Tank Waste with Spherical Resorcinol-Formaldehyde Resin.* Pacific Northwest National Laboratory, Richland Washington. 2017. ² WTP doc. no. 097893. CA Nash and CE Duffey. August 17, 2004. *Hanford RPP-WTP Alternate Resin Program - Protocol P1-RF: Spherical Resin Sampling from Containers, Resin Pretreatment, F-Factor, and Resin Loading to Column.* SRNL-RPP-2004-00058, Savannah River Technology Center, Aiken, South Carolina.

Analyte	Result	Result Units	Analysis Method
Al	1.02	М	ICP-OES
K	0.141	М	ICP-OES
Na	8.53	М	ICP-OES
OH-	1.75 ^(a)	М	Titration
¹³³ Cs	5.7E-5	М	ICP-MS
¹³⁷ Cs	$180\pm2\%^{(b)}$	µCi/mL	GEA
¹³⁷ Cs	2.07 ^(b)	μg/mL	GEA
Density	1.405 ^(c)	g/mL	Volumetric flask

 Table 2.1.
 Characterization of Sample 5AP-16-01
 Collected from Hanford Tank AP-105 (ASR 0272)

(a) Based on first inflection point; assumed to be the free (unbound) hydroxide.

(b) Reference date is 4/26/17.

(c) Measured at 27.5 °C.

ASR 0272, sample 17-0868

GEA = gamma energy analysis; ICP-MS = inductively coupled plasma mass spectrometry;

ICP-OES = inductively coupled plasma optical emission spectrometry

Samples of the AP-105 tank waste were combined and then diluted to ~5.7 M Na with 0.01 M NaOH before filtration through the cross flow filter system (Geeting et al. 2017). The sample/diluent volume in the filtration system was limited to 4 L; therefore, multiple compositing/diluting batches were prepared. The diluted and filtered AP-105 tank waste is termed AP-105DF throughout this report. The AP-105DF was provided in multiple ~1-L increments for ion exchange. Each container of material was measured for density using 10-mL volumetric flasks or the Coriolis meter installed in the cross flow filter apparatus. Densities ranged from 1.25 to 1.30 g/mL. Due to the large volume (12 L) of AP-105DF, the multiple samples were not combined into one single container for homogenization and dilution. Doing so builds risk into the process (should a container leak), and handling the massive composite would be problematic given the weight-lifting limitations of the manipulators in the hot cells.

Because the feed solutions were to be kept in separate containers, it was desired to make the density (and by inference all other chemical/physical properties) the same. To that end, samples with high density were diluted or combined with samples with low density, thus normalizing all feed to around 1.27 g/mL.

The Cs isotopic composition of the AP-105DF was determined on a peak elution sample aliquot from the first ion exchange process run. It was assumed that all samples had the same cesium isotopic distribution; Cs isotopic distribution is shown in Table 2.2. The Cs isotopic ratio was measured by ICP-MS per ASR 0329. The advantage of using the elution sample for the isotopic ratio determination is that it is relatively free from the AP-105DF high salt matrix and contains the highest Cs concentration, and thus is less likely to be affected by isobaric interferences. The total Cs concentration was calculated from the measured ¹³³Cs and ¹³⁷Cs and the isotopic composition.

Analyte ^(a)	Analysis Method	Results	Units			
		64.0	wt% ¹³³ Cs			
Cs isotopic mass ratio ^(a,b,c)	ICP-MS	19.1	wt% ¹³⁵ Cs			
		16.8	wt% ¹³⁷ Cs			
Total Cs	ICP-MS	7.92	μg/mL Cs			
(a) The peak Cs column eluated	(a) The peak Cs column eluate sample (TI014-E8-A) from the first AP-105DF ion exchange					
process cycle was analyzed	process cycle was analyzed for the Cs isotopic mass distribution by ICP-MS per ASR 0329					
sample 17-1224. The quantity of hold-over Cs from the shakedown testing was assumed to						
result in negligible ¹³³ Cs contribution compared to the Cs eluted from AP-105DF processing.						
(b) Reference date is August 9, 2017.						
(c) 134 Cs, a fission product, was not detected by GEA; with a 2.065 year half-life, it was assumed						

 Table 2.2.
 AP-105DF Cs Isotopic Composition (ASR 0329)

2.3 Ion Exchange Process Testing

to be decayed to extinction.

The ion exchange process system has been previously described (Fiskum et al. 2017); a system schematic for processing downflow lead column to lag column is reproduced in Figure 2.1. The quick disconnect valves were realigned to alter the fluid flow path downflow from lag column to lead column. Lead

column samples were collected at valve 2 and lag column samples were collected from valve 3 during the AP-105DF loading process as well as the feed displacement (FD), water rinse, elution, and elution rinse (Tests 1-3). For Tests 4-5, the FD, water rinse, and elution were collected at the effluent line, bypassing the valve 3 sample position; the elution rinse was collected from the valve 3 sample position. Tests 4-5 collected regeneration solution at the valve 3 sample position in discrete ~15-mL volumes. The valve 3 sample position was thus well rinsed before processing AP-105DF in a subsequent cycle, whereas the valve 2 sample position was not well rinsed.



Figure 2.1. Cesium Ion Exchange Process Schematic Showing Downflow Lead-to-Lag Processing

Fiskum et al. (2017) also described the out-of-column and in-column SRF resin pretreatment steps. As a reminder, the bed volume (BV) corresponds to the initial settled Na-form resin BV as measured in a

graduated cylinder prior to transferring the resin into the ion exchange column. The reference resin BVs are 9.9 mL for both the lead and lag columns.

The initial process cycle with 5.6 M Simple Simulant traced with ¹³⁷Cs was conducted in a radiologically controlled fume hood as previously described (Fiskum et al. 2017). The system was transferred to the hot cell for AP-105DF processing. A photograph of the in-cell system is shown in Figure 2.2.



Figure 2.2. Column Assembly in the Hot Cell

The AP-105DF was processed through the ion exchange resin beds, lead to lag according to PNNL test instructions. A series of AP-105DF 1.5-L feed bottles were strategically processed to allow optimal feed volume management and support unattended, off-shift (graveyard) work. Effluent was collected in two to three different bottles to better manage the consequences of Cs breakthrough from the lag column. After the AP-105DF load step, 0.1 M NaOH FD, followed by water rinse, was passed through the system in the same lead-to-lag configuration. The flow orientation was switched by rearranging the quick disconnect connections as previously described (Fiskum et al. 2017), allowing elution to occur downflow from the lag column. Once the lead column resin bed was fully converted based on visual examination of the bottom resin bed color, the flow direction was again changed back to the lead-to-lag direction.

All processing was conducted at ambient cell temperature conditions, nominally 26 to 30 °C. Test parameters, including process volumes, flowrates, and contact times, are summarized in Table 2.3 to Table 2.8. The AP-105DF flowrate and 0.45 M HNO₃ eluent volume and flowrate were adjusted from one test to the next to test process condition effects. The first AP-105DF process cycle mimics as best as possible the process flows anticipated at the LAWPS facility in terms of BV/h and total BVs. It is understood that the feed linear velocity cannot be matched in this small column configuration. As previously reported (Fiskum et al. 2017), increasing the linear velocity decreases the transition zone and

sharpens the breakthrough curve. Therefore, the load curves developed from the small-scale system are likely worst-case bounding with respect to the transition zone.

		Total Volume			Flo	wrate	Duration
Process Step	Solution	BV	AV	mL	BV/h	mL/min	h
Regeneration	1.0 M NaOH	9.6	2.4	95	4.77	0.789	2.0
Loading (lead)	AP-105DF	294	NA	2921	1.80	0.298	239 ^(b)
Loading (lag) ^(a)	AP-105DF	274	NA	2724	1.80	0.298	239 ^(b)
Feed displacement	0.1 M NaOH	6.65	1.7	66.1	3.05	0.505	2.2
Water rinse	DI water	4.17	1.04	41.4	3.08	0.511	1.4
Elution	0.45 M HNO ₃	15.9	3.98	158	1.40	0.231	11.7
Water rinse	DI water	5.79	1.44	57.5	1.47	0.243	4.0
Extended water rinse	DI water	8.74	2.18	86.8	3.14	0.520	2.8

 Table 2.3.
 Test 1 Experimental Conditions for AP-105DF Column Processing, July 10-21, 2017

(a) The feed volume through the lag column is reduced because of sampling from the lead column.

(b) Time includes the standby time over the weekend; see Section 2.3.1.

BV = bed volume (9.9 mL in the Na-form as loaded in the column).

AV = apparatus volume (nominally 40 mL).

NA = not applicable.

Processing was conducted according to PNNL test instruction TI-DFTP-014.

Table 2.4. Test 2 Ex	perimental Conditions	for AP-105DF Column	Processing, July	y 31-August 5, 2017
----------------------	-----------------------	---------------------	------------------	---------------------

		Total Volume		Flowrate		Duration	
Process Step	Solution	BV	AV	mL	BV/h	mL/min	h
Regeneration	1.0 M NaOH	7.4	1.85	73.9	2.94	0.486	2.5
Loading (lead)	AP-105DF	254	NA	2523	3.05	0.506	84
Loading (lag) ^(a)	AP-105DF	251	NA	2489	3.05	0.506	84
Feed displacement	0.1 M NaOH	6.4	0.79	63.6	3.00	0.497	2.2
Water rinse	DI water	4.7	1.18	47.1	2.94	0.486	1.7
Elution	0.45 M HNO ₃	29.9	NA	297.0	1.98	0.327	15.6
Water rinse	DI water	5.6	NA	55.7	1.99	0.330	2.8
Extended water rinse	DI water	8.7	NA	86.4	3.07	0.508	2.8

(a) The feed volume through the lag column is reduced because of sampling from the lead column.

BV = bed volume (9.9 mL in the Na-form as loaded in the column).

AV = apparatus volume (nominally 40 mL).

NA = not applicable.

Processing was conducted according to PNNL test instruction TI-DFTP-015.

		Total Volume			Flowrate		Duration
Process Step	Solution	BV	AV	mL	BV/h	mL/min	h
Regeneration	1.0 M NaOH	8.0	2.01	79.9	2.84	0.470	2.80
Loading (lead)	AP-105DF	248	NA	2459	4.53	0.746	56
Loading (lag) ^(a)	AP-105DF	241	NA	2393	4.53	0.746	56
Feed displacement	0.1 M NaOH	6.3	1.57	62.5	3.08	0.510	2.1
Water rinse	DI water	4.2	1.04	41.3	3.02	0.500	1.4
Elution	0.45 M HNO ₃	23.0	NA	228.8	1.92	0.318	12.25
Water rinse	DI water	6.3	NA	62.5	2.13	0.352	3.1
Extended water rinse	DI water	6.3	NA	89.5	2.86	0.474	3.1

Table 2.5. Test 3 Experimental Conditions for AP-105DF Column Processing, August 21-25, 2017

(a) The feed volume through the lag column is reduced because of sampling from the lead column.

BV = bed volume (9.9 mL in the Na-form as loaded in the column).

AV = apparatus volume (nominally 40 mL).

NA = not applicable.

Processing was conducted according to PNNL test instruction TI-DFTP-016.

Table 2.6. Test 4 (Take 1) Experimental Conditions for AP-105DF Column Processing,

September 12	2-15, 2017
--------------	------------

		Total Volume			Flowrate		Duration
Process Step	Solution	BV	AV	mL	BV/h	mL/min	h
Regeneration	1.0 M NaOH	9.4	2.4	93.6	2.76	0.456	3.4225
Loading (lead)	AP-105DF	53	NA	526	3.04	0.503	19.2 ^(b)
Loading (lag) ^(a)	AP-105DF	52	NA	517	3.04	0.503	19.2 ^(b)
Feed displacement	0.1 M NaOH	4.6	1.1	45.5	1.96	0.325	2.3
Water rinse	DI water	3.7	0.93	36.9	2.82	0.467	1.3
Elution ^(c)	0.45 M HNO ₃	24.9	NA	247.5	2.31	0.382	10.8
Extended water rinse	DI water	12.8	NA	127.0	2.80	0.463	4.5

(a) The feed volume through the lag column is reduced because of sampling from the lead column.

(b) The columns remained partially loaded and in contact with AP-105DF for ~27 hours; see Section 2.3.4.

(c) Erratic flow conditions occurred during elution; see Section 2.3.4.

BV = bed volume (9.9 mL in the Na-form as loaded in the column).

AV = apparatus volume (nominally 40 mL).

NA = not applicable.

Processing was conducted according to PNNL test instruction TI-DFTP-020.

		Т	`otal Volun	ne	Flo	owrate	Duration
Process Step	Solution	BV	AV	mL	BV/h	mL/min	h
Regeneration	1.0 M NaOH	8.3	2.06	82.1	2.66	0.441	3.1
Loading (lead) ^(a,b)	AP-105DF	238	NA	2370	2.95/3.6	0.489/0.597	77
Loading (lag) ^(a)	AP-105DF	235	NA	2332	2.95/3.6	0.489/0.597	77
Feed displacement	0.1 M NaOH	5.7	1.43	56.8	3.27	0.541	1.8
Water rinse	DI water	5.2	1.30	51.6	3.18	0.527	1.6
Elution	0.45 M HNO ₃	29.5	NA	292.7	2.74	0.453	10.8
Water rinse	DI water	8.1	NA	80.5	3.04	0.503	2.8
Extended water rinse	DI water	9.8	NA	97.1	3.04	0.503	3.2

Table 2.7. Test 4 (Take 2) Experimental Conditions for AP-105DF Column Processing,
September 18-22, 2017

(a) An overcorrection in the flowrate was implemented after processing 162 BVs through the lead column (159 BVs through the lag column).

(b) The feed volume through the lag column is reduced because of sampling from the lead column.

BV = bed volume (9.9 mL in the Na-form as loaded in the column).

AV = apparatus volume (nominally 40 mL).

NA = not applicable.

Processing was conducted according to PNNL test instruction TI-DFTP-020.

Table 2.8 .	Test 5 Experimental	Conditions for	AP-105DF	Column	Processing,	October 9-13, 20	17
--------------------	---------------------	----------------	----------	--------	-------------	------------------	----

		Total Volume		Flowrate		Duration	
Process Step	Solution	BV	AV	mL	BV/h	mL/min	h
Regeneration	1.0 M NaOH	9.1	2.41	90.8	3.09	0.512	3.0
Loading (lead)	AP-105DF	218	NA	2168	3.17	0.524	75 ^(b)
Loading (lag) ^(a)	AP-105DF	213	NA	2117	3.17	0.524	75 ^(b)
Feed displacement	0.1 M NaOH	7.8	2.05	77.2	4.40	0.728	1.8
Water rinse	DI water	5.1	1.34	50.4	4.12	0.682	1.2
Elution	0.45 M HNO3	29.1	NA	289.4	2.56	0.424	11.4
Water rinse	DI water	7.9	NA	78.9	3.10	0.514	2.7
Extended water rinse	DI water	8.6	NA	85.3	3.05	0.505	2.8

(a) The feed volume through the lag column is reduced because of sampling from the lead column.

(b) Time includes the 6-h standby; see Section 2.3.5.

BV = bed volume (9.9 mL in the Na-form as loaded in the column).

AV = apparatus volume (nominally 40 mL).

NA = not applicable.

Processing was conducted according to PNNL test instruction TI-DFTP-021.

During the loading phase, nominal 2-mL samples were collected from both the lead and lag columns at the sample collection ports. The solution in the lag column remained static during the lead column sampling time of about 4 min. Samples were collected after the first ~4 BVs were processed and again at nominal 10- to 20-BV increments. Feed displacement, water rinse, and elution were collected sequentially in nominal 1.1-BV increments (Tests 1-3). Bulk collection of the FD and water rinse and bulk collection

of the eluate were performed in Tests 4-5. The water rinse following elution was collected in nominal 1.1-BV increments. Finally, the bulk water rinse was conducted to better clear the system from acidic matrix, implemented to accommodate the resin idle time between tests and reduce exposure to residual acidic fluid. Aliquots of each solution were removed for GEA.

Cesium load and elution performance was determined from the ¹³⁷Cs in the collected samples relative to the native ¹³⁷Cs in AP-105DF feed. The collected samples were analyzed directly to determine the ¹³⁷Cs concentration using GEA (constant reference date of July 10, 2017). Cesium breakthrough and elution curves were generated as previously described (Fiskum et al. 2017).

Some unexpected issues occurred during processing. These are described in the following sections.

2.3.1 Test 1 Process Notes

After processing ~144 BVs of feed, one polyethylene line between the lead and lag columns broke. The break occurred off-shift and was not found until the next morning, after an estimated 147 mL of feed was calculated to have leaked from the system (passing through the lead column but not through the lag column). The broken line was replaced; the air in the replacement line and hardware displaced 3.7 mL of the fluid above the lag column. The replacement line also broke after processing another ~12 BVs of AP-105DF. At this point, the system (partially loaded) was placed in standby mode (all valves closed) over the weekend (Friday, July 14, at 5:00 p.m. until the following Monday, July 17, at 1:30 p.m.). The broken line was replaced with stainless steel tubing before resuming feed processing. The resin was in contact with AP-105DF for 239 hours, including the weekend standby period. Refer to Fiskum et al. 2017.

2.3.2 Test 2 Process Notes

Regeneration with 1 M NaOH was extended beyond the nominal 6 BV. Conversion of the lead column resin bed from H-form (orange) to Na-form (black) was slow (visually observed color change). It was clear that some channeling was occurring in the conversion process.

2.3.3 Test 3 Process Notes

Samples (15-mL) were collected from the lead column to support technetium studies (separate study) at 22, 200, and 247 BVs. The volume of AP-105DF that was processed through the lag column was proportionately lower.

2.3.4 Test 4 Process Notes

The AP-105DF feed ceased flowing after processing 52.9 BVs. The system pump head had failed. The pump was replaced along with the associated process tubing. The lead and lag columns were in static contact with the AP-105DF for 27 h while the pump was replaced. Once the pump was repaired, the AP-105DF feed process was aborted and FD, water rinse, elution, and water rinse were processed through the columns. During elution, some leakages were observed in the tubing connections. Fluid above the lead column resin bed dropped very close to the top of the bed. Corrective measures were implemented (tightened Swagelok connections, added fluid to the top of the resin bed from the column top access port).

A rapid fluid drop (5.7 mL) occurred through the lead column after establishing the fluid bed height. The knurled nut at the access port was retightened to correct this. After corrective measures, small gaps were observed at 1.0 and 1.5 cm from the bottom of the lag column resin bed, indicating that a vacuum pull from the top of the bed or upflow fluid movement through the bottom of the bed occurred. Processing continued, and the gap, as shown in Figure 2.3, was eventually no longer visible.



Figure 2.3. Gaps in Lag Column during Test 4 Take 1 Elution Processing

The test was restarted from the regeneration process step. The initial (short) Test 4 processing was termed "Take 1." The subsequent processing was termed "Take 2." After processing 162 BVs through the lead column (159 BVs through the lag column), an attempt was made to increase the AP-105DF flowrate from 0.489 to 0.50 BV/h. The correction was too high and the flowrate jumped to ~0.60 BV/h until the end of the feed processing.

2.3.5 Test 5 Process Notes

The remaining AP-105DF (2.05 L) was composited into a single, 2-L, high-density polyethylene bottle, which simplified the feed loading (no switching from one feed bottle to another feed bottle). After processing 195 BVs, the feed tube apparently was above the feed liquid and the lead column fluid headspace was displaced with air. The resin bed did not go dry, but some air contact with the top few millimeters of resin bed was likely. The feed processing was suspended for 6 h.

Lead column samples from previous process runs containing significant ¹³⁷Cs activity had been combined to form a 150-mL volume composite in a smaller, low-density polyethylene bottle (where the feed line position was more easily distinguished). This composite solution was used to re-establish the fluid height above the lead column to the more optimum full BV above the resin bed. Because the AP-105DF comprising this solution had been previously processed through the SRF column, it was depleted in Cs (27.6 μ Ci/mL ¹³⁷Cs). A 108-mL aliquot of this depleted solution was processed (combined head space adjustment and feed flow) before it was combined with the remaining AP-105DF. This final mixture (150 mL) contained an estimated 99 μ Ci/mL ¹³⁷Cs. The fluid processed after 195 BVs were fed through the lead column (190 BVs through the lag column) contained variable Cs concentrations, and examinations of the load curve will need to take this into account.

2.4 Sample Analysis

A summary of the sample collections and analyses from the various tests and process steps is provided in this section along with the cross references to ASR and Radiochemical Processing Laboratory sample identifications (IDs).

2.4.1 Cesium Load and Elution Sample Analysis

Aliquots of the various process samples' (regeneration, feed, effluent, FD, water rinse, elution, and water rinse) ¹³⁷Cs concentrations were determined by the ASO on calibrated gamma detectors. To support this analysis, all samples and sub-samples were collected and packaged in 10-mL and/or 2-mL volumes to accommodate the calibrated detector geometries. Several samples of feed material were provided in 1-mL counting geometries to better accommodate handling of the high sample activity; these were counted far away from the detector face, where geometry differences become insignificant.

Exact effluent volumes were calculated from the measured net sample mass and the known solution densities. Densities were measured using volumetric flasks and net weight or by measured mass of a known pipetted volume. Because of the high dose rate from ¹³⁷Cs, many eluate samples required dilution before removal from the hot cell. The peak eluate samples were diluted by a factor of about 10,000x in 0.45 M HNO₃. The GEA count times were adjusted to accommodate the specific sample ¹³⁷Cs concentration to target $\leq 1\%$ count uncertainty.

2.4.2 Composite Feed, Effluent, and Eluate Sample Analysis

The composite feed, effluent, and elution samples were submitted for chemical and radiochemical analyses to the ASO. The Test 1 samples underwent extensive analysis; the Tests 2-5 samples were subject to a more limited suite of analytes. The analyte concentrations are expected to duplicate from one test to the next and extensive analyses were not considered critical to all tests. An aliquot of the peak elution sample from Test 1 only was submitted for ICP-MS to determine the Cs isotopic distribution. The Cs isotopic distribution is expected to be constant for all samples. Table 2.9 provides the analysis summary for Tests 1-5 inclusive of the cross references to the project sample ID, ASO ASRs, ASO sample IDs, and analysis scope.

All analyses were conducted by the ASO according to standard operating procedures, the ASO QA Plan, and special instructions attached to the ASR. The ASO was responsible for the preparation and analysis of appropriate analytical batch and instrument quality control samples and to provide any additional processing to the sub-samples that might be required (e.g., acid digestion). Preparation by direct dilution (e.g., GEA) did not require preparative blanks and matrix spikes.

AP-105DF Test	Sample Description	Sample ID	ASR	ASO Sample ID	Analysis Scope
	Composite feed	TI014-FEED		17-1240	GEA, IC, ICP-MS, ICP-OES, total alpha/beta, U-KPA, ⁹⁰ Sr, ⁹⁹ Tc, Np, Pu, Am/Cm, TOC/TIC, free OH
1	Composite effluent	TI014-EFF Comp	0335	17-1241	GEA, IC, ICP-MS, ICP-OES, total alpha/beta, U-KPA, ⁹⁰ Sr, ⁹⁹ Tc, Np, Pu, Am/Cm, TOC/TIC, free OH
	Composite eluate	TI014-ELComp		17-1242	GEA, IC, ICP-MS, ICP-OES, total alpha/beta, U-KPA, ⁹⁰ Sr, ⁹⁹ Tc, Np, Pu, Am/Cm, TOC/TIC, H ^{+(a)}
	Peak elution sample	TI014-E8-A	0329	17-1224	Cs isotopic
	Composite feed	TI015-FEED		18-0001	GEA, ICP-OES, Pu, 99Tc, U ICP-MS
2	Composite effluent	TI015-EFFComp		18-0002	GEA, ICP-OES, Pu, 99Tc, U ICP-MS
	Composite eluate	TI015-ELComp		18-0003	GEA, ICP-OES, Pu, ⁹⁹ Tc, U ICP-MS
	Composite feed	TI016-FEED		18-0004	GEA, ICP-OES, Pu, 99Tc, U ICP-MS
3	Composite effluent	TI016-EFFComp		18-0005	GEA, ICP-OES, Pu, ⁹⁹ Tc, U ICP-MS
	Composite eluate	TI016-ELComp		18-0006	GEA, ICP-OES, Pu, 99Tc, U ICP-MS
	Composite feed	TI020-FEED-1-A		18-0007	GEA, ICP-OES, Pu, 99Tc, U ICP-MS
4 Take 1	Composite effluent	TI020-EFFComp	0372	18-0008	GEA, ICP-OES, Pu, 99Tc, U ICP-MS
	Composite eluate	TI020-ELComp		18-0009	GEA, ICP-OES, Pu, 99Tc, U ICP-MS
	Composite feed	TI020-FEED-Take2		18-0010	GEA, ICP-OES, Pu, 99Tc, U ICP-MS
4 Take 2	Composite effluent	TI020-EFFComp-Take2		18-0011	GEA, ICP-OES, Pu, ⁹⁹ Tc, U ICP-MS
	Composite eluate	TI020-EComp		18-0012	GEA, ICP-OES, Pu, 99Tc, U ICP-MS
	Composite feed	TI021-FEED		18-0013	GEA, ICP-OES, Pu, 99Tc, U ICP-MS
5	Composite effluent	TI021-EFFComp		18-0014	GEA, ICP-OES, Pu, 99Tc, U ICP-MS
	Composite eluate	TI021-EComp		18-0015	GEA, ICP-OES, Pu, 99Tc, U ICP-MS

 Table 2.9.
 Analytical Scope

(a) The large sample dilution rendered the H⁺ analysis too inaccurate for meaningful evaluation.

IC = ion chromatography; KPA = kinetic phosphorescence analysis; TIC = total inorganic carbon; TOC = total organic carbon.

3.0 Column Test Results with AP-105DF

The Cs load and elution behavior was evaluated on all AP-105DF tank waste process cycles. This section discusses the load and elution results for all tests. Raw data are provided in Appendix A.

3.1 Cs Load, Feed Displacement, and Water Rinse Results

The cesium load profiles for the Tests 1-5 are shown in Figure 3.1 through Figure 3.6. Each figure is shown with the ordinate $%C/C_0$ on a probability scale versus the abscissa BV on a linear scale. The probability scale has two advantages over a linear scale and log scale: 1) Cs breakthrough profile appears linear, allowing for backward and forward extrapolations, and 2) more detail can be observed at the onset of breakthrough and the high breakthrough. Also provided are the FD and water rinses following AP-105DF loading for Tests 1-3; FD and water rinses were not collected as individual samples for Tests 4-5. Also shown is the 10% contract limit as nominally 0.015 % C/C₀.¹ Tests 4-5 incorporate the regeneration effluent solution %C/C₀ concentrations that were collected in discrete ~15-mL increments.



Figure 3.1. Test 1 Cesium Load Curve for AP-101DF, 1.80 BV/h

¹ The contract limit was derived from the allowed curies of ¹³⁷Cs per mole of Na in the effluent to support contact handling of the final vitrified waste form—3.18E-5 Ci ¹³⁷Cs/mole Na. At 5.7 M Na and 120 μ Ci ¹³⁷Cs/mL in the feed, the contract limit is 1.5E-3 C/C₀; 10% of this value is 0.015% C/C₀.



Figure 3.2. Test 2 Cesium Load Curve for AP-101DF, 3.05 BV/h



Figure 3.3. Test 3 Cesium Load Curve for AP-101DF, 4.53 BV/h



Figure 3.4. Test 4 Take 1 Cesium Load Curve for AP-101DF, 3.04 BV/h



Figure 3.5. Test 4 Take 2 Cesium Load Curve for AP-101DF, 2.95 BV/h to 162 BVs, then 3.60 BV/h



Figure 3.6. Test 5 Cesium Load Curve for AP-101DF, 3.17 BV/h

The first three BVs of FD solution continued the Cs breakthrough profile, as was expected because AP-105DF was still in the system. Starting with the fourth BV of FD, the Cs concentration in the effluent started to drop. The Cs concentration in the effluent continued to drop as water flow started through the system. Thus, the resin was shown to continue holding the Cs well as the fluid Cs, Na, and hydroxide concentrations decreased.

The regeneration solution showed a marked drop in Cs concentration as a function of BV processed. A similar trend can be observed for the first sample collected from the lead column. This is likely associated with a combination of two factors: 1) the rinsing of the lead column sample port from the last sample collected from the previous run and 2) some Cs leakage that developed over the interim standby period between process runs.

Table 3.1 provides the Cs-decontaminated effluent composite results in terms of ¹³⁷Cs concentration and overall decontamination factor (DF). A DF of 6620 was needed to meet the 10% contract limit. Three of the effluent composites exceeded this threshold and required batch contact processing to remove additional Cs in support of follow-on vitrification work (not reported herein).

Test	Effluent Container	¹³⁷ Cs (µCi/mL)	Decontamination Factor
1	Effluent-1 (0-159 BVs)	7.07E-5	1.58E+06
	Effluent-2 (160-294 BVs)	7.25E-4	1.54E+05
2	Effluent-1 (0-154 BVs)	5.53E-3	2.14E+04
	Effluent-2 (154-254 BVs)	4.06E-3	2.91E+04
3	Effluent-1 (0-128 BVs)	2.21E-3	5.69E+04
	Effluent-2 (129-248 BVs)	7.75E-2	1.62E+03
4 Take1	Effluent-1 (0-53 BVs)	3.52E-3	3.45E+04
4 Take 2	Effluent-1 (0-144 BVs)	7.26E-4	1.68E+05
	Effluent-2 (144-238 BVs)	1.05E-1	1.16E+03
5	Effluent-1 (0-99 BVs)	3.87E-3	3.23E+04
	Effluent-2 (99-183BVs)	7.12E-3	1.75E+04
	Effluent-3 (183-218 BVs)	2.12E-2	5.90E+03
Bolded effluen	ts exceeded the 6620 DF (10% of	the waste accepta	nce criteria).

 Table 3.1. AP-105DF Effluent Composites ¹³⁷Cs Content and Decontamination Factor

3.2 Cesium Elution Results

The elution profiles for Tests 1-3 are provided in Figure 3.7. (Tests 4-5 did not assess elution profiles.) The 0.45 M HNO₃ solution flow is shown in solid symbols; the water rinse following elution is shown in open symbols. There was no substantial difference in the elution profiles for the two flowrates tested (1.4 and 2.0 BV/h). Peak Cs removal occurred between 10 and 13 BVs. Tailing was similar. Tailing is an important attribute for the next process cycle as residual Cs on the lag column can greatly influence the subsequent load cycle effluent DF. Although 99+% of Cs removal can be attained, the residual Cs can contaminate the next cycle product.



Figure 3.7. Elution Profiles for Tests 1-3

3.3 Activity Balance

The ¹³⁷Cs fractionation was determined between the effluents (collected in two to three different collection bottles), samples collected during the load processing, FD, water rinse, elution, and the final water rinses. The ¹³⁷Cs fractionation in the final extended water rinses was conducted on samples from Tests 4-5. The microcuries of ¹³⁷Cs loaded onto the lead and lag columns were calculated. Table 3.2 through Table 3.7 summarize the ¹³⁷Cs fractionations found in the various effluents as well as the Cs column loading for each process test. Overall sample handling and analytical uncertainty were estimated at \pm 3-5% and \pm 2%, respectively. Therefore, a Cs activity balance within 95% to 105% was considered excellent recovery.

Input	μCi	%
Feed Sample	328,834	100
Output		
Effluent-1 (0-159 BVs)	0.108	3.29E-05
Effluent-2 (159-294 BVs)	0.938	2.85E-04
Load samples	1740	0.529
Loss (spill)	271	0.082
Feed displacement	2.16	6.58E-04
Water rinse	0.0775	2.36E-05
Elution	351,558	107
Water rinse	4.40	1.34E-03
Total ¹³⁷ Cs recovery	353,577	108
Total ¹³⁷ Cs Column Loading		
Lead column	229,270	69.7
Lag column	97,552	29.7

 Table 3.2.
 ¹³⁷Cs Activity Balance for AP-105DF Test 1

Table 3.3.¹³⁷Cs Activity Balance for AP-105DF Test 2

Input	μCi	%
Feed Sample	294,285	100
Output		
Effluent-1 (0-154 BVs)	8.32	2.83E-03
Effluent-2 (154-254 BVs)	3.83	1.30E-03
Load samples	911	0.310
Feed displacement	1.944	6.61E-04
Water rinse	0.163	5.55E-05
Elution	313,584	107
Water rinse	0.958	3.25E-04
Total ¹³⁷ Cs recovery	314,510	107
Total ¹³⁷ Cs Column Loading		
Lead column	233,663	79.4
Lag column	59,699	20.3

Input	μCi	%
Feed Sample	312,539	100
Output		
Effluent-1 (0-128 BVs)	2.76	8.84E-04
Effluent-2 (128-248 BVs)	85.4	2.73E-02
Load samples	3775	1.21
Feed displacement	37.2	1.19E-02
Water rinse	1.89	6.05E-04
Elution	307,268	98.3
Water rinse	2.74	8.76E-04
Total ¹³⁷ Cs recovery	311,173	100
Total ¹³⁷ Cs Column Loading		
Lead column	236,739	75.7
Lag column	71,936	23.0

 Table 3.4.
 ¹³⁷Cs Activity Balance for AP-105DF Test 3

 Table 3.5.
 ¹³⁷Cs Activity Balance for AP-105DF Test 4, Take 1

Input	μCi	%
Feed Sample	66,520	100
Output		
Effluent-1 (0-53 BVs)	1.78	2.67E-03
Load samples	0.20	3.02E-04
Elution	68,063	102
Total ¹³⁷ Cs recovery	68,067	102
Total ¹³⁷ Cs Column Loading		
Lead column	66,515	100
Lag column	3.14	0.00008

Input	μCi	%
Feed Sample	294,050	100
Output		
Effluent-1 (0-144 BVs)	1.01	3.45E-04
Effluent-2 (144-238 BVs)	95.8	3.26E-02
Load samples	1286	0.437
Feed displacement and water rinse	0.5	1.81E-04
Elution	302,697	103
Water rinse	15.2	5.16E-03
Extended water rinse	3.1	1.04E-03
Total ¹³⁷ Cs recovery	304,091	103
Total ¹³⁷ Cs Column Loading		
Lead column	229,186	77.9
Lag column	64,194	21.8

 Table 3.6.
 ¹³⁷Cs Activity Balance for AP-105DF Test 4, Take 2

 Table 3.7.
 ¹³⁷Cs Activity Balance for AP-105DF Test 5

Input	μCi	%
Feed Sample	259,960	100
Output		
Effluent-1 (0-99 BVs)	3.64	1.40E-03
Effluent-2 (99-183BVs)	5.67	2.18E-03
Effluent-3 (183-218 BVs)	6.97	2.68E-03
Load samples	1942	0.747
Feed displacement and water rinse	3.8	1.46E-03
Elution	258,380	99.4
Water rinse	15.4	5.94E-03
Extended water rinse	4.1	1.56E-03
Total ¹³⁷ Cs recovery	260,357	100
Total ¹³⁷ Cs Column Loading		
Lead column	213,597	82.2
Lag column	44,867	17.3

3.4 Ion Exchange Column Absorbed Dose Calculation

The isotope ¹³⁷Cs decays to the metastable state of ¹³⁷Ba with the emission of a beta particle that has an average energy of 169.6 keV, and the metastable state de-excites to the ground state with the emission of a 662-keV photon. As the ¹³⁷Cs loads onto the resin bed, the resin is exposed to increasing decay energy (beta and gamma source terms). The accumulated dose to the SRF resin was determined from the

calculated ¹³⁷Cs load and was assumed to be distributed over the entire volume of the resin bed. More realistically, the Cs load is more concentrated at the top of the bed.

Absorbed dose was calculated on a time step basis using the grab samples taken at intervals while the ion exchange system was operated. This involved determining a volumetric ¹³⁷Cs activity concentration, C_{ν} , in each column after each grab sample. This concentration was derived from the GEA data for each of these grab samples. To first determine the gamma ray contribution to the column's absorbed dose, an exposure rate based on the volumetric geometry of the column was calculated using a volume source formula from Stabin (2007):

Exposure rate
$$\left[\frac{C}{kg h}\right] = \dot{X} = \pi \Gamma\left(\frac{C_v}{\mu}\right) (1 - e^{-\mu t}) \ln\left(\frac{r^2 + h^2}{h^2}\right)$$
 (3.1)

where Γ is the ¹³⁷Cs gamma constant 0.33 $\frac{R*m^2}{hr*Ci}$ (Saenger et al. 1972); μ is the linear attenuation coefficient of the material (assumed to be water) based on ICRU Report 44 (ICRU 1989):

$$\mu = \frac{\mu}{\rho_m} * \rho \tag{3.2}$$

where $\frac{\mu}{\rho_m}$ is the mass attenuation coefficient; ρ is the density of the material; t is the thickness of the source (ID of column 1.43 cm); r is the radius of the source $\left(\frac{t}{2}\right)$; h is the height of the source (6.3 cm).

Exposure rate was converted to exposure based on the duration of the time step. The column's exposure is converted to absorbed dose via Stabin (2007):

$$D [rad] = \frac{d\varepsilon}{dm} = 0.88 X \frac{\left(\frac{\mu_{en}}{\rho}\right)_{material}}{\left(\frac{\mu_{en}}{\rho}\right)_{air}}$$
(3.3)

where $d\varepsilon$ is the mean energy imparted by ionizing radiation to matter in a volume element of mass dm; X is the exposure (R); $\left(\frac{\mu_{en}}{\rho}\right)_{material}$ is the mass energy absorption coefficient for the specified material at the photon energy of interest (assumed to be water, from ICRU Report 44 [ICRU 1989]); while $\left(\frac{\mu_{en}}{\rho}\right)_{air}$ is the mass energy absorption coefficient for dry air at the photon energy of interest (ICRU 1989).

The mass energy absorption coefficient for the 662 keV was interpolated from a log-log plot using the following linear interpolation formula:

$$\frac{[ln(Y) - ln(Y_1)]}{[ln(X) - ln(X_1)]} = \frac{[ln(Y_2) - ln(Y_1)]}{[ln(X_2) - ln(X_1)]}$$
(3.4)

where Y is the desired mass energy absorption coefficient component; X is the gamma energy that Y is being solved for; and $(X_1, Y_1), (X_2, Y_2)$ are the known data points.

Solving for Y leads to:

$$Y = \exp\left(\frac{[ln(X) - ln(X_1)] * [ln(Y_2) - ln(Y_1)]}{[ln(X_2) - ln(X_1)]} + ln(Y_1)\right)$$
(3.5)

For the beta dose contribution, it is assumed that all beta particle energy from the ¹³⁷Cs is deposited into the resin material (Slaback and Schlein 1998):

$$\dot{D}\left[\frac{rad}{s}\right] = (1.6E - 8) A_c \bar{E}$$

$$\bar{E} \left(\beta^{-}\right) \approx \frac{1}{3} E_{max}$$
(3.6)

where \overline{E} is the average energy of the beta particle (MeV) (0.169 MeV for ¹³⁷Cs [Browne and Tuli 2007]); A_c is the radionuclide activity concentration in the source (Bq/g); the mass of the source was based on water at 30 °C for the volume of the source.

Figure 3.8 through Figure 3.19 show the cumulative column dose for each test as a function of time. In each of these figures, the last data point for Cs loading is marked with a solid vertical black line while the final data point for the FD data is noted by a vertical dashed black line. Column elution represents the data after the FD line.



Figure 3.8. Test 1 Column Absorbed Gamma Dose as a Function of Time



Figure 3.9. Test 1 Column Absorbed Beta Dose as a Function of Time



Figure 3.10. Test 2 Column Absorbed Gamma Dose as a Function of Time



Figure 3.11. Test 2 Column Absorbed Beta Dose as a Function of Time


Figure 3.12. Test 3 Column Absorbed Gamma Dose as a Function of Time



Figure 3.13. Test 3 Column Absorbed Beta Dose as a Function of Time



Figure 3.14. Test 4 Take 1 Column Absorbed Gamma Dose as a Function of Time



Figure 3.15. Test 4 Take 1 Column Absorbed Beta Dose as a Function of Time



Figure 3.16. Test 4 Take 2 Column Absorbed Gamma Dose as a Function of Time



Figure 3.17. Test 4 Take 2 Column Absorbed Beta Dose as a Function of Time



Figure 3.18. Test 5 Column Absorbed Gamma Dose as a Function of Time



Figure 3.19. Test 5 Column Absorbed Beta Dose as a Function of Time

An exposure rate measurement of the Shielded Analytical Laboratory hot cell was taken at the location of the ion exchange system with an MGP Instruments (Mirion) AMP-100 energy compensated GM-tube. This detector had a $\pm 10\%$ error associated with it and could measure from 0.5 mR/h to 1000 R/h. The background exposure rate was measured to be 14.5 R/h inside the hot cell in the location of the ion exchange system. Figure 3.20 displays this rate from the time that ion exchange system began Test 1 to the completion of Test 5.

This high background resulted in substantial dose contribution when compared to the internal column dose as a result of ¹³⁷Cs loading into the system while operating inside the hot cell. Table 3.8 summarizes the cumulative absorbed dose each column received at the end of each test due to the ¹³⁷Cs loading as well as the hot cell background radiation. The final cumulative absorbed dose the ion exchange system received from background was on the order of 2.26×10^5 rads while the final cumulative absorbed dose due to ¹³⁷Cs was 2.64×10^6 rads to the lead column and 4.92×10^5 rads to the lag column.



Figure 3.20. Column Absorbed Dose as a Function of Time during Tests 1-5 (including background from hot cell)

	Lead Column	Lag Column	Both Columns
	Cumulative Absorbed Dose to	Cumulative Absorbed Dose to	Cumulative Absorbed
	Column from ¹³⁷ Cs Loading,	Column from ¹³⁷ Cs Loading,	Dose to Columns Due
Test	rads ^(a)	rads ^(a)	to Background, rads ^(a)
1	1.04E+06	3.07E+05	8.94E+04
2	5.17E+05	5.82E+04	5.49E+04
3	6.61E+05	5.65E+04	3.51E+04
4 Take 1	9.42E+03	3.44E+00	3.70E+03
4 Take 2	3.87E+05	4.45E+04	2.45E+04
5	2.97E+04	2.57E+04	1.87E+04
(a) Cumula	tive column adsorbed dose from A	P-105DF processing at the end of e	ach test.

Table 3.8. Absorbed Dose Summary for Tests1-5

Fully loaded, full-scale columns at the LAWPS were modeled to see 300 MRad over the resin lifetime (i.e., 30 process cycles) or 10 MRad per process cycle. The gamma dose from the hot cell operations was estimated to be 2.87 MRad by the conclusion of the fifth process cycle. This was a substantial dose to SRF relative to a single process cycle at full scale, but it was a small fraction (1/100th) of the estimated full-scale resin lifetime dose. With the limited dose applied to this resin, extrapolation of synergistic effects from dose, chemical, and physical degradation mechanisms at full scale is difficult.

3.5 Chemical and Radiochemical Composition

The compositions of the AP-105DF ion exchange tests feeds, effluents, and eluates were evaluated to determine analyte fractionation. Test 1 feed, effluent, and eluate underwent extensive characterization to support follow-on work for vitrification and to understand analyte mass fractionation through the SRF ion exchange process. Table 3.9 summarizes the radioisotopic concentrations and fractionations and Table 3.10 summarizes the metals, anions, inorganic and organic carbon concentrations, and fractionations from Test 1. The analyte fractionations were calculated as ratios of the total analyte measured in the feed processed through the column and the total analyte collected in the Cs-decontaminated effluent and the Cs-bearing eluate according to Eqs. (3.7) and (3.8).

$$\frac{C_{Da} \times V_D}{C_{Fa} \times V_F} = F_{Da} \tag{3.7}$$

$$\frac{C_{Ea} \times V_E}{C_{Fa} \times V_F} = F_{Ea} \tag{3.8}$$

where C_{Da} = concentration of analyte (a) in the Cs-decontaminated effluent

- V_D = volume of Cs-decontaminated effluent
- C_{Fa} = concentration of analyte (a) in the AP-105DF feed
- $V_F = volume of AP-105DF feed$
- F_{Da} = fraction of analyte (a) in the Cs-decontaminated effluent
- C_{Ea} = concentration of analyte (a) in the Cs-bearing eluate

 V_E = volume of Cs-bearing eluate

 F_{Ea} = fraction of analyte (a) in the Cs-bearing eluate

Analyses of the feed effluents and eluates for the remaining tests (Tests 2-5) were more limited in scope: GEA, ⁹⁹Tc, ²³⁹⁺²⁴⁰Pu, metals by ICP-OES, and U (ICP-MS). Ion exchange results for anions, free hydroxide, TIC, TOC, ⁹⁰Sr, and ²³⁷Np were expected to be equivalent to those found in Test 1, i.e., no exchange whatsoever. It was noted that the ²⁴³⁺²⁴⁴Cm results were indeterminate in Test 1; however, the Cm and Am chemistry should mirror each other and Cm activity recovery was not further followed. Table 3.11 through Table 3.13 provide results for Tests 2-5 along with the percent recovery in the effluent and the eluate for each test.

Some analyte results are shown in brackets; this indicates that the analytical result was less than the estimated quantitation limit but greater than the method detection limit (MDL) and the associated analytical uncertainty could be higher than $\pm 15\%$. The fractionation result was placed in brackets where it was calculated with one or more bracketed analytical value(s) to highlight the higher uncertainty. The opportunistic analyte results measured by ICP-OES are also shown in Table 3.10; these analytes are part of the data output but have not been fully evaluated for quality control performance.

Analysis		TI014-FEED	TI014-EFF- Comp	TI014- ELComp	Fraction in Effluent	Fraction in Eluate
Method	Analyte	$\mu Ci/mL^{(a)}$	$\mu Ci/mL^{(a)}$	$\mu Ci/mL^{(a)}$	%	%
	⁶⁰ Co	<3.2E-3	5.93E-4	<8.6E-2		
	¹²⁶ Sn/Sb	<1.1E-2	1.75E-5	<1.5E-1		
Gamma Energy	¹³⁷ Cs	1.22E+2	5.30E-4	2.43E+3	0.00043%	107%
Analysis (GEA)	¹⁵² Eu	<6.3E-3	6.07E-6	<3.3E-1		
	¹⁵⁴ Eu	<1.2E-2	1.35E-4	<2.4E-1		
	²⁴¹ Am	<5.4E-1	3.72E-4	<1.3E+1		
	²³⁷ Np	9.27E-6	8.57E-6	<mdl< td=""><td>92%</td><td></td></mdl<>	92%	
	²³⁸ Pu	1.93E-5	1.39E-5	1.39E-5	72%	3.9%
Separations/	²³⁹⁺²⁴⁰ Pu	5.18E-5	4.85E-5	1.97E-5	94%	2.0%
Alpha Energy Analysis (AEA)	²⁴¹ Am	2.53E-4	2.55E-4	5.92E-5	101%	1.3%
	²⁴² Cm	[8.11E-7]	[6.04E-7]	[3.87E-7]	[74%]	[2.6%]
	²⁴³⁺²⁴⁴ Cm	1.74E-5	1.15E-5	1.00E-5	66%	3.1%
Separations/	⁹⁰ Sr	5.78E-1	5.68E-1	6.79E-2	98%	0.63%
Beta Counting	⁹⁹ Tc	9.54E-2	7.47E-2	1.05E-3	78%	0.059%
Proportional	Total Alpha	<mdl< td=""><td><mdl< td=""><td><mdl< td=""><td></td><td></td></mdl<></td></mdl<></td></mdl<>	<mdl< td=""><td><mdl< td=""><td></td><td></td></mdl<></td></mdl<>	<mdl< td=""><td></td><td></td></mdl<>		
Counting	Total Beta	1.05E+2	1.40E+0	2.58E+3	1.3%	132%

Table 3.9. AP-105DF Feed, Effluent, and Eluate Compositions (Test 1) ASR 0335, Rad
--

(a) Reference date is August 25, 2017.

"--" = not applicable.

<MDL = less than method detection limit.

Values in brackets have high uncertainty.

			TI014-EFF-		Fraction in	Fraction in
Analysis		TI014-FEED	Comp	TI014-ELComp	Effluent	Eluate
Method	Analyte	µg/mL	µg/mL	µg/mL	%	%
ICP-MS	Hg	<8.1E-5	<8.1E-5	<8.1E-5	NA	NA
	Ag ^(a)	<1.8	<1.8	< 0.3		
	Al	14,550	14,500	50.4	100%	0.019%
	As	<98	<98	<16		
	В	82.1	[55]	[8.8]	[67%]	[0.58%]
	Ba	[0.41]	[0.36]	[0.46]	[87%]	[6.0%]
	Ca	[41]	[51]	24.3	[123%]	3.2%
	Cd	<1.7	<1.7	< 0.3		
	Cr	362	362	[2.8]	100%	[0.042%]
	Fe	[4.9]	[6.25]	3.35	[128%]	3.7%
	K	3975	3920	558	99%	0.75%
ICP-OES	Li	<1.8	<1.8	< 0.3		
	Na	143,000	140,000	4770	98%	0.18%
	Ni	[30]	[33]	[1.6]	[108%]	[0.29%]
	Р	[510]	[520]	10.5	[102%]	0.11%
	Pb	<25	<25	205		
	Se	<141	[200]	<23		
	Th	<7.4	<7.4	<1.2		
	Ti	<0.9	<0.9	<0.1		
	U (total)	<42	<42	<6.9		
	Zn	[7.2]	[11]	8.33	[147%]	[6.3%]
	Zr	<1.3	<1.3	<0.2	100%	0.019%
КРА	U (total)	4.70	4.39	3.71	93%	4.2%
	Cl-	4385	4355	NA	99%	
	NO_2^-	63,600	61,800	NA	97%	
	SO4 ²⁻	1800	1950	NA	108%	
IC.	$C_2O_4^{2-}$	195	178	NA	91%	
	NO ₃ -	113,500	112,500	NA	99%	
	PO4 ³⁻	1018	967	NA	95%	
Titration	Free Hydroxide	1.05 M	1.05 M	NA	100	
Hot	TOC	2600	2770	NA	107%	
Persulfate Oxidation	TIC	5845	5430	NA	93%	

 Table 3.10.
 AP-105DF Feed, Effluent, and Eluate Compositions (Test 1) ASR 0335, Inorganic and Carbon Analytes

Analysis		TI014-FEED	TI014-EFF- Comp	TI014-ELComp	Fraction in Effluent	Fraction in Eluate
Method	Analyte	µg/mL	µg/mL	µg/mL	%	%
	Be	[0.18]	[0.18]	[0.041]	[100%]	[1.2%]
	Bi	<34	<34	[6.5]		
	Ce	<28	<28	<4.6		
	Co	<4.4	<4.4	<0.7		
	Cu	[4.3]	<2.3	11.2		[14]
	Dy	<2.1	<2.1	< 0.3		
	Eu	<0.6	<0.6	<0.1		
	La	<1.8	<1.8	< 0.3		
	Mg	<1.4	<1.4	[1.4]		
	Mn	<0.5	<0.5	[0.19]		
	Mo	[54]	58.4	<1.0	[108]	
	Nd	<12	<12	<1.9		
ICP-OES Opportunistic	Pd	<11	<11	<1.7		
Analytes	Rh	<10	<10	<1.7		
5	Ru	<9.8	<9.8	<1.6		
	S	1390	1360	[67]	98%	[0.26%]
	Sb	<57	<57	<9.4		
	Si	104	[71]	18.5	[68%]	0.96%
	Sn	[34]	<27.6	[6.1]		
	Sr	< 0.2	[0.23]	[0.10]		
	Та	<15	<15	<2.5		
	Te	<26	<26	<4.2		
	Tl	<55	<55	<9.1		
	V	<2.0	<2.0	[0.33]		
	W	[90]	[95]	<2.5		
	Y	<0.4	<0.4	<0.1		

Table 3.10 (cont.)

(a) The Ag blank spike and matrix spike recoveries were 41% and 46%, respectively, indicating a low bias in Ag analysis.

Bracketed values indicate the associated sample results were less than the estimated quantitation limit but greater than the MDL. Analytical uncertainty for these analytes is $>\pm 15\%$.

NA = not analyzed.

"--" = not applicable.

			Test 2					Test 3		
Analyte	TI015- FEED	TI015- EFFComp	TI015- ELComp	Fraction in Effluent	Fraction in Eluate	TI016- FEED	TI016- EFFComp	TI016- ELComp	Fraction in Effluent	Fraction in Eluate
Radionuclide	µCi/mL	µCi/mL	µCi/mL	%	%	µCi/mL	µCi/mL	µCi/mL	%	%
¹³⁷ Cs	109	0.00471	1110	0.0044%	121%	129	0.0401	1510	0.031%	108%
²³⁹⁺²⁴⁰ Pu	5.82E-05	5.68E-05	<mdl< td=""><td>98</td><td></td><td>5.77E-05</td><td>5.68E-5</td><td><mdl< td=""><td>99</td><td></td></mdl<></td></mdl<>	98		5.77E-05	5.68E-5	<mdl< td=""><td>99</td><td></td></mdl<>	99	
⁹⁹ Tc	1.03E-01	8.02E-02	[7.9E-06]	79%	[0.00091%]	8.92E-02	7.92E-02	7.58E-06	89%	0.00078%
Inorganic	µg/mL	µg/mL	µg/mL	%	%	µg/mL	µg/mL	µg/mL	%	%
Al	13,500	14,000	69.2	105%	0.061%	14,700	13,900	78.0	95%	0.049%
Cd	[0.85]	[1.5]	< 0.17	[178%]		[1.5]	< 0.79	< 0.17		
Cr	332	324	1.86	98%	0.067%	344	323	2.56	94%	0.068%
Cu	[3.0]	[2.4]	11.7	[81%]	[47%]	[2.7]	[1.6]	9.53	[59%]	32%
Fe	[5.6]	[3.6]	4.58	[65%]	[10%]	[5.9]	[3.5]	4.14	[59%]	6.4%
K	3735	3750	282	101%	0.90%	4130	3880	369	94%	0.82%
Мо	50.8	49.8	< 0.51	99%		50.5	47.0	< 0.51	93%	
Na	136,500	132,000	3150	98%	0.28%	142,000	134,000	4110	95%	0.27%
Ni	30.1	30.2	[2.5]	101%	[1.0%]	30.3	28.3	4.35	94%	1.3%
Pb	[17]	[8.9]	87.6	54%	63%	[20]	<7.0	111		[51%]
S	1130	1090	[27]	97%	[0.29%]	1080	1020	[35]	95%	[0.30%]
Si	125	133	739	107%	71%	103	120	625	117%	56%
Sr	[0.081]	[0.11]	0.211	[137%]	[31%]	[0.080]	[0.056]	0.414	[70%]	[0.30%]
Zn	[3.85]	[5.1]	17.6	[134%]	[55%]	<1.1	<1.1	10.2		
U ICP-MS	6.03	5.51	1.65	92%	3.3%	6.29	5.71	2.47	91%	3.6%

 Table 3.11.
 Test 2 and Test 3 Analytical Results Summary (ASR 372)

"--" = not applicable.

<MDL = less than method detection limit.

Bracketed values indicate the associated sample results were less than the estimated quantitation limit but greater than the MDL. Analytical uncertainty for these analytes is $>\pm 15\%$.

			Test 4 Take 1				,	Test 4 Take 2		
							TI020-EFF			
Analyta	TI020-	TI020-	TI020-	Fraction in	Fraction in	TI020-FEED-	Comp-Take	TI020-	Fraction in	Fraction in
Allalyte	FEED-I-A	EFFComp	ELCOMP	Efficient	Eluate	Take 2	2	EComp	Ennuent	Eluale
Radionuclide	µCi/mL	µCi/mL	µCi/mL	%	%	µCi/mL	µCi/mL	µCi/mL	%	%
¹³⁷ Cs	111	0.00352	260	0.0031%	106%	130	0.0466	1060	0.035%	99%
²³⁹⁺²⁴⁰ Pu	7.39E-05	4.39E-05	<mdl< td=""><td>57%</td><td></td><td>5.94E-05</td><td>5.18E-05</td><td><mdl< td=""><td>86%</td><td></td></mdl<></td></mdl<>	57%		5.94E-05	5.18E-05	<mdl< td=""><td>86%</td><td></td></mdl<>	86%	
⁹⁹ Tc	0.0862	0.0767	<mdl< td=""><td>86%</td><td></td><td>0.0832</td><td>0.0772</td><td>[5.2E-6]</td><td>92%</td><td>[0.00075%]</td></mdl<>	86%		0.0832	0.0772	[5.2E-6]	92%	[0.00075%]
Inorganic	µg/mL	µg/mL	µg/mL	%	%	µg/mL	µg/mL	µg/mL	%	%
Al	13,300	13,200	93.5	96%	0.32%	14,000	14,200	49.8	100%	0.043%
Cd	[1.3]	[1.1]	< 0.17	[82%]		[1.3]	[1.3]	< 0.17	[99%]	
Cr	337	301	2.63	86%	0.35%	341	334	2.50	97%	0.089%
Cu	[2.9]	<1.1	5.32		83%	[2.4]	[1.2]	11.3	[49%]	[57%]
Fe	[4.1]	[15]	[3.4]	[353%]	[37%]	[8.1]	[5.4]	4.22	[66%]	[6.3%]
K	3930	3460	289	85%	3.3%	4020	3950	268	97%	0.81%
Мо	48.8	43.9	< 0.51	87%		48.5	49.4	< 0.51	101%	
Na	140,000	128,000	3150	88%	1.0%	140,000	139,000	3020	98%	0.26%
Ni	30.5	24.8	5.86	78%	8.7%	29.8	29.0	5.48	96%	2.2%
Pb	[15]	<7.0	29.9		[90%]	[15]	<7.0	95.6		[77%]
S	1060	924	[23]	84%	[1.0%]	1050	1040	[20]	98%	[0.23%]
Si	127	152	546	115%	194%	145	172	373	117%	31%
Sr	[0.096]	[0.049]	0.148	[49%]	[70%]	[0.074]	[0.12]	0.115	[160%]	[19%]
Zn	<1.1	<1.1	13.7			[2.0]	<1.1	10.4		[63%]
U ICP-MS	5.87	4.54	1.46	75%	11%	6.08	5.52	2.22	90%	4.4%

 Table 3.12.
 Test 4 Take 1 and Test 4 Take 2 Analytical Results Summary (ASR 372)

"--" = not applicable.

<MDL = less than method detection limit.

Bracketed values indicate the associated sample results were less than the estimated quantitation limit but greater than the MDL. Analytical uncertainty for these analytes is $>\pm 15\%$.

			Test 5		
	TI021-	TI021-	TI021-	Fraction in	Fraction in
Analyte	FEED	EFFComp	EComp	Effluent	Eluate
Radionuclide	µCi/mL	µCi/mL	µCi/mL	%	%
¹³⁷ Cs	127	0.0052	860	0.0041%	89%
²³⁹⁺²⁴⁰ Pu	8.01E-05	6.42E-05	<mdl< td=""><td>76%</td><td></td></mdl<>	76%	
⁹⁹ Tc	0.0862	7.88E-02	7.87E-06	86%	0.0012%
Inorganic	µg/mL	µg/mL	µg/mL	%	%
Al	14,900	14,650	75.6	97%	0.067%
Cd	[1.1]	[0.95]	< 0.17	[85%]	
Cr	341	334	3.09	97%	0.12%
Cu	[2.4]	[1.4]	12.5	[58%]	69%
Fe	[3.4]	[2.8]	4.89	[81%]	19%
Κ	4120	4010	287	96%	0.92%
Mo	49.9	48.4	< 0.51	96%	
Na	142,000	139,000	3500	97%	0.33%
Ni	31.1	28.8	6.93	91%	2.94%
Pb	[16]	<7.08	89.1		[73%]
S	1050	1024	[22]	96%	[0.28%]
Si	94.8	77.5	718	81%	100%
Sr	[0.082]	[0.106]	0.220	[127%]	35%
Zn	<1.12	<1.96	13.1		
U ICP-MS	5.80	5.64	2.28	96%	5.2%

 Table 3.13.
 Test 5 Analytical Results Summary (ASR 372)

"--" = not applicable.

<MDL = less than method detection limit.

Bracketed values indicate the associated sample results were less than the estimated

quantitation limit but greater than the MDL. Analytical uncertainty for these analytes is $\geq \pm 15\%$.

In Test 1, the Pu partitioned mostly to the effluent (94%), with 2% to the eluate and possibly 4% remaining on the ion exchange column. However, analysis uncertainty for ^{239/240}Pu was approximately 4% (1-sigma),¹ and the possibility that all Pu could be accounted for in the effluent and eluate cannot be discounted. In subsequent tests, the Pu recovery in the effluent varied from 99% (Test 3) to 57% (Test 4 Take 1). The Test 4 Take 1 test only partially loaded the SRF with Cs because it was a short run. It is possible that Pu has some affinity for the SRF and as more Cs exchanges onto the resin, it displaces the Pu into the effluent. More studies would be required to assess this possibility. The Pu recovery in the later cycles (Test 4 Take 2 and Test 5) demonstrated reduced Pu recovery (86% and 76%, respectively) in the effluent relative to Tests 1, 2, and 3 (94%, 98%, and 99% respectively). The Pu was not detected in the eluate (Tests 2-5), indicating that some fraction of Pu may be accumulating onto the SRF from one cycle to the next.

¹ All counting uncertainty was reported as 1-sigma.

The ²³⁷Np (Test 1) results were similar to the Pu results where 92% of the loaded ²³⁷Np was accounted for in the effluent; ²³⁷Np was not detected in the eluate. The ²⁴³⁺²⁴⁴Cm chemistry behavior normally follows that of ²⁴¹Am ; therefore, the low ²⁴³⁺²⁴⁴Cm recovery (66%) in the Test 1 effluent is at odds with the quantitative recovery for ²⁴¹Am (101%) in the Test 1 effluent. Analytical uncertainties for ²⁴¹Am and ²⁴³⁺²⁴⁴Cm were 2% and 8%, respectively; these uncertainties were well below the missing 34% ²⁴³⁺²⁴⁴Cm. Virtually all ⁹⁰Sr reported to the effluent (analysis uncertainty was 2%). Nominally 78% of the ⁹⁹Tc was recovered in the effluent product; <0.1% was found in the eluate. This indicates that the ⁹⁹Tc may have reacted with the SRF resin and was retained. Analysis of the spent resin is intended to be conducted in the future.

The ⁹⁹Tc recovered similarly in all test cycles at about 79% to 92% with no up or down trend. A small amount of ⁹⁹Tc ($\leq 0.013\%$) was recovered in the eluate. Thus, the remaining ⁹⁹Tc mass balance could remain on the column.

The ICP-OES results for metals showed that virtually all analytes reported to the effluent. In some cases, the analyte percent recovery could not be calculated because the analyte was not detected in the feed, the effluent, and/or the eluate. Three transition metals (Fe, Cu, Zn), Pb, and U (U as measured by KPA or ICP-MS) recovered at >3% in the eluate composite. In the case of Test 1, Pb recovery could not be estimated because Pb was not detected in the feed. Over 50% of the Pb was detected in the eluate from Tests 2-5. These results indicate that a high proportion of Pb and some Fe, Cu, Zn, and U exchanged from the feed onto the ion exchanger during the loading phase and were eluted at least to some extent with 0.45 M HNO₃.

Nominally 100% of all anions were accounted for in the effluent. TIC is generally ascribed to carbonate and thus is ascribed as an anion. TOC includes oxalate and other organic carbon forms, usually complexants. The oxalate measured by IC, 2.2 E-3 M, was a small fraction of the TOC (0.22 M as C). The TOC recovery in the effluent was quantitative at 107%.

4.0 Discussion

This section summarizes processing trends associated with the number of cycles, flowrates, and elution parameters.

4.1 Cs Load Capacity

The Cs load capacity was calculated from the total Cs loaded onto the lead column, which was assumed to be fully saturated under the load conditions, and the dry H-form resin mass loaded into the lead column according to Eq. (4.1):

$$\frac{A_{Cs} \times CF}{M_R} = C \tag{4.1}$$

where A_{Cs} = total activity of ¹³⁷Cs, µCi loaded onto the lead column

CF = conversion factor, 6.6E-5 mg Cs/ μ Ci ¹³⁷Cs (based the nominal 7.9 μ g Cs/mL Cs and 120 μ Ci ¹³⁷Cs/mL in the AP-105DF)

 M_R = mass of dry, H-form resin, 2.45 g

C = capacity, mg Cs/g H-form resin

Table 4.1 provides the Cs capacity found on the lead column with each process test. It is clear that the capacity decreased with each process cycle. After the six load/elute cycles, the Cs capacity dropped 17% (~2.8% per process cycle). It is noted that the Cs capacity change between Test 3 and Test 4 Take 2 was minimal. A short load cycle, Test 4 Take 1, was conducted between these two runs with minimal additional Cs loading and concomitant less chemical and dose exposure from AP-105DF loading. Therefore, these sources of resin degradation (chemical and radiological) would be substantially reduced for the Test 4 Take 1 process cycle. Previous testing with SRF in simulant using the WTP-designated process conditions showed a reduction of 7% Cs capacity after processing 16 full load/elute cycles (Fiskum et al. 2006a). This corresponds to a Cs capacity reduction of ~0.4% per process cycle. The observed increased rate of SRF degradation may be attributed to the LAWPS-designated process conditions (more Cs loading) and concomitant MRad radiolytic exposure.

Test Number	Cs Capacity, mg Cs/g H-form Resin				
1	6.64				
2	6.38				
3	6.09				
4 Take 1	NA ^(a)				
4 Take 2	6.06				
5	5.53				
(a) This test did not fully load the lead column.					

 Table 4.1.
 Cesium Capacity as a Function of Process Cycle

4.2 Column Performance

The column performance parameters can be further evaluated in terms of the following:

- a. 50% breakthrough
- b. Mass transfer zone
- c. Leakage of Cs from the lag column to the product

A reduction in the 50% Cs breakthrough point indicates a reduction in the Cs load capacity of the resin bed. Increasing the mass transfer zone reduces the feed volume that can ultimately be processed before reaching the effluent contract limit. Leakage of the Cs into the effluent product will affect whether or not the product will meet the contract specification or decrease the volume that can be processed before reaching the contract specification.

4.2.1 50% Breakthrough

Table 4.2 shows the lead column Cs load values for all tests. It is clear that the 50% breakthrough was similar for Tests 1 and 2. The 50% breakthrough point decreased moderately (8%) between Tests 2 and 3. The subsequent Test 4 Take 2 dropped an additional 4% and was equivalent to Test 5. The 50% breakthrough is expected to be independent of flowrate. The decrease in the 50% breakthrough is attributed to resin degradation from chemical, radiolytic, and mechanical means.

Test Number	50% Cs Breakthrough, BV	% Difference from Test 1
1	206	0.00%
2	205	0.49%
3	189	8.3%
4 Take 1	NA ^(a)	
4 Take 2	182	12%
5	181	12%

 Table 4.2.
 50% Breakthrough as a Function of Process Cycle

4.2.2 Mass Transfer Zone

Figure 4.1 shows how the Cs breakthrough curve changes with increasing flowrate on the lead column (Tests 1-3). Test 1 processed at the slowest flowrate; it showed the least Cs leakage early in the process run because the prior run processed feed at nominally 0.1 μ C/mL ¹³⁷Cs (Fiskum et al. 2017). Examining the load curve from about 0.05% to 50% breakthrough shows that increasing flowrate decreases the breakthrough slope and thus increases the mass transfer zone.



Figure 4.1. Cesium Breakthrough Curves as a Function of Flowrate on SRF Resin

The mass transfer zone was evaluated between 1% and 90% C/C₀ for all tests (see Table 4.3). The 90% breakthrough point remained fairly constant (4% standard deviation) for all tests regardless of the flowrate and process cycle. The 1% breakthrough point varied significantly, occurring sooner with increasing flowrate as can be discerned by examination of Figure 4.1. At constant flowrate (~3.0 BV/h), the mass transfer zone had a narrow range from 112 to 130 BVs. Figure 4.2 compares the breakthrough profiles for tests processed at ~3 BV/h (Test 2, Test 4 Take 2, and Test 5). Profiles from Tests 4 and 5 are shifted left relative to Test 2, indicative of the decreased Cs loading capacity.

Test Number	Flowrate, BV/h	1% Breakthrough, BVs	90% Breakthrough, BVs	Transition Zone, BVs 1% to 90%
1	1.80	147	242	95
2	3.05	127	251	124
3	4.53	95	244	144
4 Take 1	3.04	NA ^(a)	NA ^(a)	NA ^(a)
4 Take 2	2.95	114	226	112
5	3.17	102	~230 ^(b)	130
(a) This test did i	not fully load the	lead column: only 53	BVs of AP-105DE were	processed

Table 4.3. Mass Transfer Zone (1% to 90% C/C_0)

IУ

(b) Extrapolated from 83.1% breakthrough.

Increasing flowrate while maintaining residence time (i.e., constant BV/h) was shown to sharpen the mass transfer zone by overcoming film diffusion limitations (Fiskum et al. 2017). The increased flowrate in these tests was not sufficient to overcome the film diffusion limitation.



Figure 4.2. Lead Column Cesium Breakthrough Profiles at Constant Flowrate

4.2.3 Cs Leakage

Cesium leakage from the completion of one process cycle to the next process cycle was evaluated. Cs leakage stems from residual Cs on the lag column SRF resin exchanging into the product effluent and contaminating the product. Cs leakage may interfere with the ability to process large volumes of waste and maintain the effluent Cs concentration below the 10% contract limit.

Figure 4.3 shows all lag column load curves. Prior to Test 1 with AP-105DF, one cycle had been processed with simulant spiked with 60 μ g/mL Cs and 0.1 μ Ci/mL ¹³⁷Cs (Fiskum et al. 2017). The lag column effluent ¹³⁷Cs concentration was proportionately low, resulting in low initial C/C₀ values. The Test 2 lag column effluent was the highest of the group; it followed the shortest elution (16 BVs) from Test 1. Test 3, Test 4 Take 1, and Test 5 had similar lag column Cs effluent concentrations following previous cycle elution of 30, 23, and 30 BVs, respectively. Test 4 Take 2 Cs effluent concentration started markedly lower than the others; this is attributed to the low total Cs loading associated with Test 4 Take 1, which gave the sparingly loaded lag column a chance to cycle (Na-form to H-form) once before the next load. The extra cycle is believed to remove more residual Cs from the resin bed.



Figure 4.3. Comparison of Lag Column Cs Breakthrough

The elution with 0.45 M HNO₃ varied for the different process cycles from 16 to 30 BVs. The water rinse volumes following elution were relatively constant. The relationship of the Cs concentration in the final water rinse sample following elution was evaluated relative to the following process run Cs leakage from the lag column. Table 4.4 summarizes these salient attributes. The final water rinse Cs concentration decreases approximately linearly as a function of the BVs (range 15 to 30 BVs) of eluent processed (Figure 4.4).

It is noted that the onset of Cs breakthrough from the lag column did not influence appreciably where the contract limit was crossed later in the load cycle. Increasing flowrate resulted in earlier Cs breakthrough (Tests 1-3) as expected; increasing cycles between Tests 2 and 5 resulted in a reduction of 43 BVs to reach the 10% contract limit.

The relationship between the lag column initial Cs breakthrough with respect to elution volume and water rinsing was evaluated. Table 4.4 summarizes the initial feed sample Cs % C/C₀ from the lag column for each test along with the elution volume, water rinse volume, and final 15-mL water rinse sample Cs % C/C₀. The extended water rinse was not included because the Cs concentration was not evaluated for the last segment of this rinse. The final water rinse sample Cs concentration was clearly a function of the BVs processed, as shown in Figure 4.4, where increasing the elution BVs resulted in a decreased Cs % C/C₀ in the final water rinse sample.

Test ID	Initial Lag Column Sample Cs Conc., % C/C ₀	Elution, BVs	Water Rinse, BVs	Final Water Rinse Cs Conc., % C/C ₀
Test 0 ^(a)	<1.85E-4	16.4	5.3	4.73E-1
Test 1	1.34E-4	15.9	5.8	3.84E-1
Test 2	5.06E-2	29.9	5.6	6.81E-2
Test 3	6.77E-3	23.0	6.3	2.01E-1
Test 4a	3.35E-3	24.9	NA	NA
Test 4b	8.20E-4	29.5	8.1	7.96E-2
Test 5	4.17E-3	29.1	7.9	9.61E-2

Table 4.4. Lag Column Initial Cs Leakage

(a) The initial test was conducted with simulant spiked to $60 \,\mu$ g/mL Cs and $0.1 \,\mu$ Ci/mL ¹³⁷Cs (Fiskum et al. 2017). This was the virgin run for the lag column and no effluent Cs in the initial sample was expected.



Figure 4.4. Final Water Rinse Sample Cs Concentration as a Function of the Elution BVs Processed, Tests 0 to 5

The relationship between effluent Cs coming off the lag column and eluent BVs processed in the previous ion exchange run was evaluated. Figure 4.5 shows the relationship of the initial lag column Cs effluent concentration (%C/C₀) as a function of the elution volume in the previous process cycle. It was clear that the 16 BV elution was insufficient to reduce the effluent product in the next process cycle enough to meet the 10% contract limit. Other efforts to equate Cs leakage into the subsequent process cycle were confounded by the changing elution conditions and changing Cs loading on the lag column.



Figure 4.5. Initial Cs Leakage versus Eluent BVs Processed in Previous Cycle

5.0 Resin Bed Physical Properties

The ion exchange resin bed contracts as it converts to H-form and expands when it converts to Na-form. Figure 5.1 shows the in-column contraction/expansion history starting with the in-column pretreatment operations (Fiskum et al. 2017) and continuing through the final water rinse following the Test 5 AP-105DF processing cycle. Typical 20% shrink/swell behavior was observed. The lead column appeared to increase in volume approximately 10% from the first input volume (9.9-mL Na-form) to the final volume (10.9-mL Na-form), indicating that the resin beads were likely relaxing and expanding more with the repeated process cycles. The lag column appeared to shrink and swell differently from the lead column. Starting at the Test 4 Take 1 process cycle, the resin volume increased 20% (Na-form) relative to the initial packed column. Further, the lag column Na-form resin bed was approximately 10% larger in volume relative to the lead column runs. The gap observed in the lag column during elution in Test 4 Take 1 was not observed later during processing in Test 4 Take 2 and Test 5.

Oxygen is known to attack the SRF resin (Fiskum et al. 2006a). The AP-105DF feed is not expected to contain much dissolved oxygen due to its high salt content. The feed displacement (0.1 M NaOH), water, regeneration solution (1.0 M NaOH), and eluent (0.45 M HNO₃) are anticipated to contain significantly higher dissolved oxygen concentrations. Dependent on the process step, both the lead and lag columns receive first reagent exposure of dissolved oxygen. The divergent swell behavior of the lag column seemed to start with the observed gap in the resin bed. But later cycles did not show this gap and the lag column swelling was much larger than the lead column swelling. The repeated cycling may be irreversibly leading to expanded polystyrene bead.



Figure 5.1.	Resin Bed	Expansion and	Contraction History

Feed		Feed		Feed		Feed		Feed	
ID	Matrix	ID	Matrix	ID	Matrix	ID	Matrix	ID	Matrix
1	1.0 M NaOH soak	11	1.0 M NaOH	20	0.1 M NaOH FD	30	1.0 M NaOH	40	0.45 M HNO3
2	DI water	12	AP-105DF Test 1	21	DI water	31	AP-105DF Test 4 Take 1	41	DI water
3	0.45 M HNO3	12.5	AP-105DF Test 1 end	22	0.45 M HNO3	32	0.1 M NaOH FD	42	1.0 M NaOH
4	DI water	13	0.1 M NaOH FD	23	DI water	33	DI water	43	AP-105DF Test 5
5	1.0 M NaOH	14	DI water	24	1.0 M NaOH	34	0.45 M HNO3	44	0.1 M NaOH FD
6	Simple Simulant	15	0.45 M HNO3	25	AP-105DF Test 3	35	DI water	45	DI water
7	0.1 M NaOH FD	16	DI water	26	0.1 M NaOH FD	36	1.0 M NaOH	46	0.45 M HNO3
8	DI water	17	DI water	27	DI water	37	AP-105DF Test 4 Take 2	47	DI water
9	0.45 M HNO3	18	1.0 M NaOH	28	0.45 M HNO3	38	0.1 M NaOH FD		
10	DI water	19	AP-105DF Test 2	29	DI water	39	DI water		

6.0 Multi-Cycle Testing with AP-105DF Conclusions

A radioactive waste feed Test Platform was constructed at PNNL to support tank waste filtration¹ and Cs removal by ion exchange to meet near-term waste pretreatment for delivery of feed to the WTP LAW facility. The ion exchange component of this Test Platform consisted of a small-scale column system with a pump and sampling points. Two columns were positioned in a lead-lag format, each filled with 9.9 mL of SRF resin (Na-form). At the time of testing, process parameters were prototypic of the expected LAWPS process with respect to the ion exchange media. The goal was to process feed until the ¹³⁷Cs concentration in the effluent reached 10% of the LAW contract limit. To this end, effluent samples were collected periodically during the load process and measured for ¹³⁷Cs and the load curve was developed. The combined effect of Cs leakage and the delay between sampling and analysis resulted in the processing past the 10% LAW contract limit in three of the five cycles. The processing parameters resulted in fully loading the lead column and having a significant fraction of Cs loaded onto the lag column. Regeneration solutions consisted of FD (0.1 M NaOH), water rinse, eluent (0.45 M HNO₃), and post-elution water rinse. The flowrates in terms of BV/h were also matched to the LAWPS process for the first test, but because of the small column size, 9.9-mL lab-scale testing versus 297-gallon full-scale, linear flow velocities were fundamentally different.² The results are comparable in that the small column testing provides a worst-case bounding load profile. The higher linear flow rate with constant residence time sharpens the Cs load curve. However, Cs leakage to the next process run is not likely to improve.

A total of 12.3 L of AP-105DF tank waste, consisting of 5.7 M Na and 120 µCi/mL ¹³⁷Cs, was processed through the Cs ion exchange system in six process cycles. This processing exposed the SRF resin to chemical exposure (process solution and residence time in the resin bed) and physical changes (shrink/swell) similar to what would be experienced at the WRPS LAWPS (and to some extent, the WTP). Radiolytic exposure of the small column SRF test was about 1/100th of the resin lifetime exposure expected at full scale plant conditions and about a third of the exposure expected on a single process cycle at full scale plant conditions. Feed flowrates and elution volumes were modified slightly from one test to the next to evaluate effects on the Cs load profiles. Flowrates ranged from 1.8 to 4.5 BV/h and elution flowrates ranged from 1.4 to 2.7 BV/h. The following conclusions were made as a result of this work.

- A quantity of 275 BVs of AP-105DF was processed at 1.8 BV/h through the first cycle before reaching the 10% contract limit. As flowrate increased to 4.5 BV/h, the volume processed before the limit was reached decreased to 190 BVs. Subsequent processing at ~3.0 BV/h maintained the ~200 BVs processed before reaching the 10% contract limit.
- 2. The lead column Cs capacity decreased from 6.64 (first process cycle) to 5.53 (sixth process cycle) mg Cs per g dry H-form resin, indicating ~2.8 % capacity loss per cycle. The Cs capacity decrease was attributed to combined chemical, physical, and radiolytic degradation. This degradation rate exceeded previously reported degradation of 7% over 16 process cycles (0.44% per cycle) from chemical and physical exposure.

¹ The filtration component was reported by Geeting et al. (2017).

² Increasing linear flow velocity enhances film diffusion and thus ion exchange onto the SRF resin. Small columns with lower linear flowrates have poorer film diffusion.

- 3. The Cs mass transfer zone increased with increasing flowrate, ranging from 95 BVs at 1.8 BV/h to 144 BVs at 4.5 BV/h. The increased mass transfer zone was attributed to the reduction in residence/equilibration time and the increased flowrate did not overcome limitations due to film diffusion (Fiskum et al. 2017).
- 4. The 50% Cs breakthrough generally shifted left with successive cycles, ranging from 206 BVs (first process cycle) to 181 BVs (sixth process cycle), commensurate with the decreased capacity associated with resin degradation (Fiskum et al 2006a).
- 5. Within analytical uncertainty, typically >99% of the Cs processed through the ion exchange system was collected in the eluate. However, trace Cs remained on the resin beds.
- 6. After an initial loading with AP-105DF, Cs leakage from the lag column to the effluent was a significant factor in the system's ability to satisfy the effluent decontamination target. Although virtually all Cs was accounted for in the eluate, a very small Cs fraction remained on the lag column and leaked into the next cycle product. Approximately 30 BVs of 0.45 M HNO₃ eluent were required to meet the target DF in the next process cycle.
- 7. Small fractions (3% to 15%) of Fe, Cu, U, and Zn were found in the eluate; an indeterminate fraction of Pb was found in the eluate. This indicates that some fraction of these analytes exchanged onto the resin and was eluted with the acid. Less than 1% of detectable metals were found in the eluate. All other metals and anions were found with the effluent product.
- 8. Radionuclides (⁹⁰Sr, ²³⁷Np, ²³⁹⁺²⁴⁰Pu, and ²⁴¹Am) were largely recovered in the effluent. About 2% of the Pu was found in the eluate and 4% was not accounted for. Only 66% of the ²⁴³⁺²⁴⁴Cm and 78% of the ⁹⁹Tc were recovered in in the effluent product. Less than 0.1% of the ⁹⁹Tc was recovered in the eluate, indicating that ~20% remained on the SRF. Only the first process cycle eluate was measured for ²⁴³⁺²⁴⁴Cm, where 3.1% was recovered.
- 9. The SRF resin continued to expand from one cycle to the next, increasing in volume by an additional 10% to 20% after six process cycles.
- 10. The SRF resin in two 10-mL beds maintained the Cs exchange efficacy after processing 12.3 L of AP-105DF in six cycles with reasonable efficiency despite the synergistic resin degradation effects from the combined radiological dose and chemical and physical (shrink/swell) exposures. These results support test observations and conclusions previously described in SRF literature reviews (King 2007; Brown 2014).

7.0 References

ASME. 2000. *Quality Assurance Requirements for Nuclear Facility Applications*. NQA-1-2000, American Society of Mechanical Engineers, New York, New York.

ASME. 2008. *Quality Assurance Requirements for Nuclear Facility Applications*. NQA-1-2008, American Society of Mechanical Engineers, New York, New York.

ASME. 2009. *Addenda to ASME NQA-1-2008*. NQA-1a-2009, American Society of Mechanical Engineers, New York, New York.

Browne E and JK Tuli. 2007. "Nuclear Data Sheets for A = 137." *Nuclear Data Sheets* 108(10):2173–2318. doi:10.1016/j.nds.2007.09.002.

Brown GN. 2014. *Literature Review of Spherical Resorcinol-Formaldehyde for Cesium Ion Exchange*. PNNL-23410, Pacific Northwest National Laboratory, Richland, Washington.

Duignan MR and C Nash. 2009. *Removal of Cesium from Savannah River Site Waste with Spherical Resorcinol Formaldehyde Ion Exchange Resin: Experimental Tests*. SRNL-STI-2009-00367, Savannah River National Laboratory, Aiken, South Carolina.

Fiskum SK, ST Arm, WC Buchmiller, T Trang-Le, JE Martinez, J Matyas, MJ Steele, KK Thomas, and DL Blanchard. 2006a. *Comparison Testing of Multiple Spherical Resorcinol-Formaldehyde Resins for the River Protection Project—Waste Treatment Plant*. PNWD-3785; WTP-RPT-143, Rev. 1, Battelle Pacific Northwest Division, Richland, Washington.

Fiskum SK, ST Arm, MS Fountain, MJ Steele, and DL Blanchard. 2006b. *Spherical Resorcinol-Formaldehyde Resin Testing for*¹³⁷Cs Removal from Simulated and Actual Hanford Waste Tank 241-AP-101 Diluted Feed (Envelope A) Using Small Column Ion Exchange. PNWD-3697; WTP-RPT-134, Rev. 0, Battelle Memorial Institute, Richland, Washington.

Fiskum SK, MJ Steele, and DL Blanchard. 2006c. *Small Column Ion Exchange Testing of Spherical Resorcinol-Formaldehyde Resin for* ¹³⁷Cs Removal from Pre-Treated Hanford Tank 241-AN-102 Waste (Envelope C). PNWD-3751; WTP-RPT-135, Rev. 1, Battelle Memorial Institute, Richland, Washington.

Fiskum SK, HA Colburn, MR Smoot, JR Allred, and RA Peterson. 2017. *Cesium Ion Exchange Using Spherical Resorcinol-Formaldehyde Resin in Support of Waste Qualification Testing for LAWPS*. PNNL-26837; RPT-DFTP-003, Rev. 0.0, Pacific Northwest National Laboratory, Richland, Washington.

Geeting JH, JR Allred, AM Rovira, RW Shimskey, and RA Peterson. 2017. *Crossflow Filtration of Hanford Tank AP-105 Supernatant*. RPT-DFTP-005; PNNL-27085, Pacific Northwest National Laboratory, Richland, Washington

ICRU. 1989. *Tissue Substitutes in Radiation Dosimetry and Measurement*. Report 44 of the International Commission on Radiation Units and Measurements, Bethesda, Maryland.

King WD. 2007. *Literature Reviews to Support Ion Exchange Technology Selection for Modular Salt Processing*. WSRC-STI-2007-00609, Savannah River National Laboratory, Aiken, South Carolina

Russell RL, PP Schonewill, and CA Burns. 2017. *Simulant Development for LAWPS Testing*. PNNL-26165; RPT-LPIST-001, Rev. 0, Pacific Northwest National Laboratory, Richland, Washington.

Saenger EL, FG Bloedorn, and JG Kerkeiakes. 1972. *Protection Against Radiation from Brachytherapy Sources*. Report No. 040, Appendix A, National Council on Radiation Protection and Measurements, Washington, D.C.

Slaback LA and B Schlein. 1998. *Handbook of Health Physics and Radiological Health*. Australian Radiation Laboratory. Scinta, Silver Spring, Maryland.

Stabin MG. 2007. *Radiation Protection and Dosimetry: an Introduction to Health Physics*. Springer, New York, New York.

Appendix A

Column Load and Elution Data

Appendix A

Column Load and Elution Data

The AP-105DF column loading and elution raw data for Tests 1-5 are provided in Table A.1 through Table A.12. The Test 4 cycles also include the regeneration sample data following the elution step.

	Lead (Column			Lag Co	olumn		Fe	ed Displacemen	t and Water Ri	nse
BV	μCi ¹³⁷ Cs/ mL	% C/C ₀	DF	BV	µCi ¹³⁷ Cs/ mL	% C/C ₀	DF	BV	µCi ¹³⁷ Cs/ mL	% C/C ₀	DF
4.4	1.36E-3	1.22E-3	82,238	4.4	1.50E-4	1.34E-4	743,720	FD			
10.2	1.30E-4	1.17E-4	856,213	10.0	6.05E-5	5.43E-5	1,843,199	275.3	1.71E-2	1.54E-2	6507
31.4	8.05E-5	7.21E-5	1,386,420	30.8	6.53E-5	5.86E-5	1,707,607	276.4	3.03E-2	2.72E-2	3682
50.6	7.04E-5	6.32E-5	1,583,472	49.7	6.50E-5	5.83E-5	1,716,546	277.6	3.74E-2	3.35E-2	2985
73.9	3.71E-4	3.33E-4	300,640	72.6	8.41E-5	7.54E-5	1,325,798	278.7	5.55E-2	4.97E-2	2010
93.6	3.72E-3	3.34E-3	29,978	92.1	5.09E-5	4.56E-5	2,192,986	279.7	4.91E-2	4.40E-2	2273
116.2	4.56E-2	4.09E-2	2446	114.5	6.65E-5	5.96E-5	1,676,868	280.8	7.95E-3	7.12E-3	14,040
135.8	3.65E-1	3.28E-1	305	133.8	3.85E-3	3.45E-3	28,946	DI rinse			
159.3	3.32E+0	2.98	34	141.9	3.72E-4	3.34E-4	299,567	281.9	2.77E-3	2.48E-3	40,332
171.6	1.07E+1	9.63	10	154.2	4.13E-5	3.70E-5	2,702,566	283.0	2.04E-3	1.83E-3	54,722
175.1	1.42E+1	12.73	7.9	159.9	7.80E-5	7.00E-5	1,429,522	284.0	1.44E-3	1.29E-3	77,221
200.4	4.47E+1	40.07	2.5	179.2	1.90E-4	1.70E-4	587,351	285.0	1.15E-3	1.03E-3	97,250
210.2	6.34E+1	56.8	1.8	191.9	8.02E-5	7.19E-5	1,390,128				
219.6	7.89E+1	70.7	1.41	201.0	8.63E-5	7.73E-5	1,293,150				
230.4	9.14E+1	82.0	1.22	211.7	9.25E-5	8.30E-5	1,205,396				
241.5	9.98E+1	89.5	1.12	222.6	1.28E-4	1.15E-4	870,556				
252.1	1.05E+2	94.3	1.06	232.9	2.17E-4	1.95E-4	513,854				
262.9	1.12E+2	100	1.00	243.6	4.30E-4	3.85E-4	259,499				
272.1	1.17E+2	105	0.95	252.6	1.25E-3	1.12E-3	89,109				
282.9	1.20E+2	107	0.93	263.2	4.11E-3	3.68E-3	27,169				
294.0	1.19E+2	107	0.93	274.2	1.33E-2	1.19E-2	8382				
BV = bed vc	olume; DI = deio	nized; DF = de	contamination fa	ctor; FD = fe	ed displacement	$C_0 = 112 \mu Ci$	¹³⁷ Cs/ mL.				

Table A.1. AP-105DF Test 1 Cs Load, Feed Displacement, and Water Rinse Results

	Elution			Water Rinse	
	μCi			μCi	
BV	¹³⁷ Cs/ mL	C/C_0	BV	¹³⁷ Cs/ mL	C/C_0
1.42	2.84E+0	2.54E-2	17.42	2.43E+0	2.18E-2
2.87	2.65E+0	2.37E-2	18.84	9.56E-1	8.57E-3
4.40	3.55E+0	3.18E-2	20.23	5.85E-1	5.25E-3
5.89	1.03E+1	9.21E-2	21.73	4.28E-1	3.84E-3
7.35	1.12E+1	1.01E-1			
8.77	1.33E+1	1.20E-1			
10.21	2.83E+3	2.54E+1			
11.64	1.92E+4	1.72E+2			
13.08	2.56E+3	2.29E+1			
14.53	5.84E+1	5.24E-1			
15.94	1.29E+1	1.16E-1			

Table A.2. Elution and Water Rinse Results Following AP-105DF Processing Test 1

BV = bed volume; $C_0 = 112 \ \mu Ci^{137}Cs/mL$.

	Lead (Column			Lag C	olumn		Fe	ed Displacemen	t and Water Ri	inse
BV	μCi ¹³⁷ Cs/ mL	% C/C ₀	DF	BV	μCi ¹³⁷ Cs/ mL	% C/C ₀	DF	BV	µCi ¹³⁷ Cs/ mL	% C/C ₀	DF
5.6	1.12E+1	9.49E+0	11	5.6	5.98E-2	5.06E-2	1977	FD			
14.5	1.15E-1	9.71E-2	1030	14.4	1.32E-2	1.12E-2	8934	252.1	2.94E-2	2.49E-2	4015
26.8	4.78E-2	4.04E-2	2474	26.5	8.89E-3	7.52E-3	13,300	253.7	3.23E-2	2.73E-2	3663
60.8	7.98E-2	6.75E-2	1481	60.2	5.26E-3	4.45E-3	22,456	255.4	4.34E-2	3.67E-2	2722
79.5	7.43E-2	6.29E-2	1591	78.6	3.90E-3	3.30E-3	30,288	256.9	1.63E-2	1.38E-2	7242
98.5	1.65E-1	1.39E-1	717	97.5	2.66E-3	2.25E-3	44,412	DI rinse			
116.3	5.65E-1	4.78E-1	209	115.0	2.49E-3	2.11E-3	47,476	258.5	5.00E-3	4.23E-3	23,666
134.3	1.93E+0	1.63E+0	61	132.7	1.82E-3	1.54E-3	65,121	260.0	3.07E-3	2.60E-3	38,505
154.0	6.45E+0	5.5	18	153.8	2.05E-3	1.73E-3	57,660	261.7	2.36E-3	1.99E-3	50,196
170.1	1.65E+1	14.0	7	168.0	2.20E-3	1.86E-3	53,620				
186.0	3.79E+1	32.0	3.1	185.7	2.20E-3	1.86E-3	53,784				
206.5	6.28E+1	53.1	1.9	203.6	2.45E-3	2.07E-3	48,204				
223.7	8.65E+1	73.2	1.4	220.7	3.87E-3	3.27E-3	30,549				
242.0	1.04E+2	87.9	1.14	238.8	1.22E-2	1.03E-2	9684				
254.0	1.06E+2	89.6	1.12	250.5	2.97E-2	2.51E-2	3976				

Table A.3. AP-105DF Test 2 Cs Load, Feed Displacement, and Water Rinse Results

BV = bed volume; DI = deionized; DF = decontamination factor; FD = feed displacement; $C_0 = 118 \ \mu Ci \ ^{137}Cs/mL/$

	Elution			Water Rinse	
BV	µCi ¹³⁷ Cs/ mL	C/C ₀	BV	µCi ¹³⁷ Cs/ mL	C/C ₀
1.47	1.92E+0	1.63E-2	31.37	1.30E-1	1.10E-3
2.99	1.83E+0	1.55E-2	32.88	1.14E-1	9.62E-4
4.50	3.31E+0	2.80E-2	34.36	9.24E-2	7.82E-4
6.01	1.23E+1	1.04E-1	35.50	8.05E-2	6.81E-4
7.52	9.61E+0	8.13E-2			
8.98	1.05E+1	8.89E-2			
10.47	4.04E+3	3.42E+1			
11.96	1.52E+4	1.29E+2			
13.46	1.66E+3	1.41E+1			
14.93	1.93E+2	1.63E+0			
16.43	3.29E+1	2.79E-1			
17.93	3.86E+0	3.27E-2			
19.37	1.53E+0	1.29E-2			
20.89	8.41E-1	7.11E-3			
22.38	5.72E-1	4.84E-3			
23.85	3.16E-1	2.67E-3			
25.33	2.15E-1	1.82E-3			
26.90	2.18E-1	1.84E-3			
28.37	1.73E-1	1.46E-3			
29.89	1.39E-1	1.18E-3			
BV = bed vol	plume; $C_0 = \overline{118}$	µCi ¹³⁷ Cs/mL			

Table A.4. Elution and Water Rinse Results Following AP-105DF Processing Test 2

	Lead (Column			Lag Co	olumn		Fe	ed Displacemen	t and Water Ri	nse
	μCi				μCi				μCi		
BV	¹³ /Cs/ mL	% C/C ₀	DF	BV	¹³⁷ Cs/ mL	% C/C ₀	DF	BV	¹³⁷ Cs/ mL	% C/C ₀	DF
7.5	7.04E-2	5.61E-2	1783	7.4	8.50E-3	6.77E-3	14,767	FD			
20.7	3.33E-2	2.66E-2	3764	20.4	4.42E-3	3.52E-3	28,416	242.5	6.25E-1	4.98E-1	201
34.2	2.71E-2	2.16E-2	4636	32.1	2.86E-3	2.28E-3	43,880	244.0	6.80E-1	5.42E-1	185
52.6	6.31E-2	5.03E-2	1988	50.2	2.24E-3	1.78E-3	56,115	245.6	7.40E-1	5.90E-1	170
61.5	1.22E-1	9.71E-2	1030	58.8	2.10E-3	1.67E-3	59,854	247.2	3.25E-1	2.59E-1	387
79.0	4.04E-1	0.32	311	76.1	1.81E-3	1.44E-3	69,413	DI rinse			
95.6	1.38E+0	1.10	91	92.3	1.84E-3	1.46E-3	68,300	248.7	6.55E-2	5.22E-2	1916
111.1	4.47E+0	3.56	28	107.7	1.86E-3	1.48E-3	67,386	250.3	4.09E-2	3.26E-2	3068
128.4	9.83E+0	7.83	13	128.5	2.03E-3	1.62E-3	61,788	251.3	2.32E-2	1.84E-2	5421
150.4	2.15E+1	17.15	6	146.4	2.61E-3	2.08E-3	48,015				
164.2	3.92E+1	31.26	3.2	164.2	5.00E-3	3.99E-3	25,075				
186.0	5.81E+1	46.33	2.2	181.6	1.33E-2	1.06E-2	9429				
202.5	7.98E+1	63.60	1.6	196.3	2.88E-2	2.30E-2	4355				
219.2	9.57E+1	76.27	1.31	214.3	1.07E-1	8.55E-2	1170				
232.9	1.07E+2	85.65	1.17	240.4	2.65E-1	2.11E-1	473				
247.6	1.15E+2	91.31	1.10	240.9	5.46E-1	4.35E-1	230				
BV = bed vol	olume; DI = deio	onized; DF = de	contamination fa	ctor; FD = fe	ed displacement	; $C_0 = 125 \ \mu C_1$	¹³⁷ Cs/ mL.				

 Table A.5.
 AP-105DF Test 3 Cs Load, Feed Displacement, and Water Rinse Results

	Elution			Water Rinse			Regeneration	
	μCi			μCi			μCi	
BV	¹³⁷ Cs/ mL	C/C_0	BV	¹³⁷ Cs/ mL	C/C_0	BV	¹³⁷ Cs/ mL	C/C ₀
1.44	2.85E+0	2.27E-2	24.61	4.77E-1	3.80E-3	1.8	1.36E-1	1.12E-1
2.92	2.89E+0	2.30E-2	26.23	2.86E-1	2.28E-3	3.2	1.27E-1	1.04E-1
4.45	4.03E+0	3.21E-2	27.78	3.63E-1	2.89E-3	4.7	1.14E-1	9.39E-2
5.99	7.62E+0	6.07E-2	29.33	2.52E-1	2.01E-3	6.3	2.29E-2	1.89E-2
7.50	9.95E+0	7.93E-2				7.9	7.51E-3	6.18E-3
9.00	9.57E+0	7.62E-2				9.4	7.73E-3	6.37E-3
10.49	4.26E+3	3.39E+1						
11.97	1.38E+4	1.10E+2						
13.50	2.50E+3	1.99E+1						
15.05	1.07E+2	8.51E-1						
16.57	3.51E+1	2.80E-1						
18.21	8.46E+0	6.74E-2						
19.79	1.47E+0	1.17E-2						
21.51	6.94E-1	5.53E-3						
23.03	5.22E-1	4.16E-3						
BV = bed vol	plume; $C_0 = 125$	µCi ¹³⁷ Cs/mL	<i>.</i> .					

 Table A.6.
 Elution and Water Rinse Results Following AP-105DF Processing Test 3

	Lead (Column			Lag	Column		Feed Displacement and Water Rinse				
DV	μCi	0/ C/C-	DE	DV	μCi	0/ C/C-	DE	DV	μCi	0/ C/C-	DE	
БV	US/ IIIL	% C/C0	DF	БV	Cs/ IIIL	% C/C0	DF	БV	US/ IIIL	% C/C0	DF	
4.0	2.03E-2	1.67E-2	5982	4.0	4.07E-3	3.35E-3	29,815	NA				
11.2	1.04E-2	8.53E-3	11,724	11.0	4.33E-3	3.57E-3	28,017					
18.9	6.61E-3	5.44E-3	18,379	18.5	3.53E-3	2.91E-3	34,364					
36.0	5.93E-3	4.89E-3	20,467	35.5	2.53E-3	2.08E-3	48,084					
52.9	1.10E-2	9.04E-3	11,059	52.0	2.36E-3	1.94E-3	51,503					
BV = bed v	BV = bed volume; DF = decontamination factor; FD = feed displacement; $C_0 = 121 \ \mu Ci^{137}Cs/mL$											

Table A.7. AP-105DF Test 4 Take 1 Cs Load, Feed Displacement, and Water Rinse Results

Table A.8. Elution Results Following AP-105DF Processing Test 4 Take 1

	Elution			Water Rinse			Regeneration			
BV	µCi ¹³⁷ Cs/ mL	C/C ₀	BV	µCi ¹³⁷ Cs/ mL	C/C ₀	BV	µCi ¹³⁷ Cs/ mL	C/C ₀		
24.92	2.75E+2	2.26E+0	NA	NA	NA	1.6	7.44E-2	6.09E-2		
						3.3	2.53E-2	2.07E-2		
						4.8	2.90E-2	2.38E-2		
						6.4	1.43E-2	1.17E-2		
						8.3	4.43E-3	3.62E-3		
BV – bed v	olume: $C_0 = 121$	uCi ¹³⁷ Cs/ mI								

BV = bed volume; $C_0 = 121 \ \mu C_1^{-13} \ C_s / mL$.

	Lead C	Column			Lag C	olumn		Fee	ed Displacement a	and Water Rin	se
BV	μCi	% C/Co	DE	BV	μCi	% C/C	DE	BV	μCi	% C/Co	DE
Bv	CS/ IIIL	% C/C0		DV		% C/C0		240.5		% C/C0	
4.1	5.73E-3	4.69E-3	21,324	4.1	1.00E-3	8.20E-4	121,922	240.5	4.91E-3	4.02E-3	24,866
13.3	3.23E-3	2.64E-3	37,837	13.1	8.66E-4	7.09E-4	141,082				
23.4	2.59E-3	2.12E-3	47,111	22.9	1.70E-3	1.39E-3	71,981				
41.0	2.60E-3	2.13E-3	46,961	40.4	5.58E-4	4.57E-4	218,925				
58.3	7.27E-3	5.95E-3	16,797	57.5	5.64E-4	4.62E-4	216,596				
76.1	3.50E-2	0.03	3492	74.9	5.39E-4	4.41E-4	226,611				
92.1	2.18E-1	0.18	560	90.5	5.62E-4	4.60E-4	217,423				
109.5	8.61E-1	0.70	142	107.8	5.98E-4	4.89E-4	204,353				
126.8	3.19E+0	2.61	38	124.8	7.11E-4	5.82E-4	171,674				
144.4	1.12E+1	9.18	11	142.0	7.18E-4	5.88E-4	170,002				
161.8	2.71E+1	22.20	4.5	159.0	1.22E-3	1.00E-3	99,943				
182.3	6.11E+1	50.05	2.0	179.4	5.01E-3	4.10E-3	24,375				
204.8	8.75E+1	71.64	1.4	201.5	3.80E-2	3.11E-2	3212				
226.0	1.10E+2	89.68	1.12	222.5	2.14E-1	1.75E-1	572				
238.5	1.14E+2	93.11	1.07	234.7	6.13E-1	5.02E-1	199				

Table A.9. AP-105DF Test 4 Take 2 Cs Load, Feed Displacement, and Water Rinse Results

BV = bed volume; DF = decontamination factor; FD = feed displacement; $C_0 = 122 \ \mu Ci^{137}Cs/mL$.
	Elution		Water Rinse			Regeneration		
	uCi		μCi			uCi		
BV	¹³⁷ Cs/ mL	C/C_0	BV	¹³⁷ Cs/ mL	C/C_0	BV	¹³⁷ Cs/ mL	C/C_0
29.46	1.03E+3	8.47E+0	EDi	-		1.5	8.71E-2	6.98E-2
			31.04	1.78E-1	1.46E-3	3.0	8.72E-2	6.98E-2
			32.67	1.15E-1	9.42E-4	4.5	1.15E-1	9.23E-2
			34.29	1.48E-1	1.21E-3	6.0	2.16E-2	1.73E-2
			35.93	2.13E-1	1.74E-3	7.5	6.53E-3	5.23E-3
			37.57	9.72E-2	7.96E-4	9.1	4.59E-3	3.67E-3
			EDiF					
			29.46	3.16E-2	2.59E-4			
BV = bed volume: $C_0 = 122 \text{ µCi} \frac{137}{\text{Cs}} \text{ mL}.$								

Table A.10. Elution and Water Rinse Results Following AP-105DF Processing Test 4 Take 2

Lead Column			Lag Column			Feed Displacement and Water Rinse					
BV	μCi ¹³⁷ Cs/ mL	% C/C ₀	DF	BV	μCi ¹³⁷ Cs/ mL	% C/C ₀	DF	BV	μCi ¹³⁷ Cs/ mL	% C/C ₀	DF
7.5	9.38E-2	7.52E-2	1331	7.6	5.20E-3	4.17E-3	23,998	186.8	2.98E-2	2.38E-2	4196
15.8	1.24E-2	9.90E-3	10,099	15.6	4.20E-3	3.37E-3	29,711				
29.2	8.78E-3	7.03E-3	14,223	28.7	3.79E-3	3.04E-3	32,942				
42.9	9.21E-3	7.38E-3	13,549	42.0	3.03E-3	2.43E-3	41,199				
59.5	3.30E-2	2.65E-2	3778	58.3	3.77E-3	3.02E-3	33,067				
72.9	9.76E-2	0.08	1279	71.4	2.79E-3	2.23E-3	44,774				
85.7	3.28E-1	0.26	380	84.0	2.26E-3	1.81E-3	55,171				
99.1	9.84E-1	0.79	127	97.0	2.41E-3	1.93E-3	51,864				
112.0	2.37E+0	1.90	53	109.7	2.44E-3	1.96E-3	51,136				
121.6	4.56E+0	3.66	27	119.0	2.57E-3	2.06E-3	48,637				
137.3	9.76E+0	7.82	13	134.4	2.58E-3	2.07E-3	48,420				
149.7	1.85E+1	14.80	7	146.6	3.18E-3	2.54E-3	39,310				
162.2	3.25E+1	26.03	4	158.8	3.32E-3	2.66E-3	37,582				
173.4	4.73E+1	37.93	3	169.8	3.89E-3	3.12E-3	32,063				
183.0	6.41E+1	51.39	2	179.0	4.95E-3	3.97E-3	25,204				
194.7	8.07E+1	64.66	2	190.4	9.22E-3	7.39E-3	13,534				
203.5	8.89E+1	71.20	1	198.9	1.45E-2	1.16E-2	8617				
212.1	9.62E+1	77.08	1	207.2	2.18E-2	1.75E-2	5719				
218.2	1.04E+2	83.05	1	213.1	3.42E-2	2.74E-2	3650				
BV = bed volume; DF = decontamination factor; FD = feed displacement; $C_0 = 125 \ \mu Ci \ ^{137}Cs/mL$											

 Table A.11.
 AP-105DF Test 5 Cs Load, Feed Displacement, and Water Rinse Results

	Elution			Water Rinse	
	μCi			μCi	
BV	¹³⁷ Cs/ mL	C/C_0	BV	¹³⁷ Cs/ mL	C/C_0
29.13	8.93E+2	7.15E+0	EDi		
			30.83	1.83E-1	1.46E-3
			32.42	1.45E-1	1.16E-3
			33.95	1.31E-1	1.05E-3
			35.53	1.39E-1	1.11E-3
			37.08	1.20E-1	9.61E-4
			EDiF		
			29.13	4.76E-2	3.82E-4
BV = bed v	olume; $C_0 = 125 \mu C$	i ¹³⁷ Cs/ mL.			

Table A.12. Elution and Water Rinse Results Following AP-105DF Processing Test 5

Appendix B

Analytical Reports

Appendix B

Analytical Reports

Analytical reports provided by the Analytical Support Operations (ASO) are included in this appendix. In addition to the analyte results, they define the procedures used for chemical separations and analysis, as well as quality control sample results, observations during analysis, and overall estimated uncertainties. The analyses are grouped according to Analytical Services Request (ASR) number.

Table of Contents

ASR 0272, Initial Characterization of AP-105

- ICP-OES, Metals
- ICP-MS, 133 Cs
- Titration, Free Hydroxide
- GEA

ASR 0329, Cs Isotopic Results

ASR 0335, Test 1 Results

- ICP-OES, Metals
- ICP-MS, Mercury
- Titration, Free Hydroxide
- Ion Chromatography, Anions
- TOC/TIC
- GEA
- Radionuclides (Sr, Pu, Np, total alpha, total beta)
- U

ASR 0372 Test 2 through Test 5 Results

- ICP-OES, Metals
- GEA
- Pu
- Tc
- U by ICP-MS

ASR 0272, Initial Characterization of AP-105

(Information on t Requestor Complete all field	his COVER PAGE is appl s on this COVER P	licable to all samples submitted under this ASR)
Requestor: Signature Sanctra k. fus Print Name Sandra Fiskum Phone 5-5677	MSIN <u>P7-25</u>	Project Number: <u>69833</u> Work Package: <u>N78228</u>
Matrix Type Informatio	n	QA/Special Requirements
 Liquids: Aqueous Organic Solids: Soil Soil Sludge Glass Filter Smear Organic Other: Solid/Liquid Mixture, Slurry Gas Biological Spector (If sample matrices vary, specify on R Disposition of Virgin Samples: Virgin samples are returned to requestor archiving provisions are made with recei If archiving, provide: Archiving Reference Doc: 	Multi-phase Sediment Metal Other cimen equest Page) unless ving group!	 QA Plan: ☑ ASO-QAP-001 (Equivalent to HASQARD) □ Additional QA Requirements, List Document Below: Reference Doc Number: Field COC Submitted? ☑ No □ Yes Lab COC Required? ☑ No □ Yes Sample/Container Inspection Documentation Required? ☑ No □ Yes Hold Time: ☑ No □ Yes If Yes, Contact ASO □ Use SW 846 (PNL-ASO-071, identify analytes/methods where holding times apply) submitting Samnles □ Other? Specify: Special Storage Requirements: ☑ Non □ Refrigerate □ Other, Specify:
◆ Disposition of Treated Samples: ☑ Dispose □ Return		◆ Data Requires ASO Quality Engineer Review? □ No ☑ Yes
	Data Report	ting Information
 ◆ Is Work Associated with a Fee-Based Milestone? ☑ No □ Yes If yes, milestone due date: ▲ Preliminary Results Requested, As Available? □ No ☑ Yes 	 ◆ Data Reporting L ☑ ASO-QAP-001 (E HASQARD). ☑ Minimum data rep ☑ Project Specific Re Contact ASO Lead on Document: 	evel quivalent to Requested Analytical Work Completion Date: 4/28/17 (Note: Priority rate charge for < 10 business day turn-around time) vort. equirements: r List Reference Negotiated Commitment Date: <u>5/5/17</u> (To be completed by ASO Lead)
	Waste Design	ation Information
ASO Sample Information Check List At If no, Reference Doc Attached: or, Previous ASR Number: or. Previous RPL, Number:	tached? □ No Ø Ye	S Does the Waste Designation Documentation Indicate Presence of PCBs? ☑ No □ Yes
Send Report To: Sandra Fiskum, Reic John Geeting, Jarrod Additional or Special Instructions	l Peterson, I Allred	MSIN P7-25, P7-22 MSIN K9-89, P7-25
Receiving Date Delivered: 9/25 Delivered By (optional) J. G Time Delivered: 09:4 Group ID (optional) CMC Waste Sample?	y and Login Informat	tion (to be completed by ASO staff) Received By: L. Darsell @ 09:40 ASR Number: 0272 Rev.: 00 RPL Numbers: 17-0868 (first and last)
ASO Work Accepted By:	V Pool si	ignature/Date: Jul 1. 1 al 4/25/17

Analytical Service Request (ASR)

Analytical Service Request (ASR) (REQUEST PAGE ----- Information Specific to Individual Samples)

ASO Staff Use Only RPL Number	Provide Analytes (Customer Sample ID	of Interest and Required Detectio Sample Description (& Matrix, if it varies)	on limits - Below Attached Analysis Requested	ASO Staff I Test	Jse Only Library
17-0868	SAP-16-01	AP-105 as-sampled from the Hanford	ICP-OES (AI, Na, K)		
		tank	Free hydroxide		
			GEA (Cs-137, and any other detectable		
			Cs isotopes)		
			ICP-MS (Cs-133)		
					- - - -
	AP-105 is expected	d to be 8.5 M Na			
	All analytes are ex detection limits. T	hected to be well above instrument herefore required MDLs not			
	provided.				3
				2	-
ASR # C275	P Rev.: OO		Page) of	

ASO Sample Information Checklist (SICL) Form

Sample Location, Owner and Ha To be completed by the sample custodian rel	zard Description Information inquishing the sample(s) and based on bes	t available information.				
ASO Customer Information:						
Company: PNNL						
Project #: <u>69833</u> Point of Contact (name. telephone#): <u>SK Fiskum</u> 375-5677						
Point of Contact (name. telephone#):	SK Fiskum 375-:	5677				
Comments:						
Sample Description (medium, collect	tion location, known contaminants, purpose	e of sample collection):				
Sample Collection Date: Sample Collection Time:						
Is the sample known to be radioa	Is the sample known to be radioactive?					
Comments (list known isotopes): AP-105 tank waste, G-137 ~ 200 uCi ful						
List any hazardous sample const	ituents known to be present:					
Constituent/Chemical	Concentration	Comment				
Cr	~ 460 mic Jul	TWINS				
hydroxide	hydrigide ~1,9 M TWINS					
· · · · · · · · · · · · · · · · · · ·						
Are any other comments applica						
	ble to sample receipt, storage, l	handling, or disposition?				
	ble to sample receipt, storage, l	handling, or disposition?				
	ble to sample receipt, storage, l	handling, or disposition?				
Checklist Prepared By:	ble to sample receipt, storage, l	handling, or disposition?				
Checklist Prepared By: Sandra Fiskun Sand	ble to sample receipt, storage, l	handling, or disposition?				

Battelle PNNL/RPL/Inorganic Analysis ... ICP-OES Analysis Report PO Box 999, Richland, Washington 99352

Project / WP#:	69833 / N78228		
ASR#:	0272		
Client:	S. Fiskum		
Total Samples:	1 (liquid)		

ASO Sample ID	Client Sample ID	Client Sample Description	Sample Weight (g)
17-0868	5AP-16-01	AP-105 as sampled from the Hanford Tank	NA
			_
Sample Pren	aration Client prepa	red	

a: Chent prepared. rep

Procedure: <u>RPG-CMC-211, Rev. 4</u> , "Determ Inductively Coupled Argon Plasm Spectrometry (ICP-OES)".			nation of Element Optical Emission	tal Composition n	ı by
Analyst:	J. Carter	Analysis Date:	05-01-2017	ICP File:	C0714
See Chemi	ical Measureme	nt Center 98620 file:	ICP-325-405-3 (Calibration an	d Maintenance	Records)
M&TE: PerkinElmer 5300DV ICP-OES			SN: 077N5122002		
	Sartorius N	/IE414S Balance	SN: 22406373		
	Sartorius N	/IE414S Balance		SN: 22006647	
	Sartorius F	200D Balance		SN: 39080042	
	Mettler AT	201 Balance	SN: 192720-92		

Chis Thomps

Report Preparer

11

Review and Concurrence

5/19/17

Date

5/19 17 Date

Page 1 of 4

One liquid sample submitted under Analytical Service Request (ASR) 0272 was analyzed by ICP-OES. The sample was analyzed after dilution of the "as received" liquid in 5% v/v HNO₃. Neither the sample nor any of the dilutions were filtered.

All results are reported on a mass per unit volume basis ($\mu g/mL$) for each detected analyte. The data have been adjusted for instrument dilutions.

Analytes of interest (AOI) were specified in the ASR and are listed in the upper section of the attached ICP-OES Data Report. The quality control (QC) results for the AOI have been evaluated and are presented below. Analytes other than the AOI are reported in the bottom section of the report, but have not been fully evaluated for QC performance.

Calibration of the ICP-OES was done following the manufacturer's recommended calibration procedure using multi-analyte custom standard solutions traceable to the National Institute of Standards and Technology (NIST). Midrange calibration verification standards (MCVA and MCVB) were used to verify acceptance of the two-point calibration curves obtained for each analyte and for continuing calibration verification.

The controlling document was ASO-QAP-001, Rev. 11, *Analytical Support Operations (ASO) Quality Assurance Plan.* Instrument calibrations, QC checks and blanks (e.g., ICV/ICB, CCV/CCB, LLS, ICS), post-digestion spikes, duplicate, blank spike, and serial dilution were conducted during the analysis run.

Preparation Blank (PB):

No preparation blank was required. The ICP-OES laboratory diluent (5% HNO₃) was analyzed as a blank. All AOI were within the acceptance criteria of \leq EQL (estimated quantitation level), \leq 50% regulatory decision level, or less than \leq 10% of the concentration in the samples.

Blank Spike (BS)/Laboratory Control Sample (LCS):

The MCVA solution was analyzed as the blank spike. Recovery values are listed for all analytes included in the BS that were measured at or above the EQL. All AOI were present in the MCVA solution. Recovery values for the AOI meeting this requirement ranged from 97% to 101%, and were within the acceptance criterion of 80% to 120%.

Duplicate/Replicate Relative Percent Difference (RPD)/Relative Standard Deviation (RSD): A replicate of sample 17-0868 was prepared and analyzed. RPD are listed for all analytes that were measured at or above the EQL. RPD for the AOI meeting this requirement ranged from 0.1% to 1.3% and were within the acceptance criterion of ≤20% for liquid samples.

Matrix-Spike (MS) Sample:

No matrix spike was required.

S. Fiskum ASR-0272 (Liquid Tank Waste) ICP File C0714.docPage 2 of 4

Battelle PNNL/RPL/Inorganic Analysis ... ICP-OES Analysis Report

Initial/Continuing Calibration Verification (ICV/CCV):

MCVA and MCVB solutions were analyzed immediately after calibration, after each group of not more than ten samples, and at the end of the analytical run. The concentrations of all AOI were within the acceptance criteria of 90% to 110% recovery.

Initial/Continuing Calibration Blank (ICB/CCB):

The ICB/CCB solution (5% v/v HNO₃) was analyzed immediately after the ICV solutions and after the CCV solutions (after each group of not more than ten samples and at the end of the analytical run). The concentration of all AOI were within the acceptance criteria of \leq EQL.

Low-Level Standard (LLS):

The LLS solution was analyzed immediately after the first CCB solution. The concentrations of all AOI were within the acceptance criteria of 70% to 130% recovery.

Interference Check Standard (ICS/SST):

The ICS solution was analyzed immediately after the first LLS solution and immediately prior to analyzing the final CCV solutions. The concentrations of all AOI were within the acceptance criteria of 80% to 120% recovery.

Serial Dilution (SD):

Five-fold serial dilution was conducted on sample 17-0868. Percent differences (%Ds) are listed for all analytes that had a concentration at or above the EQL in the diluted sample. The %Ds for the AOI meeting this requirement ranged from 2.2% to 5.1% and were within the acceptance criterion of $\leq 10\%$.

Post-Digestion Spike (PS-A)/Analytical Spike (AS-A) - Sample (A Component):

Because no MS sample was prepared, a post-digestion spike (A Component) was conducted on sample 17-0868. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration $\geq 25\%$ of that in the sample. Recovery values for the AOI meeting this requirement ranged from 100% to 107%, and were within the acceptance criterion of 80% to 120%.

Post-Digestion Spike (PS-B)/Analytical Spike (AS-B) - Sample (B Component):

Because no MS sample was prepared, a post-digestion spike (B Component) was conducted on sample 17-0868. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration $\geq 25\%$ of that in the sample. There were no AOI included in the spike B Component.

Other QC:

All other instrument-related QC tests for the AOI passed within their respective acceptance criteria.

S. Fiskum ASR-0272 (Liquid Tank Waste) ICP File C0714.docPage 3 of 4

Comments:

- The "Final Results" have been corrected for all laboratory dilutions performed on the samples during
 processing and analysis, unless specifically noted.
- 2) Instrument detection limits (IDL) and estimated quantitation limits (EQL) shown are for acidified water and/or fusion flux matrices as applicable. Method detection limits (MDL) for individual samples can be estimated by multiplying the IDL by the "Process Factor" for that individual sample. The estimated quantitation limit (EQL) for each concentration value can be obtained by multiplying the EQL by the "Process Factor".
- 3) Routine precision and bias is typically ±15% or better for samples in dilute, acidified water (e.g. 2% v/v HNO₃ or less) at analyte concentrations > EQL up to the upper calibration level. This also presumes that the total dissolved solids concentration in the sample is less than 5000 µg/mL (0.5 per cent by weight). Note that bracketed values listed in the data report are within the MDL and the EQL, and have potential uncertainties greater than 15%. Concentration values < MDL are listed as "--". Note, that calibration and QC standard samples are validated to a precision of ±10%.</p>
- 4) Absolute precision, bias and detection limits may be determined on each sample if required by the client. The maximum number of significant figures for all ICP measurements is two.
- 5) Analytes included in the spike A component (for the AS/PS) are; Ag, Al, As, B, Ba, Be, Bi, Ca, Cd, Co, Cr, Cu, Fe, K, Li, Mg, Mn, Mo, Na, Ni, P, Pb, Sb, Se, Si, Sn, Sr, Ta, Ti, Tl, V, W, Y, Zn, and Zr. Analytes included in the spike B component are; Ce, Dy, Eu, La, Nd, Pd, Rh, Ru, S, Te, Th, and U.

		Run Date >	5/1/2017	5/1/2017	5/1/2017
		Process Factor >	1.0	500.0	500.0
			405 411	17-0868@	17-0868@
		RPL/LAB >	405 diluent	500x	500x rep
Instr. Det	Est Quant				
Limit (IDL)	Limit (EQL)	Client ID >	Lab Diluent	SAP-	16-01
(µg/mL)	(µg/mL)	(Analyte)	(µg/mL)	(µg/mL)	(µg/mL)
0.0276	0.276	AI		20,600	20,900
0.0500	0.500	к		5,520	5,530
0.0171	0.171	Na		195,000	197,000
ther Analyt	es				
0.0030	0.030	Ag		-	
0.1576	1.576	As		-	
0.0120	0.120	В	[0.032]	71.7	68.0
0.0005	0.005	Ba		[0.60]	[0.52]
0.0002	0.002	Be		[0.19]	[0.32]
0.0546	0.546	Bi	[0.065]	[29]	
0.0233	0.233	Ca		[90]	[99]
0.0027	0.027	Cd		[1.8]	[1.9]
0.0448	0.448	Ce		-	
0.0071	0.071	Co			
0.0047	0.047	Cr		505	505
0.0036	0.036	Cu		[3.9]	[3.2]
0.0034	0.034	Dy		[1.9]	
0.0010	0.010	Eu			
0.0017	0.017	Fe	[0.0045]	(5.9)	[6.2]
0.0030	0.030	La			
0.0028	0.028	Li			
0.0023	0.023	Ma			
0.0009	0.009	Mn			
0.0094	0.094	Mo	(0.021)	83.0	84.0
0.0188	0,188	Nd			
0.0074	0.074	Ni		45.8	45.3
0.1021	1.021	Р		752	715
0.0399	0.399	Pb		[25]	[31]
0.0170	0.170	Pd			
0.0162	0.162	Rh			
0.0158	0.158	Ru		[13]	[11]
0.1773	1.773	S		1,820	1,800
0.0919	0.919	Sb			
0.2271	2.271	Se			[230]
0.0136	0.136	Si		[67]	[61]
0.0446	0.446	Sn			
0.0003	0.003	Sr		[0.28]	[0.29]
0.0242	0.242	Та			
0.0412	0.412	Te			
0.0120	0.120	Th			
0.0014	0.014	Ti			
0.0885	0.885	TI			
0.0676	0.676	ų			
0.0032	0.032	v			
0.0249	0.249	w		141	154
0.0006	0.006	Y	[0.0008]	[0,35]	[0,42]
0.0066	0.066	Zn		[28]	[27]
0.0022	0.022	Zr		[1,2]	
				1.1.01	

1) "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the ", near the top of each column. The estimated sample quantitation limit = EQL (in Column. times the "multiplier". Overall error for values \geq EQL is estimated to be within ±15%. 2) Values in brackets [] are \geq MDL but < EQL, with errors likely to exceed 15%.

ASR 0272 Final from C0714 Fiskum ASR-0272 Tank Waste.xls

Criteria >	≤ 20%	80%-120%	75%-125%	80%-120%	80%-120%	≤ 10%
QC ID >	17-0868 Rep	LCS/BS	MS (none)	17-0868 + AS-A	17-0868 + AS-B	17-0868 5-fold Serial Dil
Analytes	RPD (%)	%Rec	%Rec	%Rec	%Rec	%Diff
AI	1.3	101		104		3.3
к	0.1	100		100		2.2
Na	1.3	97		107		5.1
Other Analy	tes					
Ag		93		93		
As		99		101		
в	5.4	101		101		
Ba		98		97		
Be		99		99		
Bi		95		94		
Ca		102		101		
Cd		102		102		
Ce		.02		102	92	
Co		00		00	52	
C.	0.0	100		100		0.7
Cu	0.0	100		100		0.7
Cu		99		101		
Dy					99	
Eu					100	
Fe		99		98		
La		· /			99	
Li		102		102		
Mg		99		98		
Mn		101		101		
Mo	1.1	98		98		
Nd					99	
Ni	1,1	98		97		
Р	5.0	102		101		
РЪ		99		99		
Pd					91	
Rh					93	
Ru					95	
S	1.0				99	
Sb		96		98		
Se		103		105		
Si		99		106		
Sn		95		95		
Sr		99		98		
Ta				98		
Te					99	
Th					101	
Ti		09		07	101	
TI		30		9/		
		91		92	100	
U				00	100	
V		99		98		
W	8.8	101		100		
Y		100		100		
Zn		101		102		
Zr		99		98		

QC Performance 05/01/2017

Shaded results are outside the acceptance criteria.

nr = spike concentration less than 25% of sample concentration. Matrix effects can be assessed from the serial dilution.

ASR 0272 Final from C0714 Fiskum ASR-0272 Tank Waste.xls

4

Battelle PNNL/RPL/Inorganic Analysis ... ICP-MS Analysis Report PO Box 999, Richland, Washington 99352

Project / WP#:	69833 / N78228
ASR#:	0272
Client:	S. Fiskum
Total Samples:	1 (liquid)

ASO Sample ID	Client Sample ID	Client Sample Description	Sample Weight (g)
17-0868	5AP-16-01	AP-105 as sampled from Hanford tank	NA

Sample Preparation: Dilution of "as received" samples in 0.5 M HNO₃ and 2 performed by L. Darnell on 4/26/17 and G. Brown on 7/7/17, respectively.

Procedure	e: <u>RPG-CMC-29</u> Inductively Co	<u>92, Rev. 0</u> , "Determi oupled Argon Plasma	nation of Elemen Mass Spectrom	ntal Composition etry (ICP-MS)".	n by		
Analyst:	G. Brown	Analysis Date:	07-10-2017 ICP File: M0025				
See Chem	ical Measuremen	t Center 98620 file:	ICP-MS-325-405-1 (Calibration and Maintenance Records)				
M&TE:	PerkinElmer	r NexION TM 350X IC	CP-MS	SN: 85VN4070702			
	Sartorius M	E414S Balance		SN: 22406373			
	Mettler AT4	00 Balance		SN: 1113292667			
	Sartorius R2	200D Balance		SN: 39080042			
	Mettler AT2	201 Balance		SN: 192720-92	2		

aur

Report Preparer

1301 10 Review and Concurrence

7/24/17 Date 7/25/17

Page 1 of 5

B.12

PNNL ASO ICP-MS Analysis Results

RPL ID	Client ID	Run Date	Process Factor	Sample	Replicate	Triplicate	Units
17-0868	5AP-16-01	07/10/17	1,934,970	[8230]	[8800]	[7730]	µg/L Cs-133
17-0868	5AP-16-01	07/10/17	389,931	[8070]	[7100]	nm	µg/L Cs-133
					Average:	8.000	ug/L Cs-133

QC Performance 07/10/2017

Criteria >	≤ 20%	80%-120%	75%-125%	80%-120%	≤ 10%			
QC ID >	17-0868 Trip	LCS/BS (None)	MS (None)	17-0868 + PS	17-0868 5-fold Serial Dil	Analyte IDL	Analye EQL	<u>Lab</u> Diluent
Analytes	RSD (%)	%Rec	%Rec	%Rec	%Diff	(µg/L)	(µg/L)	(µg/L)
Cs-133	6.44%	95.3%	na	95.0%	3.38%	0.0030	0.030	

1) "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the "Process Factor" for each sample analysis. The estimated sample quantitation limit = EQL times the "Process Factor." Overall error for values \geq EQL is estimated to be within ±15%.

2) Values in brackets [] are ≥ MDL but < EQL, with errors likely to exceed 15%.

Matrix effects can be assessed from the serial dilution. Shaded results are outside the acceptance criteria. nr = spike concentration less than 25% of sample concentration.

na = not applicable (e.g., LCS, BS or MS may not be required for liquid samples).

nm = not measured (e.g., triplicate analysis may not be required by procedure).

Battelle PNNL/RPL/Inorganic Analysis ... ICP-MS Analysis Report

One liquid sample submitted under Analytical Service Request (ASR) 0272 was analyzed by ICP-MS for Cs-133. The sample was analyzed after dilution of the "as received" liquid in 0.5 <u>M</u> HNO₃ by L. Darnell on 4/26/17 and 2% v/v HNO₃ by G. Brown on 7/17/17. The sample and dilutions were not filtered.

All results are reported on a mass per unit volume basis ($\mu g/L$) for each detected analyte. The data have been adjusted for instrument dilutions.

The analyte of interest (AOI), Cs-133, was specified in the ASR. Results for Cs-133 are listed in the upper section of the attached ICP-MS Data Report. The quality control (QC) results for Cs-133 have been evaluated and are presented below. Results for other analytes are not included in the report.

Calibration of the ICP-MS was done following the manufacturer's recommended procedures using custom standard solutions traceable to the National Institute of Standards and Technology (NIST). Midrange calibration verification standards were used to verify acceptance of the sixpoint calibration curve and for initial and continuing calibration verification (ICV/CCV).

The controlling document was the analytical procedure and ASO-QAP-001, Rev. 11, *Analytical Support Operations (ASO) Quality Assurance Plan.* Instrument calibration, QC checks and blanks (e.g., ICV/ICB, CCV/CCB, LLS, ICS), post-digestion spike, triplicates, and serial dilution were conducted during the analysis run.

Internal Standard (IS):

All solutions (blanks, standards, QC checks, and samples) were mixed in-line with 10 ppb Rb-85 as an internal standard. All data were normalized using the Rb-85 mass. The IS recovery ranged from 96.8% to 103.5%, which is within the acceptance criterion of 30% to 120%.

Preparation Blank (PB) Sample:

No preparation blank (PB) sample was required to be prepared since the client sample was received as a liquid and only simple dilution was required prior to analysis. The ICP-MS laboratory diluent ($2\% v/v HNO_3$) was analyzed in place of a PB sample as the initial and continuing calibration blanks (described below).

Blank Spike (BS) Sample/Laboratory Control Sample (LCS):

A blank spike (BS) sample was prepared by spiking a 2% HNO₃ blank with an equivalent volume of the 2 ppb Cs-133 CCV standard (1:1 ratio). The Cs-133 recovery was 95.3% and was within the acceptance criteria of 80% to 120% recovery.

Triplicate/Relative Standard Deviation (RSD):

A triplicate analysis was performed on sample 17-0868. The RSD is listed for analytes that were measured at or above the EQL. The RSD for Cs-133 was 6.44% and was within the acceptance criterion of $\leq 20\%$ for liquid samples.

Matrix-Spike (MS) Sample:

No matrix spike (MS) sample was required.

Initial/Continuing Calibration Verification (ICV/CCV):

The ICV/CCV solution was analyzed immediately after calibration, after each group of not more than ten samples, and at the end of the analytical run. The Cs-133 recoveries ranged from 94.9% to 97.4% and were within the acceptance criteria of 90% to 110% recovery.

Initial/Continuing Calibration Blank (ICB/CCB):

The ICB/CCB solution (2% v/v HNO₃) was analyzed immediately after the ICV solutions and after the CCV solutions (after each group of not more than ten samples and at the end of the analytical run). The concentrations of Cs-133 in all ICB/CCBs were within the acceptance criteria of \leq EQL.

Low-Level Standard (LLS):

The LLS solution was analyzed immediately after the first CCB. The recovery for Cs-133 (97.6%) was within the acceptance criteria of 75% to 125%.

Interference Check Standard (ICS):

The ICS solution was analyzed immediately after the LLS solution and immediately after the final CCV solution. The recoveries for Cs-133 (98.7% and 99.9%) were within the acceptance criteria of 80% to 120%.

Serial Dilution (SD):

Five-fold serial dilution was conducted on sample 17-0868. Percent differences (%D) are listed for all analytes that had a concentration at or above the EQL in the diluted sample. The %D for Cs-133 was 3.38% and was within the acceptance criterion of $\leq 10\%$.

Post-Digestion Spike (PS)/Analytical Spike (AS) - Sample (Cs-133 Component):

Because no MS sample was prepared, a post-digestion spike (Cs-133 Component) was conducted on sample 17-0868. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration $\geq 25\%$ of that in the sample. The recovery for Cs-133 was 95.0% which is within the acceptance criterion of 75% to 125%.

Other QC:

All other instrument-related QC tests for the AOI passed within their respective acceptance criteria.

Comments:

- 1) The "Final Results" have been corrected for all laboratory dilutions performed on the samples during processing and analysis, unless specifically noted.
- 2) Instrument detection limits (IDL) and estimated quantitation limits (EQL) shown are for acidified water and/or fusion flux matrices as applicable. Method detection limits (MDL) for individual samples can be estimated by multiplying the IDL by the "Process Factor" for that individual sample. The estimated quantitation limit (EQL) for each concentration value can be obtained by multiplying the EQL by the "Process Factor".
- 3) Routine precision and bias is typically ±15% or better for samples in dilute, acidified water (e.g. 2% v/v HNO₃ or less) at analyte concentrations > EQL up to the upper calibration level. This also presumes that the total dissolved solids concentration in the sample is less than 500 µg/mL (0.05 per cent by weight). Note that bracketed values listed in the data report are within the MDL and the EQL, and have potential uncertainties greater than 15%. Concentration values < MDL are listed as "--". Note, that calibration and QC standard samples are validated to a precision of ±10%.</p>
- 4) Absolute precision, bias and detection limits may be determined on each sample if required by the client. The maximum number of significant figures for all ICP-MS measurements is three.



Client:	S. Fiskum			Report Date: 5/10/2017 Analysis Date: 5/1/2017
Subject:	Hydroxide	e Analyses for:		Aqueous Samples
Project:	69832	WP #	N79888	
	69833	WP #	N78228	
ASR:	0262 & 02	272 Rev-0		Procedure: RPG-CMC-228-Rev 0.1
Sample ID.	17-0791	and	17-0868	

Diluted sample aliquots of 2 aqueous samples (see above assigned RPL Sample #'s), provided on ASR's 0262 and 0272 were analyzed by manual titration for the base constituents content following procedure RPG-CMC-228. Each sample was diluted by a factor of 26x prior to titration (0.100 mL of sample added to 2.5 mL of DI water). The titrant used was 0.0957 M HCI (Standardized HCI was prepared and documented on Chem Rec 232, prepared on 3/15/2017). pH measurements were obtained using a Beckman Coulter 560 pH meter, SN 110650046. The pH meter was calibrated using 3 buffers, pH 4,7 and 10 and the calibration verified using an independent pH 7 buffer.

The initial diluted pH is reported on attached Report Summary along with the free hydroxide molarity (the 1st inflection point at pH of 11.0 - 11.1) was very weak but discernable. A 2nd inflection point was determined for both samples at pH 7.6 to 7.7. The third inflection point (at pH 4.8 to 5.0) was easily identified and reportable.

Following is the report summary and the sample results calculated from the raw data. A copy of the titration curve data for each sample is also included with this report.

Date: Prepared by: Date: Reviewed by:

ASR0262-0272-Rev-0 (Fiskum).xls

Page 1 of 1

5/10/2017

Battelle Pacific Northwest Laboratory Nuclear Chemistry and Engineering Group Chemical Measurements Center

ASR # 0262 & 0272

WP# N79888 & N78228

Hydroxide and Alkalinity Determination Procedure: RPG-CMC-228-Rev 0.1

Equip # Beckman Coulter 560, SN#110650046

		Report Summary fo	r ASR # Revision #	0262 & 0272 Rev-0	Report Da Analysis Da	te: 5/10/2017 te: 5/1/2017
				Co	ncentration, moles / Li	ter
		Diluted	011	First Point	Second Point	Third Point
RPG #	Client ID	Initial pH	UH conc ug/mL	Molarity	Molarity	Molarity
17-0791	SS07Feed A	12.57	2.51E+04	1.47	0.74	0.42
17-0868	5AP-16-01	12.50	2.98E+04	1.75	1.49	0.73

OH conc (mg/L) = M (g/L) * 17,000

Note: Results are presented for the first, second, and third inflection points on the titration curves, as applicable. The first inflection point is generally associated with the free hydroxide concentration. The second inflection point generally represents total hydroxide, or carbonate or a combination of aluminate and carbonate. The third inflection point is usually indicative of bicarbonate or other weak acids or possibly the continued protonation of alumina.

5.10.17 Analyst Reviewer:

					Molarity	base	1.474	1.751					
5/10/2017 5/1/2017	4/1/2007	10.17			Found	base	0.147	0.175					
port Date: ysis Date:	cked07.xls	2 S.		nce		μd	11.000	11.100					
Re Anal	-TemplateLo	fut	HO	1st Equivale	Point Titrant	Vol. (mL)	1.540	1.830					
Rev-0 N79888 N78228	eadsheet: OH	Josef	2	Diluted	Initial nH	reading	12.570	12.500					
0262 & 0272 S. Fiskum WP#	Spr	Analyst:			Titrator Routine	#	NA	NA					
and Rev # Client: 69832 69833					Density	g/mL	NA	NA					
ASR # Project:		301			Sample	Wt. (g)	NA	NA					
	and	Lab Loc.			Sample	Vol. (mL)	0.100	0.100					
	droxyl (OH-) es	110650046			Dilution	Factor	NA	NA					
	Determination of Hy achates and Supernat	Equip # Chem	Rec#	232									
rtthwest Laboratory • and Engineering Group	RPG-CMC-228-Rev 0.1 Ikalinity of Aqueous Solutions, Le	ц	Molarity	0.0957		ample ID	S07Feed A	AP-16-01					
Battelle Pacific No Nuclear Chemistry	Procedure:	By Manual Titratic	Titrant	HCI		RPG # Si	I7-0791 S	17-0868 5/					

	Vendor	Inorganic Ventures	Increanic Vanturae
strument Calibration	Buffer	4	r

Instrument Calibrat	ion		
Buffer	Vendor	Lot Number	Expire Date
4	Inorganic Ventures	J2-WCS02116	1-Jul-18
2	Inorganic Ventures	J2-WCS02097	1-Mar-18
10	Inorganic Ventures	H2-WCS02089	l-Jan-18
2-nd Verif	Vendor	Lot Number	Expire Date
7	Ricca	1603E33	1-Mar-18

Slope 95.2%

5/10/2017

Page 1 of 2

ASR0262-0272-Rev-0 (Fiskum).xls

Battelle Pacific Northwest Laboratory Nuclear Chemistry and Engineering Group

ASR # and Rev # 0262 & 0272

N79888 N78228

#dM

Procedure:

Alkalinity of Aqueous Solutions, Leachates and Supernates

By Manual Titration

Equip # Beckman Coulter 560, SN#110650046

Litrant	Molarity					
HCI	0.0957		2nd Equival	ence	63	
	1		Point		Found	
XPG #		Sample Vol. (mL)	Titrant Vol. (mL)	μd	millimoles base	Molarity bas
12-0791		0.100	2.310	7.700	0.074	0.73
17-0868		0.100	3.390	7.600	0.149	1.49

Srd Equivale Point Titrant Vol. (mL)	nce PH	Found millimoles base	Molarity base
2.750	4.800	0.042	0.421
4.150	5.000	0.073	0.727

ASR0262-0272-Rev-0 (Fiskum).xls

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

Gamma Energy Analysis (GEA)

Revision 1 (This report was revised to correct the sample preparation information)

Project / WP#:	69833/ N78228
ASR#:	0272.00
Client:	SK Fiskum
Total Samples:	1

RPL ID	Client Sample ID		
17-0868	SAP-16-01		

Analysis Type:	GEA- for all positively measured or non-detected isotopes			
Sample Processing Prior to Radiochemical Processing/Analysis	 None Digested as per RPG-CMC-129, Rev. 0 HNO₃-HCl Acid Extraction of Solids Using a Dry Block Heater Fusion as per RPG-CMC-115, Solubilization of Metals from Solids Using a KOH-KNO₃ Fusion Other: Preparation may also involve attaining a GEA geometry that is compatible with the calibration geometry. 			
Analysis Procedure:	RPG-CMC-450, Rev. 3 Gamma Energy Analysis (GEA) and Low-Energy Photon Spectrometry (LEPS)			
Reference Date:	Not Applicable			
Analysis Date or Date Range:	April 26, 2017			
Technician/Analyst:	T Trang-Le			
Rad Chem Electronic Data File:	17-0868 Fiskum.xls			
ASO Project 98620 File:	File Plan 5871, T4.4 Technical (Radiochemistry), Gamma Calibration, daily checks, and maintenance records; and T3 standard certificates and preparation. Also, balance calibration and Performance check records.			
M&TE Number(s):	Detectors T			

Trang-le 17/11/17 2R. Greenwood 17/11/17 Prepare Date Reviewer Date

SAMPLE RESULTS

Activities for all gamma emitters detected in this sample are presented in an attached Excel spreadsheet for ASR 0272.00. All sample results for all target isotopes are reported in units of μ Ci/sample with estimates of the total propagated uncertainty reported at the 1-sigma level.

ASO Project File, ASR 0272 has been created for this report including all appropriate supporting records which may include the Pipette Performance Verification form, Standard Certificates, Laboratory Bench Record, Shielded Analytical Laboratory Bench Sheet, and Gamma Energy Analysis printouts. Detector calibration records, control charts and balance calibration records can be found in the ASO Records.

Sample Preparation, Separation, Mounting and Counting Methods

The quality control (QC) procedures for direct GEA are discussed below.

QUALITY CONTROL RESULTS

Tracer:

Tracers are not used for ASO GEA methods.

Process Blank (PB):

The sample was sent to the counting room with 0.5ml of sample that was diluted to a total volume of 2.0ml prior to the GEA count. There is no process blank for this ASR.

Required Detection Limits

There are no required detection limits for these samples.

Blank Spike (BS)/Laboratory Control Sample (LCS)/ Matrix Spike (MS):

There are no BS, LCS or MS samples analyzed for ASO GEA analyses. Instrument performance is assessed by the analyses of daily control counts and weekly background counts, as discussed below.

Duplicate Relative Percent Difference (RPD):

There are no duplicate samples for this ASR.

Instrument Calibration and Quality Control

Gamma detectors are calibrated using multi-isotope standards that are NIST-traceable and prepared in the identical counting geometry to all samples and detectors. Counter control sources containing Am-241, Cs-137 and Co-60 are analyzed daily before the use of each detector. Procedure RPG-CMC-450 requires that a counter control source is checked daily and must be within ± 3 sigma or $\pm 3\%$ of the control value, whichever is greater. Gamma counting was not performed unless the control counts were within the required limits. Background counts are performed on all gamma detectors at least weekly for either an overnight or weekend count. The most recent background is subtracted from all sample counts.

Assumptions and Limitations of the Data

None

Interferences/Resolution

None.

Uncertainty

For gamma counting, the uncertainty in the counting data, photon abundance and the nuclear half-life, and efficiency are included in the calculation of the total uncertainty along with a systematic uncertainty for sample prep. The Canberra Genie software includes both random and systematic uncertainties in the calculation of the total uncertainties which are listed on the report. We conservatively estimate that 2% is the lowest uncertainty possible for our GEA measurements taking into account systematic uncertainties in gamma calibration standards.

Comments

None

Attachment: Data Report Sample Results for ASR 0272.

Page 3 of 3

filename 17-0868 Fiskum 4/26/2017

Radiochemical Sciences and Engineering Group **Pacific Northwest National Laboratory** PO Box 999, Richland, WA

4/28/2017 Trang-le y/a6/2017 2 Mulemord Prepared by: Technical Reviewer: Project: 69833 WP: N78228

Client: Fiskum ASR 0272

RPG-CMC-450, Rev. 3 Gamma Energy Analysis (GEA) and Low-Energy Photon Spectrometry (LEPS) Procedures:

Gamma detectors T M&TE: Count dates:

26-Apr-17

Measured Activity, μ Ci per mL ± 1s 17-0868 | 1.80E+02 ± 2% Cs-137 Lab ID SAP-16-01 Sample

B.24

ASR 0329, Cs Isotopic Results

Analytical Service Request (ASR)

PAGE, unless specified as optional or ASR is a revision
Project Number: <u>69832</u> Work Package: <u>N82456</u>
QA/Special Requirements
 ◆ QA Plan: ▲ ASO-QAP-001 (Equivalent to HASQARD) □ Additional QA Requirements, List Document Below: Reference Doc Number: ▲ Bit ACOC Sub-pit All Parks
 Field COC Submitted? IN No I Yes Lab COC Required? IN No I Yes Sample/Container Inspection Documentation Required? IN I Yes
◆ Hold Time: ⊠ No □ Yes
$\frac{\text{If Yes.}}{\text{Contact ASO}} \square \text{ Use SW 846 (PNL-ASO-071 identify}$
Lead before analytes/methods where holding times apply) submitting Other? Specify: Samples Other? Specify: Ø None Refrigerate Other, Specify:
◆ Data Requires ASO Quality Engineer Review? ☑ No □ Yes
Equivalent to Equivalent to eport. Requirements: or List Reference (10 business day turn-around time) (Note: Priority rate charge for < 10 business day turn-around time) (Note: Priority rate charge for < 10 business day turn-around time) (Note: Priority rate charge for < 10 business day turn-around time) (Note: Priority rate charge for < 10 business day turn-around time) (Note: Priority rate charge for < 10 business day turn-around time) (Note: Priority rate charge for < 10 business day turn-around time)
investion Information
es Does the Waste Designation Documentation Indicate Presence of PCBs?
MSIN <u>\$7-25</u> MSIN
nation (to be completed by ASQ staff)
ASR Number: 0329 Rev.: 00 RPL Numbers: 017-1224

PL NumberCustomer Sample IDSample Description (x Marin, i't varies)Analysis Requested -1324 $2: 4rean, AP: 105$ $7: 1MS$ $mus dishrib toth and7: 10: 14 - Eg: AP0: 4rean, AP: 1057: 1MSmus dishrib toth and10: 14 - Eg: AP0: 4rean, AP: 1050: 1330: 13310: 14 - Eg: AP0: 4rean, AP: 1050: 1330: 13310: 14 - Eg: AP0: 4rean, AP: 1050: 1330: 13410: 14 - Eg: AP1: 37G0: 1370: 1331: 37G0: 4real, AP0: 137G0: 1371: 37G0: 13F0: 13F0: 13F1: 37G0: 12F0: 12F0: 12F<$	ed ASO Sti) Staff Use Only
-1334 $T1014 - E8-A$ $C: 4cm AP.105$ $T1014 - E8-A$ $C: 134$ $C: 134$ $C: 134$ $C: 137$ C	Test	Library
CS-134 CS-137	-B	
CS-133 Intalysy in duplicate, report duplicate nearlis. Sample contains 1.8 µcl: 1345 Aample contains 1.8 µcl: 1345 13705 should be about 20-257. of the total C5 controls had		
Dralyzy in duplicate, report duplicate travilis. Dample contains 1.8 pile: 1375 ml solution = 202 yis 32 pat 13705 should be about 20-25 % of the total Cs content.		
Pralyze in deplicate, report deplicate seements. Demple contains 's reli "376/me solution. = a.o.z. justo funct. "376 should be about 20-25 & g the total C3 content.		
Indigg in deplicate, report deplicate searchs. sample contains 18 pelo 1376 ml solution = 0.02 mg 32 put		
Dample centario ¹ . ^g Jule: ¹³⁺ CS / ml solution. = a.o.2, ³ 2, / ml. ¹³⁺ CS divid be about 20-25 2 g 444 total CS centert.		
Sample centario 1.8 relei ¹³⁷ C5 ml solution = 0.02 just 325 fut. 137C5 sheeld be about 20-25 % of the total C3 centent.		
137CS divid be about 20-25 % of the total Cs centert.		
13703 should be about 20-25% of the total CS centent.		
137CS should be about 20-25 % of the total Cs content.		
		1

Analytical Service Request (ASR)

Sample Location, Owner and Has To be completed by the sample custodian reli	zard Description Information nquishing the sample(s) and based on best availe	able information.
ASO Customer Information:		
Company: <u>PNNL</u> Project #: 69832		
Point of Contact (name telephone#):	Sandra Fiskum 5-56	77
Comments:		
Sample Description (medium, collection	ion location, known contaminants, purpose of sar	mple collection):
Sample Collection Date: $\frac{7}{20}$	Sample Collection Time	23:16
Is the sample known to be radioa	ctive? Ves No	
is the sample known to be factor		
Comments (list known isotopes): C5	-137	
Cs	134	
Is the sample known to contain o List any hazardous sample consti	r have come in contact with PCBs? tuents known to be present:	Yes No
Constituent/Chemical	Concentration	Comment
Are any other comments applicat	ole to sample receipt, storage, handl	ling, or disposition?
	110	
marrix is 0.45 m p	103	
Checklist Prepared By:		3
Sandra Fiskum	Sanda Jish	7/26/17
Printed Name Signature	Date V	47 X

Exhibit 1. Example ASO Sample Information Checklist (SICL) Form

Battelle PNNL/RPL/Inorganic Analysis ... ICP-MS Analysis Report PO Box 999, Richland, Washington 99352

Project / WP#:	69832 / N82456
ASR#:	0329
Client:	S. Fiskum
Total Samples:	1 (Aqueous)

ASO Sample ID	Client Sample ID	Client Sample Description	Sample Weight (g)
17-1224	TI014-E8-A	Cs from AP-105	NA
			_

Sample Preparation: Dilution of "as received" samples in 2% v/v HNO₃ was performed by G. Brown on 8/9/17.

Procedure	e: <u>RPG-CMC-2</u> Inductively (292, Rev. 0, "Determi Coupled Argon Plasma	nation of Elemen Mass Spectrome	tal Con etry (IC	nposition P-MS)."	ı by	
Analyst:	G. Brown	Analysis Date:	08/09/2017	IC	P File:	M0026	
See Chem	ical Measureme	nt Center 98620 file:	ICP-325-405-3 (Calibration ar	<u>3</u> nd Mair	ntenance	Records)	
M&TE:	PerkinElmer 5300DV ICP-MS		SN: 077N5122002		RPL 4	RPL 405 Bench	
	Sartorius MI	Sartorius ME414S Balance		SN: 22406373 R		RPL 405 Bench	
	Sartorius R200D Balance		SN: 39080042		RPL 405 Bench		
	Mettler AT4	Mettler AT400 Balance		SN: 192720-92		RPL 405 FH #3	
	Mettler AT400 Balance		SN: 1113292667		RPL 4	RPL 420 FH #13	
	Sartorius R2	00D Balance	SN: 39080058	3	RPL 5	25 FH #9	

an

Report Preparer

8/15/17 Date 8/31/17 7

Dea fron

Review and Concurrence

Date

Page 1 of 4

Battelle PNNL/RPL/Inorganic Analysis ... ICP-MS Analysis Report

One aqueous sample submitted under Analytical Service Request (ASR) 0329 was analyzed by ICP-MS for Cs-133, Cs-135, and Cs-137. The sample was analyzed after dilution of the "as received" liquid in 2% v/v HNO₃ by G. Brown on 08/09/17. The sample and dilutions were not filtered.

All results are reported on a mass per unit volume basis (ng/mL) for each detected analyte. The data have been adjusted for instrument dilutions.

Analytes of interest (AOI) were specified in the ASR and are listed in the upper section of the attached ICP-MS Data Report. The quality control (QC) results for the AOI have been evaluated and are presented below. Analytes other than the AOI are reported in the bottom section of the report, but have not been fully evaluated for QC performance.

Calibration of the ICP-MS was done following the manufacturer's recommended calibration procedure using multi-analyte custom standard solutions traceable to the National Institute of Standards and Technology (NIST). Midrange calibration verification standards were used to verify acceptance of the six-point calibration curve and for initial and continuing calibration verification (ICV/CCV).

The controlling document was procedure RPG-CMC-292, Rev 0, *Determination of Elemental Composition by Inductively Coupled Argon Plasma Mass Spectrometry (ICP-MS)*. Instrument calibrations, QC checks and blanks (e.g., ICV/ICB, CCV/CCB, LLS, ICS), post-digestion spike, duplicate, blank spike, and serial dilution were conducted during the analysis run.

Internal Standard (IS):

All solutions (blanks, standards, QC checks, and samples) were mixed in-line with a solution containing 10 ppb each of Li-6, Sc-45, Y-89, In-115, Tb-159, and Bi-209 as the internal standard (IS). The Cs-133, Cs-135, and Cs-137 (AOI) data were normalized using the In-115 mass as the IS. The In-115 IS recoveries ranged from 99% to 103%, which were within the acceptance criterion of 30% to 120%. An anomalously high value (130%) for the Sc-45 IS recovery was observed for sample 17-1224, likely caused by >8000 ng/mL boron in the sample. Contamination of the sample with B, Na, Si and other components from the glass container was likely.

Preparation Blank (PB):

No preparation blank was required to be prepared. A diluent blank from the ICP-MS laboratory (2% HNO₃) was analyzed in place of the preparation blank. All AOI were within the acceptance criteria of <EQL (estimated quantitation level), <50% regulatory decision level, or less than $\leq 10\%$ of the concentration in the samples.

Blank Spike (BS)/Laboratory Control Sample (LCS):

A blank spike (BS) sample was prepared by spiking a 2% HNO₃ blank with an equivalent volume of the CCV-71A-2ppb standard (1:1 ratio). The recoveries were 95%, 94%, and 98% for masses 133, 135, and 137, respectively and were within the acceptance criteria of 80% to 120% recovery.

Battelle PNNL/RPL/Inorganic Analysis ... ICP-MS Analysis Report

Duplicate/Replicate Relative Percent Difference (RPD)/Relative Standard Deviation (RSD):

A replicate of sample 17-1224 was prepared and analyzed. RPD are listed for all analytes that were measured at or above the EQL. RPD for the AOI meeting this requirement ranged from 1.5% to 2.7% and were within the acceptance criterion of \leq 20% for liquid samples.

Matrix-Spike (MS) Sample:

No matrix spike sample was required to be prepared.

Initial/Continuing Calibration Verification (ICV/CCV):

The ICV/CCV solution was analyzed immediately after calibration, after each group of not more than ten samples, and at the end of the analytical run. The concentrations of all AOI were within the acceptance criteria of 90% to 110% recovery.

Initial/Continuing Calibration Blank (ICB/CCB):

The ICB/CCB solutions ($2\% \text{ v/v} \text{ HNO}_3$) were analyzed immediately after the ICV solutions and after the CCV solutions (after each group of not more than ten samples and at the end of the analytical run). The concentration of all AOI were within the acceptance criteria of <EQL.

Low-Level Standard (LLS):

The LLS solution was analyzed immediately after the first CCB solution. The concentrations of all AOI were within the acceptance criteria of 70% to 130% recovery.

Interference Check Standard (ICS):

The ICS solution was analyzed immediately after the first LLS solution and immediately prior to analyzing the final CCV solutions. The concentrations of all AOI were within the acceptance criteria of 80% to 120% recovery.

Serial Dilution (SD):

Five-fold serial dilution was conducted on sample 17-1224. Percent differences (%D) are listed for all analytes that had a concentration at or above the EQL in the diluted sample. The %Ds for the AOI meeting this requirement ranged from 0.2% to 1.8% and were within the acceptance criterion of $\leq 10\%$.

Post-Digestion Spike (PS-71A)/Analytical Spike (AS-71A) - Sample (71A Component):

Because no MS sample was required to be prepared, a post-digestion spike (PS-71A) was conducted on sample 17-1224. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration $\geq 25\%$ of that in the sample. Recovery values for the AOI meeting this requirement ranged from 96% to 99%, and were within the acceptance criterion of 80% to 120%.

Other QC:

All other instrument-related QC tests for the AOI passed within their respective acceptance criteria.

Comments:

- The "Final Results" have been corrected for all laboratory dilutions performed on the samples during
 processing and analysis, unless specifically noted.
- 2) Instrument detection limits (IDL) and estimated quantitation limits (EQL) shown are for acidified water and/or fusion flux matrices as applicable. Method detection limits (MDL) for individual samples can be estimated by multiplying the IDL by the "Process Factor" for that individual sample. The estimated quantitation limit (EQL) for each concentration value can be obtained by multiplying the EQL by the "Process Factor".
- 3) Routine precision and bias is typically ±15% or better for samples in dilute, acidified water (e.g. 2% v/v HNO₃ or less) at analyte concentrations > EQL up to the upper calibration level. This also presumes that the total dissolved solids concentration in the sample is less than 5000 µg/mL (0.5 per cent by weight). Note that bracketed values listed in the data report are within the MDL and the EQL, and have potential uncertainties greater than 15%. Concentration values < MDL are listed as "--". Note, that calibration and QC standard samples are validated to a precision of ±10%.</p>
- 4) Analytes included in the spike 71A component (for the AS/PS) are; Ag, Al, As, B, Ba, Be, Ca, Cd, Ce, Co, Cr, Cs, Cu, Dy, Er, Eu, Fe, Ga, Gd, Ho, K, La, Lu, Mg, Mn, Na, Nd, Ni, P, Pb, Pr, Rb, S, Se, Sm, Sr, Th, Tl, Tm, U, V, Yb, and Zn. Analytes included in the spike 71B component are; Ge, Hf, Mo, Nb, Sb, Si, Sn, Ta, Te, Ti, W, and Zr. Analytes included in the spike 71C component are; Ir, Os, Pd, Pt, Re, Rh, and Ru. Analytes included in the spike 71D component are; Bi, In, Li, Sc, Tb, and Y.
Battelle PNNL/RPG/Inorganic Analysis ... ICP-MS Data Report

		Run Date >	8/9/2017	8/9/2017	8/9/2017	1
		Multiplier >	1.0	10.02	10.02	1
			Lab 405	17-1224 @	17-1224 @	1
		RPL/LAB >	Blank	10x	10x rep	
Instr. Det.	Est. Quant.		Lab 405	6220015	(224 B	
Limit (IDL)	Limit (EQL)	Client ID >	Blank	<u></u>	- <u>E8-A</u>	4
(ng/mL)	(ng/mL)	(Analyte)	(ng/mL)	(ng/mL)	(ng/mL)	4
0.0021	0.021	Cs 133	[0.002]	81.7	83.0	1
0.0003	0.003	Cs 135	-	24.4	24.8	1
0.001	0.007	Cs 137		21.4	22.0]
Other Analyt	es					
0.0057	0.057	Be 9				1
0.380	3.80	B 11	8.00	8,110	8,240	(3)
0.134	1.34	Mg 25		23.4	23.7	
0.0020	0.020	V 51	6.22	0.648	0.671	
0.0125	0.125	Cr 53		6.01	5.84	
0.0061	0.061	Mn 55		1.21	1.17	
0.0053	0.053	Co 59	8 77	[0.07]	[0.10]]
0.0082	0.082	Ni 60	-	2.29	2.68]
0.0179	0.179	Cu 65	°	16.0	15.9]
0.570	5.70	Zn 66		37.1	36.6	1
0.0021	0.021	Ga 69		1.16	1.10	1
0.0070	0.070	As 75		[0.28]	[0.23]	1
0.184	1.84	Se 77				1
0.0027	0.027	Rb 85	844	1.90	1.75	1
0.0052	0.052	Sr 88		2.46	2.34	1
0.0113	0.113	Ag 107	2			1
0.0030	0.030	Cd 111	[0.003]	[0.07]	[0.06]	1
0.0039	0.039	Ba 138	-	18.0	17.9	1
0.0023	0.023	La 139	-	[0.24]	[0.25]	1
0.0027	0.027	Ce 140	-	0.613	0.636	1
0.0023	0.023	Pr 141		[0.09]	[0.08]	1
0.0027	0.027	Nd 143	-	[0.23]	[0.26]	1
0.0031	0.031	Sm 149		[0.04]	[0.04]	1
0.0025	0.025	Eu 151	-		-	1
0.0026	0.026	Gd 157		[0.04]	[0.05]	1
0.0025	0.025	Dy 161		[0.06]	[0.05]	1
0.0023	0.023	Ho 165			-	1
0.0024	0.024	Er 166	-	[0.04]	[0.04]	1
0.0025	0.025	Tm 169				1
0.0023	0.023	Yb 173		[0.05]	[0,04]	1
0.0040	0.040	Lu 175		[0.00]	[oreal	
0.0029	0.029	TI 205				
0.0051	0.051	Pb 208		98.0	99.6	
0.0039	0.039	Th 232		[0.063]	[0.055]	
0.0026	0.026	U 238		2.27	2.26	1
			1000			

Internal Standard % Recovery

Li 6 (IS)	106%	111%	107%
Sc 45 (IS)	104%	130%	128%
Y 89 (IS)	101%	104%	106%
In 115 (IS)	99%	103%	99%
Tb 159 (IS)	102%	106%	105%
Bi 209 (IS)	101%	106%	105%

 "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the "multiplier" near the top of each column. The estimated sample quantitation limit = EQL (in Column 2) times the "multiplier". Overall error for values ≥ EQL is estimated to be within ±15%.
 Values in brackets [] are ≥ MDL but < EQL, with errors likely to exceed 15%.

3) Boron-11 results were nearly 100x above the maximum calibration and may not be valid

M0026-71A ASR-0329 Fiskum 170809Re.xlsx

Page 1 of 2

Criteria >	≤ 20%	80%-120%	75%-125%	75%-125%	≤ 10%
QC ID >	17-1224 Dup	LCS/BS	MS (None)	17-1224 + CCV-71A	17-1224 5-fold Serial Dil
Analytes	RPD (%)	%Rec	%Rec	%Rec	%Diff
Cs 133	1.5	95		96	1.2
Cs 135	1.5	94		97	0.2
Cs 137	2.7	98		99	1.8
Other Analyt	es				
Be 9		96		91	
B 11	1.5			nr	4.5
Mg 25	1.4				
V 51	3.5	90		90	
Cr 53	2.8	92		87	
Mn 55	3.4	92		92	
Co 59		95		91	
Ni 60	15.6	92		96	
Cu 65	0.2	99		96	
Zn 66					
Ga 69	5.5	95		90	
As 75		99		87	
Se 77					
Rb 85	8.4	93		90	
Sr 88	5.0	96		92	
Ag 107		188		196	
Cd 111		96		96	
Ba 138	0.4	96		89	
La 139	1.9	95		92	
Ce 140	3.6	94		90	
Pr 141		94		91	
Nd 143		95		92	
Sm 149		97		92	
Eu 151		95		92	
Gd 157		95		91	
Dy 161		97		92	
Ho 165		94		90	
Er 166		97		90	
Tm 169		95		97	
Yb 173		95		90	
Lu 175		93		91	
TI 205		95		93	
Pb 208	1.6	95		95	2.0
Th 232		97		94	
U 238	0.4	94		95	
nternal Stan	dard % Reco	very			
Li 6 (IS)	106%	101%		104%	111%
Contraction and Contraction of the					

QC Performance 08/09/2017

Li 6 (IS)	106%	101%	104%	111%
Sc 45 (IS)	116%	107%	110%	130%
Y 89 (IS)	108%	104%	108%	104%
In 115 (IS)	102%	99%	102%	103%
Tb 159 (IS)	105%	102%	108%	106%
Bi 209 (IS)	107%	103%	106%	106%

Shaded results are outside the acceptance criteria.

nr = spike concentration less than 25% of sample concentration. Matrix effects can be assessed from the serial dilution.

na = not applicable; KOH flux and Ni crucible or Na2O2 flux and Zr crucible for fusion preparations, or Si for HF assisted digests.

IS = Internal Standard. The concentration of certain elements cannot be determined due to the presence of the IS in all solutions.

M0026-71A ASR-0329 Fiskum 170809Re.xlsx

.

ASR 0335, Test 1 Results

(Information on th	alytical Ser	vice Request (ASR) oplicable to all samples submitted under this ASR)
Requestor Complete all fields Requestor: Signature Print Name Phone Phone	s on this COVER	Project Number: <u>69833</u> Work Package: <u>A82.456</u>
Matrix Type Information	1	QA/Special Requirements
 ◆ Liquids: ⊠ Aqueous □ Organic ◆ Solids: □ Soil □ Sludge □ Glass □ Filter □ Smear □ Organic ◆ Other: □ Solid/Liquid Mixture Slurry 	 Multi-phase Sediment Metal Other 	 ◆ QA Plan: ▷ ASO-QAP-001 (Equivalent to HASQARD) □ Additional QA Requirements, List Document Below: Reference Doc Number: ◆ Field COC Submitted? ☑ No □ Yes ▲ Lab COC Required? ☑ No □ Yes
Gas Biological Spec	cimen	 Sample/Container Inspection Documentation Required? No I Yes
(If sample matrices vary, specify on R	equest Page)	◆ Hold Time: Ø No □ Yes
 Disposal Information Disposition of Virgin Samples: Virgin samples are returned to requestor archiving provisions are made with recei If archiving, provide: Archiving Reference Doc: 	unless ving group!	If Yes, Contact ASO Use SW 846 (PNL-ASO-071, identify Lead before analytes/methods where holding times apply) submitting Other? Specify: Samples Other? Specify: Special Storage Requirements: Mone Refrigerate Other Specify:
Disposition of Treated Samples: Dispose Return	Data Pa	Data Requires ASO Quality Engineer Review? No Yes
 Is Work Associated with a Fee-Based Milestone? ⊠ No □ Yes If yes, milestone due date: Preliminary Results Requested, As Available? □ No ☎ Yes 	Data Reporting ASO-QAP-001 HASQARD). Minimum data r Project Specific Contact ASO Lead Document:	g Level (Equivalent to (Equivalent to $8/24/17$ (Note: Priority rate charge for < 10 business day turn-around time)
	Waste Des	ignation Information
ASO Sample Information Check List At If no, Reference Doc Attached: <u>see affra</u> or, Previous ASR Number: <u>272</u> or, Previous RPL Number: Send Report To: <u>Standre</u> Fiskarda Additional or Special Instructions	tached? K No I'	Yes Does the Waste Designation Documentation Indicate Presence of PCBs? No ØYes MSIN
Receivin Date Delivered: 8 21 Delivered By (optional) 9:30 Time Delivered: 9:30 Group ID (optional) 9 CMC Waste Sample? Ø No 9	g and Login Inform	mation (to be completed by ASO staff)Received By: $IRCANG - Q$ ASR Number: 0.335 RPL Numbers: $17 - 12.40$ If $17 - 12.42$ (first and last)
ASO Work Accepted By:	Paul	Signature/Date: Juln. Val 8/21/17

17		_				
	Use Only	Library				
	ASO Staff	Test				
	1 limits - 🗆 Below 🗖 Attached	Analyses Requested	 GEA - As received solutions. (Co-60, Cs-137 and Eu-154) (Co-60, Cs-137 and Eu-154) C) IC- Anions (17-1240 and 17-1241 only) Cl, NO2, NO3, PO4, C204, S04 Acid Digestion - 128 (17-1240 and 17-1241 Only) IC-1241 Only) IC-1241 Only) IC-1240 and diluted acid digest solutions and diluted 17-1247) 	 5) ICP/OES - Ag, Al, As, B, Ba, Ca, Cd, Cr, Fe, K, Li, Na, Ni, P, Pb, Se, Th, Tl, U, Zn and Zr (all 3 samples - Use acid digest solutions and diluted 17-1242). 6) Total Alpha/Total Beta (all 3 samples) 7) U-KPA (all 3 samples) 8) Sr-00/all 3 camplec) 	9) TC-0(all 3 samples) 9) Tc-99 (all 3 samples) 10) Np-AEA (all 3 samples) 11) Pu-AEA (all 3 samples) 12) Am/Cm-AEA (all 3 samples) 13) TOC/TIC - (17-1240 and 17-1241 only) 14) Free OH - (17-1240 and 17-1241 only) 15) H+ (sample 17-1242 only)	ant sample will be pre-diluted by Hot Cell. Special instructions tions, please contact K. Pool
	of Interest and Required Detection	Sample Description (& Matrix if varies)	AP-105 Direct Feed (Caustic)	AP-105 - Cs Removed (Caustic)	AP-105 - Cs Eluant (Acidic)	e to be run in duplicate. The Cs elua tely 100x prior to loading out of the ached to the ASR. If there any ques tical work.
	Provide Analytes	Client Sample ID	TI014-FEED	TI014-EFF Comp	TI014-ELComp	Note: All 3 samples are a factor of approximat from the client are att prior to starting analyt
	ASO Staff Use Only	RPL Number	17-1240	17-1241	17-1242	ASR # 033

Analytical Services Request (ASR) (REQUEST PAGE ---- Information Specific to Individual Samples) Page 1 of 1

ASR 0335 Request Page.xls

Special Instructions for ASR 0335

Analysis of AP-105 diluted feed, effluent, and eluate from Test 1 ion exchange process run

Use the ASO QA Plan current revision

Client ID	ASO Sample ID	Location for pickup	Matrix
TI014-FEED	17-1240	SAL cell 3	Caustic/5.6 M Na, high salt
TI014-EFFComp	17-1241	Lab 305	Caustic/5.6 M Na, high salt
TI014-ELComp	17-1242	SAL cell 3	Acidic/~0.2 M HNO ₃

Please prepare each sample in duplicate so that I can have a duplicate analysis of each.

Because the TI014-ELComp sample is already acidified, acid digestion may not be needed. The high salt samples are assumed to require acid digestion. However if the analysis technique sufficiently dilutes the sample into required acid matrix, then acid digestion may be omitted.

Preparative QC samples expected for this suite include:

- 1. Process blank (digestion blank or diluent blank)
- 2. Duplicate (all samples processed in duplicate)
- 3. Blank spike or reagent spike
 - a. not applicable to GEA
- 4. Matrix spike
 - a. ICP-OES: not needed for major analytes, e.g., Na, for guidance on spiking needs see ICP-OES results from ASR 0316 report serial dilution instead per ASO QA Plan
 - b. IC: not needed for major analytes, e.g., nitrate, report serial dilution instead per ASO QA Plan
 - c. not applicable to GEA
 - d. not applicable to hydroxide analysis

Analyte	Analysis Method
⁶⁰ Co	
¹³⁷ Cs	Gamma Energy Analysis (GEA)
¹⁵⁴ Eu	(report any other observed gamma emitting isotopes)
²³⁷ Np	
²³⁸ Pu	
²³⁹⁺²⁴⁰ Pu	Acid Digestion
²⁴¹ Am	Separations/Alpha Energy Analysis (AEA)
²⁴² Cm	
²⁴³⁺²⁴⁴ Cm	
⁹⁰ Sr	Acid Digestion / Separations/Beta Scintillation Counting
⁹⁹ Tc	Inductively Coupled Plasma Mass Spectrometry (ICP-MS) or Separations/Beta Countin
Total Alpha	Acid Digestion / Alpha Scintillation or Gas Proportional Counting
Total Beta	Acid Digestion / Beta Scintillation or Gas Proportional Counting
Hg	Acid Digestion / ICP-MS
Ag	
Al	
As	
В	
Ba	
Ca	
Cd	
Cr	
Fe	
K	
Li	Acid Digestion Inductively Counled Plasma Ontical Emission Spectrometry (ICP-OES)
Na	inductively coupled r tasing optical Emission operatineary (ref 020)
Ni	
Р	
Pb	
Se	
Th	
Ti	
U (total)	
Zn	
Zr	
U (total)	Acid Digestion / Kinetic Phosphorescence Analysis (KPA)
Cl-(q)	
NO ₂ -(d)	
NO ₃ - ^(d)	Dilution
PO4 ^{3-(d)}	Ion Chromatography (IC)
$C_2 O_4^{2-(d)}$	
SO4 ^{2-(d)}	
Total organic carbon (TOC)	Hot Persulfate Ovidation

Please analyze **TI014-FEED** and **TI014-EFFComp** for the following:

Page 2 of 3

Analyze the **TI014-ELComp** for the following

Analyte	Analysis Method
⁶⁰ Co	
¹³⁷ Cs	(report any other observed gamma emitting isotopes)
¹⁵⁴ Eu	
²³⁷ Np	
²³⁸ Pu	
²³⁹⁺²⁴⁰ Pu	Acid Digestion
²⁴¹ Am	Separations/Alpha Energy Analysis (AEA)
²⁴² Cm	
²⁴³⁺²⁴⁴ Cm	
⁹⁰ Sr	Acid Digestion / Separations/Beta Scintillation Counting
⁹⁹ Tc	Inductively Coupled Plasma Mass Spectrometry (ICP-MS) or Separations/Beta Counting
Total Alpha	Acid Digestion / Alpha Scintillation or Gas Proportional Counting
Total Beta	Acid Digestion / Beta Scintillation or Gas Proportional Counting
Hg	Acid Digestion / ICP-MS
Ag	
Al	
As	
В	
Ba	
Ca	
Cd	
Cr	
Fe	
К	
Li	Acid Digestion
Na	Inductively Coupled Plasma Optical Emission Spectrometry (ICP-OES)
Ni	
Р	
РЬ	
Se	
Th	
Ti	
U (total)	
Zn	
Zr	
U (total)	Acid Digestion / Kinetic Phosphorescence Analysis (KPA)
. ()	T't ut t u

Battelle PNNL/RPL/Inorganic Analysis ... ICP-OES Analysis Report PO Box 999, Richland, Washington 99352

69833 / N82456
0335
S. Fiskum
3 (liquids)

ASO Sample ID	Client Sample ID	Client Client Sample Description	
17-1240	TI014-FEED	AP-105 Direct Feed (Caustic)	1.2794
DUP-1240	TI014-FEED	AP-105 Direct Feed (Caustic)	1.2778
17-1241	TI014-EFF Comp	AP-105 - Cs Removed (Caustic)	1.2789
DUP-1241	TI014-EFF Comp	AP-105 – Cs Removed (Caustic)	1.2804
17-1242	TI014-ELComp	AP-105 - Cs Eluant (Acidic)	NA

Sample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of Liquids for Metals Analysis Using a Dry-Block Heater", performed by L. Darnell on 8/28/17, dilution of Sample 17-1242 in deionized water performed by L. Grow on 8/17/17, and dilution of "as received" samples in 5% v/v HNO3 performed by D. Cherkasov on 8/31/17, respectively.

Procedure	: <u>RPG-CMC-211</u> Inductively Cou	<u>, Rev. 4</u> , "Determi pled Argon Plasma	nation of Elemen Optical Emissio	ntal Composition	n by (ICP-OES)."	
Analyst:	D. Cherkasov/ G. Brown	Analysis Date: Revision Date:	08/31/2017 10/02/2017	ICP File:	C0721	
See Chemi	ical Measurement	Center 98620 file:	ICP-325-405- (Calibration at	<u>3</u> nd Maintenance	Records)	
M&TE:	PerkinElmer 5300DV ICP-OES			SN: 077N5122002		
	Sartorius ME414S Balance			SN: 22406373		
	Mettler AT40	Mettler AT400 Balance			57	
	Sartorius R20	Sartorius R200D Balance			SN: 39080042	
	Mettler AT20	1 Balance		SN: 192720-92	1	
	SAL Cell 2 Balance			SN: 8033311209		
	Lab 305 Bala	Lab 305 Balance				
	Lab 309 Bala	nce		SN: 10803210		

Report Preparer

1-long

Review and Concurrence

10/02/17 Revision Date

Date

Page 1 of 4

Battelle PNNL/RPL/Inorganic Analysis ... ICP-OES Analysis Report

Three aqueous samples submitted under Analytical Service Request (ASR) 0335 were analyzed by ICP-OES. Samples 17-1240 and 17-1241 were prepared in duplicate following RPL procedure RPG-CMC-128 and diluted to approximately 25 mL. Sample 17-1242 was diluted with DI water in the hot cell. All samples were further diluted in 5% HNO₃ prior to analysis. None of the samples were filtered.

All sample results are reported on a mass per unit volume basis ($\mu g/mL$) for each detected analyte. The data have been adjusted for instrument dilutions.

Analytes of interest (AOI) were specified in the ASR and are listed in the upper section of the attached ICP-OES Data Report. The quality control (QC) results for the AOI have been evaluated and are presented below. Analytes other than the AOI are reported in the bottom section of the report but have not been fully evaluated for QC performance.

Calibration of the ICP-OES was done following the manufacturer's recommended calibration procedure using multi-analyte custom standard solutions traceable to the National Institute of Standards and Technology (NIST). Midrange calibration verification standards (MCVA and MCVB) were used to verify acceptance of the two-point calibration curves obtained for each analyte and for continuing calibration verification.

The controlling documents were procedures RPG-CMC-211, Rev. 4, *Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP-OES)*, and ASO-QAP-001, Rev. 11, *Analytical Support Operations (ASO) Quality Assurance Plan*. Instrument calibrations, QC checks and blanks (e.g., ICV/ICB, CCV/CCB, LLS, ICS), matrix spike, post-digestion spikes, duplicate, blank spike, and serial dilution were conducted during the analysis run.

Preparation Blank (PB):

A preparation blank (reagents only) was prepared for the extraction process. With the exception of iron (0.43 μ g/mL) and zinc (1.63 μ g/mL), the concentration of all AOI were within the acceptance criteria of <EQL (estimated quantitation level), \leq 50% regulatory decision level, or \leq 10% of the concentration in the samples. The iron and zinc concentrations were significant compared to the concentrations in all samples.

Blank Spike (BS)/Laboratory Control Sample (LCS):

A reagent spike (RS) sample (reagents and spikes) was prepared for the extraction process. Recovery values are listed for all analytes included in the RS that were measured at or above the EQL. Recovery values for the AOI meeting this requirement ranged from 41% to 101%, and, with the exception of silver (41%), were within the acceptance criterion of 80% to 120%. The reason for the low silver recovery is unknown. However, the silver spike is added separately from the remaining elements and may indicate reagent instability.

Duplicate/Replicate Relative Percent Difference (RPD)/Relative Standard Deviation (RSD): The ASR requested duplicate analyses of all samples. Duplicates of samples 17-1240 and 17-1241 were prepared and analyzed. Due to limited sample volume, a single preparation

Battelle PNNL/RPL/Inorganic Analysis ... ICP-OES Analysis Report

of sample 17-1242 was analyzed once undiluted and twice at a 5X dilution. RPDs are listed for all analytes that were measured at or above the EQL. RPDs for the AOI meeting this requirement ranged from 0.9% to 4.6% and were within the acceptance criterion of \leq 20% for liquid samples.

Matrix-Spike (MS) Sample:

A matrix spike (MS) sample (sample 17-1240 and spikes) was prepared for the extraction process. Recovery values are listed for all analytes included in the MS that were measured at or above the EQL. Recovery values for the AOI meeting this requirement ranged from 46% to 108%, and, with the exception of silver (46%), were within the acceptance criterion of 75% to 125%. The reason for the low silver recovery is unknown. However, the silver spike is added separately from the remaining elements and may indicate reagent instability.

Initial/Continuing Calibration Verification (ICV/CCV):

MCVA and MCVB solutions were analyzed immediately after calibration, after each group of not more than ten samples, and at the end of the analytical run. The concentrations of all AOI were within the acceptance criteria of 90% to 110% recovery.

Initial/Continuing Calibration Blank (ICB/CCB):

The ICB/CCB solution (5% v/v HNO₃) was analyzed immediately after the ICV solutions and after the CCV solutions (after each group of not more than ten samples and at the end of the analytical run). The concentration of all AOI were within the acceptance criteria of <EQL.

Low-Level Standard (LLS):

The LLS solution was analyzed immediately after the first CCB solution. The concentrations of all AOI were within the acceptance criteria of 70% to 130% recovery.

Interference Check Standard (ICS/SST):

The ICS solution was analyzed immediately after the first LLS solution and immediately prior to analyzing the final CCV solutions. The concentrations of all AOI were within the acceptance criteria of 80% to 120% recovery.

Serial Dilution (SD):

Five-fold serial dilution was conducted on sample 17-1240. Percent differences (%Ds) are listed for all analytes that had a concentration at or above the EQL in the diluted sample. The %Ds for the AOI meeting this requirement ranged from 0.4% to 8.7% and were within the acceptance criterion of $\leq 10\%$.

Post-Digestion Spike (PS-A)/Analytical Spike (AS-A) - Sample (A Component):

In addition to the MS sample, a post-digestion spike (A Component) was conducted on sample 17-1240. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration $\geq 25\%$ of that in the sample. Recovery values for the AOI meeting this requirement ranged from 92% to 106%, and were within the acceptance criterion of 80% to 120%.

Post-Digestion Spike (PS-B)/Analytical Spike (AS-B) - Sample (B Component):

In addition to the MS sample, a post-digestion spike (B Component) was conducted on sample 17-1240. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration \geq 25% of that in the sample. Recovery values for the AOI meeting this requirement ranged from 94% to 103%, and were within the acceptance criterion of 80% to 120%.

Other QC:

All other instrument-related QC tests for the AOI passed within their respective acceptance criteria.

Report Revision (10/02/2017):

A revision to the original data report was initiated on 10/02/2017 to include the 102.4X dilution of sample 17-1242 that was performed by L. Grow on 08/17/2017 in the hot cell prior to analysis. No other sample results were revised.

Comments:

- 1) The "Final Results" have been corrected for all laboratory dilutions performed on the samples during processing and analysis, unless specifically noted.
- 2) Instrument detection limits (IDL) and estimated quantitation limits (EQL) shown are for acidified water and/or fusion flux matrices as applicable. Method detection limits (MDL) for individual samples can be estimated by multiplying the IDL by the "Process Factor" for that individual sample. The estimated quantitation limit (EQL) for each concentration value can be obtained by multiplying the EQL by the "Process Factor".
- 3) Routine precision and bias is typically ±15% or better for samples in dilute, acidified water (e.g. 5% v/v HNO₃ or less) at analyte concentrations > EQL up to the upper calibration level. This also presumes that the total dissolved solids concentration in the sample is less than 5000 µg/mL (0.5 per cent by weight). Note that bracketed values listed in the data report are within the MDL and the EQL, and have potential uncertainties greater than 15%. Concentration values < MDL are listed as "--". Note, that calibration and QC standard samples are validated to a precision of ±10%.</p>
- 4) Analytes included in the spike A component (for the AS/PS) are; Ag, Al, As, B, Ba, Be, Bi, Ca, Cd, Co, Cr, Cu, Fe, K, Li, Mg, Mn, Mo, Na, Ni, P, Pb, Sb, Se, Si, Sn, Sr, Ta, Ti, Tl, V, W, Y, Zn, and Zr. Analytes included in the spike B component are; Ce, Dy, Eu, La, Nd, Pd, Rh, Ru, S, Te, Th, and U.

Battelle PNNL/RPG/Inorganic Analysis ... ICPOES Data Report

Page 1 of 2

	1	D. Data b	9/24/2047	0/24/2047	0/24/2047	0/24/2047	0/24/2047	0/24/2047	9/24/2047	9/24/2047	9/24/2047
		Run Date >	8/31/2017	8/31/2017	8/31/2017	8/31/2017	8/31/2017	8/31/2017	8/31/2017	8/31/2017	6/31/2017
		Process							100.1		
		Factor >	1.0	24.7	620.0	620.5	619.3	610.8	102.4	512.0	512.0
			105 111		17-1240	DUP-1240	17-1241	DUP-1241	17 1010	17-1242	17-1242
		RPL/LAB >	405 diluent	BLK-1240	@25x	@25X	@25x	@25x	17-1242	@5x	@5x rep
and a second											
Instr. Det.	Est. Quant.	1000									
Limit (IDL)	Limit (EQL)	Client ID >	Lab Diluent	BLK-1240	<u>TI014</u>	FEED	TI014-EF	-F Comp	1	1014-EL Com	<u>p</u>
(µg/mL)	(µg/mL)	(Analyte)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)
0.0030	0.030	Ag			**	**	**	**			
0.0276	0.276	AI	-		14,300	14,800	14,600	14,400	50.4	[44]	[47]
0.1576	1.576	As	-			••					
0.0120	0.120	В	-		82.5	81.7	[56]	[54]	[8.8]	[14]	[8.7]
0.0005	0.005	Ba		[0.032]	[0.39]	[0.43]	[0.33]	[0.38]	[0.46]	[0.44]	[0.33]
0.0233	0.233	Ca		[1.2]	[39]	[43]	[52]	[49]	24.3	[22]	[15]
0.0027	0.027	Cd				[1.8]		[2.8]			
0.0047	0.047	Cr			262	270	264	260	19 91		[4 2]
0.0047	0.047	50		0.420	15.93	570	15.91	10 71	2.25	[2 0]	[4.2]
0.0017	0.017	Fe		0.430	[0.0]	[4.0]	[5.0]	[0.7]	3.35	[3.9]	[4.0]
0.0500	0.500	ĸ		[6.1]	3,930	4,020	3,940	3,900	558	557	537
0.0028	0.028	Li			**				**		
0.0171	0.171	Na	**	[0.75]	141,000	145,000	141,000	139,000	4,770	4,710	4,610
0.0074	0.074	Ni			[31]	[29]	[35]	[30]	[1.6]		
0.1021	1.021	Р			[490]	[530]	[520]	[520]			
0.0399	0.399	Pb							205	[200]	[200]
0.2271	2.271	Se					[200]				
0.0120	0.120	Th									[7.0]
0.0014	0.014	Ti			++)						**
0.0676	0.676	11									
0.0066	0.055	70		1.63	17 01	[7 3]	[14]	17 01	8 33	[7 7]	[4.4]
0.0000	0.000	7.		1.05	[1:0]	[1:0]	[14]	[1.0]	0.00	[1.0]	[]
0.0022	0.022	21	-					-			
Other Analyti	es	-								114	
0.0002	0.002	Be			[0.18]		[0.17]	[0.19]	[0.041]		
0.0546	0.546	Bi							[6.5]		**
0.0448	0.448	Ce	-	-		-				**	
0.0071	0.071	Co						-			
0.0036	0.036	Cu		[0.098]	[4.0]	[4.6]			11.2	[11]	[10]
0.0034	0.034	Dy			**			**			
0.0010	0.010	Eu									
0.0030	0.030	La									
0.0023	0.023	Ma							[1,4]		
0.0009	0.009	Mn							[0,19]		
0.0094	0.094	Mo			[53]	[55]	58.4	[52]	10.101		
0.0499	0.499	Net			[00]	[00]	00.4	[ove]			
0.0100	0.100	Rd Rd								-	
0.0170	0.170	Pa									
0.0162	0.162	Kh						-			
0.0158	0.158	Ru			**						
0.1773	1.773	S			1,350	1,430	1,280	1,440	[67]	[92]	**
0.0919	0.919	Sb	-								
0.0136	0.136	Si			99.1	109	[70]	[72]	18.5	[16]	[12]
0.0446	0.446	Sn			[34]			[30]	[6.1]		
0.0003	0.003	Sr				[0.19]	[0.22]	[0.23]	[0.096]		-
0.0242	0.242	Та			**						
0.0412	0.412	Te	-								
0.0885	0.885	TI	-								
0.0032	0.032	Y			-		-		[0 33]		
0.0032	0.349	14/	-		1951	10.41	1961	(0.41	[0.00]		
0.0249	0.249	W			[00]	[94]	[ap]	[94]			
0.0006	0.006	Y			**						

1) "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the "multiplier" near the top of each column. The estimated sample quantitation limit = EQL (in Column 2) times the "multiplier". Overall error for values \geq EQL is estimated to be within ±15%.

2) Values in brackets [] are \geq MDL but < EQL, with errors likely to exceed 15%.

ASR 0335 Final from C0721R Fiskum ASR-0335 Normal BS & MS.xls

Criteria >	≤ 20%	≤ 20%	≤ 20%	80%-120%	75%-125%	80%-120%	80%-120%	≤ 10%
QC ID >	17-1240 Dup	17-1241 Dup	17-1242 @5x rep	LCS/BS	17-1240 MS	17-1240 + AS-A	17-1240 + AS-B	17-1240 5-fold Serial Dil
Analytes	RPD (%)	RPD (%)	RPD (%)	%Rec	%Rec	%Rec	%Rec	%Diff
Aq				41	46	92		
Al	3.2	0.9		96	nr	105		0.4
As						103		
B	1.0			97	105	102		
Ba				96	101	103		
Ca				98	102	105		
Cd				93	107	102		
Cr	4.6	1.0		94	nr	104		0.3
Fe	4.0	110		96	102	105		0.0
ĸ	2.4	1.1	3.6	96	100	103		87
11	2.7	1.1	5.0	101	104	102		0.7
Na	3.0	12	2.1	96	nr	103		0.8
Ni	5.0	1.6	2.1	96	106	104		0.0
P				90	07	104		
Ph				97	108	105		
PU				90	100	105		
Th						100	04	
Ti				0.0	101	102	94	
				90	101	103	102	
70				90	104	100	103	
20				95	103	106		
2r				97	102	103		
Other Analy	tes		· · · · · ·			07		
Be				89	97	97		
Bi				80		96		
Ce				80		100	83	
Co				07	107	103		
Cu				.97	107	105		
Dy							99	
Eu							93	
La				92	97		99	
Mg				96	103	104		
Mn				97	105	103		
Mo				95	97	101		
Nd				92	96		98	
Pd							88	
Rh							95	
Ru		11.0			407		94	
S	5.3	11.2		97	106		100	
Sb						99		
Si	9.2			88	105	105		
Sn					1.11	97		
Sr				99	107	106		
Ta						102		
Te							95	
TI						92		
V				91	96	97		_
W				96		103		
Y						100		

QC Performance 08/31/2017

Shaded results are outside the acceptance criteria.

nr = spike concentration less than 25% of sample concentration. Matrix effects can be assessed from the serial dilution,

ASR 0335 Final from C0721R Fiskum ASR-0335 Normal BS & MS xls

Battelle PNNL/RPL/Inorganic Analysis ... ICP-MS Analysis Report PO Box 999, Richland, Washington 99352

Project / WP#:	69833 / N82456
ASR#:	0335
Client:	S. Fiskum
Total Samples:	3 (Aqueous)

ASO Sample ID	Client Sample ID	Client Sample Description	Sample Weight (g)
17-1240	TI014-Feed	AP-105 Direct Feed (Caustic)	NA
DUP-1240	TI014-Feed	AP-105 Direct Feed (Caustic)	NA
17-1241	TI014-EFF Comp	AP-105 Cs Removed (Caustic)	NA
DUP-1241	TI014-EFF Comp	AP-105 Cs Removed (Caustic)	NA
17-1242	TI014-EL Comp	AP-105 Cs Eluant (Acidic)	NA
	-		

Sample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of Liquids for Metals Analysis Using a Dry-Block Heater", performed by L. Darnell on 8/28/17, dilution of Sample 17-1242 in deionized water performed by L. Grow on 8/17/17, and ICP-MS bench dilution in 2% v/v HNO₃ or 2% HCl performed by D. Cherkasov on 9/15/17, respectively.

Analyst:	D. Cherkasov G. Brown	Analysis Date:	09/2	9/2017	ICP Fil	e: M0030
See Chem	ical Measurement	Center 98620 file:	ICP (Cal	-325-405-3 libration an	d Maintena	nce Records)
M&TE:	Sartorius P200	SN: 85 V	N4070702	RPL 405		
	Mettler AT400	SN: M19	445	RPL 405 FH		
	Mettler AT400	SN: 1113	3292667	RPL 420 FH		
	Ohaus EX324	SN: 8033311209 S		SAL Cell 2		
	Sartorius BA31	05 Balance		SN: 10803210 RPL		RPL 309
	Sartorius R200	D Balance		SN: 39080058 RPL 525 FH		

Report Preparer

Review and Concurrence

Date

10/28 Date

Page 1 of 4

Battelle PNNL/RPL/Inorganic Analysis ... ICP-MS Analysis Report

Three aqueous samples submitted under Analytical Service Request (ASR) 0335 were analyzed by ICP-MS. Samples 17-1240 and 17-1241 were prepared in duplicate following RPL procedure RPG-CMC-128 by diluting 1 mL to approximately 25 mL. Approximately 0.2 mL of sample 17-1242 was diluted with approximately 20 mL DI water in the hot cell to reduce radiation dose. Prior to analysis all samples were further diluted in 2% HNO₃ or 2% HCl containing 100 ppb Au to reduce Hg carryover in the ICP-MS system. None of the samples were filtered.

All results are reported on a mass per unit volume basis (ng/mL) for each detected analyte. The data have been adjusted for instrument dilutions.

Analytes of interest (AOI) were specified in the ASR and are listed in the upper section of the attached ICP-MS Data Report. Mercury was the only AOI; all results were reported for Hg-202. The quality control (QC) results for the AOI have been evaluated and are presented below.

Calibration of the ICP-MS was done following the manufacturer's recommended calibration procedure using multi-analyte custom standard solutions traceable to the National Institute of Standards and Technology (NIST). Midrange calibration verification standards were used to verify acceptance of the six-point calibration curve and for initial and continuing calibration verification (ICV/CCV).

The controlling document was procedure RPG-CMC-292, Rev 0, *Determination of Elemental Composition by Inductively Coupled Argon Plasma Mass Spectrometry (ICP-MS)* and ASO-QAP-001, Rev. 11, *Analytical Support Operations (ASO) Quality Assurance Plan.* Instrument calibrations, QC checks and blanks (e.g., ICV/ICB, CCV/CCB, LLS, ICS), post-digestion spike, duplicate, blank spike, and serial dilution were conducted during the analysis run.

Internal Standard (IS):

All solutions (blanks, standards, QC checks, and samples) were mixed in-line with a solution containing 10 ppb each of Li-6, Sc-45, Y-89, In-115, Tb-159, and Bi-209 as the internal standard (IS). The AOI (Hg-202) data were normalized using the data for the closest IS mass (e.g., Bi-209). The Bi-209 IS recoveries ranged from 94.5% to 108%, which were within the acceptance criterion of 30% to 120%.

Preparation Blank (PB):

A preparation blank (reagents only) was prepared for the extraction process. The concentration of Hg-202 was within the acceptance criteria of \leq EQL (estimated quantitation level), \leq 50% regulatory decision level, or less than \leq 10% of the concentration in the samples. A diluent blank from the ICP-MS laboratory (2% HCl and 100 ppb Au) was also analyzed. Results for the diluent blank were also within the acceptance criteria.

Blank Spike (BS)/Laboratory Control Sample (LCS):

A blank spike (BS) sample was prepared by spiking a 2% HCl and 100 ppb Au blank with an equivalent volume of the CCV-Hg-2ppb standard (1:1 ratio). The recovery for Hg-202 was 90.6%, which is within the acceptance criteria of 80% to 120% recovery.

Battelle PNNL/RPL/Inorganic Analysis ... ICP-MS Analysis Report

Duplicate/Replicate Relative Percent Difference (RPD)/Relative Standard Deviation (RSD): Duplicates of samples 17-1240 and 17-1241, as well as a replicate of sample 17-1242, were prepared and analyzed. No RPD are listed for the AOI since all sample results were below EQL. The acceptance criterion is ≤20% RPD for liquid samples.

Matrix-Spike (MS) Sample:

No matrix spike sample was required to be prepared.

Initial/Continuing Calibration Verification (ICV/CCV):

The ICV/CCV solution was analyzed immediately after calibration, after each group of not more than ten samples, and at the end of the analytical run. The concentrations of all AOI were within the acceptance criteria of 90% to 110% recovery.

Initial/Continuing Calibration Blank (ICB/CCB):

The ICB/CCB solution (2% v/v HCl and 100 ppb Au) was analyzed immediately after the ICV solutions and after the CCV solutions (after each group of not more than ten samples and at the end of the analytical run). The concentration of all AOI were within the acceptance criteria of \leq EQL.

Low-Level Standard (LLS):

The LLS solution was analyzed immediately after the first CCB solution. The concentrations of all AOI were within the acceptance criteria of 70% to 130% recovery.

Interference Check Standard (ICS):

The ICS solution was analyzed immediately after the first LLS solution and immediately prior to analyzing the final CCV solutions. The concentrations of all AOI were within the acceptance criteria of 80% to 120% recovery.

Serial Dilution (SD):

Five-fold serial dilution was conducted on sample 17-1240. Percent differences (%D) are listed for all analytes that had a concentration at or above the EQL in the diluted sample. No %Ds are listed for the AOI since all sample results were below EQL. The acceptance criterion of is $\leq 10\%$.

Post-Digestion Spike (PS-Hg)/Analytical Spike (AS-Hg) - Sample (Hg Component):

Because no MS sample was required to be prepared, a post-digestion spike (PS-Hg) was conducted on sample 17-1240. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration \geq 25% of that in the sample. Recovery values for the AOI meeting this requirement were 90.4%, and were within the acceptance criterion of 80% to 120%.

Other QC:

All other instrument-related QC tests for the AOI passed within their respective acceptance criteria.

Comments:

- The "Final Results" have been corrected for all laboratory dilutions performed on the samples during processing and analysis, unless specifically noted.
- 2) Instrument detection limits (IDL) and estimated quantitation limits (EQL) shown are for acidified water and/or fusion flux matrices as applicable. Method detection limits (MDL) for individual samples can be estimated by multiplying the IDL by the "Process Factor" for that individual sample. The estimated quantitation limit (EQL) for each concentration value can be obtained by multiplying the EQL by the "Process Factor".
- 3) Routine precision and bias is typically ±15% or better for samples in dilute, acidified water (e.g. 2% v/v HNO₃ or less) at analyte concentrations > EQL up to the upper calibration level. This also presumes that the total dissolved solids concentration in the sample is less than 5000 µg/mL (0.5 per cent by weight). Note that bracketed values listed in the data report are within the MDL and the EQL, and have potential uncertainties greater than 15%. Concentration values < MDL are listed as "--". Note, that calibration and QC standard samples are validated to a precision of ±10%.</p>
- 4) Analytes included in the spike 71A component (for the AS/PS) are; Ag, Al, As, B, Ba, Be, Ca, Cd, Ce, Co, Cr, Cs, Cu, Dy, Er, Eu, Fe, Ga, Gd, Ho, K, La, Lu, Mg, Mn, Na, Nd, Ni, P, Pb, Pr, Rb, S, Se, Sm, Sr, Th, Tl, Tm, U, V, Yb, and Zn. Analytes included in the spike 71B component are; Ge, Hf, Mo, Nb, Sb, Si, Sn, Ta, Te, Ti, W, and Zr. Analytes included in the spike 71C component are; Ir, Os, Pd, Pt, Re, Rh, and Ru. Analytes included in the spike 71D component are; Bi, In, Li, Sc, Tb, and Y. Analytes included in the spike Hg component are; Hg.

Battelle PNNL/RPG/Inorganic Analysis ... ICP-MS Data Report

		Run Date >	9/29/2017	9/29/2017	9/29/2017	9/29/2017	9/29/2017	9/29/2017	9/29/2017	9/29/2017
		Process								
		Factor >	4.96	1.0	6.067E+06	2.983E+07	5.960E+06	2.894E+07	1.013E+04	1.015E+04
				Lab 405	17-1240 @	Dup-1240 @	17-1241 @	Dup-1241 @	17-1242 @	Rep-1242 @
		RPL/LAB >	BLK-1240	Blank	6.067E+06x	2.983E+07x	5.960E+06x	2.894E+07x	1.013E+04x	1.015E+04x
Instr. Det.	Est. Quant.			Lab 405						
Limit (IDL)	Limit (EQL)	Client ID >	PB	Blank	T1014	-Feed	T1014-EF	F Comp	TI014-E	L Comp
(ng/mL)	(ng/mL)	(Analyte)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)
0.0081	0.081	Ha 202	,	,		:		:	,	:

Internal Standard % Recovery

103%	
104%	
105%	
105%	
105%	
107%	
%2.7%	
108%	
Bi 209 (IS)	1

 "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the "multiplier" near the top of each column. The estimated sample quantitation limit = EQL (in Column 2)

times the "multiplier". Overall error for values \geq EQL is estimated to be within $\pm 15\%$. 2) Values in brackets [] are \geq MDL but < EQL, with errors likely to exceed 15%.

	E.
	L
~	
_	Ŀ
-	ь
-	г
м.	
-	
-	
	L
	Ľ
-	Ŀ
The	Ľ
-	L
-	Ŀ
-	
a	
-	
0	
_	
_	
=	L
60	L
-	г
-	
	Ł
_	
=	
0	
~	E.
-	
-	Ł
a	ε.
÷	
•	
-	۴.
	1
	L
_	L
_	~

riteria >	≤ 20%	≤ 20%	≤ 20%	80%-120%	75%-125%	75%-125%	≤ 10%
to ID >	17-1240 Dup	17-1241 Dup	17-1242 Rep	LCS/BS	MS (None)	17-1240 + PS-Hg	17-1240 5-fold Serial Dil
nalytes	RPD (%)	RPD (%)	RPD (%)	%Rec	%Rec	%Rec	%Diff
Hg 202	1	:	:	80.6%		90.4%	1

Internal Standard % Recovery

3i 209 (IS)	105%	105%	103%	101%	108%	107%
1						

Shaded results are outside the acceptance criteria.

IS = Internal Standard. The concentration of certain elements cannot be determined due to the presence of the IS in all solutions. na = not applicable; KOH flux and Ni crucible or Na2O2 flux and Zr crucible for fusion preparations, or Si for HF assisted digests. nr = spike concentration less than 25% of sample concentration. Matrix effects can be assessed from the serial dilution.

M0030-Hg ASR-0335 Fiskum 170929a.xlsx



Client:	S. Fiskum			I	Report Date:	10/9/2017
				Ar	nalysis Date:	9/26/2017
					[9/28/2017
Subject:	Hydroxide	and Hydrogen I	on Analyses for:	Aqueous Samples		
Project:	69833	WP #	N79882			
ASR:	0335	Rev-0		Procedure:	RPG-CMC-	228, Rev 0.1
Sample ID.	17-1240	thru	17-1242			

Hydroxide analysis was performed for diluted sample aliquots of 2 aqueous samples (17-1240 and 17-1241), provided on ASR 0335. Samples were analyzed by manual titration for the base constituents content following procedure RPG-CMC-228, Rev. 0.1, Determination of Hydroxyl (OH) and Alkalinity of Aqueous Solutions, Leachates & Supernates . Sample size was 0.100 mL of sample added to 2.5 mL of DI water. The titrant used was 0.0957 M HCI (Standardized HCI was prepared and documented on Chem Rec 232, prepared on 3/15/2017). pH measurements were obtained using a Beckman Coulter 560 pH meter, SN 110650046. The pH meter was calibrated using 3 buffers, pH 4,7 and 10 and the calibration verified using an independent pH 7 buffer.

The initial diluted pH is reported on attached Report Summary along with the free hydroxide molarity. Samples were analyzed in duplicate therefore two pH measuresments are given per sample for each inflection point.

Sample ID	1st Inflection point	2nd Inflection point	3rd Inflection point
17-1240	11.2	7.1	3.8
17-1240 Dup	11.1	7.2	3.7
17-1241	11.3	7.7	3.9
17-1241 Dup	11.1	7.2	3.9
17-1242	9.8	NA	NA
17-1242 Dup	10.0	NA	NA

Hydrogen lon analysis was performed using a diluted sample aliquot of 17-1242. The sample was diluted by 102.4x to lower the dose to a level that could be handled in a CA fume hood. The samples were also analyzed by manual titration for the acid constituents content following procedure RPG-CMC-228. Sample size was 0.100 mL of sample added to 2.5mL DI water. The titrant used was 0.0101M NaOH (Standardized NaOH was prepared and documented on Chem Rec 232, prepared on 3/15/17 and diluted on 9/26/17--restandardized on 9/26/17). The final H⁺ molarity result has been corrected for the Hot Cell dilution.

The only inflection point for 17-1242 was at pH 9.8 and 10.0

Following is the report summary and the sample results calculated from the raw data. A copy of the titration ^

Prepared by:	agtal Stul	Date:
Reviewed by:	Null! Pack	Date:

A

ASR 0335 (Fiskum).xls

Page 1 of 1

Battelle Pacific Northwest Laboratory Nuclear Chemistry and Engineering Group Chemical Measurements Center

Hydroxide and Alkalinity Determination Procedure: RPG-CMC-228, Rev 0.1

ASR #	0335
ASR #	0335

WP# N82456

Equip # Beckman Coulter 560, SN#110650046

		Report Summary for	r ASR # Revision #	0335 Rev-0	Report Date Analysis Date	: 10/9/2017 : 9/26/2017 9/28/2017
				Cor	centration, moles / Lite	r
		Diluted		First Point	Second Point	Third Point
RPG #	Client ID	Initial pH	OH conc ug/mL	Molarity	Molarity	Molarity
17-1240	TI014-FEED	12.27	1.79E+04	1.05	1.34	0.77
17-1240 Dup	TI014-FEED	12.19		1.05	1.24	0.77
17-1241	TI014-EFF Comp	12.18	1.79E+04	1.05	1.15	0.67
17-1241 Dup	TI014-EFF Comp	12.16		1.05	1.15	0.77
17-1242	TI014-ELComp	8.10	NA	1.03		
17-1242 Dup	TI014-ELComp	7.99	NA	1.03		

OH conc (mg/L) = M (g/L) * 17,000

Note: Results are presented for the first, second, and third inflection points on the titration curves for the hydroxide titration, as applicable. The first inflection point is generally associated with the free hydroxide concentration. The second inflection point generally represents total hydroxide, or carbonate or a combination of aluminate and carbonate. The third inflection point is usually indicative of bicarbonate or other weak acids or possibly the continued protonation of alumina.

Analyst: 30.17 Reviewer:

017 017	07	피		ound noles Molarity base hase	0.105 1.053	0.105 1.053	105 1 053	0.105 1.053			
e: 10/9/2 e: 9/26/2	ls 4/1/20	08/01	Ē	F	7	8	5	1			
Report Dat alysis Dat	ocked07.x	20	ence	Ц	11.1	11.0	11.2	11.1			
An I	I I-TemplateL	Pertu	OH 1st Equival	Point Titrant	1.10	1.10	1.10	1.10			
Rev-0 N82456	preadsheet: OF	field	Diluted	Initial pH reading	12.27	12.19	12.18	12.16			
0335 S. Fiskum WP#	S	Analyst: 🧕		Titrator Routine	NA	NA	NA	NA			
and Rev # Client: 69833				Density %/m1	NA	NA	NA	NA			
ASR # Project:		108		Sample Wt (o)	NA NA	NA	NA	NA			
) and	Lab Loc.		Sample Vol (ml.)	0.100	0.100	0.100	0.100			
	ydroxyl (OH- ites	110650046		Dilution	NA	NA	NA	NA			
	Determination of H eachates and Superna	Equip # Chem	Rec# 232								
orthwest Laboratory y and Engineering Group	RPG-CMC-228, Rev 0.1 Vlkalinity of Aqueous Solutions, L	loi	Molarity 0.0957	ample ID	1014-FEED	T014-FEED	1014-EFF Comp	T014-EFF Comp			
Battelle Pacific N Nuclear Chemistr	Procedure:	By Manual Titrat	Titrant HCI	RPG # S	17-1240 1	17-1240 Dup 1	17-1241 T	17-1241 Dup 1			

Instrument Calibration

Buffer	Vendor	Lot Number	Expire Date
4	Inorganic Ventures	J2-WCS02116	1-Jul-18
7	Inorganic Ventures	J2-WCS02097	1-Mar-18
10	Inorganic Ventures	H2-WCS02089	1-Jan-18
2-nd Verif	Vendor	Lot Number	Expire Date
7	Ricca	1603F33	1-Mar-18

Slope 96.4%

ASR 0335 (Fiskum).xls

Page 1 of 2

10/30/2017

ETCOLOR.

Battelle Pacific Northwest Laboratory Nuclear Chemistry and Engineering Group

Rev-0 ASR # and Rev # 0335

N82456

#dM

Procedure:

Alkalinity of Aqueous Solutions, Leachates and Supernates

Equip # Beckman Coulter 560, SN#110650046

By Manual Titration

		Molarity base	1.340	1.244	1.148	1.148			
		Found millimoles base	0.134	0.124	0.115	0.115			
	ence	Hq	7.090	7.170	7.690	7.160			
	2nd Equival	Point Titrant Vol. (mL)	2.500	2.400	2.300	2.300			
	1.1	Sample Vol. (mL)	0.100	0.100	0.100	0.100			
Molarity	0.0957								
Titrant	HCI	RPG #	17-1240	17-1240 Dup	17-1241	17-1241 Dup			

uivale oint rant mL)	nce pH	Found millimoles base	Molarity base
300	3.770	0.077	0.766
000	3.720	0.077	0.766
000	3.920	0.067	0.670
8	3.860	0.077	0.766
+			

ASR 0335 (Fiskum).xls

Battelle Pacific	Northwest I aboratory				A SR #	H wed pue	1225	Dav-0				
Nuclear Chemi	stry and Engineering Group				Project:	Client: 69833	S. Fiskum WP#	N82456	Re Anal	sport Date: ysis Date:	10/9/2017 9/26/2017 9/38/2017	
Procedure	: RPG-CMC-228-Rev 0.1 Alkalinity of Aqueous Solutions, L	Determination of F cachates and Supern	lydroxyl (OH. ates	-) and			S	readsheet: OH	l I-TemplateLoo	cked07.xls	4/1/2007	
By Manual Titr	ation	Equip #	110650046	Lab Loc. 3	10		Analyst.	0 210	FO	00	Filefor	
		Chem					- netrone	p.p		}	1 1/20/01	
Titrant	Molarity	Rec#	112					,	H+			
NaOH	0.0101	232						Diluted	1st Equivale	nce		
			Dilution	Sample	Sample	Density	Titrator Routine	Initial pH	Point Titrant		Found millimoles	Molarity
RPG #	Sample ID		Factor	Vol. (mL)	Wt. (g)	g/mL	#	reading	Vol. (mL)	Hd	acid	acid
17-1242	TI014-ELComp		102.4	0.100	NA	NA	NA	8.07	0.10	9.77	0.001	1.034
17-1242 Dup	TI014-ELComp		102.4	0.100	NA	NA	NA	7.99	0.10	9.95	0.001	1.034
								1.00				
Instrument Cali	ibration				Slope							
Buffer	Vendor	Lot Number	Expire Date		96.4%							
4	Inorganic Ventures	J2-WCS02116	1-Jul-18	I								
7	Inorganic Ventures	J2-WCS02097	1-Mar-18									
10	Inorganic Ventures	H2-WCS02089	1-Jan-18									
2-nd Varif	Vandor	I of Mumber	Evnira Data									

Expire Date 1-Mar-18	Lot Number 1603E33	Vendor Ricca	Verif
Expire Date	Lot Number	Vendor	erif
1-Jan-18	H2-WCS02089	Inorganic Ventures	-
1-Mar-18	J2-WCS02097	Inorganic Ventures	
1-Jul-18	J2-WCS02116	Inorganic Ventures	
Expire Date	Lot Number	Vendor	er

Page 1 of 2

ASR 0335 (Fiskum).xls

10/30/2017

B.56

Battelle Pacific Northwest Laboratory Nuclear Chemistry and Engineering Group

ASR # and Rev # 0335 Rev-0

N82456 0

WP#

Procedure:

By Manual Titration

Alkalinity of Aqueous Solutions, Leachates and Supernates

Equip # Beckman Coulter 560, SN#110650046

Point Found RPG # Yol. (mL) PH acid Vol. (mL) PH acid Nol. (mL) <th>litrant VaOH</th> <th>Molarity 0.0101</th> <th></th> <th>2nd Equiva</th> <th>lence</th> <th>i Î</th> <th></th>	litrant VaOH	Molarity 0.0101		2nd Equiva	lence	i Î	
	HG #		Sample Vol. (mL)	Point Titrant Vol. (mL)	Hd	Found millimoles acid	Molarity acid

3rd Equivale	nce	Found	
Titrant	Цч	millimoles	Molarity
V 01. (1111.)	hid	acia	acin

ASR 0335 (Fiskum).xls









B.61





B.63

Battelle - Pacific Northwest National Laboratory Analytical Support Operations – IC Report PO Box 999, Richland, Washington 99352



ASR-0335 Fiskum Report-Rev1.doc

Page 1 of 3

IC Report

Sample Results

See Attachment: Sample Results ASR 0335

Revision 1

The report has been revised to correct the sulfate results. Sulfate results were originally reported non-detect but measurable quantities of sulfate are present in the samples.

Revision 2

The report has been revised to correctly report the analytical spike results.

Sample Analysis/Results Discussion

Two liquid samples were analyzed. The sample data is reported in μ g/mL and samples required dilutions of up to 29,282 fold in order to obtain valid anion results within the calibration range. All results have been adjusted for all analytical dilutions. The analytes of interest for these samples include chloride, nitrite, nitrate, phosphate, oxalate and sulfate. The estimated method detection limits (MDL) are provided, and are based on the estimated quantitation limit (EQL), which is one-tenth of the lowest calibration standard (adjusted for the dilutions used for reporting the results).

Data Limitations

There are no limitations regarding this data. All QC sample requirements were met.

Quality Control Discussion

The method performance is evaluated against the acceptance criteria established by the analysis procedure and the Analytical Support Operations QA Plan, ASO-QAP-001.

IC Workstation QC Results

<u>Dilution Blank</u>: One dilution blank was analyzed with the sample batch. No anions were detected above the MDL.

<u>Duplicate:</u> Requested per the client, duplicate analysis was performed on both samples. The relative percent difference (RPD) is reported for all analytes which were measured at or above the EQL. The reported RPDs ranged from 0.63%-6.4%, which meets the ASO's QA Plan acceptance criteria of <20%.

<u>Instrument LCS</u>: The routine instrument LCS analyzed with the run met the acceptance criteria of 80% to 120% recovery (See attached sample results).

IC Report

<u>Analytical Spike (AS)</u> (Accuracy): Analytical spikes were prepared, as no sample preparation was required for this analysis. Aqueous samples were analyzed as received. Results are reported using the 14641x dilution of the two liquid samples by adding a known concentration of mid-range multi-mix standard, "CCV 082917". The analytical spike recoveries ranged from 98% to 105%, meeting the method acceptance criteria of 75% to 125%.

<u>Continuing Calibration Verification (CCV) / Continuing Calibration Blank (CCB)</u>: Numerous CCVs and CCBs were analyzed throughout the analysis run and the ICV/ICB are included in these results(See attachment "QC Sample Results and Performance"). All CCVs and CCBs affecting reported data produced results within the acceptance criteria of the QA Plan (i.e., 90% to 110% recovery for the CCV and results less than estimated quantitation limit or less than 5% of reported sample result for the CCB).

Deviations from Procedure

None

General Comments

- The reported "Final Results" have been corrected for all dilutions performed on the sample during processing or analysis.
- For each anion, the instrument EQL is defined as the concentration of the lowest calibration standard and the instrument MDL is set at one-tenth of the EQL. The MDLs and EQLs reported for each sample are adjusted for the sample dilution factors (processing and analysis) and assume non-complex aqueous matrices. Matrix-specific MDLs or EQLs may be determined, when requested.
- Routine precision and bias are typically ±15% or better for non-complex aqueous samples that are free of interference.

Sample Results ASR

35	
03	

		and the second second	The second		No. of the Party of the	20.1	and the second second	No. of the second se	- Inc	and the state of the	Contraction of the second s	1022			ion ³			101	
		MDL	Result		MDL	Result		MDL	Result	CIAN CAL	MDL	Result		MDL	Result		MDL	Result	
CPL Number	Client Sample ID	Jm/gui	Jm/gui	DF	Jm/ari	Jm/ani	DF	ug/mL,	ing/mL	DF	Jm/gui	ug/mL	DF	ug/mL	Jm/ani	DF	hg/mL	μg/mL	DF
7-1240	TI014-FEED	100	4,340		1,500	63,400		200	1,800		12	192		2,900	113,000		15	1,050	
7-1241	TI014-EFF Comp	001	4,380		2,900	61,300		200	006'1		12	176		5.900	114,000		15	179	
Dilution Blank	Dilution Blank (9/7/17)	0.08	0.08	n	0.10	0.10	n	0.15	0.15	n	0.10	0.10	n	0.20	0.20	n	0.12	0.12	D
Dilution Blank	Dilution Blank (09/13/17)	0.08	0.08	n	0.10	010	n	0.15	0,15	n	0.10	0,10	n	0,20	0,20	Ū	0.12	0,12	ŋ

Sample QC Results ASR

0335

Sample/Replicate Precision Results

	No. of the second se	-	5	ž	2.	ŝ	0,	C2	0	ON		Ā	"0
RPL Number	Sample ID	hg/mL	RPD	Jm/gu	RPD	Jm/gu	RPD	Jm/gu	RPD	hg/mL	RPD	Jm/gut	RPD
17-1240	Sample	4,340	:	63,400	ŧ	1,800	i	192	1	113,000	:	1,050	1
	Duplicate RPD	4,430	2.1	63,800	0.63	1,800	0	197	2.6	114,000	0.9	985	6.4
17-1241	Sample	4,380	:	61,300		1,900	4	176	1	114,000	1	126	1
	Duplicate RPD	4,330	E	62,300	1.6	2,000	5.1	179	1.7	111,000	2.7	962	0.93

Sample Spike Results - At IC Workstation

		5	ALC: NAMES OF	N) ₂	SC	04	c_2	0,	NO	3	PC	1
RPL Number	Sample ID	hg/mL	%Rec	µg/mL	%Rec	Jm/gu	%Rec	Jm/gu	%Rec	hg/mL	%Rec	Jm/gu	%Rec
17-1240	Sample	4,340		63,400	1	1,800	1	192	1	113,000		1,050	1
	AS Sample *	1.65	100	4.12	98	3.11	101	2.05	103	7.47	104	2.52	101
17-1241	Sample	4,380	1	61,300	1	1,900	Ē	176	1	114,000		1/6	1
	AS Sample *	1,68	102	4.11	101	3.15	105	2.09	105	7.33	66	2.52	101

LCS/Blank Spike Results

9/7/2017 11:00 LCS 082917 98.9 96.3 102.8 104.2 9/13/2017 13:41 LCS 082917 99.5 95.8 101.5 NA	Run ID	Sample ID	CI % Rec	NO ₂ % Rec	SO4 % Rec	C ₂ O ₄ % Rec	NO3 % Rec	P04 % Rec
9/13/2017 13:41 ILCS 082917 09 5 05 95 8 101 5 NA	9/7/2017 11:00	LCS 082917	98.9	96.3	102.8	104.2	96.5	103.8
	9/13/2017 13:41	LCS 082917	99.5	95.8	101.5	NA	100.8	1.801

AS = Analytical Spike: Spike performed at IC Workstation on Liquid Samples, LCS = Laboratory Control Sample (or Blank Spike) RPD = Relative Percent SaRec = Percent Recovery DF = Data Quality Flag

U = Not Detected Above Method Detection Limit (MDL) = Detected Result are Qualitative: Resul) >MDL but <EQL (Estimated Quantitation Limit) = = Value Not Calculated or Place Holder for Blank Cell MA = Oxalate tends to fall out of Solution after a short time period, no oxalate results are reported associated with this run date, * = Dilution factor is on applied to this result.

Sample Results ASR 0335

Project Number:	69833
Charge Code:	N79882
ASR Number:	0335
Client:	S. Fiskum
Total Samples:	2 liquid

	Sample
RPL Numbers	17-1240 and 17-1241
Client IDs	TI014-FEED and TI014-EFF Comp

Analysis Procedure	RPG-CMC-386 Rev. 1, "Carbon Measured in Solids,
	Sludge, and Liquid Matrices"
Prep Procedure	None
Analyst	C. Rutherford
Analysis Date	October 18, 2017
CCV Standards	TIC/TOC CMS# 509806 and 521011
BS/LCS/MS Standards	TIC/TOC CMS# 520858 and 520712
Excel Data File	ASR 0335
M&TE Numbers	Carbon System (WD36639, RPL/701)
	Balance : Sartorius R200D, S/N 30809774
All Analysis Records	System File TC-18-1

Juln. Pick For C. Rotho Ford 11/13/17 Prepared By Date Date Date Jul 13/17 Neviewed By Date Date
* This results is suspect - see narative for more detail

TOC in Sample 17-1241-R N 17-124	1DL (mg C/L): 240	ep (mg C/L): 5600		EQL: 1200	1DL (mg C/L): 240 EQL: 1200	41 (mg C/L): 1810* 1DL (mg C/L): 240 EQL: 1200	EQL: 1200 41 (mg C/L): 1810* 1DL (mg C/L): 240 EQL: 1200	1DL (mg C/L): 240 EQL: 1200 41 (mg C/L): 1810* 1DL (mg C/L): 240 EQL: 1200
	-	2	TIC in Sample 17-1241-R M	TIC in Sample 17-1241-R M	N TIC in Sample 17-1241-R	TOC in Sample 17-12 N TIC in Sample 17-1241-R	TOC in Sample 17-12 N TIC in Sample 17-1241-R	N TOC in Sample 17-12 N TIC in Sample 17-1241-R
		mg C/L):	mg C/L): mg C/L): 501	EQL: mg C/L): mg C/L):	mg C/L): EQL: mg C/L): mg C/L):	(mg c/L): (mg c/L): EQL: (mg c/L): (mg c/L):	EQL: mg C/L): EQL: mg C/L): mg C/L): mg C/L):	mg c/L): EQL: mg c/L): EQL: EQL: mg c/L): mg c/L): mg c/L):
601 7500		C/L): 120 MDL (mg C/L): 24	C/L): 5890 TIC in Sample 17-1241-Rep (mg C/L): 56 C/L): 120 MDL (mg C/L): 24	EQL: 480 EQL: 12 C/L): 5890 TIC in Sample 17-1241-Rep (mg C/L): 56 C/L): 120 MDL (mg C/L): 24	C/L): 96 MDL (mg C/L): 24 EQL: 480 EQL: 12 C/L): 5890 TIC in Sample 17-1241-Rep (mg C/L): 56 C/L): 120 MDL (mg C/L): 24	C/L): 2610 TOC in Sample 17-1241 (mg C/L): 181 C/L): 96 MDL (mg C/L): 24 C/L): 96 EQL: 24 C/L): 480 EQL: 12 EQL: 480 EQL: 12 C/L): 5890 TIC in Sample 17-1241-Rep (mg C/L): 56 C/L): 120 MDL (mg C/L): 24	EQL: 480 EQL: 12 C/L): 2610 TOC in Sample 17-1241 (mg C/L): 18 C/L): 96 MDL (mg C/L): 24 C/L): 96 EQL: 12 C/L): 96 EQL: 12 C/L): 5890 TIC in Sample 17-1241-Rep (mg C/L): 56 C/L): 120 TIC in Sample 17-1241-Rep (mg C/L): 56	C/L): 96 MDL (mg C/L): 24 EQL: 480 EQL: 12 C/L): 2610 TOC in Sample 17-1241 (mg C/L): 181 C/L): 96 MDL (mg C/L): 181 C/L): 96 MDL (mg C/L): 24 C/L): 96 EQL: 12 C/L): 5890 TIC in Sample 17-1241-Rep (mg C/L): 56 C/L): 120 MDL (mg C/L): 24

Results

Table 1: TOC/TIC Results for ASR 0335

B.69

Sample Analysis/Results Discussion

Two liquid samples were submitted under Analytical Service Request (ASR) 0335 for total inorganic and total organic carbon analysis. The analysis was performed by the hot persulfate wet oxidation method, with the results summarized in Table 1. The TIC is measured first with additions of heated sulfuric acid followed by the addition of a silver catalyzed acidic potassium persulfate solution for oxidation at 92-95 °C for TOC. The analyses were performed following procedure RPG-CMC-386, Rev. 1, *Carbon Analyses in Solids, Sludge and Liquid Matrices*

The two samples were analyzed in duplicate, with sample 17-1241 also run in replicate. Sample 17-1240 was selected for the analytical spike. For sample 17-1241 the original sample and the replicate sample are compared in Table 1 for RPD. The sample results are corrected for the contribution from the system blank, as per procedure RPG-CMC-386, Rev. 1. All data are reported as μ g C/mL of sample.

Data Limitations

None

Quality Control Discussion

The calibration and QC sample standards for the TOC initial/continuing calibration verification check (ICV/CCV) sample is a 1000 μ g/mL solution of total organic carbon standard. The calibration and QC sample standards for the TIC initial/continuing calibration verification check (ICV/CCV) sample is a 1000 μ g/mL total inorganic standard. The identification of the standards and their Chemical Management System (CMS) numbers are included on the raw data bench sheets for traceability.

The QC samples analyzed as part of the method include initial and continuing calibration verification samples (ICV/CCV), initial and continuing calibration blanks (ICB/CCB), laboratory duplicates for each sample, a laboratory control sample/blank spike (LCS/BS), and an analytical spike (AS). The work was performed in one batch.

Two blanks are run at the beginning of each batch and after ICV/CCV. The blanks must be <EQL. The banks run in the batch are <EQL.

Initial Calibration Check and Continuing Calibration Verification Standards:

The calibration of the coulometer analysis system was checked by calibration verification standards analyzed at the beginning and end of the analysis run. TOC results for the two ICVs runs were 103.3% and 103.2% recovery, and for the two TIC ICVs runs the results were 97.2% and 99.7% recovery, within the acceptance criterion of 90% to 110%. The TOC result for the CCV run was 99.0% recovery and the TIC CCVs run was 102.1% recovery within the acceptance criterion of 85% to 115%.

- Laboratory Control Sample/Blank Spike: One TIC and TOC LCS/BS was analyzed. The TIC LCS/BS result was 100.7% recovery and the TOC LCS/BS result was 100.1% recovery, meeting the acceptance criteria of 75% to 125%.
- <u>Duplicate/Replicate</u>: Precision of the carbon measurements is demonstrated by the relative percent difference (RPD) between sample and duplicate/replicate. Sample 17-1240 TIC RPD was 1.5% and the TOC was 0.8%. The RPD for sample 17-1241 TIC was 6.3% and the TOC RPD was 41.6%. The RPD for 17-1241 TOC was the only RPD that did not meet the acceptance criteria of $\leq 20\%$. Note: For sample 17-1241, the duplicate result (2715 mg/L) and replicate result (2769 mg/L) have an RPD of 2.0%. The result reported for 17-1241 is likely suspect. The result for 17-1241-rep reported in Table 1 is likely a more accurate result.
- <u>Analytical Spike (AS)</u>: The accuracy of the carbon measurements can be estimated by the recovery from the AS. One AS sample was run with the batch. Sample 17-1240 was selected. The results for the analytical spike for TIC is 98.8% recovery and for the TOC, 98.2% recovery. The AS recovery for the TOC and TIC results meets the acceptance criterion of 75% to 125%.

Deviation from Procedure: None

General Comments

- 1) Routine precision and bias are typically $\pm 15\%$ or better for non-complex samples that are free of interferences.
- 2) For the TIC/TOC, the analysis MDL is calculated by dividing the batch IDL by the sample mass and is therefore dependent on sample size. The estimated quantitation limit (EQL) is defined as 5x the MDL. Results <5x MDL have higher uncertainties and RPDs are not calculated if the results are <5x MDL.
- 3) Where applicable, the reported "Final Results" have been corrected for any dilution performed on the sample prior to analysis.

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

Gamma Energy Analysis (GEA)

Project / WP#:	69833/ N82456
ASR#:	0335.00
Client:	SK Fiskum
Total Samples:	3

RPL ID	Client Sample ID
17-1240	TI014-FEED
17-1241	TI014-EFF-Comp
17-1242	TI014-ELComp

Analysis Type:	GEA- for all positively measured or non-detected isotopes
Sample Processing Prior to Radiochemical Processing/Analysis	 None Digested as per RPG-CMC-129, Rev. 0 HNO₃-HCl Acid Extraction of Solids Using a Dry Block Heater Fusion as per RPG-CMC-115, Solubilization of Metals from Solids Using a KOH-KNO₃ Fusion Other: Preparation may also involve attaining a GEA geometry that is compatible with the calibration geometry.
Analysis Procedure:	RPG-CMC-450, Rev. 3 Gamma Energy Analysis (GEA) and Low-Energy Photon Spectrometry (LEPS)
Reference Date:	August 25, 2017 @ 9:00 am
Analysis Date or Date Range:	August 25, 2017
Technician/Analyst:	T Trang-Le
Rad Chem Electronic Data File:	17-1240 Fiskum.xls
ASO Project 98620 File:	File Plan 5871, T4.4 Technical (Radiochemistry), Gamma Calibration, daily checks, and maintenance records; and T3 standard certificates and preparation. Also, balance calibration and performance check records.
M&TE Number(s):	Detectors G, & T

IRang-le Prepare

1<u>9/15/17</u> <u>LR Greenwood</u> <u>19/15/17</u> Date Reviewet

Battelle PNNL/RSE/ASO Radiochemistry Analysis Report

SAMPLE RESULTS

Activities for all gamma emitters detected in this sample are presented in an attached Excel spreadsheet for ASR 0335.00. All sample results for all target isotopes are reported in units of μ Ci/mL with estimates of the total propagated uncertainty reported at the 1-sigma level.

ASO Project File, ASR 0335 has been created for this report including all appropriate supporting records which may include the Pipette Performance Check Worksheet form, standard certificates, laboratory bench records, Shielded Analytical Laboratory Bench Sheet, and Gamma Energy Analysis printouts. Detector calibration records, control charts and balance calibration records can be found in the ASO Records.

Sample Preparation, Separation, Mounting and Counting Methods

All three samples were received from the Shielded Analytical Laboratory (SAL).

Sample 17-1242 (TI014-EL-Comp) was prepared for dilution in the hot cell by filling the diluent vial with 4 aliquots of \sim 5mL each of DI water in lab 309. The diluent vial was loaded into the hot cell where \sim 0.2mL of AP-105 Cs Eluent was added to the diluent vial. The vial was then loaded out of the hot cell clean.

Samples 17-1240 and 17-1241 (TI014-FEED and TI014-EFF-Comp) were prepared by pipetting 2 mL of sample solutions provided to the ASO.

The quality control (QC) steps for direct GEA are discussed below.

QUALITY CONTROL RESULTS

Tracer:

Tracers are not used for ASO GEA methods.

Process Blank (PB):

No process blank was prepared for gamma counting.

Required Detection Limits

There are no required detection limits for these samples.

Blank Spike (BS)/Laboratory Control Sample (LCS)/ Matrix Spike (MS):

There are no BS, LCS or MS samples analyzed for ASO GEA analyses. Instrument performance is assessed by the analyses of daily control counts and weekly background counts, as discussed below.

Page 2 of 3

Battelle PNNL/RSE/ASO Radiochemistry Analysis Report

Duplicate Relative Percent Difference (RPD):

No duplicate samples were provided for gamma counting.

Instrument Calibration and Quality Control

Gamma detectors are calibrated using multi-isotope standards that are NIST-traceable and prepared in the identical counting geometry to all samples and detectors. Counter control sources containing Am-241, Cs-137 and Co-60 are analyzed daily before the use of each detector. Procedure RPG-CMC-450 requires that a counter control source is checked daily and must be within ±3 sigma or ±3% of the control value, whichever is greater. Gamma counting was not performed unless the control counts were within the required limits. Background counts are performed on all gamma detectors at least weekly for either an overnight or weekend count. The most recent background is subtracted from all sample counts. Samples 17-1240 and 17-1242 had much higher dose rates than sample 17-1241. Consequently, these two samples were counted on a remote position on our track detector (T) while sample 17-1241 was counted on the face of detector G. We are not calibrated for a 2 mL geometry on detector T at the remote position (P20) used to count samples 17-1240 and 17-1242. However, this detector counts the glass scintillation vials through the side such that the height of liquid in the vial does not have any significant impact on the calibration geometry at such remote distances. Consequently, the calibration using 10 mL of liquid should be very close to that using 2 mL of liquid.

Assumptions and Limitations of the Data

None

Interferences/Resolution

None.

Uncertainty

For gamma counting, the uncertainty in the counting data, photon abundance and the nuclear half-life, and efficiency are included in the calculation of the total uncertainty along with a systematic uncertainty for sample prep. The Canberra Genie software includes both random and systematic uncertainties in the calculation of the total uncertainties which are listed on the report. We conservatively estimate that 2% is the lowest uncertainty possible for our GEA measurements taking into account systematic uncertainties in gamma calibration standards.

Comments

None

Attachment: Data Report Sample Results for ASR 0335.

Radiochemical Sciences and Engineering Group Pacific Northwest National Laboratory PO Box 999, Richland, WA

filename 17-1240 Fiskum.xls 9/15/2017

1800-18 d/15/2017 Prepared by: Project: 69833 WP: N82456 Client: Fiskum ASR 0335

1021 2 0 Technical Reviewer: 2 R. Hullmund

RPG-CMC-450, Rev. 3, Gamma Energy Analysis (GEA) and Low-Energy Photon Spectrometry (LEPS) Procedures:

Gamma detectors G, T 25-Aug-17 M&TE:

Count dates:

Measured Activity, μCi per ml ± 1s

	41	±5%
	Am-2	< 5.4E-1 3.72E-04 < 1.3E+1
	54	± 2%
	Eu-1	< 1.2E-2 1.35E-04 < 2.4E-1
	23	+ 9%
	Eu-15	< 6.3E-3 6.07E-06 < 3.3E-1
	37	±2% ±2% ±2%
	Cs-1.	1.22E+02 5.30E-04 2.43E+03
	-126	± 2%
	Sn/Sb	<1.1E-2 1.75E-05 <1.5E-1
	20	± 2%
	C0-0	< 3.2E-3 5.93E-04 < 8.6E-2
Lab	ID	17-1240 17-1241 17-1242
	Sample	T1014-FEED T1014-EFF-Comp T1014-ELComp

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

o by Liquid	Schulation Spectronic
Project / WP#:	69833/N82456/N79882
ASR#:	0335
Client:	SK Fiskum
Total # of	2
Samples:	3

Sr-90 by Liquid Scintillation Spectrometry

RPL ID	Client Sample ID
17-1240	TI014-FEED
17-1241	TI014-EFF-Comp
17-1242	TI014-ELComp

Analysis Type:	Sr-90
Sample Processing Prior to Radiochemical Processing/Analysis	 None Digested as per RPG-CMC-128, Rev.1, HNO₃-HCL Acid extraction of Liquids for Metals Analysis Using a Dry-Block Heater Fusion as per RPG-CMC-115, Solubilization of Metals from Solids Using a KOH-KNO₃ Fusion Other:
Pre-dilution Prior to Radiochemical Processing?	□ No □ Yes
Total Alpha and Beta Preparation Procedure:	RPG-CMC-4001, Rev. 1, Source Preparation For Gross Alpha and Gross Beta Analyses.
Spike Standard ID's	R-693-b-2 (Sr-90)
Analysis Procedure Technician/Analyst:	RPG-CMC-408, Rev. 2, <i>Total Alpha and Beta Analysis</i> T. Trang-Le and CZ Soderquist, (09/21/2017)
Separation Procedure: Spike Standard ID:	RPG-CMC-476, Rev. 0, Strontium Separation Using Eichrom Strontium Resin R-693-b-2 (Sr-90)
Separation Date:	09/20/2017 @ 09:45 a.m.
Technician/Analyst:	L. Darnell
Analysis Procedure:	RPG-CMC-474, Rev. 1, Measurement of Alpha and Beta Activity by Liquid Scintillation Spectrometry
Reference Date:	09/20/2017
Analysis Date or Date Range:	09/21/17 & 09/29/17 (first count), 09/26/17 & 10/02/17 (second count)
Technician/Analyst:	T. Trang-Le & CZ Soderquist
Rad Chem Electronic Data File:	RPG-RC\PNL\Projects\Backup files\Backup 18\17-1240 Fiskum.xls
ASO Project 98620 File:	File Plan 5871: T 69833 0335: Sample preparation and analysis records; T-4.4 LSC 3100 calibration, daily checks, and maintenance records; and T3 standard certificates and preparation. Also, balance calibration and performance check records.
M&TE Number(s):	Perkin Elmer Tri-Carb 3100, Serial # DG08061340, RPL 425, Tri-Carb 2700TR software version 1.04 dated 9/99.

____/___///*\5/|7_* Date Trang-le Preparer

Jul Port 11/15/17 Reviewer

SK Fiskum Sr90 ASR-0335.doc

Page 1 of 3

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report

SAMPLE RESULTS

See attached data report, Sample Results for ASR 0335. All sample data are reported in μ Ci/mL with a 1- σ uncertainty (see Comments).

Sample preparation, separation, mounting, and counting

All three samples submitted under Analytical Service Request (ASR) 0335 were analyzed on duplicate for Sr-90 by chemical separation and liquid scintillation counting. All the samples were prepared in RPL/420. Aliquots of the acid digestions and pre-diluted TI014-EL-Comp were used for radioanalytical analyses; only Sr-90 data are included in this report.

QUALITY CONTROL RESULTS

Radioanalytical quality control (QC) samples prepared in samples prepared in RP/420 and include a lab blank (LB), and sample duplicate. Additional laboratory QC samples were prepared prior to separations; these include a laboratory separation blank, a reagent blank spike (RS), and a matrix spike (i.e., addition of Sr-90 standard to an aliquot of one of the samples).

Instrument Calibration Control

Laboratory Preparation Blank and Laboratory separations blank (LB):

There are no acceptance criteria for LB (see Comments).

Blank Spike (RS) - reagent spike:

The RS recovery of 101% and 96% meet the procedure acceptance criteria of 80% to 120% recovery.

Matrix Spike (MS):

The MS recovery of 93% and 87% meet the acceptance criterion of 75% to 125% recovery. Note: the MS sample was prepared "after" digestion (see comments), by adding a known Sr-90 standard quantity to an aliquot of 17-1240 (TI014-FEED).

Duplicate -- Relative Percent Difference (RPD):

Duplicate results are required to agree within $\leq 20\%$ RPD. The ASO QAP further specifies that the two results need to be >5 times the MDA or have individual uncertainties $\leq 20\%$. Duplicate results were 1-5% RPD; thus meeting the $\leq 20\%$ requirement.

Instrument Quality Control

The liquid scintillation counter is calibrated for tritium and C-14 using quenched standard sets that are purchased from the vendor. Daily control counts are then performed using a tritium, C-14, and a background count sample. The instrument software assesses the performance of the control counts and provides control charts to ensure the continuing calibration of the instrument. If the daily performance check fails, then the instrument is not used. Preventative maintenance and repairs are performed by the vendor under our service contract. The counting efficiency for Sr-90 is assumed to be 100%; therefore no specific Sr-90 calibration is performed. The LSC system calibration and performance is verified by assessing the recovery of a reagent spike and a matrix spike that are included in every batch of samples. A preparation blank (i.e., digestion blank) and a laboratory separations blank are also included with every

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report

batch of samples; the instrument background is subtracted from all results and the preparation and separation blanks are used to assess sample contamination during sample processing steps.

Assumption and Limitations of the Data

The 1- σ uncertainty reported for each Sr-90 result has been set at 2%. Although the calculated uncertainty values are less than 2% for all samples, the radiochemistry convention is to not report calculated uncertainties less than 2%, but to provide a more realistic estimate of the uncertainty in view of systematic uncertainties that are not fully accounted for in the uncertainty calculations.

Comments

- 1. The results have been corrected for all dilution factors resulting from sample processing.
- 2. Post-Digestion Spike (PS) A spike made after the initial sample preparation (e.g., fusion, digestion, or leach) is considered a PS. When extremely radioactive samples are analyzed, most of the radio-analytical spikes are made after the sample preparation (to avoid excessive consumption of spike and avoid creating unnecessary waste) and are post-digestion spikes. The MS prepared with this batch of sample is considered a PS, since the Sr-90 spike was not added prior to the digestion process.
- 3. Radiochemistry Electronic Systems File "RPG-RC\PNL\Projects\Backup files\Backup 18\17-1240 Fiskum.xls" has been created for this report. Supporting records such as Pipette Performance Verification forms, Laboratory Bench Record, Laboratory Sample Preparation Bench Sheet, Standards Certifications and preparation records, and balance calibration and performance check records are maintained per RS&E Group ASO File Plan 5871.
- 4. Sample results are compared to the process blank results to evaluate if the blank contains 5% or more of the measured isotope; the process blank results have been adjusted for all processing factors for the evaluation of the 5% criterion.
- 5. The stated $1-\sigma$ uncertainty represents the total propagated error associated with processing and counting operations and includes weighing errors, volume uncertainties, and counting error.

Attachment: Data Report -- Sample Results for ASR 0335

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

I THEOINGHI I	
Project / WP#:	69833/N82456/N79882
ASR#:	0335.00
Client:	SK Fiskum
Total # of Samples:	3

Plutonium 238, 239+240 Analysis

RPL ID	Client Sample ID
17-1240	TI014-FEED
17-1241	TI014-EFF-Comp
17-1242	TI014-ELComp

Analysis Type:	AEA – Pu-238, Pu-239+240
Sample Processing Prior to Radiochemical	🗌 None
Processing/Analysis	Digested as per PNL-ALO-106, Rev. 1, Acid Digestion of Waters, Soils, and Sludges for Subsequent Radiochemical Sample Analyses
	□ Fusion as per RPG-CMC-115 Rev. 0, Solubilization of Metals from Solids Using a KOH-KNO ₃ Fusion
	Digested as per RPG-CMC-128, Rev.1, HNO ₃ -HCL Acid extraction of Liquids for Metals Analysis Using a Dry-Block Heater
Pre-dilution Prior to Radiochemical Processing?	□ No
	\boxtimes Yes example 2 mL to 100 mL; 50x dilution
Total Alpha Preparation Procedure:	RPG-CMC-4001, Rev. 1, Source Preparation For Gross Alpha and Gross Beta analyses.
Technician/Analyst:	LP Darnell, (08/31/2017)
Spike Standard ID's	R-687-a-2 (Pu-239), R-693-b-2 (Sr-90)
Analysis Procedure	RPG-CMC-408, Rev. 2, Total Alpha and Beta Analysis
Technician/Analyst:	LP Darnell and T. Trang-Le (09/01/2017)
Plutonium Separation Procedure:	RPG-CMC-4017, Rev. 0, Analysis of Environmental Water Samples for Actinides and Strontium-90
Technician/Analyst:	LP Darnell, (10/16/2017)
Co-Precipitation Procedure:	RPG-CMC-496, Rev. 1, Coprecipitation Mounting of Actinides for Alpha Spectroscopy
Technician/Analyst:	LP Darnell, (10/16/2017)
Spike and Tracer Standard ID's:	R-687-a-2 (Pu-239), R-688-a-3 (Pu-242 tracer)
Analysis Procedure:	RPG-CMC-422, Rev. 2, Solutions Analysis: Alpha Spectrometry
Reference Date:	Same as analyses dates
Analysis Date or Date Range:	October 9-10, 2017
Technician/Analyst:	T. Trang-Le
Analysis Data (File):	RPG-RC\PNL\Projects\Backup files\Backup 18\17-1240 Fiskum.xls
CMC Project 98620 File:	File Plan 5871: T 69833: Sample preparation and analysis records; T-4.4 Alpha Detector calibration, calibration verification checks, and maintenance records; and T3 Standard certificates and preparation. Also balance calibration and instrument performance checks.
M&TE Number(s):	Ortec AEA counters - 32 counters - See attached M&TE list
TTrang-le / 11/	15/17 Jul/ al 11/15/17
Preparer Date	Reviewer Date

Page 1 of 3

Sample Results

See attached data report, Sample Results for ASR 0335. All data are reported in units of μ Ci per mL with a 1- σ uncertainty.

Sample Preparation, Separation, Mounting and Counting Methods

All three samples submitted under Analytical Service Request (ASR) 0335 were analyzed in duplicate for plutonium by Alpha Spectrometry. All the samples were prepared in RPL/420. Aliquots of the acid digestions and pre-diluted TI014-EL-Comp were used for radioanalytical analyses; only Pu-AEA data are included in this report.

Following the digestion process of samples TI014-FEED and TI014-EFF-Comp along with the pre-diluted TI014-EL-Comp, the Pu was separated by anion exchange using procedure RPG-CMC-4017. The separated Pu fraction was then mounted for alpha spectrometry by co-precipitation using procedure RPG-CMC-496, and then counted using alpha spectrometry using procedure RPC-CMC-422. The samples were counted on October 9-10, 2017; no decay corrections were made.

Alpha and beta analyses were performed on each sample for checking the internal consistency of the Pu alpha isotopic data. Gross alpha and gross beta activity were measured by evaporating small aliquots of leachate onto counting planchets per procedure RPG-CMC-4001 and counting per procedure RPG-CMC-408.

QUALITY CONTROL RESULTS

Quality control (QC) samples prepared in RPL/420 include a laboratory blank (LB) and sample duplicates. Additional QC samples were prepared prior to alpha counting including a laboratory blank, a reagent blank spike (RS, Pu-239), and addition of Pu-239 standard to an aliquot of the sample digestate selected as the matrix spike (MS).

The QC sample results for Pu-AEA have been evaluated and are discussed below. A summary of the Pu-AEA analysis results, including QC sample performance, is given in the attached data report.

Tracer:

The Pu-242 tracer is added to every sample after appropriate dilution and prior to plutonium separations. The use of a Pu-242 tracer corrects for radiochemical yield and mathematically removes the detector counting efficiency from the results calculations. Tracer recovery is required to be high enough to provide acceptable counting statistics. The Pu-242 tracer counting statistics were acceptable for all samples. The tracer recoveries ranged from 82% to 98%.

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report

Laboratory Preparation Blank (PB):

There is no PB blank for this batch of sample. There are no acceptance criteria for Laboratory blank.

Blank Spike (BS)/Reagent Spike (RS):

The BS recovery of 98% (Pu-239) meets the acceptance criteria of 80% to 120% recovery.

Matrix Spike (MS):

The MS recovery of 94% (Pu-239) meets the acceptance criterion of 75% to 125% recovery. Note: The MS sample was prepared after digestion, by adding a known quantity of Pu-239 standard to a diluted aliquot of the digestate. Sample number 17-1240 (TI014-FEED) was selected as the matrix spike sample.

Laboratory Duplicate - Relative Percent Difference (RPD):

The Pu-238 sample 17-1241 and duplicate pair was 9%, within DQO. The duplicate pairs for sample 17-1240 (22% RPD) and 17-1242 (40% RPD) were outside of acceptance limits. The RPD for Pu-239+240 is 3-17%, within the acceptance limit of \leq 20% RPD.

Instrument Quality Control

Alpha counters receive initial calibration with NIST traceable sources to determine the counter efficiency. When internal tracers are not used, the counter efficiency is used in calculation of final results.

Detector backgrounds are determined every 4 weeks or after the last analytical run, whichever is longer. Detector background counts are subtracted from all subsequent sample counts. A process blank is analyzed with each analytical batch to evaluate for contamination in the sample preparation process.

Assumption and Limitations of the Data

None

Attachment: Data Report -- Sample Results for ASR 0335.

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

Neptunium 237 Analysis

Project / WP#:	69833/N82456/N79882
ASR#:	0335.00
Client:	SK Fiskum
Total # of Samples:	3

Client Sample ID	RPL ID
17-1240	TI014-FEED
17-1241	TI014-EFF-Comp
17-1242	TI014-ELComp

Analysis Type:	AEA – Np-237
Sample Processing Prior to Radiochemical	□ None
Processing/Analysis	Digested as per PNL-ALO-106, Rev. 1, Acid Digestion of Waters, Soils, and Sludges for Subsequent Radiochemical Sample Analyses
	□ Fusion as per RPG-CMC-115, Rev0, Solubilization of Metals from Solids Using a KOH-KNO ₃ Fusion
	Digested as per RPG-CMC-128, Rev.1, HNO ₃ -HCL Acid extraction of Liquids for Metals Analysis Using a Dry-Block Heater
Pre-dilution Prior to Radiochemical Processing?	□ No
	Yes example 2 mL to 100 mL; 50x dilution
Total Alpha and Beta Preparation Procedure:	RPG-CMC-4001, Rev. 1, Source Preparation For Gross Alpha and Gross Beta analyses.
Technician/Analyst:	LP Darnell, (08/31/2017)
Spike Standard ID's	R-687-a-2 (Pu-239), R-693-b-2 (Sr-90)
Analysis Procedure	RPG-CMC-408, Rev. 2, Total Alpha and Beta Analysis
Technician/Analyst:	LP Darnell and T. Trang-Le (09/01/2017)
Neptunium Separation Procedure:	RPG-CMC-4017, Rev. 0, Analysis of Environmental Water Samples for Actinides and Strontium- 90
Technician/Analyst:	LP Darnell, (10/16/2017)
Spike Standard ID's:	R-686-a-39 (Np-237)
Co-Precipitation Procedure:	RPG-CMC-496, Rev. 1, Coprecipitation Mounting of Actinides for Alpha Spectroscopy
Technician/Analyst:	LP Darnell, (10/16/2017)
Analysis Procedure:	RPG-CMC-422, Rev. 2, Solutions Analysis: Alpha Spectrometry
Reference Date:	Same as counting dates
Analysis Date or Date Range:	October 16-17, 2017
Technician/Analyst:	T. Trang-Le
Analysis Data (File):	RPG-RC\PNL\Projects\Backup files\Backup 18\17-1240 Fiskum.xls
CMC Project 98620 File:	File Plan 5871: T 69833: Sample preparation and analysis records; T-4.4 Alpha Detector calibration, calibration verification checks, and maintenance records; and T3 Standard certificates and preparation. Also balance calibration and instrument performance checks.
M&TE Number(s):	Ortec AEA counters - 32 counters - See attached M&TE list
T Trang-le / 11/	15/17 Jul And 11/15/17
Preparer Date	Reviewer Date

1 Ikang-le Preparer

Reviewer -

Date

Page 1 of 4

Sample Results

See attached data report, Sample Results for ASR 0335. All data are reported in units of uCi per mL with a $1-\sigma$ uncertainty unless noted otherwise (see comments).

Sample Preparation, Separation, Mounting and Counting Methods

All three samples submitted under Analytical Service Request (ASR) 0335 were analyzed in duplicate for neptunium by Alpha Spectrometry. All the samples were prepared in RPL/420. Aliquots of the acid digestions and pre-diluted TI014-EL-Comp were used for radioanalytical analyses; only Np-AEA is reported in this report.

Following the digestion process of samples TI014-FEED and TI014-EFF-Comp along with the pre-diluted TI014-EL-Comp, the Np was separated from the sludge leachate using anion exchange chromatography using procedure RPG-CMC-4017. The separated Np fraction was mounted for alpha spectrometry by co-precipitation using procedure RPG-CMC-496, and then counted by alpha spectrometry using procedure RPC-CMC-422. The samples were counted on October 16-17, 2017; no decay corrections were made.

Alpha and beta analyses were performed on each sample for checking the internal consistency of the Np alpha isotopic data. Gross alpha and gross beta activity were measured by evaporating small aliquots of leachate onto counting planchets per procedure RPG-CMC-4001 and counting per procedure RPG-CMC-408.

QUALITY CONTROL RESULTS

Quality control (QC) samples include a laboratory blank, sample duplicate. Additional QC samples were prepared prior to separations; these include a laboratory blank, and a reagent blank spike (BS), and a matrix spike (MS) made by adding Np-237 standard to a diluted sample.

Tracer:

Tracer is not used for analyses of Np.

Laboratory Preparation Blank (PB) and Laboratory separations blank (LB):

There is no PB blank for this batch of sample. There are no acceptance criteria for Laboratory blank.

Blank Spike (BS) - reagent spike (RS):

The RS recovery of 96% meets the acceptance criteria of 80% to 120% recovery.

Matrix Spike (MS):

The MS recovery of 100% meets the acceptance criterion of 75% to 125% recovery. Note: the MS sample was prepared "after" digestion, by adding a known Np-237 standard

Page 2 of 4

quantity to an aliquot of the leachate. Sample number 17-1240 (TI014-FEED) was selected as the matrix spike sample.

Duplicate -- Relative Percent Difference (RPD):

Duplicate results are required to agree within $\leq 20\%$ RPD. Duplicate results were 1-4% RPD. The Np-237 activity detected in the samples all have 1- σ counting error of 10% or greater. 1- σ counting error at these levels indicates the activity measured in the samples is nearing the minimum detectable activity for the samples.

Instrument Quality Control

Alpha counters undergo calibration annually to determine the counter's efficiency over the normal calibration range of 3 to 6 MeV. The vendor software determines a constant detector efficiency for this energy range. Np samples are counted and results calculated using the established detector efficiency.

Detector backgrounds are determined every 4 weeks or after the last analytical run, whichever is longer. Detector background counts are subtracted from all subsequent sample counts. A process blank is analyzed with each analytical batch to evaluate for contamination in the sample preparation process.

Assumption and Limitations of the Data

None.

Comments

- 1. The results have been corrected for all dilution factors resulting from sample processing.
- 2. Post-Digestion Spike (PS) A spike made after the initial sample preparation (e.g., fusion, digestion, or leach) is considered a PS. When extremely radioactive samples are analyzed, most of the radioanalytical spikes are made after the sample preparation (to avoid excessive consumption of spike and avoid creating unnecessary waste) and are post-digestion spikes. The MS prepared with this batch of sample is considered a PS, since the Np-237 spike was not added prior to the digestion process.
- 3. The 1-sigma uncertainty represents the total propagated error associated with processing and counting operations and include; weighing errors, volume uncertainties, and counting error.
- 4. The sample results are compared to the process blank to evaluate if the blank contains 5% or more of the measured isotope; the process blank result has been adjusted for all processing factors for evaluation of the 5% criterion.

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report

5. The Laboratory Blank (LB) is prepared using laboratory reagents and provides data on the cleanliness of the radiochemistry preparation/separation processes. LB results are not normalized to processing or dilution factors associated with the samples.

_ _

6. The sample results are compared to the process blank to evaluate if the blank contains 5% or more of the measured isotope; the process blank result has been adjusted for all processing factors for evaluation of the 5% criterion.

Attachment: Data Report -- Sample Results for ASR 0335.

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

Project / WP#: 69833/N82456/N79882 ASR#: 0335.00 Client: SK Fiskum Total # of Samples: 3

Tc-99 Analysis

RPL ID	Client Sample ID
17-1240	TI014-FEED
17-1241	TI014-EFF-Comp
17-1242	TI014-ELComp

Analysis Type:	Tc-99
Sample Processing Prior to Radiochemical Processing/Analysis	 None Digested as per PNL-ALO-106, Rev. 1, Acid Digestion of Waters, Soils, and Sludges for Subsequent Radiochemical Sample Analyses Fusion as per RPG-CMC-115 Rev. 0, Solubilization of Metals from Solids Using a KOH-KNO₃ Fusion
Pre-dilution Prior to Radiochemical Processing?	 Digested as per RPG-CMC-128, Rev.1, HNO3-HCL Acid extraction of Liquids for Metals Analysis Using a Dry-Block Heater No Yes
Total Alpha and Beta Preparation Procedure:	RPG-CMC-4001, Rev. 1, Source Preparation For Gross Alpha and Gross Beta analyses.
Technician/Analyst:	LP Darnell, 08/31/2017
Spike Standard ID's	R-687-a-2 (Pu-239), R-693-b-2 (Sr-90)
Analysis Procedure	RPG-CMC-408. Rev. 2, Total Alpha and Beta Analysis
Technician/Analyst:	LP Darnell and T. Trang-Le 09/01/2017
RadioChemical Preparation Procedure:	RPG-CMC-432, Rev. 0, Technicium-99 Analysis
Technician/Analyst:	LP Darnell, 11-17-2017
Spike and Tracer Standard ID's:	R-540-b-8 (Tc-99)
Analysis Procedure:	RPG-CMC-474, Rev. 1, Measurement of Alpha and Beta Activity by Liquid Scintillation Spectrometry
Reference Date:	Not applicable
Analysis Date(s) or Date Range:	11/17/2017
Technician/Analyst:	CZ Soderquist
Analysis Data (File):	RPG-RC\PNL\Projects\Backup files\Backup 18\17-1240 Fiskum.xls
CMC Project 98620 File:	File Plan 5871: T 69833: Sample preparation and analysis records; T-4.4 Alpha Detector calibration, calibration verification checks, and maintenance records; and T3 Standard certificates and preparation. Also balance calibration and instrument performance checks.
M&TE Number(s):	Perkin Elmer 3100 TR Liquid scintillation spectrometer– See attached M&TE list

12/11/17 Date T TRang-le Preparer

Reviewer Date

Page 1 of 3

Sample Results

See attached data report, Sample Results for ASR 0335. All data are reported in units of μ Ci/g with a 1- σ uncertainty (see comments).

Sample Preparation, Separation, Mounting and Counting Methods

Three samples submitted under ASR 0335 were analyzed for Tc-99. All the samples were prepared in Laboratory 420. Aliquots of the acid digestions and pre-diluted 17-1242 and duplicate (TI014-EL-Comp) were used for radioanalytical analyses. After leaching, the leachate solution was filtered and the entire leachate was processed and analyzed using procedure RPG-CMC-432, Rev. 0, *Technicium-99 Analysis*.

The samples were counted on November 17, 2017; no decay corrections were made.

Beta analyses were performed on each sample to obtain information to estimate aliquot sizes for other analyses and for checking the internal consistency of the Tc-99 data. The beta results are included in the data tables as supplemental information only. The beta activity was measured by evaporating small aliquots of the acid digestion leachate onto counting planchets per procedure RPG-CMC-4001 and counting per procedure RPG-CMC-408.

QUALITY CONTROL RESULTS

Quality control (QC) samples prepared in laboratory 420 include a laboratory blank and sample duplicates, matrix spike, reagent spike and addition of Tc-99 standard to a separate aliquot of one of the samples.

The QC sample results for Tc-99 have been evaluated and are discussed below. A summary of the Tc-99 analysis results, including QC sample performance, is given in the attached data report.

Tracer:

There is no tracer for Tc-99 analysis.

Laboratory Preparation Blank (PB):

BLK-1240 was prepared as the PB for this batch of sample. There are no acceptance criteria for PB.

Reagent Spike (RS):

The RS recovery of 91% (Tc-99) meets the acceptance criteria of 80% to 120% recovery.

Matrix Spike (MS):

The MS recovery of 92% (Tc-99) meets the acceptance criterion of 75% to 125% recovery. Note: the MS sample was prepared after leaching, by adding a known quantity of Tc-99 standard to the leachate. Sample number 17-1240 (TI014-FEED) was selected as the matrix spike sample.

Laboratory Duplicate - Relative Percent Difference (RPD):

Duplicate results are required to agree within $\leq 20\%$ RPD. Duplicate results were 6-11% RPD. The Tc-99 activity detected in the samples all have 1- σ counting error of 10% or greater. 1- σ counting error at these levels indicates the activity measured in the samples is nearing the minimum detectable activity for the samples.

Instrument Quality Control

LSC counters receive initial calibration with NIST traceable sources to determine the counter efficiency. When internal tracers are not used, the counter efficiency is used in calculation of final results.

Detector backgrounds are determined every 4 weeks or after the last analytical run, whichever is longer. Detector background counts are subtracted from all subsequent sample counts. A process blank is analyzed with each analytical batch to evaluate for contamination in the sample preparation process.

Assumption and Limitations of the Data

None

Attachment: Data Report -- Sample Results for ASR 0335.

Page 3 of 3

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

Uranium by Kinetic Phosphorescence Analysis (KPA)

Project / WP#:	69833/N82456/N79882
ASR#:	0335.00
Client:	SK Fiskum
Total Samples:	3

Client Sample ID	RPL ID
17-1240	TI014-FEED
17-1241	TI014-EFF-Comp
17-1242	TI014-ELComp

Analysis Type:	Uranium by KPA
Sample Processing Prior to Radiochemical Processing/Analysis	 □ None ○ Digested as per RPG-CMC-128, HNO₃-HCl Acid Extraction of Liquids Using a Dry Block Heater
Spike Standard ID:	R-691-b Expiration date: 11/18/2017
Analysis Procedures:	RPG-CMC-4014, Rev. 1, Uranium by Kinetic Phosphorescence Analysis RPG-CMC-4015, Rev. 0, Analysis of Soils and Sediment Samples for Actinides and Sr-90
Analysis Date:	11/14/17
Technician/Analyst:	LP Darnell and CZ Soderquist
Rad Chem Electronic Data File:	RPG-RC\PNL\Projects\Backup files\Backup 18\17-1240 Fiskum.xls; Worksheet and Worksheet KPA standards
ASO Project 98620 File:	File Plan 5871: T 52578 8480 R1: Sample preparation and analysis records; T-4.4 Uranium KPA control charts and maintenance records; and T3 standard certificates and preparation. Also, balance calibration and performance check records.
M&TE Number(s):	Chemcheck Instruments, Inc., Model: KPA 11-R, WD59770, RPL 525; KPAWIN Software, Version 1.2.8, Release 19 July 2000.

Trang-le 12/11/17

12/11/17 Reviewer

Preparer

Date

Page 1 of 3

Sample Results

See attached data report, Sample Results for ASR 0335. The acid digestion sample results are reported in $\mu g/g$. The pre-dilution sample from the hot cell is reported in $\mu g/mL$. All results are reported with a 1- σ uncertainty (see Comments).

Sample Preparation and Analysis

All three samples were received by the Analytical Support Operations (ASO) under Analytical Service Request (ASR) 0335 for total uranium analysis by kinetic phosphorescence (KPA). Aliquots of the acid digestions and pre-diluted TI014-EL-Comp (17-1242) sample were used for this analysis.

All acid digestion aliquots and pre-diluted TI014-EL-Comp were processed through chemical separation to remove elemental interferences. Samples were diluted approximately 5-fold to ensure sample concentrations were within the calibration range.

Quality control (QC) samples prepared with the acid digestion include a process blank (PB) and sample duplicates for each samples. QC samples were not prepared with the simple dilution samples from the hot cell. The pre-diluted sample was prepared in duplicate for this analysis. Additional laboratory QC samples were prepared prior to analysis by KPA; these include a laboratory blank, a matrix spike, and two reagent blank spikes (RS/BS).

A summary of the U KPA analysis results, including QC sample performance, is given in the attached data report.

QUALITY CONTROL RESULTS

Instrument Calibration Control

Two calibration verification check standards were analyzed at the beginning of the batch analysis. The Recovery of the high range calibration verification standard is 102% and the recovery of the low range calibration standard is 99%. Both recovery values are with the acceptance criterion of 90% to 110%. Two continuing calibration verification standards were run at the end of the analytical run. The Recovery of the high range calibration verification standard is 98% and the recovery of the low range calibration verification standard is 98%. Both recovery values are with the acceptance criterion of 90% to 110%.

Process Blank (PB):

Two process blanks were prepared with the analytical run. An acid digestion blank and a lab blank (clean reagents run through the separation process). The acid digestion blank had measurable U at $3.57E-2 \ \mu g/mL$, the laboratory blank has measurable U at $1.77E-3 \ \mu g/mL$. The lowest concentration of U detected in the sample group is 3.52E+0. The U present in the process blanks is less than 5% of the concentration present in the samples.

Page 2 of 3

Battelle PNNL/RSE/ASO Radiochemistry Analysis Report

Laboratory Control Sample (LCS)/Blank Spike (BS)/Reagent Spike (RS):

A LCS/BS is required as part of the KPA batch run; i.e., prepared following after the acid digestion. This post-digestion LCS/BS has an 80% to 120% recovery acceptance criterion. No LCS/BS was prepared or analyzed with this batch run. Two reagent spike samples were samples were analyzed in this batch run, one within the higher concentration range on one in the lower concentration range. The recovery of the high range reagent spike is 97% and recovery of the low range reagent spike is 105%. Both spike recoveries are within the acceptance criterion.

Matrix Spike (MS):

Per the ASR, a matrix spike is to be prepared as part of the KPA batch run; i.e., after the acid digestion. This post-digestion matrix spike has a 75% to 125% recovery acceptance criterion. A matrix spike was prepared from TI014-FEED and the spike recovery is 82%, which meets the acceptance criterion.

Duplicate -- Relative Percent Difference (RPD):

All three samples were prepared in duplicate as requested in the instructions in the ASR. Duplicate results are required to agree within <20% RPD. The duplicate RPD for sample TI014-FEED is 0.3%, for sample TI014-EFF Comp the duplicate RPD is 5% and for sample TI014-ELComp the duplicate RPD is 5%. All duplicate RPD results meet the acceptance criterion.

Assumption and Limitations of the Data

There are no assumptions or limitations for these data.

Comments

- 1. The results have been corrected for all dilution factors resulting from sample processing.
- 2. The stated 1-σ uncertainty represents the total propagated error associated with processing and counting operations and includes weighing errors, volume uncertainties, and counting errors.

Attachment: Data Report -- Sample Results for ASR 0335

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

Alpha and Beta

Project / WP#:	69833/N82456/N79882
ASR#:	0335.00
Client:	SK Fiskum
Total # of Samples:	3

	RPL ID		Client Sample ID	
	17-1240		TI014-FEED	
	17-1241		TI014-EFF-Comp	
	17-1242		TI014-ELComp	
Analysis Type:		Alpha and E	Jeta	
Sample Processing Pr Processing/Analysis	rior to Radiochemical	 None Digester of Liqu. Fusion a Solids Using Other: 	d as per RPG-CMC-128, Rev.1, HNO3 ids for Metals Analysis Using a Dry-B as per RPG-CMC-115, Rev.0 Solubiliz g a KOH-KNO3 Fusion	-HCL Acid extraction lock Heater ation of Metals from
Pre-dilution Prior to Processing?	Radiochemical	No Yes a	example 2 mL to 100 mL; 50x dilution	
Radio Chemical Prep	paration Procedure:	RPG-CMC- Beta Analys	4001, Rev. 1, Source Preparation for is	Gross Alpha and Gross
Technician/Analyst:		LP Darnell,	(8/31/2017)	
Spike Standard ID's:		R-687-a-2 (Pu-239), R-693-b-2 (Sr-90)	
		1		
Analysis Procedure:		RPG-CMC-	408, Rev. 2, Total Alpha and Beta And	alysis
Reference Date:		Not Applica	ble	
Analysis Date(s) or D	ate Range:	Alpha and E	seta (9/1/2017); (9/13/2017)	
Technician/Analyst:		LP Darnell a	and T. Trang-Le	
		1		
Analysis Data (File):		RPG-RC\PN	IL\Projects\Backup files\Backup 18\17	-1240 Fiskum.xls
CMC Project 98620 I	File:	File Plan 58 Alpha Detect maintenance balance calil	71: T 69833: Sample preparation and a tor calibration, calibration verification records; and T3 Standard certificates bration and instrument performance ch	analysis records; T-4.4 checks, and and preparation. Also ecks.
M&TE Number(s):		LB 4100 gas	s proportional counter - See attached M	A&TE list

Irang Preparer Date

11/15/17 Date

Reviewer

Page 1 of 3

Sample Results

See attached data report, sample results for ASR 0335. All data are reported in units of μ Ci/mL with a 1- σ uncertainty (see comments).

Sample Preparation, Separation, Mounting and Counting Methods

All three samples were received from the Shielded Analytical Laboratory (SAL) and submitted under Analytical Service Request (ASR) 0335 were analyzed for alpha and beta. All samples were prepared in laboratory 420. Aliquots of the acid digestions and pre-diluted TI014-EL-Comp were mounted for alpha/beta counting using procedure RPG-CMC-4001, then counted using Ludlum, and alpha/beta gas proportional counters per procedure RPG-CMC-408.

Sample 17-1242 (TI014-EL-Comp) was prepared for dilution in the hot cell by filling the diluent vial with 4 aliquots of \sim 5mL each of DI water in lab 309. The diluent vial was loaded into the hot cell where \sim 0.2mL of AP-105 Cs Eluent was added to the diluent vial. The vial was then loaded out of the hot cell clean.

QUALITY CONTROL RESULTS

Quality control (QC) samples prepared in laboratory 420 include a laboratory duplicate sample and a preparation blank. Additional QC samples were prepared prior to alpha and beta counting including a laboratory blank, a reagent blank spike (RS, Pu-239 and Sr-90), and addition of Pu-239 and Sr-90 standard to a diluted aliquot of the sample selected as the matrix spike (MS).

A summary of the Alpha and Beta analysis results, including QC sample performance, is given in the attached data report.

Tracer:

Tracer is not used for this analysis.

Laboratory Preparation Blank (PB) and Laboratory blank (LB):

The alpha activity measured in the LB is required to be within the acceptance criteria of less than 5% of the sample isotope concentration or less than sample minimum detectable activity (MDA). The alpha was counted for 60 minutes in Ludlum, and the beta was counted for 100 minutes in LB4100 alpha beta proportional counter. The LB alpha and beta activities are less than the MDA.

Blank Spike (BS) – Reagent Spike (RS):

The RS (Pu-239) recovery of 105% and (Sr-90) recovery of 86% meet the acceptance criteria of 80% to 120% recovery.

Matrix Spike (MS):

1

The MS (Pu-239) recovery of 98% meets the acceptance criterion of 75% to 125% recovery. The MS (Sr-90) recovery is too small for the spike, and did not meet the acceptance criterion of 75% to 125%. Note: the MS sample was prepared by adding a known Pu-239 and Sr-90 standard quantity to an aliquot of the digestate. Sample number 17-1240 (TI014-FEED) was selected as the matrix spike sample.

Duplicate -- Relative Percent Difference (RPD):

Sample and duplicate sample RPD result are 2-3% for beta. RPD results are within the acceptance criterion of $\leq 20\%$. Sample and duplicate RPD results are not calculated for gross alpha measures because the results are less than the detection limit.

Instrument Quality Control

LB4100 alpha and beta counters undergo initial calibration to determine the detector efficiency. The established efficiency for each detector is used in the final calculation of the sample alpha and beta activity. Continuing calibration verification checks are performed on the detectors once per day as the system is used. Detector backgrounds are obtained once per day or as the system is used or per batch.

Assumption and Limitations of the Data

None.

Attachment: Data Report -- Sample Results for ASR 0335.

Updated 03/17/2016

1

Page 1 of 4

	And and a subscription of the local division			
	Acquisition Date	9/2006	8/2015	s/2011
	SN/ Lot#			
	Calibration Standards	Blank, H-3, and C-14 sealed sources & Quench curves (efficiency) for H-3,C-14, and Ni-63	Blank, H-3, C-14 instrument performance assessment (IPA) sources & Quench curves (efficiency) for H-3 & C-14	Blank, H-3, C-14 instrument performance assessment (IPA) sources & Quench curves (efficiency) for H-3 & C-14
'E List	Calibration Frequency	With use	With use	With use
istry M&T	M&TE Category	2	0	0
kadiochemi	Location	RPL/425	RPL/425	RPL/425
H	Serial Number/ Property Number	WD48466	WD56319	PT27127
	Model	3100TR	5110TR	3110TR
	Manufacturer/ SW	Perkin Elmer QuantaSmart 3.00/Build 5	Perkin Elmer QuantaSmart 5.00/Build 3	Perkin Elmer QuantaSmart 4.00/Build 5
	Description	Liquid Scintillation Counter	Liquid Scintillation Counter	Liquid Scintillation Counter

age 2 of 4		8/2015		1994
Pa				
		I-129 instrument performance assessment source (IPA)		Too many to list
		With use		With use
		2		
	RPL/425	RPL/93		RPL/525
	401664/ WD06783	WD56320		Serial number: 1045050203 WD59770
	2550	2480		KPA- 11R
	Packard Tri Carb Packard Tri- Carb 2700TR software version 1.04 dated 9/99. The software runs on a DOS- 2000 platform on a PC computer	Perkin Elmer Wizard 2		Chemcheck Instruments, Inc. The software used to control the instrument is KPAWIN Software, Version 1.2.9 Release 1999
ted 03/17/2016	Liquid Seintillation Counter	Automated NaI	Uranium	Uranium Analyzer/ Kinetic Phosphoresce ncc Analyzer
Updat				

z

Updated 03/17/2016

Page 3 of 4

		Ka	diochemistry M	X I E LIST		
	Number of		Model or Serial	Location, in Bldg.		Μd
ID	Detectors	Property #	#	325	Procedures using the M&TE	Contract?
Dual Alpha Beta Gas Prop	portional Count	ers				
Oxford LB4100	16	WD13066	L8285-0	425	Total alpha, total beta, Sr/Y90, Tc99	Υ
Alpha Counters						
Ludlum	10	multiple	multiple	425	Total Alpha	Z
AEA (Ortec)*	32	multiple	multiple	425	AEA (total alpha, Pu, Am/Cm,Np)	Z
Liquid Scintillation						
Perkin Elmer 3100TR	1	WD48466	DG08061340	425	H-3, C-14, Ni-63, Se-79, Pu-241, Sr-90	Υ
Perkin Elmer 3110TR	1	PT27127	DG03117614	425	H-3, C-14, Ni-63, Se-79, Pu-241, Sr-90	Υ
Perkin Elmer 5110TR	1	WD56319	SGL0344150055	425	H-3, C-14, Ni-63, Se-79, Pu-241, Sr-90	Υ
Packard 2550	t	WD06783	401664	425	H 3, C 14, Ni 63, Se 79, Pu 241, Sr 90	*
Gamma						
HpGe*	25	see list	see list	425/93	GEA	Z
LEPS (x-rays)*	4	see list	see list	425	X-ray (Fe-55, Ni-59, Nb-93m)	Z
Well HPGe*	3	see list	see list	93	GEA – In development	Z
Anti-Compton*	2	see list	see list	93	GEA – In development	z
Perkin Elmer 2480 Wizard Automated Nal	1	WD56320	SGWZ28150290	93	In development	Υ
Uranium						
Chemcheck	1	WD59770	KPA11R	525	Uranium KPA	Z
*Spectral Analysis System	1					
Canberra VAX 3000	+	WD12890	PE42AJB	425	AEA, GEA	Ż
Canberra VAX 3100	+	WC38624	KA235W0225	425	AEA, GEA (Backup VAX)	*

-	
-	
-	
C	
~	•
с.,	
-	
-	
3	
-	
-	
-	
ς.	
- 1	
+	
0	
-	
- T	

Page 4 of 4

		6					
Jame	Pronerty Tag	List of Vendor	Gamma Detectors	Serial #	Tvne	Eff. %	CMC?
υ	WD25157	EG&G Ortec	GEM70200P	33TP40378A	Ge	70	Y
D	WD06582	Princeton	IGC7022SD	2542	Ge	70	Υ
ш	WD12833	EG&G Ortec	GEM100210	34P40535A	Ge	100	Υ
ŋ	WD35719	Canberra	GC4020	9923007	Ge	40	Υ
Н	PT00941	Tennelec	CPVS3040195	6028	Ge	40	Υ
I	WD12076	EG&G Ortec	GLP3638510P	33TE187	LEPS	X-RAY	Υ
ſ	WD12075	EG&G Ortec	GLP3638510P	33TE190	LEPS	X-RAY	Υ
Х	WD06581	Princeton	NIGC6022SD	3612	P-Ge	09	Υ
Γ	WD36413	EG&G Ortec	GEM20P-Plus	44TP21884A	Ge	20	Υ
Σ	WD73130	Canberra	GC10021	9722	Ge	100	Υ
Z	WD73129	Canberra	GC10021	9716	Ge	100	Υ
а.	WD59508	EG&G Ortee	GEM30210	24P80V	Ge	30	⊁
T	WD81868	EG&G Ortec	GEM20180	35TP21056A	Ge	20	Υ
1 *	PT27133	Canberra	GC4018	6166	6 6	40	*
Ľ.	PT16418	Canberra	GC10021	10182	Ge	100	Υ
0	PT16419	Canberra	GC10021	10188	Ge	100	Υ
R	PT16420	Canberra	GC10021	10181	Ge	100	Y
S	PT16430	Canberra	GC10021	10101	Ge	100	Υ
D	PT16429	Canberra	GC10021	10178	Ge	100	Υ
>	PT16428	Canberra	GC10021	10150	Ge	100	Υ
M	PT16394	Canberra	GC10021	10129	Ge	100	Υ
\times	PT16402	Canberra	GC10021	10104	Ge	100	Υ
γ	PT16401	Canberra	GC10021	10145	Ge	100	Υ
Ζ	PT16410	Canberra	GC10021	10212	Ge	100	Υ
EPSI	WB61769	Ortec	101351600S	20M102	Ge	X-RAY	Υ
EPS2	WD48418	Canberra	GL2015R	6068240	Ge	X-RAY	Υ
ΜA	WD81452	Ortec	GEM100210	35-P-40642A	Ge	100	Υ
WB	WD81454	Ortec	GEM100210	34-P-40481A	Ge	100	Υ
WC	WD81453	Ortec	GEM100210	34-P-40522A	Ge	100	Υ
Well1	WD81450	Ortec	GEM170230-S	34-P-30914B	Ge Well	N/A	Υ
Well2	WD81449	Ortec	GEM170230-S	34-P-40544A	Ge Well	N/A	Υ
Well3	WD81451	Canberra	GCW2523-S	11051527	Ge Well	N/A	Υ
AC1	PT27479	Ortec	GEM4276LBCHJ	54P33125A	Ge-AC	45	Υ
AC2	WD56318	Canberra	GX5020	5945022	Ge-AC	50	Υ

Updated 03/17/20

ASR 0372 Test 2 through Test 5 Results

Unfortunately, ASR 0372 incorrectly mixed up the elution sample identifications following Test 4 Take 1 and Test 4 Take 2 as indicated in Table B.1. Therefore, the ASO-reported results for these two samples need to be switched for all analytes.

Table B.1. Elution Sample Identification for Test 4 Take 1 and Test 4 Take 2, ASR 0372

Test ID	Client ID on ASR	Assigned ASO Sample ID	Corrected Client ID	Corrected ASO Sample ID
Test 4 Take 1	TI020-ELComp	18-0009	TI020-EComp	18-0012
Test 4 Take 2	TI020-EComp	18-0012	TI020-ELComp	18-0009

Reque	(Informati estor Complete all	on on this COVER PAGE is fields on this COVE	s applicable to all samples submitted under this ASR) R PAGE, unless specified as optional or ASR is a revision
Requestor: Signature Print Name Phone	S. K. Fiskum_ 375-5677	MSIN	Project Number: 69832 Work Package: N79882
	Matrix Type Inform	nation	OA/Special Requirements
Liquids:	X Aqueous	ic 🛛 Multi-phase	OA Plan:
 Solids: Other: 	□ Soil □ Sludge □ Glass □ Filter □ Smear □ Organ	e	 X ASO-QAP-001 (Equivalent to HASQARD) □ Additional QA Requirements, List Document Below: Reference Doc Number: ◆ Field COC Submitted? X No □ Yes ◆ Lab COC Required? X No □ Yes
[□ Gas □ Biologica	l Specimen	◆ Sample/Container Inspection Documentation Required? X No □ Yes
(If samp	ole matrices vary, specify	on Request Page)	♦ Hold Time: X No □ Yes
	Disposal Informa	tion	$\frac{\text{If Yes.}}{\text{Contact ASO}} \square \text{Use SW 846 (PNL-ASO-071) identify}$
 Disposition Virgin san archiving 	of Virgin Samples: nples are returned to requ provisions are made with	estor unless receiving group!	Lead before analytes/methods where holding times apply) submitting Other? Specify:
If archivit Archivit	ng, provide: ing Reference Doc:		◆ Special Storage Requirements: X None □ Refrigerate □ Other, Specify:
 Disposition X Dispo 	of Treated Samples:		◆ Data Requires ASO Quality Engineer Review? □ No X Yes
		Data Re	porting Information
Hilestone? If yes Preliminar Available?	X No Yes , milestone due date: 	ASO-QAP-00 HASQARD). Minimum data X Project Specific contact ASO Lea Document:	 I (Equivalent to I report. c Requirements: d or List Reference Kequested Analytical Work Completion Date: Negotiated Commitment Date: 3/15/18 (To be completed by ASO Lead)
		Waste De	signation Information
 ASO Sampl If no, Refer or, Previous or, Previous 	e Information Check Li ence Doc Attached: ASR Number:(RPL Number:(ist Attached? X No	Yes Does the Waste Designation Documentation Indicate Presence of PCBs? X No Yes
Send Report	To:S. K. Fisku	m	MSIN
Additional or	Special Instructions		MI31N
	Rec	eiving and Login Infor	mation (to be completed by ASO staff)
Date Delivere	d:10	/24/17	Received By:T. Trang-Le
Delivered By Time Delivere Group ID (opt CMC Waste S	(optional)1 ed:1 cional) ample? X No	3:00	ASR Number: 0372 Rev.: 01 RPL Numbers: (18-0001 thru 18-0015) (first and last)
AS	O Work Accepted By:	KN Pool	Signature/Date: Jail N. Pool 3/1/18

ACID

ASO Staff Use Only	Provide Analytes of	Interest and Required Detection	<u>n limits</u> - □ Below □ Attached	ASO Staff	Use Only
RPL Number	Client Sample ID	Sample Description (& Matrix if varies)	Analyses Requested	Test	Library
	Revision 1 - ASF	R revised to remove U-KPA	v and add U by ICP/MS		
18-0001	TI015-FEED	Caustic/5.7 M Na, high Salt			
18-0002	TI015-EFF Comp	Caustic/5.7 M Na, high Salt	1) GEA - All samples (Co-60, Cs-137, Eu-154. Am-241 and any other		
18-0003	TI015-ELComp	Acidic/~0.2 M HNO ₃ - Sample Pre- Diluted in the hot cell By ~ 100x	observed gamma emitting isotopes)		
18-0004	TI016-FEED	Caustic/5.7 M Na, high Salt			
18-0005	TI016-EFF Comp	Caustic/5.7 M Na, high Salt	2) Acid Digestion- 128 - Caustic matrix camples anly		
18-0006	TI016-ELComp	Acidic/~0.2 M HNO3 - Sample Pre- Diluted in the hot cell By ~ 100x	3) ICP/OES - All samples - (Al. Cd. Cr.		
18-0007	TI020-FEED-1-A	Caustic/5.7 M Na, high Salt	Cu, Fe, K, Mo, Na, Ni, Pb, S, Si,		
18-0008	TI020-EFF Comp-Take 1	Caustic/5.7 M Na, high Salt	Sr and Zn)		
18-0009	TI020-ELComp	Acidic/~0.2 M HNO3 - Sample Pre- Diluted in the hot cell By ~ 100x	4) Pu-AEA - Pu-239/240 - All Samples		
18-0010	TI020-FEED-Take2	Caustic/5.7 M Na, high Salt	5) Tc-99 - All Samples		
18-0011	TI020-EFF Comp-Take2	Caustic/5.7 M Na, high Salt	•		
18-0012	TI020-EComp	Acidic/~0.2 M HNO3 - Sample Pre- Diluted in the hot cell By ~ 100x	6) ICP/MS - U only		
18-0013	TI021-FEED	Caustic/5.7 M Na, high Salt			
18-0014	TI021-EFF Comp	Caustic/5.7 M Na, high Salt			
18-0015	TI021-ELComp	Acidic/~0.2 M HNO3 - Sample Pre- Diluted in the hot cell By ~ 100x			
	Note: All the scidid matr	iv samulas wara nra_dilutad in th	a hat cell hu ~ 100v due to the		
	high Cs-137 content in th	ne parent sample Use the dilution	on factor from the SAL bench		
	sheet in all sample resul	t calculations.			
all and a second					
ASR # 03	72				

Analytical Services Request (ASR) (REQUEST PAGE ---- Information Specific to Individual Samples)

ASR 0372.01 Request Page.xls

Page 1 of 1

B.101

Special Instructions for ASR 0372

Analysis of AP-105 diluted feed, effluent, and eluate from Tests 2-5 ion exchange process runs

IX Test	Client ID	ASO Sample ID	Location for Pickup	Matrix
2	TI015-FEED	18-0001	SAL cell 2	Caustic/5.7 M Na, high salt
	TI015-EFFComp	18-0002	Lab 305	Caustic/5.7 M Na, high salt
	TI015-ELComp	18-0003	SAL cell 2	Acidic/~0.2 M HNO3
3	TI016-FEED	18-0004	SAL cell 2	Caustic/5.7 M Na, high salt
	TI016-EFFComp	18-0005	Lab 305	Caustic/5.7 M Na, high salt
	TI016-ELComp	18-0006	SAL cell 2	Acidic/~0.2 M HNO3
4 Take 1	TI020-FEED-1-A	18-0007	Lab 305	Caustic/5.7 M Na, high salt
	TI020-EFFComp	18-0008	Lab 305	Caustic/5.7 M Na, high salt
	TI020-ELComp	18-0009	SAL cell 2	Acidic/~0.2 M HNO3
4 Take 2	TI020-FEED-Take2	18-0010	Lab 305	Caustic/5.7 M Na, high salt
	TI020-EFFComp-Take2	18-0011	Lab 305	Caustic/5.7 M Na, high salt
	TI020-EComp	18-0012	SAL cell 2	Acidic/~0.2 M HNO3
	TI021-FEED	18-0013	SAL cell 2	Caustic/5.7 M Na, high salt
5	TI021-EFFComp	18-0014	Lab 305	Caustic/5.7 M Na, high salt
	TI021-EComp	18-0015	SAL cell 2	Acidic/~0.2 M HNO3

Use the ASO QA Plan current revision to apply batch processing and instrument QC.

Please prepare one sample in duplicate per preparative batch.

Because the TI0xx-ELComp samples are already acidified, acid digestion may not be needed. The high salt samples (FEED and EFFComp) are assumed to require acid digestion. However if the preparative technique sufficiently dilutes the sample into required acid matrix, then acid digestion may be omitted.

Preparative QC samples per preparative batch expected for this suite include:

- 1. Process blank (digestion blank or diluent blank)
- 2. Duplicate (one duplicate per analytical batch; if dilution is the only preparation, please prepare a diluent duplicate)
- 3. Blank spike or reagent spike
 - a. not applicable to GEA
- 4. Matrix spike

Page 1 of 2

- a. ICP-OES: not needed for major analytes (e.g., Na in the FEED and EFFComp samples); for guidance on spiking needs see ICP-OES results from ASR 0316. In the case where analyte spikes are not used, report serial dilution result instead per ASO QA Plan
- b. not applicable to GEA

Please analyze FEED, EFFComp, and ELComp (or EComp) samples as defined in Table 1.

Analyte	Target MDL	Analysis Method	
⁶⁰ Co	1.00E-03 µCi/mL		
¹³⁷ Cs	1.00E-05 µCi/mL	Gamma Energy Analysis (GEA) (report any other observed gamma emitting isotopes)	
¹⁵⁴ Eu	1.00E-05 µCi/mL	(report any other observed gamma emitting isotopes)	
²⁴¹ Am	2.00E-03 µCi/mL		
239+240Pu	1.00E-05 µCi/mL	Acid Digestion Coprecipitation/Alpha Energy Analysis (AEA)	
⁹⁹ Tc	1.00E-5 µCi/mL	Separations/Beta Counting	
Al	l μg/mL		
Cd	0.5 μg/mL		
Cr	0.5 μg/mL		
Cu	0.5 μg/mL		
Fe	0.5 μg/mL		
К	10 μg/mL		
Мо	l μg/mL	Acid Digestion	
Na	100 μg/mL	(ICP-OES)	
Ni	I μg/mL	···· · · · · · · · · · · · · · · · · ·	
Pb	l μg/mL		
S	1 μg/mL		
Si	l μg/mL		
Sr	0.5 μg/mL		
Zn	0.5 µg/mL		
U (total)	0.5 μg/mL	Acid Digestion / Kinetic Phosphorescence Analysis (KPA)	

Table 1.Sample Analyte List

Battelle PNNL/RPL/Inorganic Analysis ... ICP-OES Analysis Report PO Box 999, Richland, Washington 99352

Project / WP#:	69832 / N79882
ASR#:	0372
Client:	S. Fiskum
Total Samples:	15 (liquids)

ASO Sample ID	Client Sample ID	Client Sample Description	Sample Weight (g)
18-0001	TI015-FEED	Caustic/5.7 M Na, high Salt	NA
18-0002	TI015-EFF Comp	Caustic/5.7 M Na, high Salt	NA
18-0003	TI015-ELComp	Acidic/~0.2 M HNO ₃	NA
18-0004	TI016-FEED	Caustic/5.7 M Na, high Salt	NA
18-0005	TI016-EFF Comp	Caustic/5.7 M Na, high Salt	NA
18-0006	TI016-ELComp	Acidic/~0.2 M HNO ₃	NA
18-0007	TI020-FEED-1-A	Caustic/5.7 M Na, high Salt	NA
18-0008	TI020-EFF Comp-Take-1	Caustic/5.7 M Na, high Salt	NA
18-0009	TI020-ELComp	Acidic/~0.2 M HNO ₃	NA
18-0010	TI020-FEED-Take2	Caustic/5.7 M Na, high Salt	NA
18-0011	TI020-EFF Comp-Take2	Caustic/5.7 M Na, high Salt	NA
18-0012	TI020-EComp	Acidic/~0.2 M HNO ₃	NA
18-0013	TI021-FEED	Caustic/5.7 M Na, high Salt	NA
18-0014	TI021-EFF Comp	Caustic/5.7 M Na, high Salt	NA
DUP-0014	TI021-EFF Comp	Caustic/5.7 M Na, high Salt	NA
18-0015	TI021-ELComp	Acidic/~0.2 M HNO ₃	NA

Sample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of Liquids for Metals Analysis Using a Dry-Block Heater", performed by L. Darnell on Samples 18-0001, 18-0002, 18-0004, 18-0005, 18-0007, 18-0008, 18-0010, 18-0011, 18-0013, and 18-0014 on 11/28/17. Simple dilution in 5% HNO₃ was performed on Samples 18-0003, 18-0006, 18-0009, 18-0012, and 18-0015 on 10/25/17. Simple dilution of "as received" samples in 5% v/v HNO₃ was performed by J. Carter on 12/07/17.

m

Report Preparer

Review and Concurrence

Date

1/26/18

Date

Page 1 of 4

B.104
Procedure	e: <u>RPG-CMC-2</u> Inductively C	2 <u>11, Rev. 4</u> , "Determi Coupled Argon Plasma	nation of Element Optical Emission	tal Composition n Spectrometry	ı by (ICP-OES)."
Analyst:	J. Carter	Analysis Date:	12/07/2017	ICP File:	C0738
See Chem	ical Measureme	nt Center 98620 file:	ICP-325-405-3 (Calibration an	d Maintenance	Records)
M&TE:	PerkinElm	er 5300DV ICP-OES		SN: 077N5122	.002
	Mettler AT	400 Balance		SN: 11132926	67
	Sartorius R	200D Balance		SN: 39080042	
	SAL Cell 2	2 Balance		SN: 80333112	09
	🛛 Lab 201 D	enver A-160 Balance		SN: 60568	

Battelle PNNL/RPL/Inorganic Analysis ... ICP-OES Analysis Report

Fifteen aqueous samples submitted under Analytical Service Request (ASR) 0372 were analyzed by ICP-OES. Samples 18-0001, 18-0002, 18-0004, 18-0005, 18-0007, 18-0008, 18-0010, 18-0011, 18-0013, and 18-0014 were prepared following RPL procedure RPG-CMC-128 and diluted to approximately 25 mL. Sample 18-0014 was prepared in duplicate following RPL procedure RPG-CMC-128 and diluted to approximately 25 mL. Samples 18-0003, 18-0006, 18-0009, 18-0012, and 18-0015 were diluted with 5% HNO₃ in the hot cell to approximately 105x. All samples were further diluted in 5% HNO₃ prior to analysis. None of the samples were filtered.

All sample results are reported on a mass per unit volume basis ($\mu g/mL$) for each detected analyte. The data have been adjusted for instrument dilutions.

Analytes of interest (AOI) were specified in the ASR and are listed in the upper section of the attached ICP-OES Data Report. The quality control (QC) results for the AOI have been evaluated and are presented below. Analytes other than the AOI are reported in the bottom section of the report but have not been fully evaluated for QC performance.

Calibration of the ICP-OES was done following the manufacturer's recommended calibration procedure using multi-analyte custom standard solutions traceable to the National Institute of Standards and Technology (NIST). Midrange calibration verification standards (MCVA and MCVB) were used to verify acceptance of the two-point calibration curves obtained for each analyte and for continuing calibration verification.

The controlling documents were procedures RPG-CMC-211, Rev. 4, *Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP-OES)*, and ASO-QAP-001, Rev. 11, *Analytical Support Operations (ASO) Quality Assurance Plan*. Instrument calibrations, QC checks and blanks (e.g., ICV/ICB, CCV/CCB, LLS, ICS), matrix spike, post-digestion spikes, duplicate, reagent spike, blank spike, and serial dilution were conducted during the analysis run.

Preparation Blank (PB):

A preparation blank (reagents only) was prepared for the extraction process. The concentration of all AOI were within the acceptance criteria of <EQL (estimated quantitation level), \leq 50% regulatory decision level, or \leq 10% of the concentration in the samples. The silicon (1.66 µg/mL) and zinc (0.652 µg/mL) concentrations were >EQL and were significant (>10%) relative to many of the samples after taking the process factors into account.

Reagent Spike (RS)/Laboratory Control Sample (LCS):

A reagent spike (RS) sample (reagents and spikes) was prepared for the extraction process. Recovery values are listed for all analytes included in the RS that were measured at or above the EQL. Recovery values for the AOI meeting this requirement ranged from 97% to 105%, with the exception of silicon (23%), and were within the acceptance criterion of 80% to 120%. The low silicon recovery is likely due to reagent instability or volatility.

Duplicate/Replicate Relative Percent Difference (RPD)/Relative Standard Deviation (RSD):

The ASR requested duplicate analysis of one sample. A duplicate of sample 18-0014 was prepared and analyzed. RPDs are listed for all analytes that were measured at or above the EQL. RPDs for the AOI meeting this requirement ranged from 1.6% to 5.6% and were within the acceptance criterion of $\leq 20\%$ for liquid samples.

Matrix-Spike (MS) Sample:

A matrix spike was prepared for the extraction process. Recovery values are listed for all analytes included in the MS that were measured at or above the EQL. Recovery values for the AOI meeting this requirement ranged from 91% to 116%, with the exception of silicon (51%), and were within the acceptance criterion of 75% to 125%. As noted above for the RS, the reason for the low silicon recovery may be reagent instability or volatility.

Initial/Continuing Calibration Verification (ICV/CCV):

MCVA and MCVB solutions were analyzed immediately after calibration, after each group of not more than ten samples, and at the end of the analytical run. The concentrations of all AOI were within the acceptance criteria of 90% to 110% recovery, with the exception of potassium (110.1%) in the final CCV solution and sulfur (85-86%) in the last three CCV solutions.

Initial/Continuing Calibration Blank (ICB/CCB):

The ICB/CCB solution (5% v/v HNO₃) was analyzed immediately after the ICV solutions and after the CCV solutions (after each group of not more than ten samples and at the end of the analytical run). The concentration of all AOI were within the acceptance criteria of <EQL, with the exception of sodium (0.11-0.17 μ g/mL) in the final three CCB solutions.

Low-Level Standard (LLS):

The LLS solution was analyzed immediately after the first CCB solution. The concentrations of all AOI were within the acceptance criteria of 70% to 130% recovery.

Interference Check Standard (ICS/SST):

The ICS solution was analyzed immediately after the first LLS solution and immediately prior to analyzing the final CCV solutions. The concentrations of all AOI were within the acceptance criteria of 80% to 120% recovery.

Serial Dilution (SD):

Five-fold serial dilution was conducted on sample 18-0001. Percent differences (%Ds) are listed for all analytes that had a concentration at or above the EQL in the diluted sample. The %Ds for the AOI meeting this requirement ranged from 2.2% to 7.0% and were within the acceptance criterion of $\leq 10\%$.

Post-Digestion Spike (PS-A)/Analytical Spike (AS-A) - Sample (A Component):

In addition to the RS, BS, and MS samples, a post-digestion spike (A Component) was conducted on sample 18-0001. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration \geq 25% of that in the sample. Recovery values for the AOI meeting this requirement ranged from 94% to 104%, and were within the acceptance criterion of 80% to 120%.

Post-Digestion Spike (PS-B)/Analytical Spike (AS-B) - Sample (B Component):

In addition to the RS, BS, and MS samples, a post-digestion spike (B Component) was conducted on sample 18-0001. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration $\geq 25\%$ of that in the sample. The recovery value for sulfur, the only AOI meeting this requirement, was 95%, and was within the acceptance criterion of 80% to 120%.

Other QC:

All other instrument-related QC tests for the AOI passed within their respective acceptance criteria.

Comments:

- 1) The "Final Results" have been corrected for all laboratory dilutions performed on the samples during processing and analysis, unless specifically noted.
- 2) Instrument detection limits (IDL) and estimated quantitation limits (EQL) shown are for acidified water and/or fusion flux matrices as applicable. Method detection limits (MDL) for individual samples can be estimated by multiplying the IDL by the "Process Factor" for that individual sample. The estimated quantitation limit (EQL) for each concentration value can be obtained by multiplying the EQL by the "Process Factor".
- 3) Routine precision and bias is typically $\pm 15\%$ or better for samples in dilute, acidified water (e.g. 5% v/v HNO₃ or less) at analyte concentrations > EQL up to the upper calibration level. This also presumes that the total dissolved solids concentration in the sample is less than 5000 µg/mL (0.5 per cent by weight). Note that bracketed values listed in the data report are within the MDL and the EQL, and have potential uncertainties greater than 15%. Concentration values < MDL are listed as "--". Note, that calibration and QC standard samples are validated to a precision of $\pm 10\%$.
- 4) Analytes included in the spike A component (for the AS/PS) are; Ag, Al, As, B, Ba, Be, Bi, Ca, Cd, Co, Cr, Cu, Fe, K, Li, Mg, Mn, Mo, Na, Ni, P, Pb, Sb, Se, Si, Sn, Sr, Ta, Ti, Tl, V, W, Y, Zn, and Zr. Analytes included in the spike B component are; Ce, Dy, Eu, La, Nd, Pd, Rh, Ru, S, Te, Th, and U.

Battelle PNNL/RPG/Inorganic Analysis ... ICPOES Data Report

Page 1 of 3

		Run Date >	12/7/2017	12/7/2017	12/7/2017	12/7/2017	12/7/2017	12/7/2017	12/7/2017	12/7/2017	12/7/2017	12/7/2017
		Process				_						107.1
		Factor >	1.0	48.9	504.8	504.8	496.0	104.8	491.0	487.6	104.8	487.4
			405 diluont	BLK-0001	18-0001@	10v ren	18-0002@	18-0003	18-0004 @ 10x	10-0005 @	18-0006	10x
	1		405 dildent	DER-0001	104	Tox Top	104	10-0000	IVA	104		
Inetr Dat	Fet Quant						TI015-FFF	TI015-ELF		TI016-EFF	TI016-EL	TI020-Feed-
Limit (IDL)	Limit (EQL)	Client ID >	Lab Diluent	BLK-0001	TI015	-Feed	Comp	Comp	TI016-Feed	Comp	Comp	<u>1-A</u>
(ug/mL)	(ug/mL)	(Analyte)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)
0.0038	0.038	Al			13,500	13,500	14,000	69.2	14,700	13,900	78.0	13,300
0.0016	0.016	Cd			[0.83]	[0.87]	[1.5]		[1.5]	-	alle i	[1.3]
0.0016	0.016	Cr		22	335	329	324	1.86	344	323	2.56	337
0.0023	0.023	Cu			[2.5]	[3.5]	[2.4]	11.7	[2.7]	[1.6]	9.53	[2.9]
0.0033	0.033	Fe		[0.26]	[5.7]	[5.5]	[3.6]	4.58	[5.9]	[3.5]	4.14	[4.1]
0.0262	0.262	к		[4.6]	3.730	3.740	3,750	282	4,130	3,880	369	3,930
0.0048	0.048	Mo		-	51.1	50.4	49.8	245	50.5	47.0		48.8
0.0075	0.075	Na		[3.3]	138,000	135,000	132.000	3,150	142,000	134,000	4,110	140,000
0.0040	0.040	Ni			30.1	30.1	30.2	[2.5]	30.3	28.3	4.35	30.5
0.0040	0.040	Pb			[16]	[17]	[8.9]	87.6	[20]		111	[15]
0.0898	0.898	S			1,150	1,110	1.090	[27]	1,080	1,020	[35]	1,060
0.0043	0.043	Si	[0.0076]	3.33	125	125	133	739	103	120	625	127
0.0001	0.001	Sr		[0.020]	[0.083]	[0.079]	[0,11]	0.211	[0.080]	[0.056]	0.414	[0.096]
0.0023	0.023	70		1.30	[3.6]	[4,1]	[5.1]	17.6	122		10.2	
Other Analy	1 0.020				Terel	1.004	1.00					
0.0014	0.014	Ag					-	[0.38]				
0.0383	0.383	As										
0.0032	0.032	B	10 0201	106-01	123	119	68.2	555	82.3	76.1	438	213
0.0001	0.001	Ba	[0.020]	0.580	0.915	0.920	1.01	5.18	1.28	1.15	14.1	0.857
0.0001	0.001	Be	[0.000.1]		[0,19]	[0,15]	[0,16]	[0.038]	[0.20]	[0.17]	[0.046]	[0.19]
0.0220	0.220	Bi										
0.0054	0.054	Ca		6.86	45.8	46.5	41.5	30.6	50.5	45.7	23.8	50.1
0.0052	0.052	Ce				[2.9]						
0.0033	0.033	Co								See		
0.0012	0.012	Dv										
0.0006	0.006	Eu							744			
0.0010	0.010	La			12	1				1144		
0.0010	0.010	Li		[0.084]		[1.0]	[1.2]	[0.80]	[0.66]		[0.68]	[0.67]
0.0014	0.014	Ma		[0.23]				2.38			2.35	
0.0003	0.003	Mn						[0.14]		[0.13]	[0.080]	
0.0095	0.095	Nd	112	122	[5.8]	[5.0]		[2.1]	[4.9]			[6.3]
0.0369	0.369	P			476	451	455	[3.9]	465	423	[5.4]	450
0.0082	0.082	Pd			 //			-		[4.4]		
0.0108	0.108	Rh			[5.8]						<u>111</u>	
0.0068	0.068	Ru			[6.8]	[8.3]	[9.2]		[7.2]	[7.3]		[10]
0.0569	0.569	Sb		5,22							85 5	-
0.0876	0.876	Se							277		[11]	[45]
0.0195	0.195	Sn	975					-		· · · ·		
0.0109	0.109	Та										
0.0155	0.155	Te										
0.0057	0.057	Th						1000	3.55	[2.9]		
0.0004	0.004	Ti		[0.021]			[0.28]	0.474	[0.21]		0.465	[0.24]
0.0310	0.310	TI										
0.0312	0.312	U										[16]
0.0016	0.016	v	[0.0039]	[0.23]	[2.0]	[1.6]	[2.3]	[0.59]	[2.1]	[1.7]	[0.57]	[1.9]
0.0187	0.187	w			[84]	[85]	[80]		[85]	[74]		[81]
0.0003	0.003	Y										
0.0013	0.013	Zr						[0.55]		1850	[0.47]	

 "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the "multiplier" near the top of each column. The estimated sample quantitation limit = EQL (in Column 2) times the "multiplier". Overall error for values ≥ EQL is estimated to be within ±15%.
 Values in brackets [] are ≥ MDL but < EQL, with errors likely to exceed 15%.

ASR 0372 Final from C0738 ASR-0372 Fiskum AP-107 xlsm

		Run Date >	12/7/2017	12/7/2017	12/7/2017	12/7/2017	12/7/2017	12/7/2017	12/7/2017	12/7/2017	12/7/2017
		Process									
		Factor >	485.2	104.8	490.2	486.4	104.8	493.2	489.4	492.2	104.8
			18-0008 @		18-0010 @	18-0011 @		18-0013 @	18-0014 @	Dup-0014 @	
			10x	18-0009	10x	10x	18-0012	10x	10x	10x	18-0015
			TI020-EFF			TI020-EFF					
Instr. Det.	Est. Quant.		Comp-Take	TI020-EL	TI020-Feed-	Comp-	<u>TI020-</u>				TI021-EL
Limit (IDL)	Limit (EQL)	Client ID >	1	Comp	Take2	Take2	Ecomp	TI021-Feed	<u>TI021-EI</u>	F Comp	Comp
(µg/mL)	(µg/mL)	(Analyte)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)
0.0038	0.038	AI	13,200	49.8	14,000	14,200	93.5	14,900	14,500	14,800	75.6
0.0016	0.016	Cd	[1.1]		[1.3]	[1.3]	-	[1.1]	[1.1]	[0.80]	
0.0016	0.016	Cr	301	2.50	341	334	2.63	341	328	339	3.09
0.0023	0.023	Cu		11.3	[2.4]	[1.2]	5.32	[2.4]	[1.2]	[1.6]	12.5
0.0033	0.033	Fe	[15]	4.22	[8.1]	[5.4]	[3.4]	[3.4]	[2.7]	[2.9]	4.89
0.0262	0.262	K	3,460	268	4,020	3,950	289	4,120	3,970	4,050	287
0.0048	0.048	Mo	43.9	Rest Contraction	48.5	49.4		49.9	49.0	47.8	+
0.0075	0.075	Na	128,000	3.020	140.000	139,000	3,150	142,000	137,000	141,000	3,500
0.0040	0.040	Ni	24.8	5.48	29.8	29.0	5.86	31.1	28.3	29.2	6.93
0.0144	0 144	Ph		95.6	[15]		29.9	[16]			89.1
0.0808	0.898	8	924	[20]	1,050	1.040	(23)	1,050	997	1,050	[22]
0.0030	0.030	ei	150	372	1/5	172	546	94.8	76.5	78.4	718
0.0043	0.043	01	102	0.445	140	10 121	0.149	10 0821	[0, 12]	(0.0911	0.220
0.0001	0.001	Sr	[0.049]	0.115	[0.074]	[0.12]	0.140	[0.002]	[0.12]	[0.001]	12.1
0.0023	0.023	Zn	1.55	10.4	[2.0]		13.7		[2.0]		13.1
Other Analyt	es							1			
0.0014	0.014	Ag		[0.31]	-						
0.0383	0.383	As			-		-	1.75	0.72		
0.0032	0.032	В	73.1	283	167	103	409	52.7	46.4	47.5	530
0.0001	0.001	Ba	0.751	1.74	0.808	1.38	2.02	0.920	0.689	0.809	4.78
0.0001	0.001	Be	[0.14]	[0.050]	[0.20]	[0.18]	[0.037]	[0.20]	[0.17]	[0.18]	[0.060]
0.0220	0.220	Bi					575	275	1.77		
0.0054	0.054	Ca	43.3	26.4	46.0	49.1	34.7	47.5	48.4	54.1	27.2
0.0052	0.052	Ce									
0.0033	0.033	Co	(44								
0.0012	0.012	Dv	[0.73]							[1.0]	
0.0006	0.006	Eu									
0.0010	0.010	1a								[0.55]	[0.12]
0.0010	0.010	Li		f0 571	-	f0 871	[0,75]	10.991	[0.66]		[0.82]
0.0010	0.014	Ma		2.13		[cier]	3.96	0.007			2.20
0.0014	0.014	Ma		2.13			0.001	120			F0 121
0.0003	0.003	MIL		[0.031]			[0.000]		1000	Sau .	[0.15]
0.0095	0.095	NO	440	-	440	440	[1.4]	447	420	440	[4 6]
0.0369	0.369	P	412		446	446	[4.8]	44/	430	447	[4.0]
0.0082	0.082	Pd	**				5 77			1	
0.0108	0.108	Rh	5.57				-				
0.0068	0.068	Ru	[6.4]		[6.7]	[7.8]		[7.6]	[0.5]	[5.4]	
0.0569	0.569	Sb							**		
0.0876	0.876	Se	See 2		-			177	77.0		
0.0195	0.195	Sn	1.57		-			-			-
0.0109	0.109	Та		221							
0.0155	0.155	Те								100	-
0.0057	0.057	Th	10000	100			5.5%				
0.0004	0.004	Ti		[0.30]		[0.28]	[0.42]				0.485
0.0310	0.310	TI		101							
0.0312	0,312	U		***				-			
0.0016	0,016	v	[1.9]	[0.61]	[2.4]	[2.4]	[0.52]	[2.6]	[2.5]	[2.6]	[0.59]
0.0187	0.187	w	[71]	C. Maria	[87]	[83]	1	[82]	[75]	[82]	
0.0003	0.003	Y Y	61.14	100	Feed a	1					
0.0003	0.013	7.	-	[0.39]			[0.48]	-			[0,58]
0.0013	0.013	1 41		[0:00]			10.401		3.7721	1350	fereed.

1) "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the "multiplier" near the top of each column. The estimated sample quantitation limit = EQL (in Column 2) times the "multiplier". Overall error for values ≥ EQL is estimated to be within ±15%. 2) Values in brackets [] are ≥ MDL but < EQL, with errors likely to exceed 15%.

ASR 0372 Final from C0738 ASR-0372 Fiskum AP-107 xlsm

QC Perform	ance 12/7/201	7				
Criteria >	≤ 20%	80%-120%	75%-125%	80%-120%	80%-120%	≤ 10%
QC ID >	18-0014 Dup	LCS/RS	18-0001 MS	18-0001 + AS-A	18-0001 + AS-B	18-0xxx 5-fold Serial Dil
Analytes	RPD (%)	%Rec	%Rec	%Rec	%Rec	%Diff
AI	1.6	103	106	101		3.6
Cd		100	102	101		
Cr	3.1	97	92	98		2.4
Cu		104	107	104		
Fe		101	100	101		
к	1.9	104	116	102		2,4
Mo	2.6	98	94	98		
Na	2.7	107	nr	94		2.2
Ni	3.0	101	101	103		
Pb		98	99	98		
S	5.6	97	91		95	
Si	2.4	23	51	101		7.0
Sr		106	98	102		
Zn		97	98	99		
Other Analy	tes					
Ag	103	[93		
Ae		-		101		
	2.4	104	102	102		16.1
Ba	16.0	104	102	100		
Pa	10.0	100	103	99		
De		80	82	04		1
BI	44.0	108	02	104		
Ca	11,2	100	115	104	07	
Ce		99	90	00	97	
Du				33	07	
Dy					97	
EU		00	00		90	
La		99	99	100	97	
Li		118	119	108		
Mg		103	105	104		
Mn		100	100	101		
Nd		99	97		97	
P	3,1	101	95	99		
Pd					91	
Rh					94	
Ru					94	
Sb				104		
Se				102		
Sn				93		
Та				99		
Те					99	
Th					96	
Ti		103	104	102		
TI				91		
U		102	100		101	
V		100	98	98		
W		97	94	98		
Y				98		
Zr		107	107	102		

Shaded results are outside the acceptance criteria.

nr = spike concentration less than 25% of sample concentration. Matrix effects can be assessed from the serial dilution.

ASR 0372 Final from C0738 ASR-0372 Fiskum AP-107_xlsm

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

Gamma Energy Analysis (GEA)

Project / WP#:	69832/ N79882
ASR#:	0372.00
Client:	SK Fiskum
Total Samples:	15

RPL ID	Client Sample ID
18-0001	TI015-FEED
18-0002	TI015-EFF-Comp
18-0003	TI015-ELComp
18-0004	TI016-FEED
18-0005	TI016-EFF-Comp
18-0006	TI016-ELComp
18-0007	TI020-FEED-1-A
18-0008	TI020-EFF Comp-Take 1
18-0009	TI020-ELComp
18-0010	TI020-FEED-Take2
18-0011	TI020-EFF Comp-Take2
18-0012	TI020-EComp
18-0013	TI021-FEED
18-0014	TI021-EFF Comp
18-0015	TI021-ELComp

Page 1 of 4

Battelle PNNL/RSE/ASO Radiochemistry Analysis Report

Analysis Type:	GEA- for all positively measured or non-detected isotopes
Sample Processing Prior to Radiochemical Processing/Analysis	 None Digested as per RPG-CMC-129, Rev. 0 HNO₃-HCl Acid Extraction of Solids Using a Dry Block Heater Fusion as per RPG-CMC-115, Solubilization of Metals from Solids Using a KOH-KNO₃ Fusion Other:
	Preparation may also involve attaining a GEA geometry that is compatible with the calibration geometry.
Analysis Procedure:	RPG-CMC-450, Rev. 3, Gamma Energy Analysis (GEA) and Low-Energy Photon Spectrometry (LEPS)
Reference Date:	July 10, 2017 @ 10:00 am
Analysis Date or Date Range:	November 27, 2017
Technician/Analyst:	T Trang-Le
Rad Chem Electronic Data File:	18-0001 Fiskum.xls
ASO Project 98620 File:	File Plan 5871, T4.4 Technical (Radiochemistry), Gamma Calibration, daily checks, and maintenance records; and T3 standard certificates and preparation. Also, balance calibration and performance check records.
M&TE Number(s):	Detectors C,G,L,M,N,R,T

I IRang-le12/11/17Induced12/11/17PrepareDateReviewerDate Prepare

Part A

SAMPLE RESULTS

Activities for all gamma emitters detected in this sample are presented in an attached Excel spreadsheet for ASR 0372.00. All sample results for all target isotopes are reported in units of μ Ci/mL with estimates of the total propagated uncertainty reported at the 1-sigma level.

ASO Project File, ASR 0372 has been created for this report including all appropriate supporting records which may include the Pipette Performance Check Worksheet form, standard certificates, laboratory bench records, Shielded Analytical Laboratory Bench Sheet, and Gamma Energy Analysis printouts. Detector calibration records, control charts and balance calibration records can be found in the ASO Records.

Sample Preparation, Separation, Mounting and Counting Methods

Five Pre-diluted samples, 18-0003, 18-0006, 18-0009, 18-0012 and 18-0015 were received from the Shielded Analytical Laboratory (SAL). The samples were diluted ~105x by pipetting ~100 μ L of each sample into ~10 mL of 0.5M HNO₃.

All samples were prepared by pipetting 2 mL of each sample into a 22 mL glass scintillation vial (calibrated geometry) and sent to the counting room for GEA analysis.

The quality control (QC) steps for direct GEA are discussed below.

QUALITY CONTROL RESULTS

Tracer:

Tracers are not used for ASO GEA methods.

Process Blank (PB):

No process blank was prepared for gamma counting.

Required Detection Limits

Required detection limits for these samples are provided by the client in ASR0372. In most cases we were not able to meet these detection limits due to the high level of Cs-137 and remote counting geometries. However, we did report activities below the requested MDL values for the EFF Comp samples, which had much lower activity levels. For the samples with high Cs-137 activities, it is not possible to meet the requested MDL values with any reasonable counting times.

Blank Spike (BS)/Laboratory Control Sample (LCS)/ Matrix Spike (MS):

There are no BS, LCS or MS samples analyzed for ASO GEA analyses. Instrument performance is assessed by the analyses of daily control counts and weekly background counts, as discussed below.

Battelle PNNL/RSE/ASO Radiochemistry Analysis Report

Duplicate Relative Percent Difference (RPD):

No duplicate samples were provided for gamma counting.

Instrument Calibration and Quality Control

Gamma detectors are calibrated using multi-isotope standards that are NIST-traceable and prepared in the identical counting geometry to all samples and detectors. Counter control sources containing Am-241, Cs-137 and Co-60 are analyzed daily before the use of each detector. Procedure RPG-CMC-450 requires that a counter control source is checked daily and must be within ± 3 sigma or $\pm 3\%$ of the control value, whichever is greater. Gamma counting was not performed unless the control counts were within the required limits. Background counts are performed on all gamma detectors at least weekly for either an overnight or weekend count. The most recent background is subtracted from all sample counts.

Assumptions and Limitations of the Data

None

Interferences/Resolution

None.

Uncertainty

For gamma counting, the uncertainty in the counting data, photon abundance and the nuclear half-life, and efficiency are included in the calculation of the total uncertainty along with a systematic uncertainty for sample prep. The Canberra Genie software includes both random and systematic uncertainties in the calculation of the total uncertainties which are listed on the report. We conservatively estimate that 2% is the lowest uncertainty possible for our GEA measurements taking into account systematic uncertainties in gamma calibration standards.

Comments

None

Attachment: Data Report Sample Results for ASR 0372.

Page 4 of 4

Radiochemical Sciences and Engineering Group Pacific Northwest National Laboratory PO Box 999, Richland, WA

filename 18-0001 Fiskum.xls 11/29/2017

13-6-2017 12-6-1 Raing-le Technical Reviewer: Prepared by: Project: 69832 WP: N79882 Client: Fiskum ASR 0372

HORK

RPG-CMC-450, Rev. 3, Gamma Energy Analysis (GEA) and Low-Energy Photon Spectrometry (LEPS) Procedures:

Gamma detectors C,G,L,M,N, R, T 27-Nov-17 M&TE:

Count dates:

Measured Activity, μCi per ml \pm 1s

	Lab										
Sample	ID	C0-6	0	Cs-13	37	Eu-1	52	Eu-1:	54	Am-2	41
TI015-FEED	18-0001	< 4.5E-3		1.09E+02	± 2%	< 1.3E-2		< 1.1E-2		< 4.8E-1	
TI015-EFF-Comp	18-0002	6.22E-04	$\pm 2\%$	4.71E-03	$\pm 2\%$	< 6.5E-6		1.32E-04	$\pm 2\%$	3.86E-04	$\pm 5\%$
TI015-ELComp	18-0003	< 1.0E-2		1.11E+03	$\pm 2\%$	< 3.2E-2		< 2.8E-2		< 6.1E-1	
TI016-FEED	18-0004	< 8.5E-3		1.29E+02	$\pm 2\%$	< 3.3E-2		< 1.8E-2		< 2.1E-1	
TI016-EFF-Comp	18-0005	6.38E-04	$\pm 2\%$	4.01E-02	± 2%	6.01E-06	$\pm 20\%$	1.32E-04	$\pm 2\%$	3.75E-04	$\pm 17\%$
TI016-ELComp	18-0006	< 1.5E-2		1.51E+03	$\pm 2\%$	< 7.1E-2		< 7.3E-2		< 3.3E+0	
T1020-FEED-1-A	18-0007	< 4.7E-3		1.11E+02	$\pm 2\%$	< 1.6E-2		< 1.2E-2		< 4.9E-1	
TI020-EFF Comp-Take 1	18-0008	6.63E-04	$\pm 2\%$	3.52E-03	± 2%	4.08E-06	$\pm 16\%$	1.38E-04	$\pm 2\%$	3.15E-04	$\pm 6\%$
TI020-ELComp	18-0009	< 3.5E-3		1.06E+03	$\pm 2\%$	< 9.0E-3		< 8.9E-3		< 1.9E-1	
TI020-FEED-Take2	18-0010	< 7.3E-3		1.30E+02	$\pm 2\%$	< 2.0E-2		< 1.4E-2		< 2.0E-1	
TI020-EFF Comp-Take2	18-0011	5.66E-04	$\pm 2\%$	4.66E-02	$\pm 2\%$	7.12E-06	$\pm 11\%$	1.22E-04	$\pm 2\%$	3.54E-04	$\pm 8\%$
TI020-EComp	18-0012	< 4.7E-3		2.60E+02	$\pm 2\%$	< 1.8E-2		< 1.7E-2		< 7.9E-1	
TI021-FEED	18-0013	< 1.4E-2		1.27E+02	± 2%	< 6.0E-2		< 3.0E-2		< 9.2E-1	
TI021-EFF Comp	18-0014	6.55E-04	$\pm 2\%$	5.20E-03	$\pm 2\%$	< 8.4E-6		1.26E-04	$\pm 3\%$	4.73E-04	$\pm 26\%$
TI021-ELComp	18-0015	< 3.0E-3		8.60E+02	± 2%	< 1.1E-2		< 1.2E-2		< 6.4E-1	

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

Project / WP#:	69832/N79882
ASR#:	0372.00
Client:	SK Fiskum
Total # of Samples:	15

Plutonium 238, 239+240 Analysis

RPL ID	Client Sample ID
18-0001	TI015-FEED
18-0002	TI015-EFF-Comp
18-0003	TI015-ELComp
18-0004	TI016-FEED
18-0005	TI016-EFF-Comp
18-0006	TI016-ELComp
18-0007	TI020-FEED-1-A
18-0008	TI020-EFF Comp-Take 1
18-0009	TI020-ELComp
18-0010	TI020-FEED-Take2
18-0011	TI020-EFF Comp-Take2
18-0012	TI020-EComp
18-0013	TI021-FEED
18-0014	TI021-EFF Comp
18-0015	TI021-ELComp

Page 1 of 4

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report

Analysis Type:	AEA – Pu-238, Pu-239+240
Sample Processing Prior to Radiochemical Processing/Analysis	 None Digested as per PNL-ALO-106, Rev. 1, Acid Digestion of Waters, Soils, and Sludges for Subsequent Radiochemical Sample Analyses Fusion as per RPG-CMC-115 Rev. 0, Solubilization of Metals from Solids Using a KOH-KNO₃ Fusion Digested as per RPG-CMC-128, Rev.1, HNO₃-HCL Acid extraction of Liquids for Metals Analysis Using a Dry-Block Heater
Pre-dilution Prior to Radiochemical Processing?	 □ No ○ Yes All 5 EL&EComp samples were pre diluted by ~ 100x in the hot cell
Co-Precipitation Procedure:	RPG-CMC-496, Rev. 1, Coprecipitation Mounting of Actinides for Alpha Spectroscopy
Technician/Analyst:	LP Darnell, (01/2/2018 and 01/19/2018)
Spike and Tracer Standard ID's:	R-687-a-2 (Pu-239), R-688-a-10 and R700-a (Pu-242 tracer)
Plutonium Separation Procedure:	RPG-CMC-4017, Rev. 0, Analysis of Environmental Water Samples for Actinides and Strontium-90
Technician/Analyst:	LP Darnell, 1/2/18 and 1/22/18
Analysis Procedure:	RPG-CMC-422, Rev. 2, Solutions Analysis: Alpha Spectrometry
Reference Date:	Same as analyses dates
Analysis Date or Date Range:	January 2-23, 2018
Technician/Analyst:	T. Trang-Le, CZ Soderquist
Analysis Data (File):	RPG-RC\PNL\Projects\Backup files\Backup 18\18-0001 Fiskum.xlsx
CMC Project 98620 File:	File Plan 5871: T 69832: Sample preparation and analysis records; T-4.4 LSC 3100 TR calibration, daily checks, and maintenance records; and T3 standard certificates and preparation. Also balance calibration and instrument performance checks.
M&TE Number(s):	Ortec AEA counters – 32 counters – See attached M&TE list and Mettler AT400, Serial # 1113292667.

Date Date Trang-le ______ Reviewer Date

Preparer

Sample Results

See attached data report, Sample Results for ASR 0372. All data are reported in units of μ Ci/mL with a 1- σ uncertainty.

Sample Preparation, Separation, Mounting and Counting Methods

All fifteen samples submitted under Analytical Service Request (ASR) 0372 were analyzed for plutonium by Alpha Spectrometry. All the samples were prepared in RPL/420. Five of the 15 samples were pre-diluted by $\sim 105x$ with 0.5 M HNO₃ in the hotcell due to high activity of the samples and were processed as received from the hotcell in lab 420. The remaining 10 caustic matrix samples were diluted by $\sim 100x$ in lab 420 prior to the plutonium analyses.

The Pu was separated by anion exchange using procedure RPG-CMC-4017. The separated Pu fraction was then mounted for alpha spectrometry by co-precipitation using procedure RPG-CMC-496, and then counted using alpha spectrometry using procedure RPC-CMC-422. The samples were counted on January 2-23, 2018; no decay corrections were made.

QUALITY CONTROL RESULTS

Radioanalytical quality control (QC) samples prepared in RPL/420 include a lab separation blank, sample duplicate, matrix spike (i.e., addition of Pu-239 standard to an aliquot of one of the samples) and reagent spike (RS, Pu-239).

The QC sample results for Pu-AEA have been evaluated and are discussed below. A summary of the Pu-AEA analysis results, including QC sample performance, is given in the attached data report.

Tracer:

The Pu-242 tracer is added to every sample after appropriate dilution and prior to plutonium separations. The use of a Pu-242 tracer corrects for radiochemical yield and mathematically removes the detector counting efficiency from the results calculations. Tracer recovery is required to be high enough to provide acceptable counting statistics. The Pu-242 tracer counting statistics were acceptable for all samples. The tracer recoveries ranged from 93% to 107%.

Laboratory Preparation Blank (PB):

The activity level of the plutonium alpha emitters present in the blank is well below 5% of the activity present in the sample, meeting the acceptance criteria of less than 5% of the sample activity or less than the sample MDA.

Reagent Spike (RS):

The RS recoveries of 96% and 99% (Pu-239) meets the acceptance criteria of 80% to 120% recovery.

Matrix Spike (MS):

The MS recoveries of 96% and 91% (Pu-239) meet the acceptance criterion of 75% to 125% recovery. Note: The MS sample was prepared after initial dilution, by adding a known quantity of Pu-239 standard to a diluted aliquot of the diluted sample. Sample numbers 18-0005 (TI016-EFF-Comp) and 18-0014 (TI021-EFF Comp) were selected as the matrix spike samples.

Laboratory Duplicate - Relative Percent Difference (RPD):

Sample 18-0005 (TI016-EFF-Comp) and 18-0014 (TI021-EFF Comp) were selected as the duplicate samples. For sample 18-0005, the Pu-238 sample and duplicate RPD of 7% and the duplicate RPD for Pu-239+240 of 18% are within the acceptance limit of \leq 20% RPD. For sample 18-0014, the Pu-238 sample and duplicate RPD of 5% and the duplicate RPD for Pu-239+240 of 8% are within the acceptance limit of \leq 20% RPD.

Instrument Quality Control

Alpha counters receive initial calibration with NIST traceable sources to determine the counter efficiency. When internal tracers are not used, the counter efficiency is used in calculation of final results.

Detector backgrounds are determined every 4 weeks or after the last analytical run, whichever is longer. Detector background counts are subtracted from all subsequent sample counts. A process blank is analyzed with each analytical batch to evaluate for contamination in the sample preparation process.

Assumption and Limitations of the Data

None

Attachment: Data Report -- Sample Results for ASR 0372.

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

Project / WP#:	69832/N79882	
ASR#:	0372.00	
Client:	SK Fiskum	
Total # of Samples:	15	

	00	A 1	
11		Angl	TOIC
ж.	レーフフ	Alla	I V 313

RPL ID	Client Sample ID
18-0001	TI015-FEED
18-0002	TI015-EFF-Comp
18-0003	TI015-ELComp
18-0004	TI016-FEED
18-0005	TI016-EFF-Comp
18-0006	TI016-ELComp
18-0007	TI020-FEED-1-A
18-0008	TI020-EFF Comp-Take 1
18-0009	TI020-ELComp
18-0010	TI020-FEED-Take2
18-0011	TI020-EFF Comp-Take2
18-0012	TI020-EComp
18-0013	TI021-FEED
18-0014	TI021-EFF Comp
18-0015	TI021-ELComp

Page 1 of 4

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report

Analysis Type:	Tc-99
Sample Processing Prior to Radiochemical Processing/Analysis	 None Digested as per PNL-ALO-106, Rev. 1, Acid Digestion of Waters, Soils, and Sludges for Subsequent Radiochemical Sample Analyses Fusion as per RPG-CMC-115 Rev. 0, Solubilization of Metals from Solids Using a KOH-KNO₃ Fusion
	Digested as per RPG-CMC-128, Rev.1, HNO ₃ -HCL Acid extraction of Liquids for Metals Analysis Using a Dry-Block Heater
Pre-dilution Prior to Radiochemical Processing?	🗌 No
	\boxtimes Yes All 5 ELComp samples were pre diluted by ~ 100x in the hot cell
RadioChemical Preparation Procedure:	RPG-CMC-432, Rev. 0, Technicium-99 Analysis
Technician/Analyst:	LP Darnell, 01/10/2018
Spike and Tracer Standard ID's:	R-540-b-8 (Tc-99)
Analysis Procedure:	RPG-CMC-474, Rev. 1, Measurement of Alpha and Beta Activity by Liquid Scintillation Spectrometry
Reference Date:	Not applicable
Analysis Date(s) or Date Range:	01/11/2018
Technician/Analyst:	CZ Soderquist
Analysis Data (File):	RPG-RC\PNL\Projects\Backup files\Backup 18\18-0001 Fiskum.xlsx
CMC Project 98620 File:	File Plan 5871: T 69832: Sample preparation and analysis records; T-4.4 Alpha Detector calibration, calibration verification checks, and maintenance records; and T3 Standard certificates and preparation. Also balance calibration and instrument performance checks.
M&TE Number(s):	Perkin Elmer Tri-Carb 3100 TR, Serial # DG08061340, RPL 425, Tri-Carb 2700TR software version 1.04 dated 9/99, Mettler AT400, Serial # 1113292667

Preparer

<u>n. Pale / 2/14/18 Trang-le / 2/14/18</u> er Date Reviewer Date

Page 2 of 4

Sample Results

See attached data report, Sample Results for ASR 0372. All data are reported in units of μ Ci/mL with a 1- σ uncertainty (see comments).

Sample Preparation, Separation, Mounting and Counting Methods

All fifteen samples submitted under Analytical Service Request (ASR) 0372 were analyzed for Tc-99 by chemical separation and beta counting. All the samples were prepared in RPL/420. Five of the 15 samples were pre-diluted by $\sim 100x$ with 0.5 M HNO₃ in the hotcell due to high activity of the samples and were processed as received from the hotcell in lab 420. The remaining 10 caustic matrix samples were diluted by $\sim 100x$ in lab 420 prior to Tc-99 analyses by procedure RPG-CMC-432, Rev. 0, *Technicium-99 Analysis*.

The samples were counted on January 11, 2018; no decay corrections were made.

QUALITY CONTROL RESULTS

Radioanalytical quality control (QC) samples prepared in RPL/420 include a lab separation blank (LB), sample duplicate, matrix spike (i.e., addition of Tc-99 standard to an aliquot of one of the samples) and reagent spike. (RS, Tc-99).

The QC sample results for Tc-99 have been evaluated and are discussed below. A summary of the Tc-99 analysis results, including QC sample performance, is given in the attached data report.

Tracer:

There is no tracer for Tc-99 analysis.

Laboratory Preparation Blank (PB):

The activity level of the Tc-99 present in the lab is very near the instrument background and below the MDA for all samples. The lab blank meets the acceptance criteria for being below the MDA for all samples.

Reagent Spike (RS):

The RS recovery of 89% (Tc-99) meets the acceptance criteria of 80% to 120% recovery.

Matrix Spike (MS):

The MS recovery for Tc-99 could not be calculated. The activity of Tc-99 added to the matrix spike samples was too low relative to the activity present in the sample. Sample 18-0001 (TI015-FEED) was selected for the matrix spike analyses. Note: The MS sample was prepared after initial dilution, by adding a known quantity of Tc-99 standard to a an aliquot of the diluted sample.

Page 3 of 4

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report

Laboratory Duplicate - Relative Percent Difference (RPD):

Sample 18-0014 (TI021-EFF Comp) was selected as the duplicate sample. The Tc-99 sample and duplicate RPD of 10% is within the acceptance limit of \leq 20% RPD.

Instrument Quality Control

LSC counters receive initial calibration with NIST traceable sources to determine the counter efficiency. When internal tracers are not used, the counter efficiency is used in calculation of final results.

Detector backgrounds are determined every 4 weeks or after the last analytical run, whichever is longer. Detector background counts are subtracted from all subsequent sample counts. A process blank is analyzed with each analytical batch to evaluate for contamination in the sample preparation process.

Assumption and Limitations of the Data

None

Attachment: Data Report -- Sample Results for ASR 0372.

Page 4 of 4

Pacific Northwest Natio Richland, WA Radiochemical Sciences	nal Laboratory and Engineerir	1g Group			Ţ	ilename 1 F	8-0001 Fisl /24/2018 'artial Data:	kum.xls uranium results to follow later
Client: S. Fiskum ASR 372	Project: 69832 WP: N79882		Pre Technical I	pared by: T	freng	els o	2113118	
Procedures:	RPG-CMC-43 RPG-CMC-47- RPG-CMC-49 RPG-CMC-42 RPG-CMC-40	2, Rev 0, T 4, Rev 1, M 6, Rev 1, C 2, Rev 2, S 14, Rev 1,	echnetium- feasuremen oprecipitat olutions Ar Uranium by	99 Analysis t of Alpha and Beta A ion Mounting of Actii ialysis: Alpha Spectro / Kinetic Phosphoresc	vctivity by I nides for Al metry (Np, ence Analy	.iquid Scin pha Spectro Pu, Am, Ci	tillation Spe ometry n)	cctrometry (Tc-99, Sr-90)
M&TE:	Perkin Elmer J Chem-Chek In	TriCarb mo	del 3100TF model KPA	<pre>L liquid scintillation, / -11R kinetic phospho</pre>	Alpha Spect rescence an	rometry co alyzer	unting syste	E
Count dates:	January 11-23,	2018				•		
	Lab		M	easured activities, µ($Ci/mL \pm 1s$	uncertaint	v	
Sample	D	Tc-	66	Total U, μg/mL	Pu-2	38	Pu-239	+240
T1015-FEED	18-0001	1.03E-1	± 2%		1.68E-5	± 14%	5.82E-5	$\pm 8\%$
TI015-EFF-Comp	18-0002	8.02E-2	= 2%		2.80E-5	$\pm 11\%$	5.68E-5	= 8%
TI015-ELComp	18-0003	7.87E-6	$\pm 23\%$		5.84E-7	± 122%	2.33E-7	± 285%
TI016-FEED	18-0004	8.92E-2	$\pm 2\%$		1.26E-5	$\pm 17\%$	5.77E-5	$\pm 8\%$
TI016-EFF-Comp	18-0005	7.92E-2	± 2%		1.39E-5	$\pm 16\%$	6.20E-5	± 7%
	18-0005 Dup	1			1.30E-5	$\pm 16\%$	5.16E-5	$\pm 8\%$
	RPD				7%		18%	
TI016-ELComp	18-0006	7.58E-6	± 23%		-1.29E-7	± 725%	1.16E-6	± 81%
T1020-FEED-1-A	18-0007	8.62E-2	$\pm 2\%$		2.01E-5	$\pm 13\%$	7.39E-5	= 7%
TI020-EFF Comp-Take 1	18-0008	7.67E-2	$\pm 2\%$		1.52E-5	$\pm 16^{0/0}$	4.39E-5	$\pm 9\%$
TI020-ELComp	18-0009	5.15E-6	$\pm 33\%$		3.14E-7	$\pm 224\%$	-6.28E-8	$\pm 1025\%$
T1020-FEED-Take2	18-0010	8.32E-2	± 2%		2.16E-5	$\pm 13\%$	5.94E-5	\pm 8%
TI020-EFF Comp-Take2	18-0011	7.72E-2	$\pm 2\%$		1.65E-5	$\pm 14\%$	5.18E-5	\pm 8%
TI020-EComp	18-0012	2.66E-5	± 49%		-1.25E-7	± 725%	8.12E-7	± 105%

Page 1 of 2

B.124

Pacific Northwest National Laboratory Richland, WA Radiochemical Sciences and Engineering Group

filename 18-0001 Fiskum.xls 1/24/2018 Partial Data; uranium results to follow later

Sample ID IC-39 IOMALO, R0140 IC-39 IOMALO, R0140 IC-39 IOMALO, R0140 IC-39 R01E-5 IOMALO, R0150 R01E-5 IOMALO, R000 R000E+0 R017E-5 IOMALO, R000E+0 R000E+0		E	E				000	OCC - U	0701
$ \begin{array}{cccccccccccccccccccccccccccccccccccc$	Sample	ID	5	22	10tal O, µg/mL	7-n J	003	KC7-NJ	1740
TI021-EFF Comp18-00147.50E-2 $\pm 2\%$ $5.4E-5$ $\pm 9\%$ $6.17E-5$ 18-0014 Dup8.26E-2 $\pm 2\%$ $3.74E-5$ $\pm 9\%$ $6.66E-5$ 18-0014 Dup8.26E-2 $\pm 2\%$ $3.74E-5$ $\pm 9\%$ $6.66E-5$ 7.87E-6 $\pm 23\%$ $0.00E+0$ $8.57E-7$ TI021-ELComp18-0015 $7.87E-6$ $\pm 23\%$ $0.00E+0$ $-8.57E-7$ Reagent spike89% $$ $$ $99\%, 96\%$ Matrix spike 18-0014 $$ $$ $$ $99\%, 96\%$ Matrix spike 18-0005 -106% $-1.55E-6$ $-1.55E-6$ $-1.55E-6$	TI021-FEED	18-0013	8.61E-2	± 2%		5.39E-5	$\pm 8\%$	8.01E-5	± 7%
18-0014 Dup8.26E-2 $\pm 2\%$ $3.74E-5$ $\pm 9\%$ $6.66E-5$ 71021-ELComp18-00157.87E-6 $\pm 23\%$ 0.00E+0 $-8.57E-7$ Reagent spike89% $\pm 23\%$ 0.00E+0 $-8.57E-7$ $-8.57E-7$ Matrix spike 18-0014 $$ 99%, 96% $-1.55E-6$ Matrix spike 18-0005 $3.74E-6$ $\pm 106\%$ $-7.43E-7$ $\pm 65\%$ $-1.55E-6$	TI021-EFF Comp	18-0014	7.50E-2	± 2%		3.94E-5	+ 9%	6.17E-5	$\pm 8\%$
Ti021-ELComp 18-0015 7.87E-6 $\pm 23\%$ 5% 8% Reagent spike 89% 99%,96% 91% Matrix spike 18-0014 99% 91% 91% 91% Lab blank 3.74E-6 $\pm 106\%$ -7.43E-7 $\pm 65\%$ -1.55E-6		18-0014 Dup	8.26E-2	± 2%		3.74E-5	± 9%	6.66E-5	± 7%
T1021-ELComp 18-0015 7.87E-6 ±23% 0.00E+0 -8.57E-7 Reagent spike 89% 99%,96% 99%,96% 96% 91% 91% 91% 91% 1.55E-6			10%			5%		8%	
Reagent spike89%99%, 96%Matrix spike 18-001496%Matrix spike 18-000591%Lab blank $3.74E-6 \pm 106\%$ $-7.43E-7 \pm 65\%$ $-1.55E-6$	T1021-ELComp	18-0015	7.87E-6	± 23%		0.00E+0		-8.57E-7	± 93%
Matrix spike 18-0014 96% Matrix spike 18-0005 -7.43E-7 $\pm 65\%$ -1.55E-6 -5.65% -1.55E-6 -7.06-7 $\pm 115\%$ -6.66%		Reagent spike	89%			l		96%, 96%	
Matrix spike 18-0005 91% $-7.43E-7 \pm 65\% -1.55E-6$ 7.06% 7.07 $-7 \pm 11.5\%$ -6.60 $+7$	K	Matrix spike 18-0014	ł			1		6%	
Lab blank 3.74E-6 $\pm 106\%$ -7.43E-7 $\pm 65\%$ -1.55E-6 7.70E-7 $\pm 115\%$ -6.60E-7	K.	Matrix spike 18-0005						91%	
7.20E-7 + 11.5% -6.60E-7		Lab blank	3.74E-6	$\pm 106\%$		-7.43E-7	$\pm 65\%$	-1.55E-6	± 45%
						7.20E-7	$\pm 115\%$	-6.60E-7	± 93%

Page 2 of 2

Battelle PNNL/RPL/Inorganic Analysis ... ICP-MS Analysis Report PO Box 999, Richland, Washington 99352

Project / WP#: ASR#: Client: **Total Samples:** 69832 / N79882 0372.01 S. Fiskum 15 (Aqueous)

ASO Sample ID	Client Sample ID	Client Sample Description	Sample Weight (g)
18-0001	TI015-FEED	Caustic/5.7 M Na, high Salt	NA
18-0002	TI015-EFF Comp	Caustic/5.7 M Na, high Salt	NA
18-0003	TI015-ELComp	Acidic/~0.2 M HNO ₃	NA
18-0004	TI016-FEED	Caustic/5.7 M Na, high Salt	NA
18-0005	TI016-EFF Comp	Caustic/5.7 M Na, high Salt	NA
18-0006	TI016-ELComp	Acidic/~0.2 M HNO ₃	NA
18-0007	TI020-FEED-1-A	Caustic/5.7 M Na, high Salt	NA
18-0008	TI020-EFF Comp-Take-1	Caustic/5.7 M Na, high Salt	NA
18-0009	TI020-ELComp	Acidic/~0.2 M HNO ₃	NA
18-0010	TI020-FEED-Take2	Caustic/5.7 M Na, high Salt	NA
18-0011	TI020-EFF Comp-Take2	Caustic/5.7 M Na, high Salt	NA
18-0012	TI020-EComp	Acidic/~0.2 M HNO ₃	NA
18-0013	TI021-FEED	Caustic/5.7 M Na, high Salt	NA
18-0014	TI021-EFF Comp	Caustic/5.7 M Na, high Salt	NA
18-0015	TI021-ELComp	Acidic/~0.2 M HNO ₃	NA

Sample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of Liquids for Metals Analysis Using a Dry-Block Heater", performed by L. Darnell on Samples 18-0001, 18-0002, 18-0004, 18-0005, 18-0007, 18-0008, 18-0010, 18-0011, 18-0013, and 18-0014 on 11/28/17. Simple dilution in 5% HNO3 was performed on Samples 18-0003, 18-0006, 18-0009, 18-0012, and 18-0015 by J. Turner on 10/25/17. Prior to analysis all samples were further diluted in 2% v/v HNO3 by G. Brown on 02/28/18.

a

Report Preparer

Review and Concurrence

Date

4/11/18

Date

Page 1 of 5

Procedure	RPG-CMC-2 Inductively C	92, Rev. 1, "Determi oupled Argon Plasma	nation 1 Mass	of Elemental Spectrometry	Composi y (ICP-M	tion by S)."	
Analyst:	G. Brown	Analysis Date:	03/02	2/2018	ICP Fil	e: M005	8A
See Chemi	ical Measuremen	t Center 98620 file:	<u>ICP</u> (Cal	- <u>325-405-3</u> ibration and	Maintena	nce Record	s)
M&TE:	PerkinElmer	SN: 85VN4	070702	RPL 405			
	Sartorius R20	0D Balance		SN: 390800)42	RPL 405	
	Mettler AT40	0 Balance		SN: M1944	RPL 405	FH	
	Mettler AT40	0 Balance		SN: 111329	RPL 420	FH	
	Ohaus EX324	Balance		SN: 803331	1209	SAL Cell	2
Ī	Denver A-160) Balance		SN: 60568		RPL 201	
Ī	Sartorius R20	0D Balance		SN: 390800)58	RPL 525	FH

Battelle PNNL/RPL/Inorganic Analysis ... ICP-MS Analysis Report

Fifteen aqueous samples submitted under Analytical Service Request (ASR) 0372.01 were analyzed by ICP-MS. Samples 18-0001, 18-0002, 18-0004, 18-0005, 18-0007, 18-0008, 18-0010, 18-0011, 18-0013, and 18-0014 were prepared following RPL procedure RPG-CMC-128 and diluted to approximately 25 mL. Sample 18-0014 was prepared in duplicate following RPL procedure RPG-CMC-128 and diluted to approximately 25 mL. Samples 18-0003, 18-0003, 18-0006, 18-0009, 18-0012, and 18-0015 were diluted with 5% HNO₃ in the hot cell to approximately 105x. All samples were further diluted in 2% HNO₃ prior to analysis. None of the samples were filtered.

All results are reported on a mass per unit volume basis (ng/mL) for each detected analyte. The data have been adjusted for instrument dilutions.

Analytes of interest (AOI) were specified in the ASR and are listed in the upper section of the attached ICP-MS Data Report. Uranium-238 was the only AOI. The quality control (QC) results for the AOI have been evaluated and are presented below.

Calibration of the ICP-MS was done following the manufacturer's recommended calibration procedure using multi-analyte custom standard solutions traceable to the National Institute of Standards and Technology (NIST). Midrange calibration verification standards were used to verify acceptance of the six-point calibration curve and for initial and continuing calibration verification (ICV/CCV).

The controlling document was procedure RPG-CMC-292, Rev. 1, *Determination of Elemental Composition by Inductively Coupled Argon Plasma Mass Spectrometry (ICP-MS)*. Instrument calibrations, QC checks and blanks (e.g., ICV/ICB, CCV/CCB, LLS, ICS), post-digestion spike, duplicate, blank spike, and serial dilution were conducted during the analysis run.

Detection Limit:

The ASR specified a target detection limit of $0.5 \ \mu g/mL$ for total uranium. The instrument detection limit (IDL) and estimated quantitation limit (EQL) were determined from

replicate analyses of a 2% HNO₃ blank solution. Both values (IDL = 0.0001 ng/mL, EQL = 0.001 ng/mL) were significantly below the requested target detection limit.

Internal Standard (IS):

All solutions (blanks, standards, QC checks, and samples) were mixed in-line with a solution containing 10 ppb each of Li-6, Sc-45, Y-89, In-115, Tb-159, and Bi-209 as the internal standard (IS). The AOI (U-238) data were normalized using the data for the closest IS mass (i.e., Bi-209). With one exception, the Bi-209 IS recoveries ranged from 91% to 103% and were within the acceptance criterion of 30% to 120%. The second interference check standard had a Bi-209 IS recovery of 131% that was caused by insufficient volume of the ICS standard in the autosampler (no ICS standard was available to mix in-line with the IS solution, resulting in a higher IS recovery).

Preparation Blank (PB):

A preparation blank (reagents only) was prepared for the extraction process. Results for the PB was 1.18 ng/mL U-238, which, after accounting for process factors, exceeded the acceptance criteria of <EQL (estimated quantitation level), less than $\leq 10\%$ of the concentration in the samples, or <50% regulatory decision level. The uranium concentration in the PB at 243x dilution was greater than 10% of the concentrations in several of the diluted samples with process factors of ~270,000 (18-0001, 18-0002, 18-0004, 18-0005, 18-0007, 18-0008, 18-0010, 18-0011, 18-0013, and 18-0014). Trace levels of various environmental contaminants (e.g., calcium, iron, uranium, zinc, etc.) are to be expected when analyzing highly diluted samples that have been prepared in a radiological laboratory. In addition to the PB, a diluent blank from the ICP-MS laboratory (2% HNO₃) was analyzed and the result was <EQL (<0.001 ng/mL U-238).

Reagent Spike (RS)/Laboratory Control Sample (LCS):

A reagent spike (RS) sample (reagent and spikes) was prepared for the extraction process. The recovery for the AOI was 90%, which is within the acceptance criteria of 80% to 120% recovery.

Blank Spike (BS)/Laboratory Control Sample (LCS):

A blank spike (BS) sample (lab diluent and spikes) was prepared by spiking the 2% HNO₃ lab diluent with an equivalent volume of a 2ppb standard (1:1 ratio). The recovery for the AOI was 96%, which is within the acceptance criteria of 80% to 120% recovery.

Duplicate/Replicate Relative Percent Difference (RPD):

Duplicates of sample 18-0014 were prepared and analyzed. RPD are listed for all analytes that were measured at or above the EQL. The RPD for the AOI meeting this requirement was 12% and was within the acceptance criterion of \leq 20% for liquid samples.

Matrix-Spike (MS) Sample:

A matrix spike (MS) sample (18-0001 and spikes) was prepared for the extraction process. Recovery values are listed for all analytes included in the MS that were measured at or above the EQL. The recovery value for the AOI meeting this requirement was 90% and was within the acceptance criterion of 75% to 125%.

Initial/Continuing Calibration Verification (ICV/CCV):

The ICV/CCV solution was analyzed immediately after calibration, after each group of generally not more than ten samples, and at the end of the analytical run. Due to analyst oversight, twelve samples were analyzed between the last two groups of CCV solutions. The concentrations of all AOI were within the acceptance criteria of 90% to 110% recovery.

Initial/Continuing Calibration Blank (ICB/CCB):

The ICB/CCB solution ($2\% v/v HNO_3$) was analyzed immediately after the ICV solutions and after the CCV solutions (after each group of generally not more than ten samples and at the end of the analytical run). Due to analyst oversight, twelve samples were analyzed between the penultimate and final CCB solutions. The concentration of all AOI were within the acceptance criteria of <EQL. Following an additional rinse time, the final CCB met the acceptance criteria.

Low-Level Standard (LLS):

The LLS solution was analyzed immediately after the first CCB solution. The recovery for the AOI (99.7%) was within the acceptance criteria of 70% to 130%.

Interference Check Standard (ICS):

The ICS solution was analyzed immediately after the first LLS solution and immediately prior to analyzing the final CCV solutions. The concentration of the AOI in the first ICS corresponded to 98% recovery, which is within the acceptance criteria of 80% to 120%. However, U-238 recovery in the second ICS was 55% due to insufficient volume of solutions loaded into the autosampler.

Serial Dilution (SD):

Five-fold serial dilution was conducted on samples 18-0003. Percent differences (%D) are listed for all analytes that had a concentration at or above the EQL in the diluted sample. The %Ds for the AOI meeting this requirement was 0.6% and was within the acceptance criterion of $\leq 10\%$.

Post-Digestion Spike (PS)/Analytical Spike (AS) - Sample (71A Component):

In addition to the MS sample, a post-digestion spike was conducted on samples 18-0003. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration $\geq 25\%$ of that in the sample. Recovery value for the AOI meeting this requirement was 96%, and was within the acceptance criterion of 80% to 120%.

Other QC:

All other instrument-related QC tests for the AOI passed within their respective acceptance criteria.

Comments:

- The "Final Results" have been corrected for all laboratory dilutions performed on the samples during processing and analysis, unless specifically noted.
- 2) Instrument detection limits (IDL) and estimated quantitation limits (EQL) shown are for acidified water and/or fusion flux matrices as applicable. Method detection limits (MDL) for individual samples can be estimated by multiplying the IDL by the "Process Factor" for that individual sample. The estimated quantitation limit (EQL) for each concentration value can be obtained by multiplying the EQL by the "Process Factor".
- 3) Routine precision and bias is typically ±15% or better for samples in dilute, acidified water (e.g. 2% v/v HNO₃ or less) at analyte concentrations > EQL up to the upper calibration level. This also presumes that the total dissolved solids concentration in the sample is less than 5000 µg/mL (0.5 per cent by weight). Note that bracketed values listed in the data report are within the MDL and the EQL, and have potential uncertainties greater than 15%. Concentration values < MDL are listed as "--". Note, that calibration and QC standard samples are validated to a precision of ±10%.</p>
- 4) Analytes included in the spike 71A component (for the AS/PS) are; Ag, Al, As, B, Ba, Be, Ca, Cd, Ce, Co, Cr, Cs, Cu, Dy, Er, Eu, Fe, Ga, Gd, Ho, K, La, Lu, Mg, Mn, Na, Nd, Ni, P, Pb, Pr, Rb, S, Se, Sm, Sr, Th, Tl, Tm, U, V, Yb, and Zn. Analytes included in the spike 71B component are; Ge, Hf, Mo, Nb, Sb, Si, Sn, Ta, Te, Ti, W, and Zr. Analytes included in the spike 71C component are; Ir, Os, Pd, Pt, Re, Rh, and Ru. Analytes included in the spike 71D component are; Bi, In, Li, Sc, Tb, and Y. Analytes included in the spike Hg component are; Hg.

Battelle PNNL/RPG/Inorganic Analysis ... ICP-MS Data Report

		Run Date >	3/2/2018	3/2/2018	3/2/2018	3/2/2018	3/2/2018	3/2/2018	3/2/2018	3/2/2018	3/2/2018	3/2/2018	3/2/2018
		Process Factor >	1	242.59	277,636	272,399	10,217	269,723	267,856	10,257	268,115	266,807	10,295
		RPL/LAB >	BLK-2% HNO3	BLK-0001 @ 5x	18-0001 @ 5625x	18-0002 @ 5625x	18-0003 @ 100x	18-0004 @ 5625x	18-0005 @ 5625x	18-0006 @ 100x	18-0007 @ 5625x	18-0008 @ 5625x	18-0009 @ 100x
Instr. Det. Limit (IDL)	Est. Quant. Limit (EQL)	Client ID >	BLK-2% HNO3	Prep Reagents	TI015-FEED	TI015-EFF Comp	TI015- ELComp	TI016-FEED	TI016-EFF Comp	TI016- ELComp	TI020-FEED- 1-A	TI020-EFF Comp- Take1	TI020- ELComp
(ng/mL)	(ng/mL)	(Analyte)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)
0.0001	0.001	U 238		1.18	6025	5511	1653	6294	5712	2470	5867	4535	2216

Internal Standard % Recovery

101%	125%	104%	102%	105%	100%	100%	102%	100%	101%	104%
102%	107%	105%	105%	105%	104%	103%	102%	102%	101%	104%
100%	100%	101%	100%	99%	102%	101%	102%	100%	101%	97%
101%	99%	101%	101%	97%	101%	100%	98%	99%	97%	100%
96%	100%	98%	99%	98%	99%	99%	98%	98%	98%	99%
99%	97%	96%	99%	98%	96%	98%	97%	97%	98%	96%
	101% 102% 100% 101% 96% 99%	101% 125% 102% 107% 100% 100% 101% 99% 96% 100% 99% 97%	101% 125% 104% 102% 107% 105% 100% 100% 101% 101% 99% 101% 96% 100% 98% 99% 97% 96%	101% 125% 104% 102% 102% 107% 105% 105% 100% 100% 101% 100% 101% 99% 101% 101% 96% 100% 98% 99% 99% 97% 96% 99%	101% 125% 104% 102% 105% 102% 107% 105% 105% 105% 100% 100% 101% 100% 99% 101% 99% 101% 101% 97% 96% 100% 98% 99% 98% 99% 97% 96% 99% 98%	101% 125% 104% 102% 105% 100% 102% 107% 105% 105% 105% 104% 100% 100% 101% 100% 99% 102% 101% 99% 101% 100% 99% 102% 101% 99% 101% 101% 97% 101% 96% 100% 98% 99% 98% 99% 99% 97% 96% 99% 98% 96%	101% 125% 104% 102% 105% 100% 100% 102% 107% 105% 105% 105% 104% 103% 100% 100% 101% 100% 99% 102% 101% 101% 99% 101% 100% 99% 102% 101% 96% 100% 98% 99% 98% 99% 99% 99% 99% 97% 96% 99% 98% 96% 98%	101% 125% 104% 102% 105% 100% 100% 102% 102% 107% 105% 105% 105% 104% 103% 102% 100% 100% 101% 100% 99% 102% 101% 102% 100% 100% 101% 100% 99% 102% 101% 102% 101% 99% 101% 101% 97% 101% 100% 98% 96% 100% 98% 99% 98% 99% 98% 97% 99% 97% 96% 99% 98% 96% 98% 97%	101% 125% 104% 102% 105% 100% 100% 102% 100% 102% 107% 105% 105% 105% 104% 103% 102% 102% 100% 100% 101% 100% 99% 102% 101% 102% 102% 100% 100% 101% 100% 99% 102% 100% 100% 100% 99% 101% 101% 97% 101% 100% 98% 99% 99% 98% 99% 98% 98% 98% 98% 98% 97% <t< td=""><td>101% 125% 104% 102% 100% 100% 102% 100% 101% 102% 107% 105% 105% 105% 104% 103% 102% 102% 101% 100% 100% 101% 105% 104% 103% 102% 101% 100% 100% 101% 102% 101% 102% 101% 101% 99% 101% 101% 97% 101% 100% 98% 99% 97% 96% 100% 98% 99% 98% 99% 98% <t< td=""></t<></td></t<>	101% 125% 104% 102% 100% 100% 102% 100% 101% 102% 107% 105% 105% 105% 104% 103% 102% 102% 101% 100% 100% 101% 105% 104% 103% 102% 101% 100% 100% 101% 102% 101% 102% 101% 101% 99% 101% 101% 97% 101% 100% 98% 99% 97% 96% 100% 98% 99% 98% 99% 98% <t< td=""></t<>

1) "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the "multiplier"

near the top of each column. The estimated sample quantitation limit = EQL (in Column 2)

times the "multiplier". Overall error for values \ge EQL is estimated to be within ±15%.

Values in brackets [] are ≥ MDL but < EQL, with errors likely to exceed 15%.

IS = Internal Standard. The concentration of certain elements cannot be determined due to the presence of the IS in all solutions.

Page 1 of 3

Page 2 of 3

Battelle PNNL/RPG/Inorganic Analysis ... ICP-MS Data Report

		Run Date >	3/2/2018	3/2/2018	3/2/2018	3/2/2018	3/2/2018	3/2/2018	3/2/2018
		Process Factor >	269,263	268,038	10,252	271,700	268,660	482,058	10,240
		RPL/LAB >	18-0010 @ 5625x	18-0011 @ 5625x	18-0012 @ 100x	18-0013 @ 5625x	18-0014 @ 5625x	Dup-0014 @ 10,000x	18-0015 @ 100x
Instr. Det. Limit (IDL)	Est. Quant. Limit (EQL)	Client ID >	TI020-FEED- Take 2	TI020-EFF Comp- Take2	TI020- EComp	TI021-FEED	TI021-E	FF Comp	TI021- ELComp
(ng/mL)	(ng/mL)	(Analyte)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)
0.0001	0.001	U 238	6077	5516	1457	5804	5305	5984	2276

Internal Stan							
Li 6 (IS)	99%	104%	104%	104%	99%	93%	105%
Sc 45 (IS)	100%	103%	104%	102%	106%	96%	107%
Y 89 (IS)	102%	97%	100%	99%	100%	90%	101%
In 115 (IS)	99%	98%	98%	100%	99%	92%	100%
Tb 159 (IS)	96%	98%	97%	99%	99%	94%	96%
Bi 209 (IS)	96%	96%	97%	97%	98%	91%	96%

1) "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the "multiplier"

near the top of each column. The estimated sample quantitation limit = EQL (in Column 2)

times the "multiplier". Overall error for values \geq EQL is estimated to be within ±15%.

2) Values in brackets [] are \geq MDL but < EQL, with errors likely to exceed 15%.

IS = Internal Standard. The concentration of certain elements cannot be determined due to the presence of the IS

QC Performance	03/02/2018
----------------	------------

Criteria >	≤ 20%	80%-120%	80%-120%	75%-125%	75%-125%	≤ 10%
QC ID >	18-0014 Dup	LCS/RS	LCS/BS	18-0001 MS	18-0003 + CCV	18-0003 5-fold Serial Dil
Analytes	RPD (%)	RPD (%)	RPD (%)	%Rec	%Rec	%Diff
U 238	12%	90%	96%	90%	96%	0.6%

Internal Stan

treet treet						
Li 6 (IS)	93%	105%	104%	93%	103%	105%
Sc 45 (IS)	96%	103%	104%	95%	102%	105%
Y 89 (IS)	90%	98%	99%	89%	97%	99%
In 115 (IS)	92%	99%	98%	93%	97%	97%
Tb 159 (IS)	94%	95%	96%	93%	96%	98%
Bi 209 (IS)	91%	101%	94%	97%	94%	98%

Shaded results are outside the acceptance criteria.

nr = spike concentration less than 25% of sample concentration. Matrix effects can be assessed from the serial dilution.

na = not applicable; KOH flux and Ni crucible or Na2O2 flux and Zr crucible for fusion preparations, or Si for HF assisted digests.

IS = Internal Standard. The concentration of certain elements cannot be determined due to the presence of the IS in all solutions.

NM = Not measured. The isotope was not measure due to method or molecular interference limitations.

Updated 03/17/2016

ĺ

Page 1 of 4

P	and a second			
	Acquisition Date	9/2006	8/2015	5/2011
	SN/ Lot#			
	Calibration Standards	Blank, H-3, and C-14 sealed sources & Quench curves (efficiency) for H-3,C-14, and Ni-63	Blank, H-3, C-14 instrument performance assessment (IPA) sources & Quench curves (efficiency) for H-3 & C-14	Blank, H-3, C-14 instrument performance assessment (IPA) sources & Quench curves (efficiency) for H-3 & C-14
'E List	Calibration Frequency	With use	With use	With use
istry M&T	M&TE Category	2	0	0
ladiochemi	Location	RPL/425	RPL/425	RPL/425
H	Serial Number/ Property Number	WD48466	WD56319	PT27127
	Model	3100TR	5110TR	3110TR
	Manufacturer/ SW	Perkin Elmer QuantaSmart 3.00/Build 5	Perkin Elmer QuantaSmart 5.00/Build 3	Perkin Elmer QuantaSmart 4.00/Build 5
	Description	Liquid Scintillation Counter	Liquid Scintillation Counter	Liquid Scintillation Counter

age 2 of 4		8/2015		1994
P				
		I-129 instrument performance assessment source (IPA)		Too many to list
		With use		With use
		2		
	RPL/425	RPL/93		RPL/525
	401664/ WD06783	WD56320		Serial number: 1045050203 WD59770
	2550	2480		KPA- 11R
	Packard Tri Carb Packard Tri- Carb 2700TR software version 1.04 dated 9/99. The software runs on a DOS- 2000 platform on a PC computer	Perkin Elmer Wizard 2		Chemcheck Instruments, Inc. The software used to control the instrument is KPAWIN Software, Version 1.2.9 Release 1999
ted 03/17/2016	Liquid Scintillation Counter	Automated NaI	Uranium	Uranium Analyzer/ Kinetic Phosphoresce nce Analyzer
Updat				

z

Updated 03/17/2016

Page 3 of 4

		Ka	diochemistry M.	WIELLIST		
	Number of		Model or Serial	Location, in Bldg.		PM
ID	Detectors	Property #	#	325	Procedures using the M&TE	Contract?
Dual Alpha Beta Gas Prop	portional Count	ters				
Oxford LB4100	16	WD13066	L8285-0	425	Total alpha, total beta, Sr/Y90, Tc99	Y
Alpha Counters						
Ludlum	10	multiple	multiple	425	Total Alpha	Z
AEA (Ortec)*	32	multiple	multiple	425	AEA (total alpha, Pu, Am/Cm,Np)	Z
Liquid Scintillation						
Perkin Elmer 3100TR	1	WD48466	DG08061340	425	H-3, C-14, Ni-63, Se-79, Pu-241, Sr-90	Y
Perkin Elmer 3110TR	I	PT27127	DG03117614	425	H-3, C-14, Ni-63, Se-79, Pu-241, Sr-90	Υ
Perkin Elmer 5110TR	1	WD56319	SGL0344150055	425	H-3, C-14, Ni-63, Se-79, Pu-241, Sr-90	Υ
Packard 2550	+	WD06783	401664	425	H 3, C 14, Ni 63, Se 79, Pu 241, Sr 90	*
Gamma						
HpGe*	25	see list	see list	425/93	GEA	Z
LEPS (x-rays)*	4	see list	see list	425	X-ray (Fe-55, Ni-59, Nb-93m)	Z
Well HPGe*	3	see list	see list	93	GEA – In development	Z
Anti-Compton*	2	see list	see list	93	GEA – In development	z
Perkin Elmer 2480 Wizard Automated Nal	1	WD56320	SGWZ28150290	93	In development	γ
Uranium						
Chemcheck	I	WD59770	KPA11R	525	Uranium KPA	Z
*Spectral Analysis System	u					
Canberra VAX 3000	Ŧ	WD12890	PE42AJB	425	AEA, GEA	z
Canberra VAX 3100	+	WC38624	KA235W0225	425	AEA, GEA (Backup VAX)	Ż

10
4
-
0
0
-
5
~
C
-
0
5
-

Page 4 of 4

						_				_										_		_			-				_						
	CMC?	Υ	γ	Υ	γ	Υ	γ	Υ	Υ	Υ	Υ	Υ	\star	Υ	*	Υ	Υ	Y	Υ	Υ	Υ	Υ	Υ	Υ	Υ	Υ	Υ	Υ	Υ	Υ	Υ	Υ	Υ	Υ	Υ
	Eff. %	70	70	100	40	40	X-RAY	X-RAY	60	20	100	100	30	20	40	100	100	100	100	100	100	100	100	100	100	X-RAY	Χ-RΑΥ	100	100	100	N/A	N/A	N/A	45	50
	Type	Ge	Ge	Ge	Ge	Ge	LEPS	LEPS	P-Ge	Ge	Ge	Ge	Ge	Ge	6 e	Ge	Ge	Ge	Ge	Ge	Ge Well	Ge Well	Ge Well	Ge-AC	Ge-AC										
	Serial #	33TP40378A	2542	34P40535A	9923007	6028	33TE187	33TE190	3612	44TP21884A	9722	9716	24P80V	35TP21056A	6166	10182	10188	10181	10101	10178	10150	10129	10104	10145	10212	20M102	6068240	35-P-40642A	34-P-40481A	34-P-40522A	34-P-30914B	34-P-40544A	11051527	54P33125A	5945022
Gamma Detectors	Model	GEM70200P	IGC7022SD	GEM100210	GC4020	CPVS3040195	GLP3638510P	GLP3638510P	NIGC6022SD	GEM20P-Plus	GC10021	GC10021	GEM30210	GEM20180	GC4018	GC10021	101351600S	GL2015R	GEM100210	GEM100210	GEM100210	GEM170230-S	GEM170230-S	GCW2523-S	GEM4276LBCHJ	GX5020									
List of	Vendor	EG&G Ortec	Princeton	EG&G Ortec	Canberra	Tennelec	EG&G Ortec	EG&G Ortec	Princeton	EG&G Ortec	Canberra	Canberra	EG&G Ortee	EG&G Ortec	Canberra	Canberra	Ortec	Canberra	Ortec	Ortec	Ortec	Ortec	Ortec	Canberra	Ortec	Canberra									
	Property Tag	WD25157	WD06582	WD12833	WD35719	PT00941	WD12076	WD12075	WD06581	WD36413	WD73130	WD73129	WD59508	WD81868	PT27133	PT16418	PT16419	PT16420	PT16430	PT16429	PT16428	PT16394	PT16402	PT16401	PT16410	WB61769	WD48418	WD81452	WD81454	WD81453	WD81450	WD81449	WD81451	PT27479	WD56318
	Name	C	D	Щ	IJ	Н	I	ſ	К	Γ	М	Z	а.	F	1 *	ц	0	R	S	N	>	W	×	γ	Ζ	LEPS1	LEPS2	MA	WB	WC	Well1	Well2	Well3	AC1	AC2

Updated 03/1

Distribution

No. of	No. of
<u>Copies</u>	<u>Copies</u>

3 Washington River Protection Solutions

ST Arm	(PDF)
KA Colosi	(PDF)
MR Landon	(PDF)
JG Reynolds	(PDF)

10 Pacific Northwest National Laboratory

	SK Fiskum	(PDF)
	JR Allred	(PDF)
]	HA Colburn	(PDF)
]	M Fountain	(PDF)
]	RA Peterson	(PDF)
	AM Rovira	(PDF)
	SN Schlahta	(PDF)
]	MR Smoot	(PDF)
]	DM Wellman	(PDF)
]	Information Release	(PDF)





Proudly Operated by Battelle Since 1965

902 Battelle Boulevard P.O. Box 999 Richland, WA 99352 1-888-375-PNNL (7665)

www.pnnl.gov