

PNNL-28958, Rev. 0 RPT-DFTP-013, Rev. 0

Cesium Ion Exchange Testing Using a Three-Column System with Crystalline Silicotitanate and Hanford Tank Waste 241-AP-107

August 2019

SK Fiskum AM Rovira HA Colburn AM Carney RA Peterson



Prepared for the U.S. Department of Energy under Contract DE-AC05-76RL01830

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor Battelle Memorial Institute, nor any of their employees, makes **any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights**. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or Battelle Memorial Institute. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

PACIFIC NORTHWEST NATIONAL LABORATORY operated by BATTELLE for the UNITED STATES DEPARTMENT OF ENERGY under Contract DE-AC05-76RL01830

Printed in the United States of America

Available to DOE and DOE contractors from the Office of Scientific and Technical Information, P.O. Box 62, Oak Ridge, TN 37831-0062; ph: (865) 576-8401 fax: (865) 576-5728 email: reports@adonis.osti.gov

Available to the public from the National Technical Information Service 5301 Shawnee Rd., Alexandria, VA 22312 ph: (800) 553-NTIS (6847) email: orders@ntis.gov <<u>https://www.ntis.gov/about</u>> Online ordering: <u>http://www.ntis.gov</u>

PNNL-28958, Rev. 0 RPT-DFTP-013, Rev. 0

Cesium Ion Exchange Testing Using a Three-Column System with Crystalline Silicotitanate and Hanford Tank Waste 241-AP-107

August 2019

SK Fiskum AM Rovira HA Colburn AM Carney RA Peterson

Prepared for the U.S. Department of Energy under Contract DE-AC05-76RL01830

Pacific Northwest National Laboratory Richland, Washington 99354

Executive Summary

Washington River Protection Solutions (WRPS) is working to support initial production of immobilized low-activity waste by feeding Hanford tank supernate from tank farms to the Hanford Waste Treatment and Immobilization Plant (WTP) Low-Activity Waste (LAW) facility. This goal incorporates the design of a Tank-Side Cesium Removal (TSCR) system, which in effect filters tank waste supernate and processes it through a cesium ion exchange medium to remove ¹³⁷Cs. The ¹³⁷Cs-depleted product is sent to the WTP for vitrification. Rovira et al. (2018)¹ reported the successful Cs removal from Hanford tank 241-AP-107 waste (hereafter called AP-107) using crystalline silicotitanate (CST) ion exchange media manufactured by Honeywell UOP, LLC (product IONSIVTM R9140-B, Batch 2081000057)² in a lead-lag column system. Since that testing, WRPS has changed the column design to a three-column system. WRPS requested a repeated study with AP-107 tank waste using a different batch of IONSIVTM R9140-B, Lot 2002009604, in a three-column format.

A small-scale test platform was constructed to demonstrate solids filtration, Cs removal, and LAW vitrification and installed at Pacific Northwest National Laboratory. This report describes the small-scale ion exchange testing system component of the test platform and the processing of approximately 8.55 L of filtered AP-107 supernate using CST, Lot 2002009604, that passed through a 25-mesh filter. This report also describes the Cs ion exchange batch contact testing to determine the Cs distribution coefficient and the maximum Cs loading capacity in the AP-107 tank waste matrix.

Batch contact testing helps to evaluate CST performance on tank waste supernate and is often used prior to processing tank waste through ion exchange columns to assess some performance metrics (e.g., 50% Cs breakthrough). Batch contact tests were performed with the filtered AP-107 tank waste at four Cs concentrations at a phase ratio of 200 (liquid volume to exchanger mass). The distribution coefficient (K_d) at the equilibrium condition of 9.2 μ g Cs/mL (AP-107 feed condition) was determined to be 806 mL AP-107/g CST. With a CST bed density of 1.01 g/mL (<25 mesh CST), this K_d corresponded to a predicted 50% Cs breakthrough of 814 bed volumes (BVs). The Cs load capacity at the equilibrium feed condition was determined to be 8.3 mg Cs/g dry CST.

The three-column format testing was prototypic to the intended TSCR operations, albeit on a small scale with 10-mL CST beds. Flowrate was adjusted to match the CST contact time expected for the full-scale operation, i.e., matched bed volumes per hour (BV/h) flow rate. The feed was processed downflow through the lead column, then through the middle column, and then through the polish column. Loading continued at an average of 1.88 BV/h (1.83 to 2.04 BV/h range) until the entire available AP-107 feed was processed. The Cs-decontaminated product will be provided for vitrification testing.

The lead column only reached 45% Cs breakthrough after processing 855 BVs of feed. The 50% Cs breakthrough was extrapolated to occur at ~900 BVs. This extrapolated 50% Cs breakthrough value differed from the batch contact estimate (814 BVs) by 11%. Given the extrapolation from column processing and the overall measurement uncertainties, the agreement within 11% was considered reasonable.

¹ Rovira AM, SK Fiskum, HA Colburn, JR Allred, MR Smoot, and RA Peterson. 2018. *Cesium Ion Exchange Testing Using Crystalline Silicotitanate with Hanford Tank Waste 241-AP-107*. PNNL-27706. Pacific Northwest National Laboratory, Richland, Washington.

² The term "Batch" was provided by UOP in paperwork accompanying the CST provided to PNNL February 2018; this term was used on earlier test reports. The term "Batch" and "Lot" may be used interchangeably. For consistency with previous reports, the term "Batch" is used herein.

The waste acceptance criteria (WAC) for the WTP LAW vitrification facility is <3.18E-5 Ci ¹³⁷Cs per mole Na.¹ For the AP-107 tank waste, only 0.127% of the influent ¹³⁷Cs concentration may be delivered to the WTP; this required a Cs decontamination factor of 787. The Cs effluent from the middle column reached the WAC after processing 590 BVs. Cs breakthrough from the polish column began at 590 BVs, reaching 6.3E-3 μ Ci/mL or 4.3E-3% Cs breakthrough after processing 844 BVs of feed. Table ES.1 summarizes the observed column performance and relevant Cs loading characteristics.

	WACLimit	Extrapolated 50%	¹³⁷ Ca Londod	Caloadad (mg/g
	WAC LIIIII	es breakinough	Cs Loaded	Cs Loaded (Ing/g
Column	Breakthrough (BVs)	(BVs)	(µCi)	CST)
Lead	200	900 ^(a)	1.08E+6	6.82
Middle	590	NA	1.52E+5	0.96
Polish	1010 ^(a)	NA	2.75E+5	0.017
$(\cdot) \mathbf{E} + \cdots + 1$	4.1.1.			

Table	ES.1. A	P-107 0	Column	Performance	Summarv	with	CST
1 4010	L D.11.11	107	Joranni	I UII UII UII UII UII U	Sammary	** 1011	~~ 1

(a) Extrapolated value.

BV = bed volume, 10 mL

The time weighted average flowrate was 1.88 BV/h.

The Cs capacity for IONSIVTM R9140-B Lot 2002009604 was superior to that of Batch 2081000057. Column testing with AP-107 tank waste showed that the Cs breakthrough profile for Lot 2002009604 was delayed about ~100 to 200 BVs (0% to 25% Cs breakthrough range) relative to that of Batch 2081000057. The total Cs load capacity determined from batch contact testing was 44% higher for Lot 2002009604 (0.72 vs. 0.50 mmoles Cs/g CST). The achieved Cs load capacity at the AP-107 feed condition was 9% greater for Lot 2002009604 (0.062 vs. 0.057 mmoles Cs/g CST).

The AP-107 composite feed and composite effluent were characterized to understand the fractionation of selected metals and radionuclides. Concentrations and recoveries of the selected analytes are summarized in Table ES.2; those with low recovery were assumed to be adsorbed onto CST. Six Resource Conservation and Recovery Act (RCRA) elements were in the suite of analytes of interest for this study (Ag, As, Ba, Cd, Cr and Pb). Cd and Cr partitioned mostly to the effluent. Lead (Pb) and barium (Ba) were detected in the feed (with concentration errors likely to exceed 15%) but were below the method detection limit (MDL) in the effluent; this was indicative of uptake by the CST. The Ag and As were not detected in the feed or effluent and their fractionations could not be discerned. In addition to Cs removal, large fractions of calcium (Ca), uranium (U), ⁹⁰Sr, ²³⁷Np, and ²³⁹⁻²⁴⁰Pu also significantly fractionated to the CST.

¹ From *ICD 30 – Interface Control Document for Direct LAW Feed*, 24590-WTP-ICD-MG-01-030, Rev. 1, 2017, Bechtel National, Inc. (River Protection Project Waste Treatment Plant), Richland, Washington.

	Analyte	Feed Concentration (M)	Effluent Concentration (M)	Fraction in Effluent
	Ag	<4.4E-6	<4.4E-6	
	Al	3.71E-01	3.61E-01	98%
	As	<1.8E-4	<1.8E-4	
	Ba	[3.0E-6]	<5.7E-7	
	Ca	[8.6E-4]	[4.5E-4]	[52%]
	Cd	5.79E-05	[4.7E-5]	[81%]
	Cr	1.01E-02	9.90E-03	98%
	Fe	2.64E-04	[2.2E-4]	[83%]
Metals /	К	1.20E-01	1.15E-01	97%
Non-metals	Na	5.97E+00	5.71E+00	96%
	Р	2.82E-02	2.67E-02	95%
	Pb	[3.9E-5]	<1.9E-5	
	S	6.03E-02	5.75E-02	96%
	Sr	[1.5E-6]	<1.4E-6	
	Ti	<2.3E-6	[3.9E-6]	
	U	7.51E-5	4.53E-5	61%
	Zn	<1.6E-5	[2.0E-5]	
	Zr	<5.2E-6	[2.3E-5]	
	Analyte	Feed Concentration (µCi/mL)	Effluent Concentration (µCi/mL)	Fraction in Effluent
	⁹⁰ Sr	4.52E-01	<1E-3	<0.2%
	⁹⁹ Tc	1.00E-01	9.82E-02	99%
	¹³⁷ Cs	1.54E+02	5.29E-04	0.00035%
Radionuclides	²³⁷ Np	4.76E-05	2.02E-05	43%
	²³⁸ Pu	7.54E-05	2.75E-05	37%
	²³⁹⁺²⁴⁰ Pu	5.60E-04	2.01E-04	36%

Table ES.2. Recoveries of Analytes of Interest in the AP-107 Effluent

Notes:

"<" values were < MDL, sample-specific MDL provided. "--" indicates effluent recovery could not be calculated.

Values in brackets [] were \geq MDL but < EQL, with errors likely to exceed 15%.

EQL = estimated quantitation limit.

MDL = method detection limit

Acknowledgments

The authors thank the Shielded Facilities Operations staff Jarrod Turner and Michael Rojas for in-cell sample handling and systems manipulations. We thank the Analytical Support Operations (ASO) count room staff Mike Cantaloub for rapid ¹³⁷Cs analysis for the column load and batch contact sample analyses. We thank the ASO staff Karl Pool, Lori Darnell, Jenn Carter, Denis Cherkasov, Sam Morrison, Andrew Carney, Crystal Rutherford, and Chuck Soderquist for the sample analysis, data processing, and reporting. We thank Renee Russell for reviewing the calculation files, test data packages, and this technical report. We also thank Matt Wilburn and Bill Dey for technical editing and the quality review, respectively, of this report.

Acronyms and Abbreviations

AEA	alpha energy analysis
ASO	Analytical Support Operations
ASR	Analytical Service Request
AV	apparatus volume
BV	bed volume
CST	crystalline silicotitanate
DF	decontamination factor
DI	deionized (water)
EQL	estimated quantitation limit
FD	feed displacement
GEA	gamma energy analysis
IC	ion chromatography
ICP-MS	inductively coupled plasma mass spectrometry
ICP-OES	inductively coupled plasma optical emission spectrometry
ID	identification (number)
LAW	low-activity waste
LSC	liquid scintillation
MDL	method detection limit
PNNL	Pacific Northwest National Laboratory
PSD	particle size distribution
PTFE	polytetrafluoroethylene
QA	quality assurance
QC	quality control
R&D	research and development
RCRA	Resource Conservation and Recovery Act
SEM	scanning electron microscopy
SV	system volume
TIC	total inorganic carbon
TOC	total organic carbon
TSCR	Tank-Side Cesium Removal
WAC	waste acceptance criteria
WRPS	Washington River Protection Solutions
WTP	Hanford Waste Treatment and Immobilization Plant
WWFTP	WRPS Waste Form Testing Program

Contents

Execut	ive Sum	mary		ii			
Ackno	wledgme	ents		v			
Acrony	yms and	Abbrevia	ations	vi			
Conter	nts			vii			
1.0	Introdu	ction		1.1			
2.0	Quality	Assurar	nce	2.1			
3.0	Test Co	onditions					
	3.1	CST M	edia				
	3.2	AP-107	Tank Waste Sample				
	3.3	Batch C	Contact Conditions				
	3.4	Ion Exc	hange Process Testing				
		3.4.1	Ion Exchange Column System				
		3.4.2	AP-107 Tank Waste Process Conditions				
	3.5	Sample	Analysis				
4.0	Batch (Contact F	Results	4.1			
5.0	Colum	n Process	sing	5.1			
	5.1	Cs Load	ding for AP-107, Feed Displacement, and Water Rinse				
	5.2	Cesium Activity Balance					
	5.3	WAC Limit					
	5.4	Transiti	on Zone				
	5.5	Chemic	al and Radiochemical Composition				
6.0	Conclu	sions		6.1			
	6.1	Batch C	Contact Testing	6.1			
	6.2	Column	1 Testing				
	6.3	Analyte	Fractionation				
7.0	Referen	nces		7.1			
Appen	dix A – 0	Column I	Load Data	A.1			
Appen	dix B – A	Analyte C	Concentrations As a Function of Loading	B.1			
Appen	dix C – S	Solids Ch	naracterization	C.1			
Appen	dix D – 0	CST Lot	Comparison	D.1			
Appen	dix E – A	Analytica	l Reports	E.1			

Figures

Figure 3.1. PSD Comparison for Two CST Lots	3.2
Figure 3.2. Ion Exchange System Schematic	3.7
Figure 3.3. Photograph of Ion Exchange System Outside of the Hot Cell	3.7
Figure 3.4. Closeup of Lead Chromaflex® Column Loaded with CST	3.9
Figure 3.5. Ion Exchange Assembly in the Hot Cell, May 1, 2019, ~744 BVs Processed	3.10
Figure 3.6. AP-107 Flowrate as a Function of Time	3.11
Figure 3.7. Suspended Solids at the Top of the Lead Column During Backflow Processing	3.12
Figure 4.1. Equilibrium Cs K _d Curves for AP-107 with Two Lots of CST	4.2
Figure 4.2. Isotherm for AP-107 Tank Waste with Two Lots of CST	4.3
Figure 5.1. Lead, Middle, and Polish Column Cs Load Profiles of AP-107 at 1.88 BV/h, Linear- Linear Plot	5.2
Figure 5.2. Lead, Middle, and Polish Column Cs Load Profiles of AP-107 at 1.88 BV/h, Probability-Log Plot	5.2
Figure 5.3. Lead Column Extrapolated to 50% Breakthrough	5.3
Figure 5.4. Polish Column Extrapolated to WAC Limit	5.4
Figure 5.5. Polish Column Cs Load Profile with Feed Displacement, Water Rinse, and Column Flush Solution	5.5
Figure 5.6. AP-107 (Two Tests) and AW-102 Load Profiles Comparisons	5.6
Figure 5.7. Cs Load Profile Comparisons at ~1.85 BV/h	5.7
Figure 5.8. Cs Load Comparison at ~0.65 SV/h	5.8
Figure 5.9. System Volume to WAC Limit vs. Flowrate	5.11
Figure 5.10. Al, Ca, Fe, and U Load Profiles onto the Lead Column	5.17
Figure 5.11. RCRA Metal Recoveries during Lead Column Processing; Ag, As, Ba Were Not	c 17
	5.17
Figure 5.12. ⁹⁰ Sr, ¹³ /Cs, ²³ /Np, and ²³⁹⁺²⁴⁰ Pu Load Profiles onto the Lead Column	5.18

Tables

Table 3.1. Physical Properties of <25 Mesh, Washed R9140-B CST (Two Production Batches)
Table 3.2. Characterization of Samples 7AP-18-01 and 7AP-18-36 Collected from Hanford Tank AP-107 January 2019 (ASR 0747)
Table 3.3. Cs Concentrations in Stock Contact Solutions 3.4
Table 3.4. Experimental Conditions for AP-107 Column Processing, April 15 to May 4, 2019
Table 3.5. Analytical Scope for Column Process, ASR 0787
Table 4.1. Equilibrium Results for AP-107 Batch Contact Samples with CST Lot 2002009604 4.1
Table 4.2. α_i and β Parameter Summary
Table 5.1. ¹³⁷ Cs Activity Balance for AP-107 5.9
Table 5.2. Cs Loading onto CST Comparison 5.10
Table 5.3. Bed Volumes Processed to Reach WAC Limit
Table 5.4. Transition Zone Comparison 5.12
Table 5.5. AP-107 Feed and Effluent Radionuclide Concentrations and Fractionations (ASR 0787)
Table 5.6. AP-107 Feed and Effluent Inorganic Analyte Concentrations and Fractionation (ASR 0787)
Table 5.7. AP-107 Feed Free Hydroxide and Carbon Composition (ASR 0787)

1.0 Introduction

The U.S. Department of Energy (DOE) is working to expedite processing of Hanford tank waste supernate at the Hanford Waste Treatment and Immobilization Plant (WTP). To support this goal, Washington River Protection Solutions (WRPS) is designing a system for suspended solids and cesium (Cs/¹³⁷Cs) removal from Hanford tank waste supernate. The effluent will then be sent to the WTP Low-Activity Waste (LAW) facility for vitrification. The Cs removal is critical for eliminating the high dose rate associated with ¹³⁷Cs and facilitating a contact maintenance philosophy for the LAW Facility. The maximum ¹³⁷Cs concentration in the LAW sent to the WTP is targeted to be below 3.18E-5 Ci ¹³⁷Cs/mole of Na waste acceptance criteria (WAC) limit.¹ The filtration and ion exchange systems will be placed near the Hanford tanks and are collectively termed the Tank-Side Cesium Removal (TSCR) system. The ion exchange media selected for Cs removal at TSCR is crystalline silicotitanate (CST) that is manufactured in a nearly spherical form by Honeywell UOP LLC (UOP; Des Plaines, IL) as product IONSIVTM

R9140-B.

Recent testing with CST was reported in support of TSCR design inputs with 5.6 M Na simulant using two different CST production lots, Batch² 2081000057 and Lot number 2002009604 (Fiskum et al. 2018; 2019, respectively). Fiskum et al. (2018) evaluated CST Batch 2081000057 for physical properties, Cs capacity from batch contacts, and processing through lead-lag column systems (10-mL beds) at three different flowrates. The authors found that increasing the flowrate decreased the Cs loading; the Cs capacity and 50% breakthrough predicted by batch contact testing [6.16 mg Cs per gram of CST and 770 bed volumes (BVs), respectively] agreed within 10% of the results of the column tests (~6.3 mg Cs per gram of CST and 800 BVs). Fiskum et al. (2019) also tested CST Lot 2002009604 for CST physical properties, Cs kinetic exchange performance from batch contact testing, and column processing. The column processing evaluated Cs loading at full-height lead/lag columns and individual smaller-scale columns (27-mL beds) at a variety of flow rates and feed conditions. Long Cs exchange transition zones were reported (1120 BVs from 5% to 95% Cs breakthrough), indicating slow Cs kinetic exchange. Additionally, CST Lot 2002009604 was shown to load Cs more effectively during column processing (6.95 mg Cs/g CST and ~1000 BVs at 50% breakthrough) than CST Batch 2081000057. Appendix D provides a summary comparison of the physical properties and ion exchange performances of these two CST production lots.

Testing of cesium (¹³⁷Cs) removal from defense wastes using CST has been previously reported (King 2007; Walker et al. 1998; Hendrickson et al. 1996; Brown et al. 1996). However, limited data are available on actual Hanford tank wastes processed with more recent CST production lots. Rovira et al. (2018) described testing with CST IONSIVTM R9140-B, Batch 2081000057, on a Hanford tank waste sample collected from 241-AP-107 (hereafter called AP-107). They assessed column loading using a lead/lag column format, processing at 2.2 bed volumes per hour (BV/h). Column testing was conducted such that the lead column was removed after ~25% Cs breakthrough (470 BVs) because the lag column effluent had reached the WAC limit; the lag column was placed in the lead position and a fresh lag column was installed. Processing continued, but the lead column only reached 26% Cs breakthrough before feed was exhausted and lag column effluent again exceeded the WAC. Rovira et al. (2018) also

¹ From *ICD 30 – Interface Control Document for Direct LAW Feed*, 24590-WTP-ICD-MG-01-030, Rev. 1, 2017, Bechtel National, Inc. (River Protection Project Waste Treatment Plant), Richland, Washington.

² The term "Batch" was provided by UOP in paperwork accompanying the CST provided to PNNL February 2018; this term was used on earlier test reports. The term "Batch" and "Lot" may be used interchangeably. For consistency with previous reports, the term "Batch" is used herein.

reported the effective Cs load capacity at equilibrium feed condition was 7.2 mg Cs per g CST based on batch contact testing.

Rovira et al. (2019) tested Cs loading using the newer CST production Lot 2002009604 with AW-102 tank waste that was diluted to 5.8 M Na. They again used 10-mL CST beds in a lead/lag format. In their tests, only 4.3% Cs breakthrough (451 BVs) was obtained during column testing before depleting the available feed. Batch contact testing with the diluted and filtered AW-102 showed that the effective Cs capacity at the equilibrium feed condition was 6.8 mg Cs per g CST. Matrix effects such as higher K concentration in AW-102 (0.15 M K) relative to AP-107 (0.096 M K) possibly reduced the Cs load capacity (from 7.2 to 6.8 mg Cs per g CST).

WRPS funded Pacific Northwest National Laboratory (PNNL) to conduct additional testing with AP-107 tank waste. The goals of this current study are as follows.

- 1. Evaluate the effectiveness of Cs removal from AP-107 using a three-column ion exchange system (lead, middle, and polish) with the newer production batch of CST, Lot 2002009604.
- 2. Determine the Cs load capacity of CST, Lot 2002009604, using batch contact testing with AP-107.
- 3. Compare the AP-107 Cs load profile to the previously reported AP-107 load curve (Rovira et al. 2018).
- 4. Analyze the AP-107 effluent to derive the fates of key analytes (⁹⁰Sr, ⁹⁹Tc, ¹³⁷Cs, ²³⁹⁺²⁴⁰Pu, ²³⁷Np, Al, As, Ba, Ca, Cd, Cr, Fe, K, Na, P, Pb, S, Sr, Ti, Zn, Zr).
- 5. Provide Cs-decontaminated effluent for vitrification studies (to be conducted at a later date).
- 6. Provide Cs-loaded CST for potential follow-on analysis (to be conducted at a later date).

The efficacy of loading higher amounts of Cs onto the lead column CST while maintaining a product below the WAC limit from the polish column was of prime interest to support the evolving WRPS TSCR design. This test design further exposes the CST to higher feed volume through the individual column beds, allowing for a more representative assessment of the fractionations of analytes of interest.

2.0 Quality Assurance

The work described in this report was conducted with funding from WRPS contract 36437/273, *DFLAW Radioactive Waste Test Platform*. This contract was managed under PNNL Project 73312. All research and development (R&D) work at PNNL is performed in accordance with PNNL's Laboratory-Level Quality Management Program, which is based on a graded application of NQA-1-2000, *Quality Assurance Requirements for Nuclear Facility Applications* (ASME 2000), to R&D activities. To ensure that all client quality assurance (QA) expectations were addressed, the QA controls of the PNNL's WRPS Waste Form Testing Program (WWFTP) QA program were also implemented for this work. The WWFTP QA program implements the requirements of NQA-1-2008, *Quality Assurance Requirements for Nuclear Facility Applications* (ASME 2008), and NQA-1a-2009, *Addenda to ASME NQA-1-2008* (ASME 2009), and consists of the WWFTP Quality Assurance Plan (QA-WWFTP-001) and associated QA-NSLW-numbered procedures that provide detailed instructions for implementing NQA-1 requirements for R&D work.

The work described in this report was assigned the technology level "Applied Research" and was planned, performed, documented, and reported in accordance with procedure QA-NSLW-1102, Scientific Investigation for Applied Research. All staff members contributing to the work received proper technical and QA training prior to performing quality-affecting work.

3.0 Test Conditions

This section describes the CST media, AP-107 tank waste, batch contact conditions, and column ion exchange conditions. All testing was conducted in accordance with a test plan prepared by PNNL and approved by WRPS.¹

3.1 CST Media

WRPS purchased ten 5-gallon buckets (149 kg total) of IONSIV TM R9140-B², Lot number 2002009604, material number 8056202-999, from Honeywell UOP LLC (Des Plaines, IL). This CST production lot was screened by the manufacturer to achieve an 18 x 50 mesh size product. The product was requested to be delivered to WRPS in a series of 5-gallon buckets (as opposed to a 50-gallon drum) to aid in material distribution, handling, and sampling at PNNL. The CST was transferred from WRPS to PNNL on September 20, 2018, under chain of custody. Once received, the CST was maintained at PNNL in environmentally controlled spaces. One of the 5-gallon buckets of CST was delivered to the PNNL Radiochemical Processing Laboratory. The handling and splitting of the CST were previously described (Fiskum et al. 2019). A 117-g subsample split was passed through a 25-mesh sieve (ASTM E11 specification) as previously described (Fiskum et al. 2019). Of this starting mass, 61.6 g or 52 wt% passed through the sieve and was collected for batch contact testing and column testing; this was close to the 59 wt% <25 mesh mass fraction reported by Fiskum et al. (2019). The <25 mesh CST fraction was pretreated by contacting with 200 mL of 0.1 M NaOH three successive times. The 0.1 M NaOH rinse solution and colloidal fines from the CST were decanted. The rinsed CST was maintained with an overburden of 0.1 M NaOH.

Table 3.1 compares the physical properties measured on the Batch 2081000057 CST (used on the AP-107 two-column test in 2018) and the Lot 2002009604 CST (used on the current three-column test). The physical properties were essentially the same, within experimental uncertainty of a couple of percent, except for the post-sonication particle size distribution (PSD) d_{90} , which was 29 µm larger for the current test batch. Figure 3.1 shows the PSD profiles for the two CST production batches at the indicated sieve cuts. Additional comparisons between the two CST lots are provided in Appendix D.

¹ Fiskum SK. 2019. TP-DFTP-058, Rev. 0. *DFLAW Cesium Ion Exchange Testing with AP-107 Tank Waste for the Tank Side Cesium Removal (TSCR) Project*. Pacific Northwest National Laboratory, Richland Washington. Not publicly available.

 $^{^{2}}$ R9140-B is provided in the sodium form by the vendor.

Parameter	Batch 2081000057 (Fiskum et al. 2018)	Lot 2002009604 (Fiskum et al. 2019)
Sieve cut for testing	<25 mesh/+60 mesh	<25 mesh
Bulk density, g/mL	1.00	1.02
CST bed density, g/mL	1.00	1.01
Settled bed void volume, %	65.6	66.2
Cumulative particle undersize fractions, microns ^(a)	d ₁₀ : 412 d ₅₀ : 563 d ₉₀ : 765	d ₁₀ : 406 d ₅₀ : 567 d ₉₀ : 795

Table 3.1. Physical Properties of <25 Mesh, Washed R9140-B CST (Two Production Batches)

(a) Volume basis, post-sonication



Figure 3.1. PSD Comparison for Two CST Lots

Figure notes: CST Batch 2081000057 <25 mesh/ >60 mesh; Lot 2002009604 <25 mesh, Post-sonication results

3.2 AP-107 Tank Waste Sample

Multiple samples (36 each at nominally 270 mL) were collected from the AP-107 Hanford tank in January 2019. The samples were delivered to PNNL's Radiochemical Processing Laboratory and placed into the Shielded Analytical Laboratory hot cells. The first and last samples collected, 7AP-18-01 and 7AP-18-36, were subsampled for a limited analysis suite to confirm density, Na, K, Cs, ¹³³Cs, and ¹³⁷Cs concentrations. The densities were measured in-cell using 10-mL volumetric flasks. All other measurements were conducted by the Analytical Support Operations (ASO) according to Analytical Service Request (ASR) 0747; results are provided in Table 3.2. The results of the two samples agreed within 7%, indicating that the 36 samples were essentially homogenous, within analytical uncertainty (±10% to 15%). The results for the AP-107 samples collected in 2017 and used for the two-column test (Rovira et al. 2018) are also shown in Table 3.2 for direct comparison. The average results for the 2017 sample collection agreed within 3% of the average results of the 2019 collection (except for Cs).

	AP-107 Average (Rovira et al.	7AP-18-01	7AP-18-36		
Analyte	2018)	(19-1227) ^(a)	$(19-1228)^{(a)}$	Units	Analysis Method
Al	0.364	0.361	0.385	Μ	ICP-OES
К	0.097	0.093	0.099	Μ	ICP-OES
Na	5.62	5.61	5.98	Μ	ICP-OES
OH-	0.92 ^(b)	Not measured	Not measured	М	Titration
¹³³ Cs	5.63	5.96	6.14	µg/mL	ICP-MS
¹³⁷ Cs	156 ^(c)	140 ^(d)	141 ^(d)	µCi/mL	GEA
¹³⁷ Cs	1.80 ^(c)	1.61 ^(d)	1.62 ^(d)	µg/mL	GEA
Density	1.266 ^(e)	1.268 ^(f)	1.259 ^(f)	g/mL	Volumetric flask

Table 3.2. Characterization of Samples 7AP-18-01 and 7AP-18-36 Collected from Hanford Tank AP-107 January 2019 (ASR 0747)

(a) ASO sample ID, see Appendix E for ASO reports.

(b) Based on first inflection point; assumed to be the free (unbound) hydroxide.

(c) Reference date is 11/27/17.

(d) Reference date is 3/18/19.

(e) Measured at 26.6 °C using a 10-mL volumetric flask.

(f) Measured at 22.8 °C using a 10-mL volumetric flask.

GEA = gamma energy analysis; ICP-MS = inductively coupled plasma mass spectrometry;

ICP-OES = inductively coupled plasma optical emission spectrometry.

The ¹³³Cs results for the 2019 sample collection (average 6.05 μ g/mL) were 14% higher than those from the 2017 sample collection (5.30 μ g/mL); the 14% difference was within the overall analytical uncertainty. The 2019 ¹³⁷Cs results were ~5% lower than the 2018 sampling results, accounting for decay correction. The ¹³⁵Cs isotopic composition of the AP-107 samples was assumed to be equivalent to that of the 2018 measured value of 17.2 wt% (Rovira et al. 2018). The total Cs concentration (from all Cs isotopes) was thus assigned to 9.19 μ g/mL (reference date is April 12, 2019) for this report; the Rovira et al. (2018) results were based on 8.57 μ g/mL total Cs (7.2 % lower value). This difference was likely driven by analytical uncertainties and will not significantly affect the Cs load profile test conclusions or comparisons between tests. The difference may affect the calculated total Cs load capacity from column processing but not batch contact processing where ¹³³Cs is added.

The AP-107 samples were first processed through either of two dead-end filtration systems (Geeting et al. 2019). The AP-107 samples were filtered in batches and provided in multiple ~0.7- to 1.4-L increments for ion exchange. Each container of material was measured for density using 10-mL volumetric flasks or the Coriolis meter installed in the filtration apparatus. Effluent densities ranged from 1.26 to 1.28 g/mL. It was assumed that, within analytical uncertainty, all samples provided for column ion exchange processing were equivalent and the average density of 1.266 g/mL was applied. Due to the large volume (~8.6 L) of AP-107, the multiple samples were not combined into a single container for homogenization. Doing so builds risk into the process (should a container leak), and handling the massive composite would be problematic given the weight-lifting limitations of the manipulators in the hot cells.

3.3 Batch Contact Conditions

A primary metric of an ion exchanger's performance is its distribution coefficient (K_d) measured in a batch contact experiment. The distribution coefficient is a quantitative measure of a material's capability to remove an ion from a specific solution matrix and is the ratio of the concentration of the ion remaining in solution at equilibrium to the concentration of the ion sorbed on the ion exchange material.

Batch contact solutions consisted of the AP-107 tank waste plus various amounts of added ¹³³Cs as $C_{s}NO_{3}$ solution. The equilibrium Cs concentrations were determined after batch contacts to assess the effective Cs loading capacity on the CST and the Cs K_d in the AP-107 feed matrix. The preparation and batch contacts were processed in accordance with a test instruction.¹

Aliquots of Cs spike solutions (140 mg/mL or 14.0 mg/mL) were added to three centrifuge tubes in small volumes (1.06, 6.97, and 49.1 mg added Cs). The Cs-spiked centrifuge tubes were transferred to the hot cell and ~33-mL aliquots of filtered AP-107 supernate were transferred to each of the Cs spike solution containers. Exact masses transferred were determined by difference from the measured masses before and after Cs spikes and AP-107 transfers; the added volumes were calculated from the solution densities and net sample masses. The three vessels of AP-107 plus added Cs and the unspiked AP-107 are termed "stock contact solutions." The stock contact solutions were shaken to mix AP-107 thoroughly with the Cs spike. Table 3.3 shows the calculated initial Cs concentrations in the stock contact solutions. The Cs spike solutions were equilibrated with AP-107 matrix ~10 days.

	Cs Concentration	Cs Concentration
Solution ID	(mg/L)	(M)
TI062-S0	9.19	6.91E-5
TI062-S1	37.5	2.82E-4
TI062-S2	229	1.73E-3
TI062-S3	1462	1.10E-2

Table 3.3. Cs Concentrations in Stock Contact Solutions

An aliquot of the washed CST, sufficient to apply to all batch contact tests, was allowed to air-dry overnight at ambient temperature to a free-flowing form. However, the air-dried CST still contained water. The F-factor, ratio of dry mass exchanger to sampled mass exchanger, was determined to correct for water content. A small fraction of the air-dried CST was removed for nominal F-factor evaluation. The F-factor sample aliquot was dried at ~100 °C overnight to determine the nominal water content remaining in the air-dried CST. This nominal F-factor was used to determine the target CST aliquot mass to collect for the batch contact samples. This partially dried CST contained ~15% water by mass.

A precisely weighed quantity of the washed and air-dried CST (targeted to be 0.088 g "wet" and 0.075 g dry) was aliquoted into a 20-mL liquid scintillation (LSC) vial for each batch contact sample. The air-dried CST mass was determined to an accuracy of $\leq 1\%$.

Two F-factor samples were also collected, one at the beginning of CST aliquoting process, one at the end of CST aliquoting process in a tight subsampling time window (≤ 10 min). The initial F-factor sample masses were designated M_I. The F-factor samples were dried to constant mass at 100 °C. The final F-

¹ Fiskum SK. 2019. TI-DFTP-062, *Batch Contacts with Crystalline Silicotitanate in AP-107 Tank Waste Matrix—* 2019. Pacific Northwest National Laboratory, Richland, Washington. Not publicly available. Implemented April 2019.

factor sample masses were designated M_F . The F-factors were calculated according to Eq. (3.1). The average of the two F-factor samples (first and last from the weighing series, 0.8546 ±0.40% relative percent difference) was used to calculate the dry CST masses contacted with solution.

$$\frac{M_F}{M_I} = F \tag{3.1}$$

The CST aliquots were transferred to the hot cell and then contacted with 15 mL of the various stock contact solutions (see Table 3.3) in duplicate. The AP-107 volume was transferred by pipet, and the actual volume delivered was determined by mass difference and solution density. The targeted phase ratio (liquid volume to dry exchanger mass) was 200 mL/g CST. The obtained ratio varied between 179 and 198 mL/g CST.

The primary batch contact vials along with a temperature sentinel vial (15 mL of deionized [DI] water) were placed upright onto a Thermo LP vortex mixer¹ initially set to ~400 revolutions per minute. Agitation continued for 104 h; the sentinel temperature upon completion was 29.7 °C. After standing in static contact with the CST for 104 h, the duplicate set of batch contact samples were agitated at ~400 revolutions per minute for 71 h; the sentinel temperature was 28.3 °C. In each case, some cloudiness in the contact solution vial was observed, indicating a bit of particle abrasion. However, the CST beads did settle and appeared intact, indicating particle abrasion was minor.

After the batch contact time, the CST was settled and ~2 to 5 mL of the aqueous fraction was removed from the hot cell. The samples were filtered through 0.45-µm pore size nylon-membrane syringe filters. Filtered sample aliquots (0.5 to 2 mL) were collected for gamma counting; the sampled aliquot masses were measured and the exact volumes determined by dividing by the solution density.

Aliquots of the AP-107 stock solutions and batch contacted samples were analyzed by gamma spectrometry to determine ¹³⁷Cs concentrations. The batch contact Cs K_d value was determined for each sample using the relationship shown in Eq. (3.2):

$$\frac{(C_0 - C_1)}{C_1} \times \frac{V}{M \times F} = K_d$$
(3.2)

where:

- C_0 = initial ¹³⁷Cs concentration (μ Ci/mL) in the stock contact solution
- C_1 = final ¹³⁷Cs concentration (μ Ci/mL) in the batch contacted solution
- V = volume of the batch contact liquid (mL)
- M = measured mass CST (g)
- F = F-factor, mass of the dried CST divided by the mass of the sampled CST
- K_d = batch-distribution coefficient (mL/g)

Final Cs concentrations (Cs_F) were calculated relative to the 137 Cs recovered in the batch contacted samples according to Eq. (3.3):

$$Cs_0 \times \left(\frac{C_1}{C_0}\right) = Cs_F \tag{3.3}$$

¹ The Thermo LP vortex mixer was selected for hot cell use because of its small size $(15.4 \times 21.0 \times 8.3 \text{ cm})$ and small mass (3.1 kg).

where:

 Cs_0 = initial Cs concentration in solution (µg/mL or M)

 Cs_F = final Cs concentration in solution (µg/mL or M)

The equilibrium Cs concentrations loaded onto the CST (Cs_{IX} in units of mg Cs per g of dry CST mass) were calculated according to Eq. (3.4):

$$\frac{Cs_0 \times V \times \left(1 - \frac{C_1}{C_0}\right)}{M \times F \times 1000} = Cs_{IX}$$
(3.4)

where:

 Cs_{IX} = equilibrium Cs concentration in the CST (mg Cs/g CST)

 $1000 = \text{conversion factor to convert } \mu \text{g to mg}$

3.4 Ion Exchange Process Testing

This section describes the ion exchange column system and AP-107 process conditions. The preparations and column testing were conducted in accordance with a test instruction.¹

3.4.1 Ion Exchange Column System

Figure 3.2 provides a schematic of the ion exchange process system; Figure 3.3 shows a photograph of the system before installation in the hot cell. Flow through the system was controlled with a Fluid Metering Inc. (FMI) positive displacement pump. Fluid was pumped past an Ashcroft pressure gage and a Swagelok pressure relief valve with a 10-psi trigger point. The ¹/₈-inch outside diameter / ¹/₁₆-inch inside diameter polyethylene tubing was purchased from Polyconn (Plymouth, MN). The ¹/₈-inch outside diameter / ¹/₁₆-inch inside diameter stainless steel tubing was used in conjunction with the manifold. Valved quick disconnects were purchased from Cole Parmer (Vernon Hills, IL). Use of the quick disconnects were color-coded to ease correct re-installation.

The quick disconnects had the added advantage of enabling column replacement and accommodating alternate fluid processing options. At the end of the AP-107 feed processing, the lead column fluid height decreased to the top of the CST bed. Fluid replacement was possible by releasing quick disconnects m1/f1 and m2/f2 and using standby quick disconnects at positions m1 and f2 that allowed for fluid backflow into the lead column to replace the lost fluid head and expel the gas void. The original m1 and f2 quick disconnects were re-installed and fluid processing continued.

¹ Rovira AM. 2019. *Cesium Removal from AP-107 Using Crystalline Silicotitanate in a Three-Column* Format. TI-DFTP-061. Pacific Northwest National Laboratory, Richland, Washington. Not publicly available. Implemented March 2019.

PNNL-28958, Rev. 0 RPT-DFTP-013, Rev. 0



Figure 3.2. Ion Exchange System Schematic



Figure 3.3. Photograph of Ion Exchange System Outside of the Hot Cell

Chromaflex[®] column assemblies were custom-ordered from Kimble Chase (www.kimble-chase.com). Each column assembly included the column plus the standard top and bottom end fittings. Each column was made of borosilicate glass; the straight portion of the column was 9 cm tall with an inside diameter of 1.5 cm (corresponding to a CST volume of 1.77 mL/cm). The columns flared at each end to support the off-the-shelf column fittings and tubing connectors that were composed of polytetrafluoroethylene (PTFE). The CST was supported by an in-house constructed support consisting of a 200-mesh stainless steel screen tack welded onto a stainless-steel O-ring. With a rubber O-ring, the fitting was snug-fitted into place in the column (as previously described by Fiskum et al. 2018). After packing with CST, a small number of CST particles were observed to have slipped into the narrow gap between the stainless-steel support and glass column barrel; they were blocked from passage by the O-ring. The flared cavity at the bottom of each column was filled to the extent possible with 4-mm-diameter glass beads to minimize the mixing volume below the CST bed (volume reduction from ~5.6 to ~2.5 mL). An adhesive centimeter scale with 1-mm divisions (Oregon Rule Co. Oregon City, OR) was affixed to the column with the 0-point coincident with the top of the support screen.

Four Swagelok valves were installed in the valve manifold. Valve 1 was used to isolate the columns from the system (when in the closed position) and purge the tubing from the inlet to valve 1 (when placed in the sampling position). Lead column samples were collected at valve 2, the middle column samples were collected at valve 3, and the polish column samples were collected at valve 4. The gross AP-107, feed displacement (FD), water rinse, and flushed fluid effluents were collected at the effluent line. Samples of FD and water rinse were collected from valve 4.

The system was filled with water and then slightly pressurized to confirm system leak tightness. The pressure relief valve was confirmed to trigger at the manufacturer set point (10 psig). Water was removed from the columns and replaced with 0.1 M NaOH. Triplicate 10.0-mL aliquots of settled CST (pretreated, <25 mesh) were measured using a graduated cylinder and quantitatively transferred to each of the three columns. The CST was allowed to settle through the 0.1 M NaOH solution, thus mitigating gas bubble entrainment. The columns were tapped with a rubber bung until the CST height no longer changed.

The CST BV corresponded to the settled CST media volume as measured in the graduated cylinder prior to transferring the media into the ion exchange column. The reference CST BV was 10.0 mL; each of the three columns contained 10.0 mL CST. The settled CST bed heights in the columns were ~5.6 cm. This CST height corresponded to 2.4% of the full height column (234 cm).

Figure 3.4 provides a closeup image of the lead column loaded with CST, the fluid headspace, the CST bed support, and glass beads filling the void space below the bed. Note that the centimeter scale 0-point is positioned at the CST support screen and some CST particles slipped into the small gap between the column wall and the rubber screen support ring.



Figure 3.4. Closeup of Lead Chromaflex® Column Loaded with CST

The entire fluid-filled volume of the assembly was ~71 mL, inclusive of fluid in the CST beds. The bed void volume was 66% (Fiskum et al. 2019); therefore, each CST bed held 6.6 mL fluid. The fluid-filled mixing space above each CST bed averaged 5 mL (lead column 4 mL, middle column 6 mL, polish column 5 mL). The fluid mixing volume below each CST bed ranged from 2.1 to 2.6 mL. Thus, ~59% of the total fluid holdup volume was unavoidably associated with the geometry of the three columns. These fluid mixing volume fractions are not likely to be representative of plant-scale operations.

Figure 3.5 is a photograph of the ion exchange system during in-cell AP-107 processing (after processing \sim 744 BVs of AP-107). Note the lead column has a dark grey appearance whereas the companion columns appear white. The lead column darkening increased over time due to radiolytic damage from the ¹³⁷Cs on the glass column.



Figure 3.5. Ion Exchange Assembly in the Hot Cell, May 1, 2019, ~744 BVs Processed

3.4.2 AP-107 Tank Waste Process Conditions

Once installed in the hot cell, a flow of 0.1 M NaOH was used to verify the system integrity and calibrate the pump. The various containers from the filtration process (Geeting et al. 2019) were incrementally transferred to a 2-L rectangular polyethylene bottle during the loading phase. The 2-L rectangular bottle was less top-heavy than the tall 1.5-L bottles and offered more volume reserve for unattended operations. The AP-107 feed was processed downflow through the ion exchange media beds, lead to middle to polish. Effluent was collected in ~1.5-L increments. The volume limitation was intended to minimize the impact of potential polish column Cs breakthrough exceeding the WAC and to allow for safe transfer out of cell in 1.5-L polyethylene bottles. After the AP-107 loading was completed, 12 BVs of 0.1 M NaOH FD followed by 12 BVs of DI water rinse were passed downflow through the system to rinse residual feed out of the columns and process lines. The 12 BVs was equivalent to 1.7 times the fluid-filled apparatus volume (AV).

All processing was conducted at ambient cell temperature conditions, nominally 24 to 29 °C. Test parameters, including process volumes, flowrates, and CST contact times, are summarized in Table 3.4. The pump head stroke length was close to the minimum at which it could be set. The stroke rate was toggled between 12.9 and 13.0 (maximum fidelity of 0.1 units) to maintain the flowrate between 1.83 and 2.04 BV/h, respectively. Greater fidelity with the stroke rate controller could not be obtained to center on the target 1.9 BV/h. Figure 3.6 shows the achieved flowrate as a function of time. The 1.88 BV/h flowrate reported in Table 3.4 is a time-weighted average.

			Volume		Flo	wrate	Duration
Process Step	Solution	(BV)	(AV)	(mL)	(BV/h) ^(c)	(mL/min) ^(c)	(h)
Loading lead column	AP-107	855	NA	8550	1.88	0.313	452
Loading middle column ^(a)	AP-107	849	NA	8490	1.88	0.313	452
Loading polish column ^(b)	AP-107	844	NA	8440	1.88	0.313	452
Feed displacement	0.1 M NaOH	12.8	1.79	128	3.66	0.610	3.5
Water rinse	DI water	12.3	1.72	123	4.54	0.757	2.8

Table 3.4. Experimental Conditions for AP-107 Column Processing, April 15 to May 4, 2019

(a) The feed volume through the middle column was reduced relative to that of the lead column because samples collected from the lead column did not enter the middle column.

(b) The feed volume through the polish column was reduced relative to that of the middle column because samples from the lead and middle columns did not enter the polish column.

(c) Time-weighted average flowrates.

BV = bed volume (10.0 mL as measured in graduated cylinder).

AV = apparatus volume (nominally 71 mL).

NA = not applicable.



Figure 3.6. AP-107 Flowrate as a Function of Time

The total cumulative volume of AP-107 processed was 8.55 L (855 BVs). The AP-107 process cycle mimicked, as best as possible, the process flow anticipated at the TSCR facility in terms of BV/h (i.e., contact time), FD, and water rinse. It was understood that the feed linear flow velocity in this small-column configuration (0.18 cm/min) could not begin to match that of the full-height processing configuration (7.3 cm/min, Fiskum et al. 2019).

During the loading phase, nominal 2-mL samples were collected from the lead, middle, and polish columns at the sample collection ports (see Figure 3.2, valves 2, 3, and 4). Sampling from the columns necessitated brief (10-minute) interruptions of flow to the downstream columns. Samples were collected after the first ~12 BVs were processed and again at nominal 20- to 57-BV increments. The feed was processed for 452 h nearly continuously. Only brief (~5 min) interruptions were associated with changing the feed bottle near the end of feed processing.

At the end of the AP-107 loading phase, the fluid level in the lead column dropped to the height of the

CST bed. Approximately 5 mL of 0.1 M NaOH was back flowed into the lead column from quick disconnect *m2* (see Figure 3.2) to re-establish the headspace fluid such that the CST would not dry as the FD processing began. During backflow processing, brown/black solids were observed at the top of the CST bed. They were briefly suspended with the backflow processing, whereas the CST did not move. Figure 3.7 is an image of the suspended solids observed during the backflow operation. A small sample of the solids was later collected for transmission electron microscopy analysis from the parent bottle where they were found to originate. Results are discussed in Appendix C.

The FD effluent was collected in a polyethylene bottle; four samples were collected during the FD process from valve 4 in nominal 3-BV increments. The water rinse was collected similarly to the FD. The fluid-filled volume was expelled with compressed air connected at the first quick disconnect in the system f0 (see Figure 3.2); the collected volume did include the interstitial fluid space between the CST beads, but did not include fluid in the CST pore space. Hours of additional gas flow were required to dry the CST enough to be free-flowing and effectively pour out of the columns for later use.



Figure 3.7. Suspended Solids at the Top of the Lead Column During Backflow Processing

Cesium load performance was determined from the ¹³⁷Cs measured in the collected samples relative to the native ¹³⁷Cs in AP-107 feed. The collected samples were analyzed directly to determine the ¹³⁷Cs concentration using GEA. Cesium loading breakthrough curves for both the lead and lag columns were generated based on the feed ¹³⁷Cs concentration (C₀) and the effluent Cs concentration (C) in terms of % C/C_0 .

3.5 Sample Analysis

A composite feed sample was prepared by sampling 2 mL from each filtered sample bottle into one polyethylene vial. An effluent composite sample was generated by collecting a pro-rated volume from each effluent bottle and combining in a polyethylene vial. Selected effluent samples from the lead column were measured for selected radionuclides and cations in an effort to assess the exchange behavior for these analytes. Table 3.5 summarizes the specific sample collections and targeted analytes along with the cross references to the ASO sample identification numbers (IDs).

Samples were submitted to the ASO on ASR 0787. The ASO was responsible for the preparation and analysis of appropriate analytical batch and instrument quality control samples and for providing any additional processing to the sub-samples that might be required (e.g., acid digestion, radiochemical

separations, dilutions). All analyses were conducted by the ASO according to standard operating procedures, the ASO QA Plan, and the ASR. Samples were analyzed directly (no preparation) by GEA.

Sample ID	ASO Sample ID	Analysis Scope
TI061-Comp-FEED	19-1773	GEA (⁶⁰ Co, ¹³⁷ Cs, ¹⁵⁴ Eu) IC anions (F ⁻ , Cl ⁻ , NO ₂ ⁻ , NO ₃ ⁻ , PO4 ³⁻ , C ₂ O4 ²⁻ , SO4 ²⁻) ^(a) TOC/TIC free OH ICP-OES (Ag, Al, As, Ba, Ca, Cd, Cr, Fe, K, Na, Ni, P, Pb, S, Sr, Ti, Zn, Zr) ICP-MS (²³⁸ U) Radioanalytical (⁹⁰ Sr, ⁹⁹ Tc, ²³⁷ Np, ²³⁸ Pu, ²³⁹⁺²⁴⁰ Pu)
TI061-Comp-EFF	19-1774	GEA (⁶⁰ Co, ¹³⁷ Cs, ¹⁵⁴ Eu) ICP-OES (Ag, Al, As, B, Ba, Ca, Cd, Cr, Fe, K, Li, Mg, Mo, Na, Ni, P, Pb, S, Si, Sr, Ti, W, Zn, Zr) ICP-MS (²³⁸ U) Radioanalytical (⁹⁰ Sr, ⁹⁹ Tc, ²³⁷ Np, ²³⁸ Pu, ²³⁹⁺²⁴⁰ Pu)
TI061-L-F1-A (12.6 BVs)	19-1775	
TI061-L-F4-A (87.7 BVs)	19-1776	
TI061-L-F8-A (180 BVs)	19-1777	
TI061-L-F11-A (280 BVs)	19-1778	
TI061-L-F14-A (366 BVs)	19-1779	ICP-OES (Al, Ba, Ca, Cd, Cr, Fe, K, Na, Ni, P, Pb, S, Sr, Ti, Zn, Zr),
TI061-L-F18A (457 BVs)	19-1780	Radioanalytical (90 Sr. 237 Nn. $^{239+240}$ Pu)
TI061-L-F22-A (558 BVs)	19-1781	
TI061-L-F25-A (656 BVs)	19-1782	
TI061-L-F27-A (724 BVs)	19-1783	
TI061-L-F29-A (815 BVs)	19-1784	
GEA – gamma energy analysis		

Table 3.5. Analytical Scope for Column Process, ASR 0787

rgy - В

IC = ion chromatography

ICP-MS = inductively coupled plasma mass spectrometry

ICP-OES = inductively coupled plasma optical emission spectrometry

TIC = total inorganic carbon

TOC = total organic carbon

(a) The IC anion concentrations will be provided in a later report supporting the Cs-decontaminated AP-107 vitrification testing.

4.0 Batch Contact Results

This section discusses the batch contact results for the AP-107 filtered tank waste with <25 mesh CST Lot 2002009604.

Equilibrium Cs concentrations and K_d results for the batch contact samples are provided in Table 4.1. Note that the primary samples were contacted for 104 h and the duplicates were contacted for 71 h. There were no measurable differences in K_d values between the 71- and 104-h contact times, which confirmed that the CST equilibrium condition had been reached by the 71-h contact time.

Sample ID	Initial [Cs] (µg/mL)	Final [Cs] (µg/mL)	Δ Time (h)	K _d (mL/g)	Equilibrium Cs in CST (mg Cs/g)
TI062-S0-CST	9.19	1.94	104	736	1.42
TI062-S1-CST	37.5	7.12	104	793	5.62
TI062-S2-CST	229	53.4	104	642	34.3
TI062-S3-CST	1460	994	104	89	88.6
TI062-S0-CST-d	9.19	1.97	71 ^(a)	720	1.41
TI062-S1-CST-d	37.5	7.13	71 ^(a)	826	5.89
TI062-S2-CST-d	229	48.4	71 ^(a)	668	32.3
TI062-S3-CST-d	1460	1030	71 ^(a)	83	85.8

Table 4.1. Equilibrium	Results for AP-107 Ba	tch Contact Samples with	CST Lot 2002009604
1		1	

Note that the AP-107 tank waste also contained 0.096 M K and 0.89 M free hydroxide.

(a) The duplicate samples remained in static contact with AP-107 while the primary sample set was agitated 104 h.

The K_d values versus Cs concentrations are plotted in Figure 4.1 on a log-log scale. Also shown are the K_d values from a different lot of CST using AP-107 tank waste (Rovira et al. 2018). In both cases, a curve maximum appeared around 7.1 μ g/mL Cs.

The K_d value vs. the log of the Cs equilibrium concentration was fit to a third order polynomial equation to calculate the K_d at the feed concentration. At the feed concentration of 9.19 µg Cs/mL, the curve resulted in 806 mL AP-107/g CST. Rovira et al. (2018) reported the K_d for CST Batch 2081000057 at 670 mL AP-107/g CST. Thus, the current test batch of CST was demonstrated to have higher Cs capacity than that of Batch 2081000057 in the AP-107 feed condition.

The theoretical 50% Cs breakthrough on the ion exchange column (λ) can be predicted from the product of the K_d value and the ion exchanger bed density (ρ_b) according to Eq. (4.1). The CST bed density is the dry CST mass divided by the volume in the column. With a CST bed density of 1.01 g/mL (Fiskum et al. 2019), the theoretical 50% breakthrough (λ) for AP-107 with 9.19 µg/mL Cs is predicted to be 814 BVs.

$$\mathbf{K}_{\mathrm{d}} \times \boldsymbol{\rho}_{\mathrm{b}} = \boldsymbol{\lambda} \tag{4.1}$$



Equilibrium Cs concentration, µg/mL

Figure 4.1. Equilibrium Cs K_d Curves for AP-107 with Two Lots of CST

Figure 4.2 provides the isotherms for the AP-107 batch contact test samples with both batches of CST. In this case, the abscissa equilibrium Cs concentration is expressed in terms of molarity and the ordinate is expressed in terms of Q (mmoles Cs/g CST). The isotherm was fit to the Freundlich/Langmuir Hybrid equilibrium isotherm model (see Hamm et al. 2002) according to Eq. (4.2). The expected Cs loading onto the CST at a given Cs concentration can be determined from the isotherm.

$$\frac{\alpha_{i} \times [Cs]}{(\beta + [Cs])} = Cs_{IX}$$
(4.2)

where:

[Cs] = equilibrium Cs concentration (mmoles Cs/mL)

 Cs_{IX} = equilibrium Cs loading on the CST (mmoles Cs per g CST)

 α_i = isotherm parameter constant (mmoles Cs/g CST)¹

 β = isotherm parameter constant (mmoles Cs/mL)²

 $^{^1}$ The α_i parameter represents the maximum Cs capacity in the CST (Hamm et al. 2002).

² The β parameter incorporates the selectivity coefficients, making it dependent on temperature and liquid composition of all the ionic species in solution; the larger the beta parameter, the less favorable (and lower loadings) an isotherm will be (Hamm et al. 2002).



Figure 4.2. Isotherm for AP-107 Tank Waste with Two Lots of CST

Figure notes: At the equilibrium Cs concentration of $9.19 \ \mu g \ Cs/mL$ (6.85E-5 M), the equilibrium Cs loading on CST Lot 2002009604 corresponded to 0.062 mmole Cs per g dry CST (8.3 mg Cs/g CST).

The α_i and β parameters for current and past testing are summarized in Table 4.2. Both the α_i and the β parameters were significantly higher than those reported by Hamm et al. (2002) (α_i of 0.3944 mmoles Cs/g CST and average β value of 2.8552E-4 M Cs for Envelope A tank waste). The Cs capacity α_i parameter conformed to 0.72 mmoles Cs/g CST, which was consistent with the 0.70 mmoles Cs/g CST capacity measured with AW-102 diluted tank waste (Rovira et al. 2019). The AP-107 β parameter conformed to 7.25E-4 mmoles Cs/mL, which was higher than that of AW-102 test (5.84E-4 mmoles Cs/mL). The CST Lot 2002009604 demonstrated 44% higher total Cs capacity than that of Batch 2081000057 (α_i 0.72 vs 0.50); the β parameters (matrix effect) differed by 36% (7.24E-4 vs 5.3 E-4 M Cs). However, at the equilibrium condition of 6.85E-5 M Cs, Q (Cs loading) values in the AP-107 matrix only differed by 9% between the two different CST lots (0.062 vs. 0.057 mmoles Cs/g CST, Lot 2002009604 and Batch 2081000057, respectively).

Matrix	CST Lot or Batch No.	α _i , (mmoles Cs/g CST)	β, (Cs M)	Reference			
AP-107	2002009604	0.72	7.25E-4	This report (Fiskum et al.)			
AP-107	2081000057	0.50	5.3E-4 ^(a)	Rovira et al. 2018			
AW-102	2002009604	0.70	5.84E-4	Rovira et al. 2019			
Envelope A	Not defined	0.3944	2.8552E-4	Hamm et al. 2002			
(a) Calculated fro	m reported raw data.						

Table 4.2. α_i and β Parameter Summary

5.0 Column Processing

This section discusses the Cs exchange behavior during the load, FD, water rinse, and final solution flush from the system. Raw data are provided in Appendix A.

5.1 Cs Loading for AP-107, Feed Displacement, and Water Rinse

The AP-107 feed was processed at nominally 1.88 BV/h through the lead, middle, and polish columns. Figure 5.1 shows a linear-linear plot of the cesium load profile for feed processed through each column. The x-axis shows the BVs processed and the y-axis shows the effluent Cs concentration (C) relative to the feed concentration (C₀) in terms of % C/C₀. The C₀ value for ¹³⁷Cs was determined to be 145 μ Ci/mL (average of the seven filtered feeds, relative standard deviation of 2.5%). In this graphing layout, the Cs breakthrough from the lead column appeared to start at ~280 BVs and continued to 41% C/C₀ after processing 815 BVs when the last sample was collected from the lead column. Feed processing continued for 40 BVs past this last sample collection point for a total of 855 BVs. The final feed loading was extrapolated to 44% C/C₀ before FD was initiated. Because the polish column was in place to further remove Cs from the effluent, the system could process well beyond the 25% C/C₀ implemented on the previous AP-107 testing where the two-column system was used (Rovira et al. 2018). It is obvious that the middle and polish column Cs breakthrough performances are not discernable at this linear scale.

Figure 5.2 shows the same Cs load data provided in Figure 5.1, but with the ordinate % C/C₀ on a probability scale and the abscissa BVs processed on a log scale. Under normal load processing conditions, these scales provide a predictable straight-line Cs breakthrough curve and provide greater fidelity of load characteristics at low and high % C/C₀ values (Buckingham 1967). In contrast to Figure 5.1, the Cs breakthrough from the lead column was observed to occur between 60 and 90 BVs processed. In addition to the 50% C/C₀ indication line, the WAC at 0.127% C/C₀ is also apparent (dotted red line).¹ The WAC Cs breakthrough for the lead column occurred at 200 BVs. The middle column WAC Cs breakthrough occurred at ~590 BVs. The polish column did not reach the WAC limit after processing 844 BVs.

¹ The WAC limit was derived from the allowed curies of ¹³⁷Cs per mole of Na in the effluent to support contact handling of the final vitrified waste form—3.18E-5 Ci ¹³⁷Cs/mole Na. At 5.6 M Na and 145 μ Ci ¹³⁷Cs/mL in the feed, the WAC limit is 0.127% C/C₀.



Figure 5.1. Lead, Middle, and Polish Column Cs Load Profiles of AP-107 at 1.88 BV/h, Linear-Linear Plot



Figure 5.2. Lead, Middle, and Polish Column Cs Load Profiles of AP-107 at 1.88 BV/h, Probability-Log Plot

The 50% Cs breakthrough on the 10.0-mL lead column was not reached because there was insufficient AP-107 feed available at the Radiochemical Processing Laboratory for processing. The lead column data were evaluated to estimate the BVs to 50% breakthrough. The breakthrough curve can be estimated by the error function (erf) (Hougen and Marshall 1947; Klinkenberg 1994):

$$\frac{C}{C_0} = \frac{1}{2} \left(1 + \operatorname{erf}(\sqrt{k_1 t} - \sqrt{k_2 z}) \right)$$
(5.1)

where k_1 and k_2 are parameters dependent on column conditions and ion exchange media performance, t is time (or BVs processed), and z is the length of the column. Using this model, a fit was generated to the experimental data. The 50% breakthrough can then be estimated by multiplication of k_1 and k_2 to give an estimate of 900 BV.

This value was ~11% higher than the 814 BVs Cs λ value predicted from batch-contact studies. Given the extrapolation to 50% breakthrough from column processing and the overall measurement uncertainties, the agreement within 11% was considered reasonable.



Figure 5.3. Lead Column Extrapolated to 50% Breakthrough

The polish column data from 591 to 844 BVs were evaluated to estimate the BVs processed to WAC limit. The error function fit has some simplifying assumptions that were not valid at these concentrations; therefore, the log of the % C/C₀ values was plotted against the BVs processed (see Figure 5.4). The curve was fitted to a polynomial function ($R^2 = 0.999$), which was then extrapolated to the WAC limit. This evaluation showed that the polish column would reach the WAC limit after processing ~1010 BVs.



Figure 5.4. Polish Column Extrapolated to WAC Limit

Figure 5.5 shows the Cs breakthrough profile of the polish column with the FD, water rinse, and the final flushed fluid from the column system on a log-linear plot. These scales provide better Cs concentration resolution of the various effluent solutions relative to graphing on a linear-linear scale or probability-log scale. As previously observed, the FD exhibited declining Cs concentration with increased process volume through the first 9.5 BVs (1.3 AVs) processed, then a slight increase in Cs effluent concentration was observed. As water rinsing commenced, the Cs concentration in the effluent continued to increase, approaching the Cs concentration in the last load effluent sample. The Cs concentration in the final flush solution exceeded the last sample Cs concentration collected from the polish column. In all cases, these rinse solutions remained below the WAC limit. If AP-107 processing continued until the polish column reached the WAC limit, it is possible that parts of the water rinse could also approach the WAC limit and the final solution flush could exceed the WAC limit.

The full height column test resulted in a similar Cs leakage characteristic during the water rinsing step (Fiskum et al. 2019). The flowrate was higher (shorter residence time) for the AP-107 test relative to the full height test at 4.54 BV/h (12.3 BVs) and 3.3 BV/h (10.7 BVs), respectively. The corresponding linear flow velocity was much reduced for the AP-107 test relative to the full height test, 0.43 cm/min and 12.3 cm/min, respectively. In the case of the full height test, the Cs leakage appeared to taper after processing 9 BVs (last sample collected). In contrast, the Cs leakage from the AP-107 test showed no signs of tapering after processing 9 BVs. This issue may have ramification for how the rinse water is dispositioned post-ion exchange processing.



Figure 5.5. Polish Column Cs Load Profile with Feed Displacement, Water Rinse, and Column Flush Solution

Figure 5.6 compares the AP-107 lead and lag/middle column Cs load profiles for two CST materials, Lot 2002009604 (current test data set) and Batch 2081000057 (Rovira et al. 2018). Rovira et al. (2018) Cs load profile results were clearly offset to the left for both the lead and lag columns relative to the current testing lead and middle columns. The CST capacity for CST Lot 2002009604 was shown to be higher than that of Batch 2081000057 from batch contact testing. The Cs loading profile will shift right as the media Cs capacity increases. The kinetic exchange rate was not appreciably different between the two tests based on the similarity in the slopes of Cs loading curves and despite the difference in flowrates (2.2 BV/h compared to 1.9 BV/h).

Figure 5.6 also compares the AP-107 load profiles with the AW-102 load profiles (Rovira et al. 2019) where the same CST Lot 2002009604 was used. The Cs capacities appeared to be similar for these two matrices based on the general congruency of the lead column load profiles. The middle column from the current study and lag column from AW-102 processing were compared. In this case, the Cs load profiles showed a wider gap in performance where the AW-102 Cs load data resulted in a delayed breakthrough (~100 BVs) relative to the AP-107 Cs load data. This effect may be consistent with the higher β parameter for AP-107 than that of AW-102; the β parameter increases with increasing adverse matrix effects on Cs loading (see Section 4.0 of this report and Hamm et al. 2002). It is noted that the K concentration was 58% higher in AW-102 than in AP-107; K is a competitor for Cs on CST.



Figure 5.6. AP-107 (Two Tests) and AW-102 Load Profiles Comparisons

Figure notes: Rovira et al. (2018) used CST Batch 2081000057; Rovira et al. (2019) used CST Lot 2002009604

The AP-107 lead column Cs breakthrough data were compared with full-height and medium-height testing of 5.6 M Na simulant previously reported (Fiskum et al. 2018). The CST production lots used in testing were the same, and the residence times in the lead columns were similar at 1.88 BV/h for AP-107 and 1.82 and 1.83 BV/h for the 5.6 M Na simulant tests. The full-height column replicated the anticipated TSCR column height (234 cm); the medium column height was 12% of the full height column. The efficacy of using the small-scale column (5.6 cm tall/10-mL CST bed) to accurately predict large-scale column performance has not been substantiated. Therefore, the differences in load profiles (i.e., onset of Cs breakthrough and achieved Cs load capacity) cannot be totally separated from column geometry differences. Differences in the feed compositions may also affect the comparative Cs load performances. The feed Cs concentrations differed slightly from 7.9 μ g/mL for the simulant to 9.19 μ g/mL for AP-107. The AP-107 contained additional components not intentionally added to the simulant (such as Ca, Cr, Fe, U, Pu, and other trace components). The Na concentrations were equivalent within analytical uncertainty.

Figure 5.7 and Figure 5.8 compare the Cs load curves for the 5.6 M Na simulant tests and the AP-107 test. Figure 5.7 shows the lead column Cs load profiles only from AP-107 processing and the Blue (full height) and Red (12% full-height) columns previously described (Fiskum et al. 2018). The AP-107 Cs broke through the lead column sooner than the Cs from simulant tests. Divergence in the load profiles increased from ~50 to ~100 BVs as the processed BVs increased from 160 to 800 BVs.


Figure 5.7. Cs Load Profile Comparisons at ~1.85 BV/h

Figure notes:

- AP-107 processed at 1.88 BV/h, 9.19 μ g/mL Cs, 0.096 M K, 0.89 M OH; lead column test data
- Fiskum et al. (2018) processed 5.6 M Na simulant at 1.83 BV/h, 7.9 μg/mL Cs, 0.12 M K, 1.4 M OH; lead column test data

Similarly, Figure 5.8 compares the system volumes (SVs) and system flowrates for the Orange (full-height lead plus lag) column (Fiskum et al. 2019) and the small (10.0 mL), combined lead, middle, polish columns from the AP-107 test. The AP107 system BV was 30 mL (3 x 10-mL beds) and flow rate for the system was one third of the baseline flow rate at 0.63 SV/h (1.88 BV/h \div 3). Similarly, Fiskum et al. (2018) described the Orange column system of 1.18 L BV (2 x 592-mL beds) at a flow rate for the system one half of the baseline flowrate at 0.65 SV/h (1.13 BV/h \div 2). In each case, the system flow rate nominally corresponded to 0.64 SV/h and can be compared based on the common residence times in the system CST beds. The C/C₀ detection limit was much improved for the AP-107 test relative to the Orange column test simply from the increased ¹³⁷Cs concentration. However, the Cs breakthrough curves join at nominally 0.002% C/C₀, Orange column ¹³⁷Cs detection limit. At this juncture, the AP-107 Cs breakthrough is shown to precede the Orange column Cs breakthrough by ~43 system BVs, similar to that found for the 1.88 BV/h comparison discussed above.



Normalized SystemVolumes

Figure 5.8. Cs Load Comparison at ~0.65 SV/h

Figure notes:

- AP-107 processing through the polish column, normalized to a 30-mL SV (three columns at 10-mL per column), and system flowrate 0.63 SV/h (1.88 BV/h ÷ 3)
- 5.6 M Na simulant processing through the full height system Orange lag column, normalized to a 1.18 L SV (two columns at 0.592 L columns), and system flowrate of 0.65 SV/h (1.30 BV/h ÷ 2)

These results indicated the CST Cs exchange capacity was consumed sooner in AP-107 processing, likely from a combination of factors including slightly higher Cs concentration in AP-107 (9.19 ug/mL vs 7.9 ug/mL); other competitors such as Sr, U, Pu, and Ca; and matrix effects from different K^+ and OH^- concentrations.

In both evaluations (Figure 5.7 and Figure 5.8), load curves were essentially parallel by observation, indicating common Cs exchange kinetics between the different matrices. Thus, despite the Cs breakthrough offset, the 5.6 M Na simulant tests were reasonable predictors of the AP-107 Cs exchange performance.

5.2 Cesium Activity Balance

The Cs fractionations to the effluents and the columns were determined based on the known input ¹³⁷Cs and the measured ¹³⁷Cs in the various effluent streams. The quantities of Cs loaded onto the lead, middle, and polish columns were determined by subtracting the Cs recovered in the samples and effluents from the Cs fed to each column. Table 5.1 summarizes the ¹³⁷Cs fractions found in the various effluents as well

as the calculated ¹³⁷Cs column loadings. About 87% of the total Cs loaded onto the lead column, 12% loaded onto the middle column, and 0.22% loaded onto the polish column. Sample and effluent collection amounted to less than 1% of the input Cs. Although the polish column did not accumulate a significant fraction of the Cs (0.22%), its presence was crucial to preventing the effluent from exceeding the WAC limit while loading over 590 BVs (middle column WAC limit processing volume) of AP-107 feed.

Input	μCi	%
Feed sample	1.23E+06	100
Output		
Effluent-1 (0-131 BVs)	0.136	1.10E-05
Effluent-2 (131-280 BVs)	0.166	1.35E-05
Effluent-3 (280-411 BVs)	0.136	1.10E-05
Effluent-4 (411-558 BVs)	0.155	1.26E-05
Effluent-5 (558-680 BVs)	0.240	1.94E-05
Effluent-6 (680-786 BVs)	0.936	7.58E-05
Effluent-7 (786-855 BVs)	2.59	2.10E-04
Load samples	939	7.61E-02
Feed displacement, water rinse and flush	1.45	1.17E-04
Total ¹³⁷ Cs recovery	945	0.077
Total ¹³⁷ Cs Column Loading		
Lead column Cs loading	1.08E+06	87.4
Middle column Cs loading	1.52E+05	12.3
Polish column Cs loading	2.75E+03	0.22
Column total	1.23E+06	99.9

Table 5.1. ¹³⁷Cs Activity Balance for AP-107

The total Cs loaded per g CST was calculated from the total Cs loaded onto the lead column, which was nearly fully saturated under these load conditions (estimated 44% Cs breakthrough), and the dry CST mass loaded into the lead column according to Eq. (5.2).

$$\frac{A_{Cs} \times CF}{M} = C$$
(5.2)

where

 A_{Cs} = activity of ¹³⁷Cs, µCi on the lead column

 $CF = conversion factor, mg Cs/\mu Ci^{137}Cs$

M = mass of dry CST (10.1 g)

C = capacity, mg Cs/g CST

A total 6.76 mg Cs/g CST (0.0504 mmoles Cs/g CST) was loaded onto the lead column. This represented ~81% of the total capacity found from batch contact testing (see Section 4.0); the lower Cs loading was consistent with the fact that the 50% Cs breakthrough was not quite attained. The total Cs loading onto CST following AP-107 processing at the conclusion of testing was similar to those of the simulant tests (see Table 5.2).

	CST Cs loading
Test	(mg Cs/g CST)
AP-107, 2.4% full height	6.76
5.6 M Na simulant, 12% full height	6.95
5.6 M Na simulant, 100% full height	6.60

Table 5.2. Cs Loading onto CST Comparison

5.3 WAC Limit

The amount of feed that can be processed before the effluent reaches the WAC limit is directly affected by the contact time the feed has with the CST bed. Thus, the effect of flowrate on BVs processed to WAC limit is of special interest. Fiskum et al. (2018, 2019) examined this effect using 5.6 M Na simulant solutions and found the data fit a power curve. In these evaluations, the test column systems were evaluated in segmented units or SVs.

- The first column was evaluated as a unit with the SV defined as the first column BV.
- The first/second columns were evaluated as a unit with the SV defined as the combined CST volume in the two columns.
- The first/second/third columns were evaluated as a unit with the SV defined as the combined CST volume in the three columns, and so on.

The flowrates were calculated based on the residence time in the unit in terms of SV/h. Fiskum et al. (2019) showed that 400 SVs could be processed if the flowrate was reduced to 0.42 SV/h using the 5.6 M Na simulant. In a two-column system, this would equate to 0.84 BV/h, where BV is defined as the lead column CST volume. In a three-column system, the corresponding flowrate would equate to 1.26 BV/h.

The current TSCR approach is to process three columns at 1.9 BV/h, which equates to 0.63 SV/h. Given the simulant results, this was expected to be sufficient for the third column to maintain effluent below the WAC limit. The three columns used to test AP-107 were evaluated in a similar approach. SVs, flowrate, and SVs to WAC limit are provided in Table 5.3. These data are also superimposed on the previously reported graphed data set in Figure 5.9. The AW-102 lead column datum (Rovira et al. 2019) was also incorporated (WAC limit reached at 214 BVs at a flowrate of 1.81 BV/h). The lead column AP-107 and AW-102 data points fall on the curve established from simulant testing. However, the second and third data points for AP-107 define a different trajectory with fewer SVs processed to reach WAC limit than predicted by testing with the 5.6 M Na simulant. This indicated that components in the tank waste were consuming or otherwise affecting exchange sites that were not well modeled by the simulant.

AP-107 Systems	SV (mL)	Flowrate (SV/h)	SVs to WAC Limit
Lead column	10	1.88	200
Lead/middle column	20	0.94	594
Lead/middle/polish column	30	0.63	337 ^(a)
(a) Extrapolated value from 281 B	Vs and 4.35E-3%	6 C/C ₀ .	

Table 5.3. Bed Volumes Processed to Reach WAC Limit



Figure 5.9. System Volume to WAC Limit vs. Flowrate

Figure notes:

- Fiskum et al. 2018, 5.6 M Na simulant test matrix CST Batch 2081000057, two serial 10-mL CST beds.
- Fiskum et al. 2019, 5.6 M Na simulant test matrix, CST Lot 2002009604.
 - Orange and Blue column data collected from four serial ~0.592-L CST beds.
 - Medium column data were collected from single 44-mL CST beds.
- AW-102 datum collected from single 10-mL CST bed (lead column), CST Lot 2002009604 (Rovira et al. 2019).
- The left most data point from AP-107 (current test) was extrapolated, CST Lot 2002009604.

5.4 Transition Zone

The transition (or exchange) zone is defined as the volume processed from the onset of Cs breakthrough to the full saturation of the ion exchanger where the effluent Cs concentration equals the influent Cs concentration. The 50% Cs breakthrough point is the inflection point around which the transition zone pivots. In the current study, the lead column was only loaded to 44% Cs breakthrough; the 50% breakthrough was extrapolated to ~919 BVs (see Section 5.1). The number of BVs processed between 20% and 50% Cs breakthrough was calculated from the load curve. This value was doubled to determine the 20% to 80% Cs breakthrough transition zone: 598 BVs. Similarly, the transition zone between 5% and 95% Cs breakthrough was calculated: 1000 BVs. Table 5.4 compares the transition zones for AP-107 to those measured with 5.6 M Na simulant; comparison to Rovira et al. (2018) AP-107 results was not included because Cs loading only resulted in 25% Cs breakthrough. The AP-107 transition zone was ~100 BVs shorter than that determined with the 5.6 M Na simulant (CST Lot 2002009604). The fundamental reason for a decreased transition zone is not understood at this time. It is noted that the simulant hydroxide concentration was ~58% higher than that of AP-107; increased hydroxide concentration improves Cs exchange onto CST supporting the observed delayed Cs breakthrough. Note that the simulant and AP-107 K concentrations were essentially identical (0.122 and 0.120 M, respectively) which minimizes difference from the K competitor effect.

	Flowrate	BVs	to Cs Breakth	rough	Transition	Zone, BVs
Test	(BV/h)	(5%)	(20%)	(50%)	(20-80%)	(5 to 95%)
AP-107 ^(a)	1.88	400	620	~900	~560	~1000
Blue (Fiskum et al. 2019) ^(a,b)	1.82	492	700	~1050	~700	~1120
Blue (Fiskum et al. 2018) ^(b,c)	1.99	330	490	1130	640	1600
 (a) CST Lot 2002009604 (b) 5.6 M Na simulant matrix (c) CST Batch 2081000057 						

Table 5.4. Transition Zone Comparison

5.5 Chemical and Radiochemical Composition

The AP-107 composite feed and composite effluent samples underwent extensive characterization to better define waste characteristics and assess analyte fractionation to the CST. Ten lead column samples were also selected for metal and radionuclide analysis to assess analyte load characteristics (12.6, 87.7, 180, 280, 365, 457, 558, 656, 723, and 815 BVs). Results from the extensive effluent characterization also helped support follow-on vitrification glass formulation.

Table 5.5 summarizes the feed and effluent radioisotope concentrations and fractionations to the effluent. Table 5.6 summarizes the feed and effluent metals concentrations and fractionations to the effluent. The anions, free hydroxide, inorganic carbon, and organic carbon concentrations in the feed are provided in Table 5.7; they were not measured in the effluent because their concentrations were shown not to be

affected by the CST processing (Rovira et al. 2018). Analytical reports along with result uncertainties and quality control (QC) discussions are provided in Appendix E.

By inference, the analytes not found in the effluent were assumed to be retained on the CST. Analyte fractionation was calculated as the ratio of the total analyte measured in the feed processed through the columns and the total analyte collected in the Cs-decontaminated effluent according to Eq. (5.3):

$$\frac{C_{Da} \times V_D}{C_{Fa} \times V_F} = F_{Da}$$
(5.3)

where:

 C_{Da} = concentration of analyte *a* in the Cs-decontaminated effluent

 V_D = volume of Cs-decontaminated effluent

 C_{Fa} = concentration of analyte *a* in the AP-107 feed

 V_F = volume of AP-107 feed

 F_{Da} = fraction of analyte *a* in the Cs-decontaminated effluent

Some analyte results are shown in brackets; this indicates that the analytical result was less than the estimated quantitation limit (EQL) but greater than or equal to the method detection limit (MDL) and the associated analytical uncertainty could be higher than $\pm 15\%$. The fractionation result was placed in brackets where it was calculated with one or more bracketed analytical values to highlight the higher uncertainty. The opportunistic analyte results measured by inductively coupled plasma optical emission spectrometry (ICP-OES) are also shown in Table 5.6; these analytes are part of the ICP-OES data output but have not been fully evaluated for quality control performance. The composite feed sample results in Table 5.5 and Table 5.6 (Al, K, Na, and ¹³⁷Cs) were compared to the initial characterization sample results (Table 3.2). All results agreed within 10% except for K, which was 20% higher as measured via ASR 0787; this was considered an anomaly.

Analysis Method	Analyte	TI061-Comp-Feed (µCi/mL)	TI061-Comp-Eff (µCi/mL)	Fraction in Effluent (%)	
Gamma energy	⁶⁰ Co	[5.0E-04]	5.29E-04	107%	
analysis (GEA) ^(a)	¹²⁶ Sb	<2E-3	1.37E-04		
	¹³⁷ Cs	1.54E+02	5.29E-04	0.0003%	
	¹⁵⁴ Eu	<2E-3	4.65E-05		
	¹⁵⁵ Eu	<3E-2	1.04E-05		
	²⁴¹ Am	<1E-1	2.45E-04		
Separations/	²³⁷ Np	4.76E-05	2.02E-05 ^(c)	43%	
Alpha energy	²³⁸ Pu	7.54E-05	2.75E-05	37%	
analysis (AEA) ^(b)	²³⁹⁺²⁴⁰ Pu	5.60E-04	2.01E-04	36%	
Separations/	⁹⁰ Sr	4.52E-01	<1E-3	<0.2%	
Beta counting ^(b)	⁹⁹ Tc	1.00E-01	9.82E-02	99%	

Table 5.5. AP-107 Feed and Effluent Radionuclide Concentrations and Fractionations (ASR 0787)

(a) Reference date is June 10, 2019.

(b) Reference date is August 2019.

(c) The effluent was analyzed in duplicate and the RPD was 25% exceeding the QC criterion of 20%, see Appendix E.

"--" = not applicable; value not reported or fractionation cannot be calculated with a less-than value.

Analysis Method	Analyte	TI061-Comp-Feed (M)	TI061-Comp-Eff (M)	Fraction in Effluent
	Ag ^(a)	<4.4E-6	<4.4E-6	
	Al	3.71E-01	3.61E-01	98%
	As	<1.8E-4	<1.8E-4	
	Ba	[3.0E-6]	<5.7E-7	
	Ca	[8.6E-4]	[4.5E-4]	[52%]
	Cd	5.79E-05	[4.7E-5]	[81%]
	Cr	1.01E-02	9.90E-03	98%
	Fe	2.64E-04	[2.2E-4]	[83%]
ICP-OES	K ^(b)	1.20E-01	1.15E-01	97%
	Na	5.97E+00	5.71E+00	96%
	Ni	4.19E-04	3.95E-04	95%
	Р	2.82E-02	2.67E-02	95%
	Pb	[3.9E-5]	<1.9E-5	
	S	6.03E-02	5.75E-02	96%
	Sr	[1.5E-6]	<1.4E-6	
	Ti	<2.3E-6	[3.9E-6]	
	Zn	<1.6E-5	[2.0E-5]	
	Zr	<5.2E-6	[2.3E-5]	
ICP-MS	²³⁸ U	7.51E-05	4.53E-05	61%

Table 5.6. AP-107 Feed and Effluent Inorganic Analyte Concentrations and Fractionation (ASR 0787)

(a) QC samples recovered only 50% of the Ag.

(b) The K result was $\sim 20\%$ higher in the composite feed analysis than measured in the individual sample aliquots collected from the first and last samples collected from the Hanford tank.

Bracketed results were less than the estimated quantitation limit but greater than or equal to the method detection limit.

"--" indicates the recovery could not be calculated.

Analysis Method	Analyte	TI061-Comp-Feed (M)	TI061-Comp-Eff (M)	Fraction in Effluent
	В	2.90E-03	2.03E-03	70%
	Be	[1.5E-5]	[1.2E-5]	[82%]
	Bi	<5.9E-5	<5.9E-5	
	Ce	<1.2E-5	<1.2E-5	
	Со	<1.5E-5	<1.5E-5	
	Cu	[1.8E-5]	<6.4E-6	
	Dy	<4.4E-6	<4.5E-6	
	Eu	<2.2E-6	<2.2E-6	
	La	<3.9E-6	<3.9E-6	
	Li	[3.3E-5]	<3.3E-5	
	Mg	<6.9E-5	<6.9E-5	
	Mn	<9.8E-7	<9.8E-7	
	Mo	4.62E-04	4.33E-04	94%
ICP-OES	Nd	<1.5E-5	<1.5E-5	
Opportunistic	Pd	<1.8E-5	<1.8E-5	
Analytes	Rh	<4.0E-5	<4.0E-5	
	Ru	[7.0E-5]	[6.5E-5]	[94%]
	Sb	<1.1E-4	<1.1E-4	
	Se	<5.0E-4	<5.0E-4	
	Si	2.53E-03	[1.5E-3]	[58%]
	Sn	[1.2E-4]	[9.4E-5]	[82%]
	Та	<3.0E-5	<3.0E-5	
	Te	[6.6E-5]	<6.1E-5	
	Th	<2.0E-5	<2.0E-5	
	Tl	<6.5E-5	<6.5E-5	
	U	[6.4E-5]	<3.8E-5	
	V	[1.5E-5]	<1.0E-5	
	W	3.86E-04	3.69E-04	96%
	Y	<3.4E-6	<3.4E-6	

Table 5.6 (cont.)

Bracketed values indicate the associated sample results were less than the EQL but greater than or equal to the MDL. Analytical uncertainty for these analytes is $> \pm 15\%$. "--" indicates the recovery could not be calculated.

		TI061-Comp-Feed	TI061-Comp-Feed
Analysis Method	Analyte	$(\mu g/mL)$	(M)
Titration	Free Hydroxide	1.51E4	0.89
Hot persulfate	Total organic C	2,170	0.18
oxidation	Total inorganic C ^(a)	7,749	0.65
(a) Assumed to be c	arbonate.		

	Table 5.7. AP-107 Feed Free H	ydroxide and Carbon	Composition (ASR 0787)
--	-------------------------------	---------------------	------------------------

In addition to Cs, the CST quantitatively removed ⁹⁰Sr (Sr) with a decontamination factor of \geq 450. The ⁹⁰Sr removal agreed with previous testing (Rovira et al. 2018 and 2019). About 57% of the Np and 64% of the Pu were also removed. The Np and Pu removal factors were consistent with those reported by Rovira et al. (2018). Assuming the difference in total Np and Pu µCi content between the feed and effluent remained with the CST, the CST would contain 123 nCi/g TRU, exceeding the 100 nCi/g threshold defining TRU waste. The ⁶⁰Co (Co) and ⁹⁹Tc were not affected by the CST and reported to the effluent.

The feed composite and effluent composite ICP-OES results showed that the majority of analytes remained in the effluent. (See Table 5.6 and Appendix E for analytical reports.) The Al, Cr, K, Na, Ni, P (phosphate), and S (sulfate) partitioned exclusively to the effluent (>95% recovery). Ba, Pb, and Sr were detected in the feed (with high errors likely exceeding 15%) but were below the MDL in the effluent composite. Ca and U were partially removed by the CST with only ~52% and 61% reporting to the effluent, respectively. By difference, ~3.5 mmoles of Ca were removed from the 8.55 L AP-107 feed, resulting in a loading of ~0.12 mmoles Ca per gram CST (30 g total CST). Similarly, 0.25 mmoles of U were removed from the feed, resulting in a loading of ~0.0084 mmoles U per gram CST.

The load behaviors of selected analytes were examined as a function of BVs processed from the lead column. (Raw data are provided in Appendix B.) Figure 5.10 shows the Al, Ca, Fe, and U breakthrough results. Al breakthrough can serve as an internal standard for the ICP-OES analysis and Fe and Al breakthrough track similarly (notwithstanding the two high flier Fe values). Initially, U is shown to break through slowly on the CST with full recovery occurring starting at about 200 BVs (thus, a 200-BV transition zone). Ca recovery was essentially flat at ~66% breakthrough with no evidence of a breakthrough curve. Both the feed and effluent Ca concentrations were reported with high uncertainty; however, the results were generally similar to those previously reported (Rovira et al. 2018). It is possible that the Ca is present in a multiplicity of chemical forms in the solution such as oxalate, carbonate or sulfate complexes. Additionally, 60-70% of Ca as $Ca(OH)_{2(aq)}$ has been reported in high (e.g., 1 M) NaOH solutions (Kutus et al., 2016). The complexed Ca and $Ca(OH)_{2(aq)}$ would likely not exchange onto CST. It is probable that 34% of the Ca in AP-107 exists in a form amenable to ion exchange or as a suspended colloid form that filters onto the media.

Small concentrations of Ti (0.23 μ g/mL) and Zr (1.4 μ g/mL) were observed in each of the selected lead column effluent samples (see also Table 5.6 for feed and effluent sample concentrations). Neither Ti nor Zr were detected in the feed; therefore, it was inferred that some small, continuous breakdown of the CST may have occurred during processing. These results were contrary to those reported by Rovira et al. (2018) with AP-107 testing (two columns). However, in that case, a different lot of CST was used and small levels of Ti and Zr were identified in the feed (0.8 and 1.4 μ g/mL, respectively) and were equivalent to that of the effluent. Thus, at best, attrition of Ti and Zr should be considered inconclusive.



Figure 5.10. Al, Ca, Fe, and U Load Profiles onto the Lead Column

Figure Notes:

- Analytes with dashed lines indicate data were \geq MDL but < EQL, with errors likely to exceed 15%.
- Two Fe results at 179% and 230% recoveries were likely contaminated in the analysis process.

Resource Conservation and Recovery Act (RCRA) elements were examined and are graphed in Figure 5.11 (Al recovery used as a general recovery marker). Except for the first sample Cd result, there seemed to be relatively no difference in any of the analyte percentage recoveries in the effluent with increased loading. Ba and Pb were detected in the feed (with errors likely to exceed 15%) but were not detected in the lead column effluent samples. Ag and Cd were not detected in the feed or effluent samples. Cr and Ni, like Al, reported solely to the effluent. Data from Figure 5.11 are shown in tabular form in Appendix B.



Figure 5.11. RCRA Metal Recoveries during Lead Column Processing; Ag, As, Ba Were Not Detected

Similarly, the selected lead column samples were analyzed for ⁹⁰Sr, ²³⁷Np, ²³⁸Pu and ²³⁹⁺²⁴⁰Pu. Results from both Pu isotopes were equivalent. Figure 5.12 shows the load profiles for these isotopes in comparison with that of ¹³⁷Cs. The ⁹⁰Sr clearly did not break through the CST after processing 815 BVs feed. The anomalously high ⁹⁰Sr result at 280 BVs was likely associated with in-cell contamination. The Np breakthrough profile showed increasing effluent concentration from 10% to 60% in the 12 to 300 BVs range at which point a steady state was reached. It is unclear why the breakthrough profile did not increase in a more normal manner, extending toward 100% breakthrough. The Pu demonstrated an initial 40% breakthrough and very slowly increased to 60% breakthrough to the end of testing. The Pu exchange may be limited by the Pu oxidation state (e.g., mixture of III and IV) and/or the nature and extent of a Pu complexant.



Figure 5.12. ⁹⁰Sr, ¹³⁷Cs, ²³⁷Np, and ²³⁹⁺²⁴⁰Pu Load Profiles onto the Lead Column

6.0 Conclusions

Cesium ion exchange batch contact and column testing with CST Lot 2002009604 was conducted to assess Cs exchange performance with AP-107 tank waste. By necessity, testing was conducted at a small scale given the high radiological dose rate of the Hanford tank waste matrix. WRPS established five objective outcomes for this testing; all objectives for CST testing were met.

- 1. Conduct batch contact testing with CST to determine the Cs load capacity of AP-107.
- 2. Process filtered AP-107 through a lead-lag-polish column configuration and establish a Cs load profile.
- 3. Compare the AP-107 Cs load curve results to the previously reported AP-107 load curve results (Rovira et al. 2018).
- 4. Analyze the feed and effluent to discern the fates of key analytes.
- 5. Provide the decontaminated AP-107 tank waste (8.55 L) and the Cs-loaded CST for follow-on testing (vitrification and analysis, respectively).

The results summary is provided below.

6.1 Batch Contact Testing

Batch contact testing with CST was conducted to determine the Cs K_d value at equilibrium AP-107 conditions and the total Cs load capacity in the AP-107 matrix. Duplicate tests were conducted; the primary test matrix was mixed for 104 h and the duplicate was mixed for 71 h. The following conclusions were made as a result of this work.

- 1. Equilibrium was achieved by the 71-h contact time (conditions: orbital shaker at ~400 rpm, 15-mL solution in a glass 20-mL LSC vial, phase ratio ~200). This was consistent with previously reported studies where equilibrium was reached between 46- and 123-h contact time (Fiskum et al. 2019; Rovira et al. 2019).
- The calculated ¹³⁷Cs K_d value of 806 mL AP-107/g CST at Cs equilibrium condition of 9.19 μg Cs/mL corresponded to a predicted 50% Cs breakthrough of 814 BVs. Although 50% Cs breakthrough was not achieved with column testing, 50% breakthrough was extrapolated to ~900 BVs. The column test 50% breakthrough was 11% higher than predicted by batch contact. This differed from previously reported data on a different lot of CST where the extrapolated 50% breakthrough agreed within 6% of the batch contact derived value (Rovira et al. 2018).
- 3. The Cs load capacity at 9.19 μg Cs/mL equilibrium condition was 0.062 mmoles Cs/g CST (8.3 mg Cs/g CST). This value was over 22% higher than measured on the lead column; however, the lead column was not loaded to full capacity (44% Cs breakthrough).

Batch contact testing cannot be used to determine the slope of the Cs load curve. Thus, it cannot predict the WAC limit breakthrough in terms of BVs.

6.2 Column Testing

AP-107 tank waste was processed through three columns sequentially positioned in a lead-middle-polish format; each column was filled with 10.0 mL of CST ion exchanger. A total of 8.55 L of AP-107 tank waste, consisting of 5.6 M Na and 145 μ Ci/mL ¹³⁷Cs, was processed through the Cs ion exchange system at 1.88 BV/h. Effluent samples were collected periodically from each column during the load process and measured for ¹³⁷Cs to establish the Cs load curves. The flowrate was increase to 3.7 BV/h and 4.5 BV/h to process 0.1 M NaOH FD and water rinse, respectively. The following conclusions were made as a result of this work.

- 1. Testing showed that 855 BVs of AP-107 tank waste, processed at 1.88 BV/h, can be treated before reaching 50% Cs breakthrough on the lead column. The WAC limit was reached on the middle column when 590 BVs of AP-107 feed was processed. The effluent from the polishing column (4.53E-3% C/C₀) was substantially lower than the WAC limit.
- 2. Extrapolation of Cs effluent concentration data from the polish column showed the WAC limit would be reached after processing 1010 BVs.
- 3. FD results showed decreasing Cs concentration coming off the polish column, but water rinse resulted in increasing effluent Cs concentration. The change in solution matrix from high salt to 0.1 M NaOH/water mix appeared to result in some Cs leakage from the column system. This Cs leakage was previously observed (Fiskum et al. 2019).
- 4. AP-107 processing breakthrough profile on the lead column was similar to that observed for AW-102 (Rovira et al. 2019).
- 5. The higher Cs load capacity of CST Lot 2002009604 relative to CST Batch 2081000057 with AP-107 was clearly demonstrated by comparison of the Cs load profiles. Rovira et al. (2018) tested AP-107 with Batch 2081000057; its Cs load profile was offset (earlier Cs breakthrough) ~100 to 200 BVs relative to the current test using CST Lot 2002009604.
- 6. The Cs load profile for AP-107 was offset left (earlier Cs breakthrough) of the Cs load profiles generated with 5.6 M Na simulant (Fiskum et al. 2019). This could be due in part to the slightly higher Cs concentration in the AP-107 tank waste (9.2 ug/mL) relative to the simulant (7.9 ug/mL). Additionally, this may indicate that other matrix factors in actual tank waste AP-107 could be limiting Cs loading onto the CST relative to components in the simulant.
- 7. The total Cs loading onto the lead column (6.76 mg Cs/g CST) was similar to those from the simulant testing (6.60 and 6.95 mg Cs/g CST) at similar processing flowrates.
- 8. The AP-107 SVs processed to reach the WAC limit as a function of flowrate were evaluated. The curve established with AP-107 veered away from the trajectory established by the simulant testing. This was again indicative that other components may be consuming exchange sites or the matrix is more limiting relative to the simulant.
- 9. The transition zones for Cs were calculated to be 560 BVs (20% to 80% Cs breakthrough range) and 1000 BVs (5% to 95% Cs breakthrough range).

6.3 Analyte Fractionation

- 1. Major components Al, Cr, K, Na, P, and S partitioned exclusively to the effluent.
- 2. Ba, Pb, and Sr were detected in the feed but not in the effluent. This indicated that a major fraction of these analytes exchanged onto the CST.
- 3. Substantial fractions of Ca and U (52% to 61%, respectively) were found in the effluent composite, indicating they were partially removed by the CST from the feed. The U load curve through the lead column indicated a transition zone of ~200 BVs. The Ca breakthrough curve was essentially flat.
- 4. Ti and Zr were not detected in the feed. Yet Zr and Ti were measured in the effluent, indicating that potentially a small fraction of CST may be degrading.
- 5. ⁹⁰Sr was quantitatively removed by CST; the ⁹⁰Sr decontamination factor was \geq 450 (the decontamination factor was limited by the effluent ⁹⁰Sr detection limit).
- 6. Only 40% of the feed Np and Pu were measured in the effluent, demonstrating ~60% of the Np and Pu was retained on the CST. Assuming these isotopes were bound to the CST, the CST would contain 123 nCi/g TRU, exceeding the 100 nCi/g threshold defining TRU waste.
- 7. ⁹⁹Tc did not exchange onto the CST.

7.0 References

ASME. 2000. *Quality Assurance Requirements for Nuclear Facility Applications*. NQA-1-2000, American Society of Mechanical Engineers, New York, New York.

ASME. 2008. *Quality Assurance Requirements for Nuclear Facility Applications*. NQA-1-2008, American Society of Mechanical Engineers, New York, New York.

ASME. 2009. *Addenda to ASME NQA-1-2008*. NQA-1a-2009, American Society of Mechanical Engineers, New York, New York.

Brown GN, LA Bray, CD Carlson, KJ Carson, JR DesChane, RJ Elovich, FV Hoopes, DE Kurath, LL Nenninger, and PK Tanaka. 1996. *Comparison of Organic and Inorganic Ion Exchangers for Removal of Cesium and Strontium from Simulated and Actual Hanford 241-AW-101 DSSF Tank Waste*. PNNL-11120, Pacific Northwest National Laboratory, Richland, Washington.

Buckingham J. 1967. *Waste Management Technical Manual*. ISO-100, Isochem, Incorporated. Richland, Washington.

Fiskum SK, AM Rovira, JR Allred, HA Colburn, MR Smoot, AM Carney, TT Trang-Le, MG Cantaloub, EC Buck, and RA Peterson. 2019. *Cesium Removal from Tank Waste Simulants Using Crystalline Silicotitanate at 12% and 100% TSCR Bed Heights*. PNNL-28527, Rev. 0; RPT-TCT-001, Rev. 0, Pacific Northwest National Laboratory, Richland, Washington.

Fiskum SK, HA Colburn, RA Peterson, AM Rovira, and MR Smoot. 2018. *Cesium Ion Exchange Using Crystalline Silicotitanate with 5.6 M Sodium Simulant*. PNNL-27587, Rev. 0; RPT-DFTP-008, Rev. 0, Pacific Northwest National Laboratory, Richland, Washington.

Geeting JGH, JR Allred, AM Rovira, and RA Peterson. 2019. *Fiscal Year 2019 Filtration of Hanford Tank AP-107 Supernatant*. PNNL-28780, Rev. 0; RPT-DFTP-015, Rev. 0, Pacific Northwest National Laboratory, Richland, Washington.

Hamm LL, T Hang, DJ McCabe, and WD King. 2002. *Preliminary Ion Exchange Modeling for Removal of Cesium from Hanford Waste Using Hydrous Crystalline Silicotitanate Material*. WSRC-TR-2001-00400; SRT-RPP-2001-00134, Westinghouse Savannah River Company, Aiken, South Carolina.

Hendrickson DW, RK Biyani, and MA Beck. 1996. *Hanford Tank Waste Supernatant Cesium Removal Test Report*. WHC-SD-RE-TRP-018, Rev. 0A, Westinghouse Hanford Company, Richland, Washington.

Hougan OA and WR Marshall, Jr. 1947. "Adsorption from a Fluid Stream Flowing Through a Stationary Granular Bed." *Chemical Engineering Progress* 43(4):197-208.

King WD. 2007. *Literature Reviews to Support Ion Exchange Technology Selection for Modular Salt Processing*. WSRC-STI-2007-00609, Washington Savannah Company, Aiken, South Carolina.

Klinkenberg A. 1994. "Numerical Evaluation of Equations Describing Transient Heat and Mass Transfer in Packed Solids." *Industrial and Engineering Chemistry* 40(10):1992.

Kutus, B., A. Gácsi, A. Pallagi, I. Pálinkó, G Peintler, and P. Sipos. 2016 A comprehensive study on the dominant formation of the dissolved $Ca(OH)_{2(aq)}$ in strongly alkaline solutions saturated by Ca(II). *RSC Adv.* 2016. **6**, 45231-45240.

Rovira AM, SK Fiskum, HA Colburn, JR Allred, MR Smoot, and RA Peterson. 2018. *Cesium Ion Exchange Testing Using Crystalline Silicotitanate with Hanford Tank Waste 241-AP-107*. PNNL-27706, RPT-DFTP-011, Rev. 0, Pacific Northwest National Laboratory, Richland, Washington.

Rovira AM, SK Fiskum, JR Allred, JGH Geeting, HA Colburn, AM Carney, TT Trang-Le, and RA Peterson. 2019. *Dead-End Filtration and Crystalline Silicotitanate Cesium Ion Exchange with Hanford Tank Waste AW-102*. PNNL-28783, Rev. 0; RPT-TCT-003, Rev. 0, Pacific Northwest National Laboratory. Richland, Washington.

Walker, Jr. JF, PA Taylor, RL Cummins, BS Evans, SD Heath, JD Hewitt, RD Hunt, HL Jennings, JA Kilby, DD Dee, S Lewis-Lambert, SA Richardson, and RF Utrera. 1998. *Cesium Removal Demonstration Utilizing Crystalline Silicotitanate Sorbent for Processing Melton Valley Storage Tank Supernate: Final Report*. ORNL/TM-13503, Oak Ridge National Laboratory, Oak Ridge, Tennessee.

Appendix A – Column Load Data

The AP-107 lead, middle, and polish column loading raw data are provided in Table A.1. The feed displacement, water rinse, and final fluid expulsion raw data are provided in Table A.2. The raw data include the processed bed volumes (BVs) and corresponding ¹³⁷Cs concentration in the collected sample, % C/C₀, and the decontamination factor (DF).

	Lead C	Column			Middle Column Polish Column						
	µCi ¹³⁷ Cs/				µCi ¹³⁷ Cs/				µCi ¹³⁷ Cs/		
BV	mL	% C/C ₀	DF	BV	mL	% C/C ₀	DF	BV	mL	% C/C ₀	DF
12.6*	1.04E-4	7.14E-5	1.40E+6	42.6	1.96E-4	1.35E-4	7.39E+5	42.6	1.06E-4	7.30E-5	1.37E+6
42.8	3.73E-4	2.57E-4	3.89E+5	87.1	2.04E-4	1.40E-4	7.12E+5	86.9	1.97E-4	1.36E-4	7.37E+5
58.8	1.85E-4	1.27E-4	7.86E+5	130.5	1.48E-4	1.02E-4	9.81E+5	130.1	1.10E-4	7.58E-5	1.32E+6
87.7*	2.52E-3	1.74E-3	5.76E+4	178.6	2.54E-4	1.75E-4	5.70E+5	178.0	1.11E-4	7.63E-5	1.31E+6
107.7	3.77E-3	2.60E-3	3.85E+4	235.4	1.80E-4	1.24E-4	8.04E+5	234.6	1.14E-4	7.88E-5	1.27E+6
131.6	1.32E-2	9.07E-3	1.10E+4	277.9	3.08E-4	2.12E-4	4.71E+5	277.0	1.27E-4	8.72E-5	1.15E+6
151.7	3.54E-2	2.44E-2	4.10E+3	317.7	1.32E-3	9.13E-4	1.10E+5	316.5	1.09E-4	7.54E-5	1.33E+6
180.2*	1.05E-1	7.21E-2	1.39E+3	362.6	2.28E-3	1.57E-3	6.37E+4	361.3	1.73E-4	1.19E-4	8.39E+5
200.1	1.98E-1	1.36E-1	7.34E+2	408.0	7.77E-3	5.35E-3	1.87E+4	406.5	1.02E-4	7.04E-5	1.42E+6
237.4	4.83E-1	3.33E-1	3.01E+2	453.3	1.67E-2	1.15E-2	8.67E+3	451.6	1.32E-4	9.08E-5	1.10E+6
280.1*	1.35E+0	9.32E-1	1.07E+2	475.4	3.01E-2	2.08E-2	4.81E+3	496.9	1.07E-4	7.36E-5	1.36E+6
320.1	2.42E+0	1.67E+0	6.00E+1	498.9	4.07E-2	2.81E-2	3.56E+3	550.9	1.26E-4	8.66E-5	1.16E+6
341.5	3.56E+0	2.45E+0	4.08E+1	520.6	6.78E-2	4.67E-2	2.14E+3	590.7	1.33E-4	9.13E-5	1.09E+6
365.5*	4.50E+0	3.10E+0	3.22E+1	553.4	1.07E-1	7.36E-2	1.36E+3	628.0	2.14E-4	1.47E-4	6.79E+5
385.2	6.12E+0	4.22E+0	2.37E+1	593.5	1.82E-1	1.26E-1	7.96E+2	670.9	3.44E-4	2.37E-4	4.22E+5
411.2	7.65E+0	5.27E+0	1.90E+1	630.9	2.88E-1	1.98E-1	5.04E+2	690.2	5.03E-4	3.47E-4	2.89E+5
433.5	9.78E+0	6.74E+0	1.48E+1	650.4	3.74E-1	2.57E-1	3.88E+2	714.4	7.30E-4	5.03E-4	1.99E+5
457.0*	1.09E+1	7.53E+0	1.33E+1	674.3	4.87E-1	3.36E-1	2.98E+2	733.0	9.74E-4	6.71E-4	1.49E+5
479.3	1.49E+1	1.02E+1	9.77E+0	693.8	6.32E-1	4.35E-1	2.30E+2	757.4	1.42E-3	9.77E-4	1.02E+5
503.0	1.67E+1	1.15E+1	8.70E+0	718.2	7.89E-1	5.44E-1	1.84E+2	775.9	1.95E-3	1.34E-3	7.46E+4
524.8	1.90E+1	1.31E+1	7.62E+0	736.9	9.85E-1	6.79E-1	1.47E+2	804.2	2.93E-3	2.02E-3	4.95E+4
557.9*	2.27E+1	1.56E+1	6.41E+0	761.6	1.20E+0	8.30E-1	1.20E+2	825.7	4.54E-3	3.13E-3	3.20E+4
598.2	2.75E+1	1.89E+1	5.28E+0	780.3	1.50E+0	1.03E+0	9.70E+1	844.1	6.31E-3	4.35E-3	2.30E+4
635.8	3.23E+1	2.23E+1	4.49E+0	808.8	1.89E+0	1.30E+0	7.68E+1				
655.6*	3.63E+1	2.50E+1	4.00E+0	830.5	2.41E+0	1.66E+0	6.02E+1				
679.6	3.81E+1	2.63E+1	3.81E+0	849.0	2.84E+0	1.95E+0	5.12E+1				
723.8*	4.55E+1	3.14E+1	3.19E+0								
767.4	5.14E+1	3.54E+1	2.82E+0								
814.8*	5.99E+1	4.13E+1	2.42E+0								
BV = bed v	olume. 10 mL/	/BV						•			

Table A.1. Lead, Middle, and Polish Column Cs Breakthrough Results with AP-107

DF = decontamination factor

 $C_0 = 145 \ \mu Ci^{137} Cs/mL$

* = samples submitted for additional analysis

	Feed Disp	lacement		Water Rinse				Final Fluid Flush			
	μCi				μCi				μCi		
BV	¹³⁷ Cs/ mL	% C/C ₀	DF	BV	¹³⁷ Cs/ mL	% C/C ₀	DF	BV	¹³⁷ Cs/ mL	% C/C ₀	DF
2.7	7.85E-3	5.41E-3	1.85E+4	2.5	3.14E-4	2.16E-4	4.62E+5	5.1	1.76E-2	1.21E-2	8.23E+3
5.9	1.28E-3	8.79E-4	1.14E+5	6.0	5.41E-4	3.73E-4	2.68E+5				
9.5	1.88E-4	1.29E-4	7.74E+5	9.3	2.49E-3	1.72E-3	5.83E+4				
12.8	2.09E-4	1.44E-4	6.96E+5	12.3	4.71E-3	3.25E-3	3.08E+4				
BV = bed v	volume, 10 mL							-			
DF = decontamination factor											
$C_0 = 145 \ \mu$	Ci ¹³⁷ Cs/ mL										

Table A.2. Feed Displacement, Water Rinse, and Final Flush Results Following AP-107 Processing

Appendix B – Analyte Concentrations As a Function of Loading

The load behaviors of selected analytes in AP-107 were evaluated from selected samples collected from the lead column. Analysis results of these samples are summarized in Table B.1.

BV Processed>	NA	12.6	87.7	180	280	365	457	558	656	724	815
Sample ID>	TI061- Comp-Feed	TI061-L- F1-A	TI061-L- F4-A	TI061-L- F8-A	TI061-L- F11-A	TI061-L- F14-A	TI061-L- F18-A	TI061-L- F22-A	TI061-L- F25-A	TI061-L- F27-A	TI061-L- F28-A
Analyte						µg/mL					
Al	10,014	8,519	9,134	8,960	9,033	8,772	9,432	9,527	9,598	9,577	9,636
Ba	[0.41]	< 0.1	< 0.1	< 0.1	< 0.1	< 0.1	< 0.1	< 0.1	< 0.1	< 0.1	< 0.1
Ca	[34]	[29]	[24]	[23]	[20]	[21]	[24]	[22]	[20]	[21]	[22]
Cd	6.51	[2.4]	5.51	[5.4]	6.00	[5.5]	5.95	6.39	5.96	6.50	5.98
Cr	527	487	495	488	495	477	498	519	505	509	515
Fe	14.8	15.3	26.4	[12]	14.1	[12]	13.8	13.2	14.2	13.4	33.9
К	4,685	4,144	4,359	4,187	4,328	4,159	4,512	4,514	4,492	4,459	4,554
Na	137,330	122,716	128,844	126,428	127,623	125,402	128,473	131,486	132,019	130,987	133,132
Ni	24.6	20.6	21.6	21.0	20.9	21.2	21.5	21.6	22.7	23.0	22.2
Р	873	775	804	775	804	764	792	824	813	820	822
Pb	[8.1]	<4.0	<4.0	<4.0	<4.0	<4.1	<4.0	<4.0	[4.2]	<4.0	[6.0]
S	1,933	1,761	1,831	1,744	1,852	1,787	1,847	1,906	1,831	1,848	1,855
Sr	[0.13]	< 0.1	< 0.1	< 0.1	< 0.1	< 0.1	< 0.1	< 0.1	< 0.1	< 0.1	< 0.1
Ti	< 0.11	[0.30]	[0.26]	[0.23]	[0.29]	[0.21]	[0.18]	[0.24]	[0.14]	[0.21]	[0.21]
²³⁸ U	17.9	0.623	11.6	14.9	15.4	15.6	15.7	16.7	17.1	17.0	17.0
Zn	<1.1	[1.1]	<1.1	<1.1	[1.3]	<1.1	[1.2]	[1.4]	<1.1	<1.1	<1.1
Zr	< 0.48	[1.4]	[1.5]	[1.7]	[1.5]	[1.1]	[1.4]	[1.7]	[1.3]	[1.4]	[1.1]
Analyte						$\mu Ci/mL^{(a)}$					
⁹⁰ Sr	4.52E-01	<1E-3	1.24E-03	<1E-3	1.51E-02	<1E-3	1.19E-03	<1E-3	1.68E-03	2.14E-03	2.73E-03
²³⁷ Np	4.76E-05	7.43E-06	2.52E-05	3.98E-05	4.51E-05	4.61E-05	4.58E-05	4.60E-05	4.35E-05	4.16E-05	4.22E-05
²³⁸ Pu	7.54E-05	2.09E-05	4.05E-05	4.10E-05	4.98E-05	4.56E-05	5.00E-05	4.89E-05	5.10E-05	5.44E-05	5.07E-05
²³⁹⁺²⁴⁰ Pu	5.60E-04	1.49E-04	2.72E-04	3.07E-04	3.42E-04	3.42E-04	3.77E-04	3.68E-04	3.73E-04	3.96E-04	3.95E-04

Table B.1. Analyte Concentrations of Selected Samples from the Lead Column

(a) Reference date is August 2019.

BV = bed volume, 10 mL

Bracketed values indicate the associated sample results were less than the estimated quantitation limit but greater than or equal to the MDL. Analytical uncertainties for these analytes are $> \pm 15\%$.

Additional analyte concentrations may be found in Appendix E.

Appendix C – Solids Characterization

Backflow processing was briefly applied to the lead column before loading was completed. Brown/black solids were observed at the top of the crystalline silicotitanate (CST) bed. These solids were briefly suspended in the tank waste fluid with the backflow processing, whereas the CST did not move. Figure 3.7 in the main body of this report is an image of the suspended solids observed during the backflow operation. Small samples of the solids were collected for transmission electron microscopy analysis from the parent bottle where they were found to originate.

The materials removed from the hot cells were examined using a Quanta[™] 250FEG (Thermo Fisher Scientific, Inc., Hillsboro, OR) SEM equipped with an EDAX Inc. (Mahwah, NJ) energy dispersive spectrometer (EDS) for detection of major elements. SEM images using backscattered electrons (BSE) were taken from a few of the samples that were prepared. The first sample analyzed had visible white particles on the stub. These were mainly salt-type material. The second sample did not have any clearly visible particles but had significant activity indicating the presence of materials from the hot cell experiments; high contrast particles using BSE imaging were found indicating materials possessing an average higher atomic number than the salt. The SEM images were obtained under low-vacuum conditions (i.e., no conductive coating was used).

Figure C.1 shows a multitude of different morphologies and particle shapes and compositions. There were several areas where spherical particle agglomerates were found. In Figure C.1(C), in the lower left-hand corner of the image, these round particles are visible. Scattered throughout the sample, bright (high-Z) angular particles were observed. Examples of these are shown in both Figure C.2 and Figure C.3.



Figure C.1. Low-Magnification Images of the Solids Indicating a Heterogeneous Mixture of Particle Types

Figure Note: The salt-like particles that were visible on other samples were not found on this sample.



Figure C.2. SEM Images of High-Z Particles in the Specimen with Low-magnification Images (C and D) and High-magnification images (A and B) of Angular Particles



Figure C.3. Low- and High-Magnification SEM Images of High-Z Particles within the Same Region in the Sample

EDS analysis of particles revealed a variety of compositions. Various Ca and Ba-bearing phases were found (see Figure C.4 through C.6). In Figure C.7, an EDS map shows the presence of a tungsten-bearing phase within a matrix of the sodium salt. This was confirmed in Figure C.8 by multi-point EDS analysis.





Figure C.4. EDS Analysis and SEM Image of High Z Particles Indicating the Presence of a Ca-rich Phase Containing P and Si



Figure C.5. EDS Analysis and SEM Image of High Z Particles Indicating the Presence of a Ba Sulfate Phase



Figure C.6. EDS Analysis and SEM Image of High Z Particles Indicating the Presence of a Ca-rich Phase, Possibly Calcite



Figure C.7. EDS Mapping Showing the Presence of Na-phase and a W-containing Phase, Together with Particles of Fe and Cr



Figure C.8. Multi-point EDS Analysis and SEM Image of High Z Particles Indicating the Presence of W-rich Phases (1 and 2) and Na-P Bearing Phases (3)

Appendix D – CST Lot Comparison

Crystalline silicotitanate (CST) is manufactured by Honeywell UOP, LLC (UOP); product IONSIVTM R9140-B is provided by UOP already converted to the sodium form. Pacific Northwest National Laboratory recently tested both CST production lots with 5.6 M Na simulant (Russell et al. 2017) and Hanford tank waste collected from tank AP-107. The two CST production lots were identified as:

- Batch 2081000057
- Lot number 2002009604, material number 8056202-999

Summaries of the key characterization attributes are provided in the following data tables and figures for the two CST production lots. A sample of Batch 2081000057 CST was passed through a 25-mesh sieve and collected on a 60-mesh sieve to support batch contact testing and small-scale column processing in 1.44 cm inside diameter ion exchange columns. A sample of Lot 2002009604 was passed through a 25-mesh sieve to support batch contact testing and small-scale column processing removed large particles that could potentially lead to wall effects and flow maldistribution in the ion exchange column. Most characterizations were conducted on the sieved product materials. Results are also reported where characterization was conducted on unsieved material.

Table D.1 summarizes the CST chemical properties as measured by Campbell et al. (2019). The Nb concentration in Lot 2002009604 was notably higher (27 wt%) compared to that of Batch 2081000057. Chitra et al. (2013) demonstrated significant Cs uptake differences in CST as a function of the Nb content. The corresponding Si concentration in Lot 2002009604 was significantly lower than that of Batch 2081000057. The other analyte concentrations were essentially equal within analytical uncertainty.

	Batch 2081000057,	Lot 2002009604,	% difference
Analyte	>60/<25 mesh	<25 mesh	between lots
Na, wt%	9.76	9.38	-4%
Nb, wt%	9.86	12.5	+27%
Si, wt%	3.0	2.1	-30%
Ti, wt%	18.9	17.9	-5%
Zr, wt%	12.0	11.4	-5%
Per UOP product is wt%.	nformation, Si content should	be 7.8 to 8.2 wt% and Nb	content be 12 to 15

Table D.1. Chemical Composition (Campbell et al. 2019)

Table D.2 provides an overview of physical properties. Lot 2002009604 underwent more physical property testing than Batch 2081000057. Batch 2081000057 appeared to be provided as a 25 x 40 mesh sieve cut whereas Lot 2002009604 was more consistent with an 18 x 40 mesh sieve cut. Figure D.1 shows the volume percent particle size distribution (PSD) for two sieved samples. There were no appreciable differences in the sieved material PSDs.

	Batch 208	31000057	Lot 2002	2009604
Physical Properties	>60/<25 mesh	As-received	<25 mesh	As-received
wt% >25 mesh ^(a)	0	7.7	0	59
Particle density, g/mL	nm	nm	3.77	3.33
Pre-sonication cumulative	$d_{10}418$	$d_{10}419$	$d_{10}428$	$d_{10}467$
particle undersize fractions,	$d_{50}571$	$d_{50} 573$	$d_{50}589$	$d_{50}708$
volume basis, microns	d ₉₀ 775	$d_{90}779$	$d_{90} \ 809$	$d_{90} 1098$
Post-sonication cumulative	$d_{10}412$	$d_{10}412$	$d_{10}406$	$d_{10} 394$
particle undersize fractions,	$d_{50}563$	$d_{50} 563$	$d_{50}567$	$d_{50} 633$
volume basis, microns	$d_{90}765$	d90 765	d ₉₀ 795	d90 954
Surface area, $m^2/g^{(b)}$	nm	nm	108	104
Bulk density, g/mL ^(c)	1.00	nm	1.02	1.01
CST bed density, g/mL ^(c)	1.00	nm	1.01	1.00
Settled bed void volume, %	65.6	nm	66.2	67.6

Table D.2. Physical Properties

(a) 1.7 wt% passed through the 60-mesh sieve.

(b) Samples were dried to 105 °C prior to conducting measurement.

(c) Mass based on dry weight at 105 °C.

nm = not measured





Figure notes: CST Batch 2081000057 <25 mesh/ >60 mesh; Lot 2002009604 <25 mesh, Post-sonication results

The thermal gravimetric analysis was completed in duplicate for both lots of CST. Figure D.2 provides a comparison of the two CST lots tested. Rapid mass loss to 100 °C was observed and was associated directly with water loss. Per King et al. (2018), $Zr(OH)_4$ loses mass with the reorganization of $Zr(OH)_4$ to ZrO_2 and two H₂O when heated to 250 °C. Soliz et al. (2016) describe the mass losses of $Zr(OH)_4$ in phases where the initial mass loss up to 100 °C is associated with water loss and peaks "around 175 °C and between 400 to 475 °C in the derivative weight loss plot correlate to the loss of strongly bound hydroxyls and the formation of bridging oxygens, respectively, as the material begins to crystallize into ZrO_2 ." Thus, the extra mass loss associated with the peak around 200 °C is likely water associated with the decomposition of the $Zr(OH)_4$ binder to ZrO_2 and two waters. Given the CST formulation proposed by

Nyman et al. (2001), $H_2NaSi_2(Nb_{0.30}Ti_{0.7}O)_4O_{13}(OH) \cdot 4H_2O + 0.93Zr(OH)_4$, (FW 840.6 g/mole), an extra 4.3 wt% loss would be expected when the $Zr(OH)_4$ converts to water (loss of 36 g/mole) and ZrO_2 .

Both lots of CST were measured in duplicate for mass loss by thermogravimetric analysis. The temperature ramp was 10 °C/min from 20 to 800 °C under an Ar cover gas. Sample analysis size ranged from 20 to 30 mg. Figure D.2 provides the mass loss as a function of temperature along with the derivative of the mass loss for both lots of CST. Each lot was analyzed in duplicate and the thermogravimetric analysis scans for the duplicate measures were essentially identical. Three exotherms were found in equivalent positions for each production batch. The major exotherm occurred at 100 °C, a secondary exotherm at ~230 °C and a small, sharp exotherm at ~350 °C. The first exotherm was considered water loss from entrained water easily evaporated from the CST. The origins of the second exotherm may be associated with the destruction of $Zr(OH)_4$ binder. The genesis of the final exotherm is not understood at this time.



Figure D.2. Thermal Gravimetric Analysis of CST

The batch contact performances with a 5.6 M Na simulant [produced according to formulation provided by Russell et al. (2017)] and AP-107 tank waste were used to compare Cs load performance. Kinetic testing was not conducted with Batch 2081000057; however, >24 h was shown to be needed at low Cs concentrations to reach equilibrium. Lot 2002009604 was tested for kinetic exchange in batch contacts but not capacity testing. Direct comparison with the AP-107 tank waste matrix could be made between the two CST lots. At the equilibrium AP-107 tank waste matrix, the achieved capacity for Lot 2002009604 was shown to be 10% higher than that of Batch 2081000057. Table D.3 summarizes the batch contact performances.

	Batch 2081000057		Lot 2002009604	
Parameter	>60/<25 mesh	Reference	<25 mesh	Reference
5.6 M Na Simulant				
Time to equilibrium	Estimated 45 h	Fishum et al	Between 45 and 123 h	Fiskum et al. 2019
K _d at feed condition, mL/g	770 (45 h contact)	riskulli et al.	nm	
Achieved Cs load capacity at feed conditions, mg/g	6.2	2018	nm	
AP-107 Tank Waste				
Time to equilibrium	Estimated 48 h		Reached by 71 h	This report
K _d at feed condition, mL/g	669 (48-91 h)	Rovira et al.	806 (71 h)	This report
Achieved Cs load capacity at feed conditions, mL/g	7.5	2018	8.3	This report
Dry mass basis for CST is 100	°C.			

Table D.3. Batch Contact Summary

Column testing was conducted with both CST production lots in two matrix test solutions: 5.6 M Na simulant and AP-107 tank waste. Table D.4 provides a summary of the salient column performance parameters with effort to match general flow rate or residence time in terms of bed volumes per hour (BV/h).

	Batch		Lot		
	2081000057		2002009604		
Parameter	>60/<25 mesh	Reference	<25 mesh	Unseived	Reference
5.6 M Na Simulant					
Flowrate, BV/h	1.99		1.83	1.82	Eighum at
WAC limit breakthrough, BV	162	Fiskum et	240	240	riskulli et
50% Cs breakthrough, BV	800	al. 2018	~1000	~1050	al. 2019
Cs loading, mg/g	6.38		6.95	6.60	
AP-107 Tank Waste					
Flowrate, BV/h	2.20		1.88	Not tested	
WAC limit breakthrough, BV	120	Rovira et	200	Not tested	In this
50% Cs breakthrough, BV	~615	al. 2018	900	Not tested	report
Cs loading, mg/g	Not comparable		6.76	Not tested	
Values containing "~" are extrapol	ated values.				

Table D.4. Column Performance Summary

Radiolytic gas generation testing was recently conducted on the two CST lots in a side-by-side test (Colburn et al. 2019). The CST production lots were pretreated as follows:

- Wet CST: CST was soaked in 0.1 M NaOH for 1 h, soaked in 3 M NaOH overnight followed by a 1-h 0.1 M NaOH soak, and then soaked in deionized water for 1 h. The excess liquid was drained from the CST and the CST was sealed in a jar until the vessels were loaded for testing.
- Dry CST: The CST received the same treatment as the wet, free-drained CST, after which it was airdried on the lab bench until it reached a free-flowing state.

Nominal 30-mL aliquots of the CST samples were exposed to 588 krad/h for 549 h with an estimated total accumulated dose of 323 Mrad. Samples were maintained at 25 °C during the exposure period. Table D.5 provides the cumulative G-values (molecules/100 eV) for the two CST lots. It was not expected that

differences in PSD (sieved vs. unsieved) would alter gas production rates. Clearly, the CST Batch 2081000057 resulted in much higher H_2 and O_2 production rates than that of Lot 2002009604.

	Batch 2081000057	Lot 2002009604	Parameter Ratio Batch 2081000057
Parameter	Unsieved	Unsieved	/ Lot 2002009604
Wet CST			
Moisture, wt % (105 °C)	38.38	39.57	0.97
G(H ₂)	0.1259	0.0191	6.59
G(CH ₄)	No CH ₄ observed	$9.47 imes10^{-6}$	
G(O ₂)	0.0194	0.004	4.85
$G(N_2O)$	No N ₂ O observed	No N ₂ O observed	
Dry CST			
Moisture, wt % (160 °C)	22.45	11.12	2.02
G(H ₂)	0.3291	0.0894	3.68
G(CH ₄)	No CH ₄ observed	0.0003	
G(O ₂)	0.0689	No O2 observed	
$G(N_2O)$	0.0004	$9.1 imes 10^{-5}$	4.40

Table D.5. G Values, Molecules/100 eV (Colburn et al. 2019)

D.1 References

Chitra S, R Sudha, S Kalavathi, AGS Mani, SVS Rao, and PK. Sinha. 2013. "Optimization of Nb-substitution and Cs+/Sr+2 ion exchange in crystalline silicotitanates (CST)." *J. Radioanal. Nucl. Chem.* 295:607-613.

Campbell, EL, AM Rovira, F Colon-Cintron, D Boglaienko, TG Levitskaia, RA Peterson. 2019. *Characterization of Cs-Loaded CST Used for Treatment of Hanford Tank Waste in Support of Tank-Side Cesium Removal*. PNNL-28945, Rev. 0, RPT-TCT-005, Rev. 0, Pacific Northwest National Laboratory, Richland, Washington.

Colburn HA, AM Rovira, SR Adami, DM Camaioni, SA Bryan and PP Schonewill. 2019. *Gas Generation Testing of Crystalline Silicotitanate Ion Exchange Media*. PNNL-28842, Rev. 0; RPT-LPTTS-002, Rev. 0, Pacific Northwest National Laboratory, Richland, Washington.

Fiskum SK, HA Colburn, RA Peterson, AM Rovira and MR Smoot. 2018. *Cesium Ion Exchange Using Crystalline Silicotitanate with 5.6 M Sodium Simulant*. PNNL-27587, Rev. 0; RPT-DFTP-008, Rev. 0, Pacific Northwest National Laboratory, Richland, Washington.

Fiskum SK, AM Rovira, JR Allred, HA Colburn, MR Smoot, AM Carney, T Trang-Le, MG Cantaloub, EC Buck and RA Peterson. 2019. *Cesium Removal from Tank Waste Simulants Using Crystalline Silicotitanate at 12% and 100% TSCR Bed Heights*. PNNL-28527, Rev. 0; RPT-TCT-001, Rev. 0, Pacific Northwest National Laboratory, Richland, Washington.

King G, JR Soliz and WO Gordon. 2018. "Local Structure of Zr(OH)₄ and the Effect of Calcination Temperature from X-ray Pair Distribution Function Analysis." *Inorganic Chemistry* 57: 2797-2803.

Nyman M, TM Nenoff and TJ Headley. 2001. *Characterization of UOP IONSIV IE-911*. SAND2001-0999, Sandia National Laboratories, Albuquerque, New Mexico.

Rovira AM, SK Fiskum, HA Colburn, JR Allred, MR Smoot and RA Peterson. 2018. *Cesium Ion Exchange Testing Using Crystalline Silicotitanate with Hanford Tank Waste 241-AP-107*. PNNL-27706; RPT-DFTP-011, Rev. 0, Pacific Northwest National Laboratory, Richland, Washington.

Russell RL, PP Schonewill, and CA Burns. 2017. *Simulant Development for LAWPS Testing*. RPT-LPIST-001, Rev. 0; PNNL-26165, Rev. 0, Pacific Northwest National Laboratory, Richland, Washington.

Soliz JR, AD Klevitch, CR Harris, JA Rossin, A Ng, RM Stroud, AJ Hauser and GW Peterson. 2016. "Structural Impact on Dielectric Properties of Zirconia." *J. Phys. Chem. C* 120:26834-26840.

Appendix E – Analytical Reports

Analytical reports provided by the Analytical Support Operation (ASO) are included in this appendix. In addition to the analyte results, they define the procedures used for chemical separations and analysis, as well as quality control sample results, observations during analysis, and overall estimated uncertainties. The analyses are grouped according to Analytical Service Request (ASR) number. Cross reference of ASO sample IDs to test description are provided in the body of the report (see Table 3.2 and Table 3.5).

Appendix E Table of Contents

ASR 0747, Initial Characterization of AP-107, As-Received

•	ASR 0747 Rev. 2	E.2
•	ICP-OES, Metals	E.4
•	GEA, ¹³⁷ Cs	E.10
•	ICP-MS, ¹³³ Cs	E.14

ASR 0787, AP-107 Filtered Feed, Effluent, and Selected Lead Column Samples

•	ASR 0	787 E.19
•	ICP-OI	ES, Metals E.22
•	ICP-M	S, ²³⁸ U E.29
•	Titratio	on, Free Hydroxide E.34
•	TOC/T	E.42
•	GEA	
•	Radion	uclides
	0	⁹⁰ Sr Narrative E.50
	0	⁹⁹ Tc Narrative E.54
	0	²³⁷ Np Narrative E.57
	0	²³⁸ Pu, ^{239/240} Pu Narrative
	0	⁹⁰ Sr, ⁹⁹ Tc, ²³⁷ Np, ²³⁸ Pu, ^{239/240} Pu Data Summary E.63

PNNL-28958, Rev. 0 RPT-DFTP-013, Rev. 0

Analytical Service Request (ASR) (Information on this COVER PAGE is applicable to all samples submitted under this ASR)

Requestor Complete all fields on this CO	VER PAGE, unless specified as optional or ASR is a revision
Requestor:SignatureAmyRauiraPrint NameAmy RoviraPhone371-7908MSIN	Project Number: 73312 Work Package: NC4186
Matrix Type Information	QA/Special Requirements
 ◆ Liquids: ⊠ Aqueous □ Organic □ Multi-phas ◆ Solids: □ Soil □ Sludge □ Sediment □ Glass □ Filter □ Metal □ Smear □ Organic □ Other ◆ Other: □ Solid/Liquid Mixture, Slurry □ Con □ Dialogical Speciment 	 Se QA Plan: Se ASO-QAP-001 (Equivalent to HASQARD) □ Additional QA Requirements, List Document Below: Reference Doc Number: ♦ Field COC Submitted? INO □ Yes ♦ Lab COC Required? INO □ Yes
Gas Biological Specimen	◆ Sample/Container Inspection Documentation Required?
(If sample matrices vary, specify on Request Page)	◆ Hold Time: ⊠/No □ Yes
 Disposal Information Disposition of Virgin Samples: Virgin samples are returned to requestor unless archiving provisions are made with receiving group! If archiving, provide: Archiving Reference Doc: Disposition of Treated Samples: Dispose 	If Yes. Contact ASO □ Use SW 846 (PNL-ASO-071, identify Lead before analytes/methods where holding times apply) submitting
	ta Reporting Information
 ◆ Is Work Associated with a Fee-Based Milestone? ☑ No □ Yes If yes, milestone due date: △ Minimun □ Project S ◆ Data Rep ☑ ASO-QA HASQAI ○ Minimun □ Project S Contact ASC Document: 	Requested Analytical Work Completion Date: P-001 (Equivalent to RD). n data report. pecific Requirements: D Lead or List Reference (Note: Priority rate charge for < 10 business day turn-around time) Negotiated Commitment Date: (To be completed by ASO Lead)
Was	ste Designation Information
 ◆ ASO Sample Information Check List Attached? ▲N If no, Reference Doc Attached:	Io □ Yes Does the Waste Designation Documentation Indicate Presence of PCBs? Image: No □ Yes
Additional or Special Instructions	MSIN
Receiving and Login	Information (to be completed by ASO staff)
Date Delivered: 3 18 9 Delivered By (optional) 1:00 pm Time Delivered: 1:00 pm Group ID (optional) Yes	ASR Number: RPL Numbers: (first and last) Received By: ASR Number: RPL Numbers: (first and last)
ASO Work Accepted By: KN Pal	Signature/Date: Lull. Fack 3/18/19

PNNL-28958, Rev. 0 RPT-DFTP-013, Rev. 0

Analytical Service Request (ASR)

(REQUEST PAGE ----- Information Specific to Individual Samples)

ASO Staff Use Only	Provide Analytes of	Interest and Required Detection	on limits - 🗆 Below 🗖 Attached	ASO Staf	f Use Only
RPL Number	Customer Sample ID	Sample Description (& Matrix, if it varies)	Analysis Requested	Test	Library
19-1227	7AP-18-36	AP-107 tank supernate	(CS-137, AM-241) 2) acia digestion - 128 a) ICP/OES-AI, K, and Na b) ICP/MS-CS-133 (isotopes)		
	 	Project expectations for ASO to ma copy records showing evidence of A off and Technical Reviewer signatu have been suspended for this ASR. relieved from meeting these require the project assumes responsibility f records.	intain hard Analyst sign- re approval ASO is ments and or the		
	· · · · · · · · · · · · · · · · · · ·				

Battelle PNNL/RPL/Inorganic Analysis ... ICP-OES Analysis Report PO Box 999, Richland, Washington 99352

Project / WP#:	73312 / NC4186
ASR#:	0747
Client:	A. Rovira
Total Samples:	2 (liquids)
Sample Description	AP-107 Tank Waste

Sample ID		Client Sample D	D		Sample Weight (g)
19-1227	7AP-18-1				NA
19-1228	7AP-18-36				NA
Sample Pro Metals Ana dilution of '	eparation: RPG-CM lysis Using a Dry-Blo 'as received" samples	IC-128, Rev. 1. "Hock Heater", perfo	HNO3-HCl Aci	d Extraction of L rnell on 04/17/19.	iquids for Simple
	-	5 111 5 70 474 111 103	performed by	J. Carter on 04/23	0/19.
Procedure:	RPG-CMC-211, I Inductively Coupl	Rev. 4, "Determined Argon Plasma	nation of Elema Optical Emiss	ental Composition ion Spectrometry	n by (ICP-OES)."
Procedure: Analyst:	RPG-CMC-211, I Inductively Coupl J. Carter	Rev. 4, "Determined Argon Plasma Analysis Date:	nation of Elemo Optical Emiss 04/25/19	ental Composition ion Spectrometry ICP File:	n by (ICP-OES)." C0789
Procedure: Analyst: See Chemie	RPG-CMC-211, I Inductively Coupl J. Carter cal Measurement Co	Rev. 4, "Determined Argon Plasma Analysis Date: enter 98620 file:	nation of Elemo Optical Emiss 04/25/19 <u>ICP-325-405</u> (Calibration	ion Spectrometry ICP File: i-3 and Maintenance	r by (ICP-OES)." C0789 Records)
Procedure: Analyst: See Chemic M&TE:	RPG-CMC-211, I Inductively Coupl J. Carter cal Measurement Co	Rev. 4, "Determined Argon Plasma Analysis Date: enter 98620 file: 00DV ICP-OES	nation of Elemo Optical Emiss 04/25/19 <u>ICP-325-405</u> (Calibration	ion Spectrometry ICP File: i-3 and Maintenance SN: 077N5122	n by (ICP-OES)." C0789 Records) 002
Procedure: Analyst: See Chemie M&TE:	RPG-CMC-211, I Inductively Coupl J. Carter Cal Measurement Co PerkinElmer 53	Rev. 4, "Determined Argon Plasma Analysis Date: enter 98620 file: 00DV ICP-OES Balance	nation of Elemo Optical Emiss 04/25/19 <u>ICP-325-405</u> (Calibration	ion Spectrometry ICP File: 3-3 and Maintenance SN: 077N5122 SN: 111329266	(ICP-OES)." (ICP-OES)." C0789 Records) 002 67
Procedure: Analyst: See Chemic M&TE:	RPG-CMC-211, I Inductively Coupl J. Carter Cal Measurement Ce PerkinElmer 53 Mettler AT400 Ohaus PA224C	Rev. 4, "Determined Argon Plasma Analysis Date: enter 98620 file: 00DV ICP-OES Balance Balance	nation of Elemon Optical Emiss 04/25/19 <u>ICP-325-405</u> (Calibration	ion Spectrometry ICP File: ion Maintenance SN: 077N5122 SN: 111329266 SN: B7252877	n by (ICP-OES)." C0789 Records) 002 67 90
Procedure: Analyst: See Chemie M&TE:	RPG-CMC-211, I Inductively Coupl J. Carter cal Measurement Ce PerkinElmer 53 Mettler AT400 Ohaus PA224C Sartorius ME41	Rev. 4, "Determined Argon Plasma Analysis Date: enter 98620 file: 00DV ICP-OES Balance Balance 4S Balance	nation of Elemo Optical Emiss 04/25/19 <u>ICP-325-405</u> (Calibration	intal Composition ion Spectrometry ICP File: 3-3 and Maintenance SN: 077N5122 SN: 111329266 SN: B7252877 SN: 21308482	n by (ICP-OES)." C0789 Records) 002 67 90

Report Preparer

ens

Review and Concurrence

Date

05 01

Date
Two aqueous samples submitted under Analytical Service Request (ASR) 0747 were analyzed by ICP-OES. The samples were prepared following RPL procedure RPG-CMC-128, Rev. 1, *"HNO3-HCl Acid Extraction of Liquids for Metals Analysis Using a Dry-Block Heater"*. All samples were diluted in 5% HNO₃ prior to analysis. None of the samples were filtered.

All sample results are reported on a mass per unit volume basis (μ g/mL) for each detected analyte. The data have been adjusted for instrument dilutions.

Analytes of interest (AOI) were specified in the ASR and are listed in the upper section of the attached ICP-OES Data Report. There were two analyte lists requested. Samples were divided between those two lists and a separate data report is provided for each. The quality control (QC) results for the AOI have been evaluated and are presented below. Analytes other than the AOI are reported in the bottom section of the report but have not been fully evaluated for QC performance.

Calibration of the ICP-OES was done following the manufacturer's recommended calibration procedure using multi-analyte custom standard solutions traceable to the National Institute of Standards and Technology (NIST). Midrange calibration verification standards (MCVA and MCVB) were used to verify acceptance of the two-point calibration curves obtained for each analyte and for continuing calibration verification.

The controlling documents were procedures RPG-CMC-211, Rev. 4, *Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP-OES)*, and ASO-QAP-001, Rev. 11, *Analytical Support Operations (ASO) Quality Assurance Plan*. Instrument calibrations, QC checks and blanks (e.g., ICV/ICB, CCV/CCB, LLS, ICS), matrix spike, post-digestion spikes, duplicate, reagent spike, blank spike, and serial dilution were conducted during the analysis run.

Preparation Blank (PB):

A preparation blank (reagents only) was prepared for the extraction process. All analytes were within the acceptance criteria of \leq EQL (estimated quantitation level), \leq 50% regulatory decision level, or \leq 10% of the concentration in the samples.

Blank Spike (BS)/Laboratory Control Sample (LCS):

A reagent spike (RS) sample (reagents and spikes) was prepared for the extraction process. Recovery values are listed for all analytes included in the RS that were measured at or above the EQL. Recovery values for the AOI meeting this requirement and within the acceptance criterion of 80% to 120%, ranged from 99% to 104%.

Duplicate/Replicate Relative Percent Difference (RPD)/Relative Standard Deviation (RSD):

A duplicate of sample 19-1227 was prepared and analyzed. RPD's are listed for all analytes that were measured at or above the EQL. RPDs for the AOI meeting this requirement ranged from 2.5% to 4.1% and were within the acceptance criterion of $\leq 20\%$ for liquid samples.

Matrix-Spike (MS) Sample:

A matrix spike (MS) of sample 19-1228 was prepared for the extraction process. Recovery values are listed for all analytes included in the MS that were measured at or above the EQL. Recovery values for the AOI meeting this requirement was 107%, within the acceptance criterion of 75% to 125%, except for K (144%). The recovery for Na is not reported (nr) as the spike concentration was less than 25% of the sample.

Initial/Continuing Calibration Verification (ICV/CCV):

MCVA and MCVB solutions were analyzed immediately after calibration, after each group of not more than ten samples, and at the end of the analytical run. The concentrations of all AOI were within the acceptance criteria of 90% to 110% recovery.

Initial/Continuing Calibration Blank (ICB/CCB):

The ICB/CCB solution (5% v/v HNO₃) was analyzed immediately after the ICV solutions and after the CCV solutions (after each group of not more than ten samples and at the end of the analytical run). The concentration of all AOI were within the acceptance criteria of \leq EQL.

Low-Level Standard (LLS):

The LLS solution was analyzed immediately after the first CCB solution. The concentrations of all AOI were within the acceptance criteria of 70% to 130% recovery.

Interference Check Standard (ICS/SST):

The ICS solution was analyzed immediately after the first LLS solution and immediately prior to analyzing the final CCV solutions. The concentrations of all AOI were within the acceptance criteria of 80% to 120% recovery.

Serial Dilution (SD):

Five-fold serial dilution was conducted on sample 19-1227. Percent differences (%Ds) are listed for all analytes that had a concentration at or above the EQL in the diluted sample. The %Ds for the AOI meeting this requirement ranged from 1.4% to 9.5% and were within the acceptance criterion of $\leq 10\%$.

Post-Digestion Spike (PS-A) - Sample (A Component):

In addition to the BS sample, a post-digestion spike (A Component) was conducted on sample 19-1227. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration $\geq 25\%$ of that in the sample. Recovery values for the AOI meeting this requirement ranged from 109% to 112% and were within the acceptance criterion of 80% to 120%.

Post-Digestion Spike (PS-B) - Sample (B Component):

In addition to the MS sample, a post-digestion spike (B Component) was conducted on sample 19-1227. No AOIs are present in this spike.

Other QC:

All other instrument-related QC tests for the AOI passed within their respective acceptance criteria.

Comments:

- 1) The "Final Results" have been corrected for all laboratory dilutions performed on the samples during processing and analysis, unless specifically noted.
- 2) Instrument detection limits (IDL) and estimated quantitation limits (EQL) shown are for acidified water and/or fusion flux matrices as applicable. Method detection limits (MDL) for individual samples can be estimated by multiplying the IDL by the "Process Factor" for that individual sample. The estimated quantitation limit (EQL) for each concentration value can be obtained by multiplying the EQL by the "Process Factor".
- 3) Routine precision and bias is typically ±15% or better for samples in dilute, acidified water (e.g. 5% v/v HNO₃ or less) at analyte concentrations > EQL up to the upper calibration level. This also presumes that the total dissolved solids concentration in the sample is less than 5000 µg/mL (0.5 per cent by weight). Note that bracketed values listed in the data report are within the MDL and the EQL, and have potential uncertainties greater than 15%. Concentration values < MDL are listed as "--". Note, that calibration and QC standard samples are validated to a precision of ±10%.</p>
- 4) Analytes included in the spike A component (for the AS/PS) are; Ag, Al, As, B, Ba, Be, Bi, Ca, Cd, Co, Cr, Cu, Fe, K, Li, Mg, Mn, Mo, Na, Ni, P, Pb, Sb, Se, Si, Sn, Sr, Ta, Ti, Tl, V, W, Y, Zn, and Zr. Analytes included in the spike B component are; Ce, Dy, Eu, La, Nd, Pd, Rh, Ru, S, Te, Th, and U.

PNNL-28958, Rev. 0 RPT-DFTP-013, Rev. 0

Battelle PNNL/RPG/Inorganic Analysis ... ICPOES Data Report

Page 1 of 2

		Run Date >	4/25/2019	4/25/2019	4/25/2019	4/25/2019	4/25/2019
		Factor >	1.0	49.2	1230.0	1227.5	1226.0
			405 diluent	BLK-1227	19-1227 @ 25x	19-1228 @ 25x	Dup-1228 @ 25x
Instr. Det. Limit (IDL)	Est. Quant. Limit (EQL)	Client ID >	Lab Diluent	Process Blank	<u>7AP-18-1</u>	7AP-	18-36
(µg/mL)	(µg/mL)	(Analyte)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)
0.0122	0.122	Al		[5.0]	9.730	10.200	10.600
0.0240	0.240	к		[5.9]	3,620	3,820	3,920
0.0246	0.246	Na			129.000	135,000	140.000
Other Analyte	es						
0.0019	0.019	Aq		[0.10]			[2.6]
0.0529	0.529	As			[110]	[68]	
0.0067	0.067	В	[0.036]	[2,1]	[60]	[54]	[56]
0.0003	0.003	Ba		[0,15]	10,901	[0.80]	[0.65]
0.0001	0.001	Be			[0,19]	[0.13]	IO 161
0.0494	0.494	Ri			(arred	Ter tel	Ferred
0.0177	0.177	Ca		[1.1]	[36]		
0.0022	0.022	Cd		Lod	17 61	[5.6]	10.91
0.0069	0.022	Co	10,00801	-	[7.0]	[3.5]	[0.3]
0.0035	0.005	Co	[0.0000]		1		
0.0035	0.035	00			502	E47	£22
0.0025	0.025	Ci		100.001	502	517	532
0.0016	0.016	Du		[0.20]	[2.3]	[3.0]	[2.1]
0.0029	0.029	Dy					
0.0013	0.013	EU					
0.0052	0.052	Fe		[0.48]	[13]	[14]	[13]
0.0022	0.022	La					
0.0009	0.009	Li		[0.28]	[1.4]		
0.0068	0.068	Mg		[0.46]			
0.0002	0.002	Mn		[0.031]			[0.33]
0.0045	0.045	Мо		-	[43]	[39]	[40]
0.0086	0.086	Nd					
0.0049	0.049	Ni			[21]	[20]	[24]
0.0554	0.554	Р			800	866	825
0.0162	0.162	Pb		-			
0.0077	0.077	Pd		[0.41]			
0.0165	0.165	Rh					**
0.0097	0.097	Ru					
0.1186	1.186	S			1,730	1,880	1,900
0.0518	0.518	Sb					
0.1595	1.595	Se			inte		
0.0237	0.237	Si			[84]	[94]	[100]
0.0313	0.313	Sn					
0.0005	0.005	Sr					
0.0217	0.217	Та					
0.0311	0.311	Те					
0.0189	0.189	Th	-				**
0.0004	0.004	Ti		-			-
0.0530	0.530	TI					
0.0360	0.360	U			****		
0.0021	0.021	V					[2.7]
0.0216	0.216	W			[61]	[60]	[66]
0.0012	0.012	Y					
0.0043	0.043	Zn		3.81			
0.0019	0.019	Zr					

1) "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the "multiplier"

near the top of each column. The estimated sample quantitation limit = EQL (in Column 2)

times the "multiplier". Overall error for values \geq EQL is estimated to be within ±15%. 2) Values in brackets [] are \geq MDL but < EQL, with errors likely to exceed 15%.

Battelle PNNL/RPG/Inorganic Analysis ... ICPOES Data Report

Page 2 of 2

QC Performance 4/25/2019							
Criteria >	≤ 20%	80%-120%	75%-125%	80%-120%	80%-120%	≤ 10%	
QC ID >	19-1227 Dup	BS	19-1227 MS	19-1227 + AS-A	19-1227 + AS-B	19-1227 5-fold Serial Dil	
Analytes	RPD (%)	%Rec	%Rec	%Rec	%Rec	%Diff	
Al	2.5	104	107	109		2.1	
к	4.1	99	144	112		9.5	
Na	2.6	104	nr	112		1.4	
Other Analy	tes						
Ag				92			
As				103			
В				111			
Ba		101	102	104			
Be		103	105	106			
Bi		84		97			
Ca		107	123	111			
Cd		103	107	105			
Ce		97	96		104		
Co				101			
Cr	0.4	95	96	104		3.3	
Cu		100	102	106			
Dy					103		
Eu					103		
Fe		94	107	105			
La		97	101		101		
Li		117	136	121			
Mg		101	105	106			
Mn		101	106	108			
Mo				101			
Nd		97	100		102		
Ni		97	103	103			
Р	3.2			104			
Pb		100	104	101			
Pd					98		
Rh					103		
Ru					100		
S	2.1				101		
Sb				104			
Se				102			
Si				108		•	
Sn				104			
Sr		105	106	107			
Ta				102			
Te					105		
Th					101		
Ti				105			
TI				95			
		99	102		106		
v		94	99	101			
w				105			
				102			
Zn		98	107	105			
7r			107	106			
L. <u>-</u>				100			

Shaded results are outside the acceptance criteria.

nr = spike concentration less than 25% of sample concentration. Matrix effects can be assessed from the serial dilution.

Battelle PNNL/RPL/ASO Radiochemistry Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

Gamma Energy Analysis (GEA)

Project / WP#:	73312/ NC4186
ASR#:	0747.00
Client:	A Rovira
Total Samples:	2

RPL ID	Client Sample ID
19-1227	7AP-18-1
19-1228	7AP-18-36

Analysis Type:	GEA- for all positively measured or non-detected isotopes			
Sample Processing Prior to	⊠ None			
Radiochemical Processing/Analysis	Digested as per RPG-CMC-129, Rev. 0 HNO ₃ -HCl Acid Extraction of Solids Using a Dry Block Heater			
	Fusion as per RPG-CMC-115, Solubilization of Metals from Solids Using a KOH-KNO ₃ Fusion			
	Other:			
	Preparation may also involve attaining a GEA geometry that is compatible with the calibration geometry.			
Analysis Procedure:	RPG-CMC-450, Rev. 3 Gamma Energy Analysis (GEA) and Low-Energy Photon Spectrometry (LEPS)			
Reference Date:	3/18/2019			
Analysis Date or Date Range:	March 18, 2019			
Technician/Analyst:	T Trang-Le			
Rad Chem Electronic Data File:	19-1227 Rovira.xls			
ASO Project 98620 File:	File Plan 5872, Technical (Radiochemistry), Gamma Calibration, daily checks, and maintenance records; and standard certificates and preparation. Also, balance calibration and performance check records.			
M&TE Number(s):	Detectors T			

3/21/19 I IRang-le Prepare Date

<u>ARGuenard 1 3/21/19</u> Reviewer Day

Battelle PNNL/RSE/ASO Radiochemistry Analysis Report

SAMPLE RESULTS

Activities for all gamma emitters detected in these samples are presented in an attached Excel spreadsheet for 0747.00. Cs-137 was the main activities for these samples.

ASO Project File, ASR 0747.00 has been created for this report including all appropriate supporting records which may include the Pipette Performance Check Worksheet form, standard certificates, laboratory bench records, Shielded Analytical Laboratory Bench Sheet, and Gamma Energy Analysis printouts. Detector calibration records, control charts and balance calibration records can be found in the ASO Records.

Sample Preparation, Separation, Mounting and Counting Methods

The quality control (QC) sample results for each of the isotopes measured above background have been evaluated and are discussed below. A summary of the GEA analysis results, including QC sample performance, is given in the attached Data Report.

QUALITY CONTROL RESULTS

Tracer:

Tracers are not used for ASO Gamma Energy Analysis (GEA) methods.

Process Blank (PB):

No process blank was prepared for gamma counting.

Required Detection Limits

There are no required detection limit requirements for these samples.

Blank Spike (BS)/Laboratory Control Sample (LCS)/ Matrix Spike (MS):

There are no BS, LCS or MS QC samples analyzed for ASO GEA analyses. Instrument performance is assessed by the analyses of daily control counts and weekly background counts, as discussed below.

Duplicate Relative Percent Difference (RPD):

No duplicate samples were provided for gamma counting.

Instrument Calibration and Quality Control

Gamma detectors are calibrated using multi-isotope standards that are NIST-traceable and prepared in the identical counting geometry to all samples and detectors, if possible. Counter control sources containing Am-241, Cs-137 and Co-60 are analyzed daily before the use of each detector. Procedure RPG-CMC-450 requires that a counter control source is checked daily and

Battelle PNNL/RSE/ASO Radiochemistry Analysis Report

must be within ± 3 sigma or $\pm 3\%$ of the control value, whichever is greater. Gamma counting was not performed unless the control counts were within the required limits. Background counts are performed on all gamma detectors at least weekly for either an overnight or weekend count within operational limits, and/or when system background is thought to have changed. The most recent background is subtracted from all sample counts.

The samples were prepared in a 2 mL geometry from client.

Assumptions and Limitations of the Data

None

Interferences/Resolution

None.

Uncertainty

For gamma counting, the uncertainty in the counting data, photon abundance and the nuclear half-life, and efficiency are included in the calculation of the total uncertainty along with a systematic uncertainty for sample prep. The Canberra Genie software includes both random and systematic uncertainties in the calculation of the total uncertainties which are listed on the report. We conservatively estimate that 2% is the lowest uncertainty possible for our GEA measurements taking into account systematic uncertainties in gamma calibration standards. This uncertainty assumes the sample geometry matches the calibration geometry, and is undoubtedly larger when the sample geometry does not match a calibration geometry.

Comments

None.

Attachment: Data Report Sample Results for ASR 0747.00.

Pacific Northwest National Laboratory PO Box 999, Richland, WA Radiochemical Sciences and Engineering Group

2.80E+02

± 2%

2.82E+02

filename: 19-1227 Rovira 3/20/2019

Client: Rovira	Project: 73312	Prepared by:	1 Irang-le 3/20/19
ASR 0747	NC4186	Technical Reviewer	2R 4 100 3/20/19
			o voo camero o o o o
Procedures:	RPG-CMC-450, Rev. 3 G Spectrometry	amma Energy Analysis (GE	A) and Low-Energy Photon Spectrometry (LEPS)
M&TE:	Gamma detectors T		
Reference dates:	3/18/2019 @ 1:00 pm		
	Measured Activity,	μCi per Sample ± 1s	
RPL ID:	19-1227	19-1228	
Sample ID:	7AP-18-1	7AP-18-36	
Isotope			

 $\pm 2\%$

Cs-137

Battelle PNNL/RPL/Inorganic Analysis ... ICP-MS Analysis Report PO Box 999, Richland, Washington 99352

Project / WP#:	73312 / NC4186
ASR#:	0747.00
Client:	A. Rovira
Total Samples:	2 (Aqueous)

ASO Sample ID	Client Sample ID	Client Sample Description	Sample Weight (g)
19-1227	7AP-18-1	AP-107 Tank Supernate	NA
19-1228	7AP-18-36	AP-107 Tank Supernate	NA

Sample Preparation: Acid digestion was performed according to RPG-CMC-128, "HNO₃-HCl Acid Extraction of Liquids for Metals Analysis Using a Dry-Block Heater", performed by L. Darnell between 04/16/19. Sample were diluted prior to ICP-MS analysis in 2% v/v HNO₃ performed by S.S. Morrison on 04/19/19.

Procedure	rocedure: <u>RPG-CMC-292, Rev. 1</u> , "Determination of Elemental Composition by Inductively Coupled Argon Plasma Mass Spectrometry (ICP-MS)."								
Analyst:	S.S. Morrison	9/2019	ICP Fil	e: M0154					
See Chemi	- <u>325-405-3</u> ibration and l	Maintena	nce Records)						
M&TE:	PerkinElmer Nex	LION TM 350X ICP	-MS	SN: 85VN4	070702	RPL 405			
	Ohaus PA224C			SN: B72528	87790	RPL 405			
	Mettler AT400 E	Balance		SN: M1944	5	RPL 405 FH			
. [Mettler AT400 E	Balance		SN: 1113292667 RJ		RPL 420 FH			
	Ohaus EX324 Balance				1209	SAL Cell 2			
	Sartorius BA310	5 Balance		SN: 108032	10	RPL 309			
	Sartorius R200D	Balance		SN: 390800	58	RPL 525 FH			

Report Preparer

CHIS

Review and Concurrence

04/26/2019

Date

04 2019 26

Date

Two aqueous samples submitted under Analytical Service Request (ASR) 0747.00 were analyzed by ICP-MS. Prior to analysis all samples were diluted in 2% HNO₃. None of the samples were filtered.

All results are reported on a mass per unit volume basis (ng/mL) for each detected analyte. The data have been adjusted for instrument dilutions.

Analytes of interest (AOI) were specified in the ASR and are listed in the upper section of the attached ICP-MS Data Report. Cesium-133, Cesium-135, and Cesium-137 were the AOI. The quality control (QC) results for the AOI have been evaluated and are presented below.

Calibration of the ICP-MS was done following the manufacturer's recommended calibration procedure using multi-analyte custom standard solutions traceable to the National Institute of Standards and Technology (NIST). Midrange calibration verification standards were used to verify acceptance of the five-point calibration curve and for initial and continuing calibration verification (ICV/CCV).

The controlling document was procedure RPG-CMC-292, Rev 1, *Determination of Elemental Composition by Inductively Coupled Argon Plasma Mass Spectrometry (ICP-MS.* Instrument calibrations, QC checks and blanks (e.g., ICV/ICB, CCV/CCB, LLS, ICS), post-digestion spike, duplicate, blank spike, and serial dilution were conducted during the analysis run.

Internal Standard (IS):

All solutions (blanks, standards, QC checks, and samples) were mixed in-line with a solution containing 10 ppb of Bi-209 as the internal standard (IS). The AOI (m/z 133, 135, and 137) data were normalized using the data for the IS Bi-209). The Bi-209 IS recoveries ranged from 91.0% to 110.4% for the entire analysis sequence, which were within the acceptance criterion of 30% to 120%.

Preparation Blank (PB):

A preparation blank was prepared according to RPG-CMC-128, and labeled as BLK-1227 and diluted in 2% HNO₃, this solution was analyzed as a preparation blank. Results for the diluent blank were within the acceptance criteria of less than $\leq 10\%$ of the concentration in the samples.

Blank Spike (BS)/Laboratory Control Sample (LCS):

A blank spike (BS) sample was prepared by spiking the preparation blank (BLK-1227) with an equivalent volume of the CCV-71A-2ppb standard (1:1 ratio). The recovery for the m/z 133 (98%), 135 (95%), 137 (95%) were within the acceptance criteria of 80% to 120% recovery.

Duplicate/Replicate Relative Percent Difference (RPD)/Relative Standard Deviation (RSD): Duplicate of sample 19-1228 was prepared during dissolution according to RPG-CMC-128, the solution was diluted and analyzed. RPD are listed for all analytes that were measured at or above the EQL. RPD for the m/z 133, 135, and 137 were 4%, 2%, and 4%, respectively; which were within the acceptance criterion of ≤20% for liquid samples.

Matrix-Spike (MS) Sample:

The sample was not analyzed because there is no Cs tracer in the matrix spike solution and the analysis would not speak to matrix effects within the sample. Instead, a post spike sample was prepared and analyzed results reported in the post digestion spike section.

Initial/Continuing Calibration Verification (ICV/CCV):

The ICV/CCV solution was analyzed immediately after calibration, after each group of not more than ten samples, and at the end of the analytical run. The concentrations of all AOI were within the acceptance criteria of 90% to 110% recovery.

Initial/Continuing Calibration Blank (ICB/CCB):

The ICB/CCB solution ($2\% \text{ v/v} \text{HNO}_3$) was analyzed immediately after the ICV solutions and after the CCV solutions (after each group of not more than ten samples and at the end of the analytical run). The concentration of all AOI were within the acceptance criteria of <5% of sample.

Low-Level Standard (LLS):

The LLS solution was analyzed immediately after the first CCB. The recovery for m/z 133, 135, and 137 were 103%, 102%, and 99% respectfully; all were within the acceptance criteria of 75% to 125%.

Interference Check Standard (ICS):

The ICS solution was analyzed immediately after the LLS solution and immediately after the final CCV solution. The recoveries for m/z 133 (105%, 105%), m/z 135 (102%, 102%), and m/z 137 (100%, 101%) which were within the acceptance criteria of 80% to 120%.

Serial Dilution (SD):

Five-fold serial dilution was conducted on samples 19-1228. Percent differences (%D) are listed for all analytes that had a concentration at or above the EQL in the diluted sample. The %D for the AOI m/z 133 (6%), m/z 135 (2%), and m/z 137 (9%) were within the acceptance criterion of \leq 10%, when analytes were at >10xEQL.

Post-Digestion Spike (PS-71A)/Analytical Spike (AS-71A) - Sample (71A Component):

Because no MS sample was analyzed, a post-digestion spike (PS-71A) was conducted on samples 19-1228. The recovery is listed for all analytes in the spike that was measured at or above the EQL, and that had a spike concentration $\geq 25\%$ of that in the sample. Recovery values for m/z 133, 135 and 137 were 102%, 106%, and 104% respectively, and were within the acceptance criterion of 75% to 125%.

Other QC:

All other instrument-related QC tests for the AOI passed within their respective acceptance criteria.

Comments:

- 1) The "Final Results" have been corrected for all laboratory dilutions performed on the samples during processing and analysis, unless specifically noted.
- 2) Instrument detection limits (IDL) and estimated quantitation limits (EQL) shown are for acidified water and/or fusion flux matrices as applicable. Method detection limits (MDL) for individual samples can be estimated by multiplying the IDL by the "Process Factor" for that individual sample. The estimated quantitation limit (EQL) for each concentration value can be obtained by multiplying the EQL by the "Process Factor".
- 3) Routine precision and bias is typically ±15% or better for samples in dilute, acidified water (e.g. 2% v/v HNO₃ or less) at analyte concentrations > EQL up to the upper calibration level. This also presumes that the total dissolved solids concentration in the sample is less than 5000 µg/mL (0.5 per cent by weight). Note that bracketed values listed in the data report are within the MDL and the EQL, and have potential uncertainties greater than 15%. Concentration values < MDL are listed as "--". Note, that calibration and QC standard samples are validated to a precision of ±10%.</p>
- 4) Analytes included in the spike 71A component (for the AS/PS) are; Ag, Al, As, B, Ba, Be, Ca, Cd, Ce, Co, Cr, Cs, Cu, Dy, Er, Eu, Fe, Ga, Gd, Ho, K, La, Lu, Mg, Mn, Na, Nd, Ni, P, Pb, Pr, Rb, S, Se, Sm, Sr, Th, Tl, Tm, U, V, Yb, and Zn. Analytes included in the spike 71B component are; Ge, Hf, Mo, Nb, Sb, Si, Sn, Ta, Te, Ti, W, and Zr. Analytes included in the spike 71C component are; Ir, Os, Pd, Pt, Re, Rh, and Ru. Analytes included in the spike 71D component are; Bi, In, Li, Sc, Tb, and Y. Analytes included in the spike Hg component are; Hg.

		Run Date >	04/19/19	04/19/19	04/19/19	04/19/19
		Process				
		Factor >	43207	42808	42660	42346
		RPL/LAB >	BLK-1227	19-1227	19-1228	DUP-1228
Instr. Det. Limit (IDL)	Est. Quant. Limit (EQL)	Client ID >	Process Reagent Blank	7AP-18-1	7AP-18-36	7AP-18-36
(ng/mL)	(ng/mL)	(Analyte)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)
2.35E-04	2.35E-03	Cs 133	1.45E+01	5.96E+03	6.04E+03	6.21E+03
6.41E-05	6.41E-04	Cs 135	4.74E+01	1.71E+03	1.76E+03	1.78E+03
7.56E-05	7.56E-04	Cs 137	7.67E+01	1.96E+03	2.03E+03	2.09E+03

Internal Standard % Recovery

-					
	Bi 209 (IS)	95%	102%	99%	97%

1) "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the "multiplier"

near the top of each column. The estimated sample quantitation limit = EQL (in Column 2)

times the "multiplier". Overall error for values \geq EQL is estimated to be within ±15%.

Values in brackets [] are ≥ MDL but < EQL, with errors likely to exceed 15%.

IS = Internal Standard. The concentration of certain elements cannot be determined due to the presence of the IS in all solutions.

QC Performance 04/19/2019

Criteria >	≤ 20%	80%-120%	75%-125%	75%-125%	≤ 10%
QC ID >	DUP-1228	BS/LCS	MS (None)	19-1228 PSA	19-1228 5-fold Serial Dil
Analytes	RPD (%)	%Rec	%Rec	%Rec	%Diff
Cs 133	3.7%	98%	NA	102%	5.7%
Cs 135	2.0%	95%	NA	106%	2.2%
Cs 137	3.4%	95%	NA	104%	9.3%

Standard % Recovery

Bi 209 (IS)	97%	101%	NA	102%	99%	
						7

Shaded results are outside the acceptance criteria.

nr = spike concentration less than 25% of sample concentration. Matrix effects can be assessed from the serial dilution.

na = not applicable; KOH flux and Ni crucible or Na2O2 flux and Zr crucible for fusion preparations, or Si for HF assisted digests.

IS = Internal Standard. The concentration of certain elements cannot be determined due to the presence of the IS in all solutions.

NM = INot measured. The isotope was not measure due to method or molecular interference limitations.

PNNL-28958, Rev. 0 RPT-DFTP-013, Rev. 0

An (Information on thi	alytical Servies s COVER PAGE is appli	ce Request	(ASR) submitted under this ASR)				
Requestor Complete all fields	on this COVER PA	AGE, unless spe	ecified as optional of	or ASR is a revision				
Requestor:SignaturePrint NamePhone375-5677	r 5. Fiskum MSINP7-25	Project Nun Work Packa	nber:73312 age:1	NC4189				
Matrix Type Information			OA/Special Requi	irements				
 ◆ Liquids: X Aqueous □ Organic ◆ Solids: □ Soil □ Sludge □ Glass □ Filter □ Smear □ Organic ◆ Other: □ Solid/Liquid Mixture, Slurry □ Gas □ Biological Spec 	□ Multi-phase □ Sediment □ Metal □ Other imen	 ◆ QA Plan: XASO-QAP-001 (Equivalent to HASQARD) □ Additional QA Requirements, List Document Below: Reference Doc Number: ← Field COC Submitted? X No □ Yes ◆ Lab COC Required? X No □ Yes ◆ Sample/Container Inspection Documentation Required? 						
(If sample matrices vary, specify on Re	equest Page)	◆ Hold Time: X	No I Yes					
Disposal Information		If Yes,						
 Disposition of Virgin Samples: Virgin samples are returned to requestor to archiving provisions are made with receive If archiving, provide: Archiving Reference Doc: Disposition of Treated Samples: 	inless /ing group!	Contact ASO Lead before submitting Samples ◆ Special Storag X None □ ◆ Data Require	□ Use SW 846 (PNL analytes/methods whe □ Other? Specify: ge Requirements: Refrigerate □ Other, es ASO Quality Engin	-ASO-071, identify re holding times apply) Specify: neer Review? X No Yes				
X Dispose Return								
· · · · · ·	Data Report	ing Information						
 Is Work Associated with a Fee-Based Milestone? □ No X Yes If yes, milestone due date: Preliminary Results Requested, As Available? □ No X Yes 	Data Reporting Log X ASO-QAP-001 (Eq HASQARD). Minimum data report Project Specific Report Contact ASO Lead or Document:	evel juivalent to ort. equirements: List Reference	Requested Analytical (Note: Priority rate charge fo Negotiated Commi 7/23 (To be comm	Work Completion Date: r < 10 business day turn-around time) tment Date: /19 leted by ASO Lead)				
	Waste Designa	tion Information	(ro oc comp	leten oy noo benn)				
ASO Sample Information Check List Att If no, Reference Doc Attached: or, Previous ASR Number:0521 or, Previous RPL Number: Send Report To:	ached? X No Yes	MSIN	Does the Waste Designation Indicate Presen X P7-25 P7-25	ntion Documentation tee of PCBs? No □Yes				
D	and Locks Inf		ted by ASO -t-M					
Receiving Date Delivered: 5/7/19 Delivered By (optional) S. Fiskum Time Delivered:	and Login Informati	ASR Number:	<u>K. Pool 5/7/19</u> 0787	Rev.: 02				
Group ID (optional)		RPL Numbers	: <u>19-1773 thru</u> (first a	19-1784 and last)				
CMC Waste Sample? X No Ye	es		(1101)					

ASO Work Accepted By: _

Signature/Date:

KN Pool

enpel 6/26/19

Analytical Services Request (ASR)

(REQUEST PAGE ---- Information Specific to Individual Samples)

ASO Staff Use Only	Provide Analytes of	Interest and Required Detect	ion limits - 🛛 Below 🗖 Attached	ASO Sta	ff Use Only
RPL Number	Client Sample ID	Sample Description (& Matrix if varies)	Analyses Requested	Test	Library
	Revision 1 - ASR revise	ed to add F to the IC analyses on sample 1 sample 19-1774 ICP/OES anal	19-1773 and add Si, B, Li, Mg, Mo and W to lyses		
	Revis	on 2 - ASR revised to remove the requi	irement for QE Review		
19-1773	TI061-COMP-FEED	AP-107 Tank Waste	 GEA - All samples (Cs-137, Co-60 and Eu-154 and any other observed gamma emitting isotopes) Tc-99 IC-Anions - F, Cl, NO₂, NO₃, PO₄, C₂O₄ and SO₄ TOC/TIC - Hot Pursulfate OH Acid Digestion- 128 - Prep Lab a) ICP/OES - Ag, Al, As, Ba, Ca, Cd, Cr, Fe, K, Na, Ni, P, Pb, S, Sr, Ti, Zn, Zr ICP/MS - U-238 Sr-90 Np-AEA, Np-237 Pu-AEA, Pu-238, Pu-239/240 		
19-1774	TI061-COMP-EFF	AP-107 Tank Waste - Cs Removed	 GEA - All samples (Cs-137, Co-60 and Eu-154 and any other observed gamma emitting isotopes) Tc-99 Acid Digestion- 128 - Prep Lab a) ICP/OES - Ag, Al, As, <u>B</u> Ba, Ca, Cd, Cr, Fe, K,<u>Li</u>, <u>Mg</u>, <u>Mo</u>, Na, Ni, P, Pb, S, <u>Si</u>, Sr, Ti, <u>W</u>, Zn, Zr ICP/MS - U-238 c) Sr-90 Pu-AEA, Pu-238, Pu-239/240 Np-AEA, Np-237 		

<u>ASR: 0787</u>_____Rev: 02

ASR 0787.02 Request Page

Page 1 of 2

Analytical Services Request (ASR)

(REQUEST PAGE ---- Information Specific to Individual Samples)

ASO Staff Use Only	Provide Analytes of	vide Analytes of Interest and Required Detection limits - Below Attached										
RPL Number	Client Sample ID	Sample Description (& Matrix if varies)	Test	Library								
19-1775	TI061-L-F1-A											
19-1776	TI061-L-F4-A											
19-1777	TI061-L-F8-A		1) Acid Digest - 128 - Prep Lab									
19-1778	TI061-L-F11-A		a) ICP/OES - Ba, Ca, Cd, Cr, Fe,									
19-1779	TI061-L-F14-A	AD 107 Topk Masta Co Domoved	K, Ni, Pb, Ti, Zn, Zr									
19-1780	TI061-L-F18-A	AP-107 Tallk Waste - Cs Removed	c) Sr-90									
19-1781	TI061-L-F22-A		c) Pu-AEA, Pu-238, Pu-239/240									
19-1782	TI061-L-F25-A		d) Np-AEA, Np-237									
19-1783	TI061-L-F27-A											
19-1784	TI061-L-F29-A											

ASR: 0787 Rev: 02

Battelle PNNL/RPL/Inorganic Analysis ... ICP-OES Analysis Report PO Box 999, Richland, Washington 99352

ASR#:0787 Client:S. Fiskum I 2 (liquids) AP-107 TauterSet is is is is imple to be cription AP-107 TauterASOClient Sample DAP-107 TauterASOClientSample MAP-107 TauterASOClientIs is		110jeet/ WI#.	73312 / NO	24189							
Client: S. Fiskum Total Samples: 12 (liquids) Sample Description AP-107 Tark Waste Sample ID Sample ID Sample ID Sample ID Sample ID Sample ID Sample ID 9-1773 T1061-COMP-FEED NA 19-1779 T1061-L-F14-A 19-1774 T1061-L-F1-A NA 19-1780 T1061-L-F18-A 19-1775 T1061-L-F4-A NA 19-1782 T1061-L-F22-A 19-1776 T1061-L-F1-A NA 19-1783 T1061-L-F2-A 19-1777 T1061-L-F1-A NA 19-1783 T1061-L-F2-A IP-1778 T1061-L-F1-A NA 19-1784 T1061-L-F2-A IP-1778 T1061-L-F1-A NA 19-1784 T1061-L-F2-A Sample ID A <th <="" colspan="4" th=""><th></th><th>ASR#:</th><th>0787</th><th></th><th></th><th></th><th></th></th>	<th></th> <th>ASR#:</th> <th>0787</th> <th></th> <th></th> <th></th> <th></th>					ASR#:	0787				
Total Samples: 12 (liquids)Sample DescriptionAP-107 Tark WasteASOClientSampleASOClientSample IDSample IDMSample IDSample IDSample IDSample IDSample IDM919-1773T1061-COMP-FEEDNA19-1779T1061-L-F14-AA19-1774T1061-COMP-EFFNA19-1780T1061-L-F18-AA19-1775T1061-L-F4-ANA19-1782T1061-L-F22-AA19-1776T1061-L-F8-ANA19-1783T10-1-L-F27-AA19-1777T1061-L-F11-ANA19-1783T10-1-L-F27-AA19-1778T1061-L-F11-ANA19-1784T10-1-L-F29-AA19-1778T1061-L-F11-ANA19-1784T10-1-L-F29-AA19-1778T1061-L-F11-ANA19-1784T10-1-L-F29-AA19-1779T1061-L-F11-ANA19-1784T10-1-L-F29-AA19-1780T1061-L-F11-ANA19-1784T0-1-L-F29-AA19-1781T1061-L-F11-ANA19-1784T0-1-L-F29-AA19-1799T1061-L-F11-ANA19-1784T0-1-L-F29-AA19-179T1061-L-F11-ANA19-1784T0-1-L-F29-AA19-1781T1061-L-F11-ANA19-1784T0-1-L-F29-AA19-179Simple dilution of "as received" samples in 5% v/v HNO-y erformed by L. DaDa19-179J. CarterAnalysis Date		Client:	S. Fiskum								
Sample DescriptionAP-107 Tank WasteASOClientSample IDSample IDSample IDSample IDMSample IDSample IDNA19-177Sample IDSample IDW19-1773T1061-COMP-FEEDNA19-1770T1061-L-F14-ANA19-1774T1061-COMP-EFFNA19-1780T1061-L-F18-ANA19-1775T1061-L-F1-ANA19-1781T1061-L-F22-ANA19-1776T1061-L-F4-ANA19-1783T1061-L-F22-ANA19-1777T1061-L-F8-ANA19-1783T1061-L-F22-ANA19-1778T1061-L-F1-ANA19-1783T1061-L-F22-ANA19-1778T1061-L-F1-ANA19-1784T1061-L-F22-ANA19-1778T1061-L-F1-ANA19-1784T1061-L-F22-ANA19-1778T1061-L-F1-ANA19-1783T1061-L-F22-ANA19-1778T1061-L-F1-ANA19-1784T1061-L-F22-ANA19-1778T1061-L-F1-ANA19-1784T1061-L-F22-ANA19-178T1061-L-F1-ANA19-1784T1061-L-F22-ANA19-1779T1061-L-F1-ANA19-1784T1061-L-F22-ANA19-178T1061-L-F1-ANA19-1784T1061-L-F22-ANA19-178T1061-L-F1-ANA19-1784T1061-L-F22-ANA10/202/19Simple dilution of "as received" samples in 5% v/v HNO3 performed by J. CaCa <tr< th=""><th></th><th>Total Samples:</th><th>12 (liquids</th><th>)</th><th></th><th></th><th></th></tr<>		Total Samples:	12 (liquids)							
ASO Sample IDClient Sample IDSample ID Weight (g)ASO Sample IDClient Sample IDH19-1773T1061-COMP-FEEDNA19-1779T1061-L-F14-AV19-1774T1061-COMP-EFFNA19-1780T1061-L-F18-AV19-1775T1061-L-F1-ANA19-1780T1061-L-F22-AV19-1776T1061-L-F4-ANA19-1782T1061-L-F22-AV19-1777T1061-L-F8-ANA19-1783T1061-L-F22-AV19-1778T1061-L-F11-ANA19-1783T1061-L-F22-AV19-1778T1061-L-F11-ANA19-1784T1061-L-F22-AV19-1778T1061-L-F11-ANA19-1784T1061-L-F22-AV19-1779T1061-L-F11-ANA19-1784T1061-L-F22-AV19-1778T1061-L-F11-ANA19-1784T1061-L-F22-AV19-1779T1061-L-F11-ANA19-1784T1061-L-F22-AV19-1779T1061-L-F11-ANA19-1784T1061-L-F22-AV19-1779T1061-L-F11-ANA19-1784T1061-L-F22-AV19-1779T1061-L-F11-ANA19-1784T1061-L-F22-AV19-1779T1061-L-F11-ANA19-1784T1061-L-F22-AV19-1779T1061-L-F11-ANA19-1784T1061-L-F14-AV10-12-E228TE28TE28TE28TE28TE2810-12-E228TE28TE28TE28TE28TE28 <th></th> <th>Sample Descrip</th> <th>tion AP-107 Ta</th> <th>nk Waste</th> <th></th> <th></th> <th></th>		Sample Descrip	tion AP-107 Ta	nk Waste							
Sample IDSample IDWeight (g)Sample IDSample IDSample IDW19-1773T1061-COMP-FEEDNA19-1779T1061-L-F14-A119-1774T1061-COMP-EFFNA19-1780T1061-L-F18-A119-1775T1061-L-F1-ANA19-1781T1061-L-F22-A119-1776T1061-L-F4-ANA19-1782T1061-L-F22-A119-1777T1061-L-F8-ANA19-1783T1061-L-F22-A119-1778T1061-L-F11-ANA19-1783T1061-L-F22-A119-1778T1061-L-F11-ANA19-1784T1061-L-F29-A119-1778T1061-L-F11-ANA19-1784T1061-L-F29-A119-1778T1061-L-F11-ANA19-1784T1061-L-F29-A119-1779T1061-L-F11-ANA19-1784T1061-L-F29-A119-1778T1061-L-F11-ANA19-1784T1061-L-F29-A119-1779T1061-L-F11-ANA19-1784T1061-L-F29-A119-1778T1061-L-F11-ANA19-1784T1061-L-F29-A119-1779T1061-L-F11-ANA19-1784T1061-L-F29-A119-1779T1061-L-F11-ANA19-1784T1061-L-F29-A119-1779T1061-L-F11-ANA19-1784T1061-L-F29-A110-10-10-10-10-10-10-10-10-10-10-10-10-1	ASO	Client	Sample	ASO		Client	Sample				
19-17/73 T1061-COMP-FEED NA 19-17/9 T1061-L-F14-A 19-1774 T1061-COMP-EFF NA 19-1780 T1061-L-F18-A 19-1775 T1061-L-F1-A NA 19-1781 T1061-L-F22-A 19-1776 T1061-L-F4-A NA 19-1782 T1061-L-F22-A 19-1777 T1061-L-F8-A NA 19-1783 T1061-L-F22-A 19-1778 T1061-L-F11-A NA 19-1784 T1061-L-F22-A Sample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCI Acid Extraction of Liquids Metals Analysis Using a Dry-Block Heater" using Nitric Acid only, performed by L. Da 07/02/19. O7/02/19. Simple dilution of "as received" samples in 5% v/v HNO3 performed by J. Ca O7/03/19. Procedure: RPG-CMC-211, Rev. 4, "Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP- Analyst: J. Carter Analysis Date: 07/03/19 ICP File:	Sample ID	Sample ID	Weight (g)	Sample ID	S	Sample ID	Weight (g)				
19-1774TI061-COMP-EFFNA19-1780TI061-L-F18-A19-1775TI061-L-F1-ANA19-1781TI061-L-F18-A19-1776TI061-L-F4-ANA19-1782TI061-L-F22-A19-1777TI061-L-F8-ANA19-1783TI061-L-F27-A19-1778TI061-L-F11-ANA19-1784TI061-L-F29-A19-1778TI061-L-F11-ANA19-1784TI061-L-F29-ASample Preparation:RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of LiquideMetals Analysis Using a Dry-Block Heater" using Nitric Acid only, performed by L. Da07/02/19.Simple dilution of "as received" samples in 5% v/v HNO3 performed by J. Ca07/03/19.Procedure:RPG-CMC-211, Rev. 4, "Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP-Analyst:J. CarterAnalysis Date:07/03/19ICP File:C08See Chemical Measurement Center 98620 file:ICP-325-405-3 (Calibration and Maintenance RecoSN: 077N5122002Mettler AT400 BalanceSN: 1113292667SN: 8725287790Ohaus PA224C BalanceSN: B725287790	9-1773	TI061-COMP-FEED	NA	19-1779	TIO	51-L-F14-A	NA				
19-1775 TI061-L-F1-A NA 19-1781 TI061-L-F22-A 19-1776 TI061-L-F4-A NA 19-1782 TI061-L-F25-A 19-1777 TI061-L-F8-A NA 19-1783 TI061-L-F27-A 19-1778 TI061-L-F11-A NA 19-1784 TI061-L-F29-A Sample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of Liquids Metals Analysis Using a Dry-Block Heater" using Nitric Acid only, performed by L. Da 07/02/19. Simple dilution of "as received" samples in 5% v/v HNO3 performed by J. Ca 07/03/19. Procedure: RPG-CMC-211, Rev. 4, "Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP- Analyst: J. Carter Analysis Date: 07/03/19 ICP File: C08 See Chemical Measurement Center 98620 file: ICP-325-405-3 (Calibration and Maintenance Record) Maintenance Record) M&TE: PerkinElmer 5300DV ICP-OES SN: 077N5122002 SN: 1113292667 Mettler AT400 Balance SN: 111329267 SN: B725287790 SN: 07019	9-1774	TI061-COMP-EFF	NA	19-1780	TIO	51-L-F18-A	NA				
19-1776TI061-L-F4-ANA19-1782TI061-L-F25-A19-1777TI061-L-F8-ANA19-1783TI061-L-F27-A19-1778TI061-L-F11-ANA19-1784TI061-L-F29-ASample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of Liquids Metals Analysis Using a Dry-Block Heater" using Nitric Acid only, performed by L. Da 07/02/19. Simple dilution of "as received" samples in 5% v/v HNO3 performed by J. Ca 07/03/19.Procedure:RPG-CMC-211, Rev. 4, "Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP- Analyst:ICP File:C08See Chemical Measurement Center 98620 file:ICP-325-405-3 (Calibration and Maintenance RecoSN: 077N5122002M&TE:PerkinElmer 5300DV ICP-OESSN: 077N5122002Mettler AT400 BalanceSN: 1113292667 SN: B72528770SN: B72528770	9-1775	TI061-L-F1-A	NA	19-1781	TIO	51-L-F22-A	NA				
19-1777 TI061-L-F8-A NA 19-1783 TI061-L-F27-A 19-1778 TI061-L-F11-A NA 19-1784 TI061-L-F29-A Sample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of Liquide Metals Analysis Using a Dry-Block Heater" using Nitric Acid only, performed by L. Da 07/02/19. Simple dilution of "as received" samples in 5% v/v HNO3 performed by J. Ca 07/03/19. Procedure: RPG-CMC-211, Rev. 4, "Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP-Analyst: J. Carter Analysis Date: 07/03/19 See Chemical Measurement Center 98620 file: ICP-325-405-3 (Calibration and Maintenance Reconstruction Construction Balance) M&TE: PerkinElmer 5300DV ICP-OES SN: 077N5122002 Mettler AT400 Balance SN: B725287790 Ohaus PA224C Balance SN: B725287790	9-1776	TI061-L-F4-A	NA	19-1782	TIO	51-L-F25-A	NA				
19-1778TI061-L-F11-ANA19-1784TI061-L-F29-AISample Character StructureIndextructureIndextructureSample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of LiquidsMetals Analysis Using a Dry-Block Heater" using Nitric Acid only, performed by L. Da07/02/19.Simple dilution of "as received" samples in 5% v/v HNO3 performed by L. Da07/03/19.Procedure:RPG-CMC-211, Rev. 4, "Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP-Analyst:J. CarterAnalysis Date: $07/03/19$ ICP File:C08Coefficient Center 98620 file:ICP-325-405-3 (Calibration and Maintenance Reconstruction Spectrometry SpectrometryM&TE:PerkinElmer 5300DV ICP-OESSN: 077N5122022Mettler AT400 BalanceSN: 1113292667Ohaus PA224C BalanceSN: B725287790	9-1777	TI061-L-F8-A	NA	19-1783	TIO	61-L-F27-A	NA				
Sample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of Liquids Metals Analysis Using a Dry-Block Heater" using Nitric Acid only, performed by L. Da 07/02/19. Simple dilution of "as received" samples in 5% v/v HNO3 performed by J. Ca 07/03/19. Procedure: RPG-CMC-211, Rev. 4, "Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP- Analyst: J. Carter Analysis Date: 07/03/19 ICP File: C08 See Chemical Measurement Center 98620 file: ICP-325-405-3 (Calibration and Maintenance Record M&TE: PerkinElmer 5300DV ICP-OES SN: 077N5122022 M&TE: PerkinElmer 5300DV ICP-OES SN: 077N5122022 SN: 1113292667	9-1778	TI061-L-F11-A	NA	19-1784	TI06	61-L-F29-A	NA				
Sample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of Liquids Metals Analysis Using a Dry-Block Heater" using Nitric Acid only, performed by L. Da 07/02/19. Simple dilution of "as received" samples in 5% v/v HNO3 performed by J. Ca 07/03/19. Procedure: RPG-CMC-211, Rev. 4, "Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP- Analyst: J. Carter Analysis Date: 07/03/19 ICP File: C08 See Chemical Measurement Center 98620 file: ICP-325-405-3 (Calibration and Maintenance Recording) ICP File: C08 M&TE: PerkinElmer 5300DV ICP-OES SN: 077N5122002 SN: 1113292667 Mettler AT400 Balance SN: 8725287790 SN: 8725287790											
Sample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of Liquids Metals Analysis Using a Dry-Block Heater" using Nitric Acid only, performed by L. Da 07/02/19. Simple dilution of "as received" samples in 5% v/v HNO3 performed by J. Ca 07/03/19. Procedure: RPG-CMC-211, Rev. 4, "Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP- Analyst: J. Carter Analysis Date: 07/03/19 See Chemical Measurement Center 98620 file: ICP-325-405-3 (Calibration and Maintenance Record M&TE: PerkinElmer 5300DV ICP-OES Mettler AT400 Balance SN: 077N5122002 Mettler AT400 Balance SN: 1113292667 Ohaus PA224C Balance SN: B725287790											
Analyst: J. Carter Analysis Date: 07/03/19 ICP File: C08 See Chemical Measurement Center 98620 file: ICP-325-405-3 (Calibration and Maintenance Recond) M&TE: PerkinElmer 5300DV ICP-OES SN: 077N5122002 Mettler AT400 Balance SN: 1113292667 Ohaus PA224C Balance SN: B725287790		simple dilution of "as	s received" sample	es in 5% v/v HN	NO ₃ p	erformed by	J. Carter on				
See Chemical Measurement Center 98620 file: ICP-325-405-3 (Calibration and Maintenance Reco M&TE: PerkinElmer 5300DV ICP-OES SN: 077N5122002 Mettler AT400 Balance SN: 1113292667 Ohaus PA224C Balance SN: B725287790)7/03/19. Procedure:	RPG-CMC-211, I Inductively Coup	s received" sample Rev. 4, "Determin led Argon Plasma	es in 5% v/v HN nation of Eleme Optical Emissi	NO ₃ p ental (ion Sp	Composition	J. Carter on by (ICP-OES)."				
M&TE: PerkinElmer 5300DV ICP-OES SN: 077N5122002 Mettler AT400 Balance SN: 1113292667 Ohaus PA224C Balance SN: B725287790)7/03/19. Procedure:	RPG-CMC-211, Inductively Coup	s received" sample Rev. 4, "Determin led Argon Plasma Analysis Date:	es in 5% v/v HN nation of Eleme Optical Emissi 07/03/19	NO ₃ p ental (ion S ₁	Composition pectrometry (by (ICP-OES)."				
☐Mettler AT400 BalanceSN: 1113292667⊠Ohaus PA224C BalanceSN: B725287790⊠00SN: D12021200)7/03/19. Procedure: Analyst: See Chemic	RPG-CMC-211, Inductively Coup J. Carter	s received" sample Rev. 4, "Determin led Argon Plasma Analysis Date: enter 98620 file:	es in 5% v/v HN nation of Eleme Optical Emissi 07/03/19 <u>ICP-325-405</u> (Calibration a	NO ₃ p ental (ion Sp -3 and N	Composition pectrometry (ICP File: Maintenance 1	2. Darnell on J. Carter on by (ICP-OES)." C0804 Records)				
Ohaus PA224C Balance SN: B725287790	97/03/19. Procedure: Analyst: See Chemic	RPG-CMC-211, Inductively Coup J. Carter cal Measurement Co PerkinElmer 53	s received" sample Rev. 4, "Determin led Argon Plasma Analysis Date: enter 98620 file: 00DV ICP-OES	es in 5% v/v HN nation of Eleme Optical Emissi 07/03/19 <u>ICP-325-405</u> (Calibration a	NO ₃ p ental (ion Sj <u>-3</u> and N	Composition pectrometry (ICP File: Maintenance 1 : 077N5122(2. Darnell on J. Carter on by (ICP-OES)." C0804 Records)				
)7/03/19. Procedure: Analyst: See Chemic M&TE:	RPG-CMC-211, J Inductively Coup J. Carter Cal Measurement Co PerkinElmer 53 Mettler AT400	s received" sample Rev. 4, "Determin led Argon Plasma Analysis Date: enter 98620 file: 00DV ICP-OES Balance	es in 5% v/v HN nation of Eleme Optical Emissi 07/03/19 <u>ICP-325-405</u> (Calibration a	-3 and M SN	Composition pectrometry (ICP File: //aintenance 1 : 077N5122(: 111329266	2. Darnell on J. Carter on by (ICP-OES)." C0804 Records) 002 7				
X Sartorius ME414S Balance SN: 21308482	97/03/19. Procedure: Analyst: See Chemic M&TE:	RPG-CMC-211, Inductively Coup J. Carter Cal Measurement Co PerkinElmer 53 Mettler AT400 Ohaus PA224C	s received" sample Rev. 4, "Determin led Argon Plasma Analysis Date: enter 98620 file: 00DV ICP-OES Balance Balance	es in 5% v/v HN nation of Eleme Optical Emissi 07/03/19 <u>ICP-325-405</u> (Calibration a	NO ₃ p ental (ion Sp -3 and M SN SN	Composition pectrometry (ICP File: //aintenance 1 : 077N5122(: 111329266 : B72528779	2. Darnell on J. Carter on by (ICP-OES)." C0804 Records) 002 7 00				
SAL Cell 2 Balance SN: 8033311209)7/03/19. Procedure: Analyst: See Chemic	RPG-CMC-211, J Inductively Coup J. Carter Cal Measurement Co PerkinElmer 53 Mettler AT400 Ohaus PA224C Sartorius ME41	s received" sample Rev. 4, "Determin led Argon Plasma Analysis Date: enter 98620 file: 00DV ICP-OES Balance Balance 4S Balance	es in 5% v/v HN nation of Eleme Optical Emissi 07/03/19 <u>ICP-325-405</u> (Calibration a	-3 and N SN SN SN	Composition pectrometry (ICP File: //aintenance l : 077N5122(: 111329266 : B72528779 : 21308482	2. Darnell on J. Carter on by (ICP-OES)." C0804 Records) 002 7 00				

Report Preparer

Review and Concurrence

7/10/19 Date 7/12/19 Date

Twelve aqueous samples submitted under Analytical Service Request (ASR) 0787 were analyzed by ICP-OES. The samples were prepared following RPL procedure RPG-CMC-128, Rev. 1, "*HNO3-HCl Acid Extraction of Liquids for Metals Analysis Using a Dry-Block Heater*". This process was conducted without the use of HCl. All samples were diluted in 5% HNO₃ prior to analysis. None of the samples were filtered.

All sample results are reported on a mass per unit volume basis (μ g/mL) for each detected analyte. The data have been adjusted for instrument dilutions.

Analytes of interest (AOI) were specified in the ASR and are listed in the upper section of the attached ICP-OES Data Report. There were two analyte lists requested, one being a shortened version. Samples were reported using the most extensive analyte list. The quality control (QC) results for the AOI have been evaluated and are presented below. Analytes other than the AOI are reported in the bottom section of the report but have not been fully evaluated for QC performance.

Calibration of the ICP-OES was done following the manufacturer's recommended calibration procedure using multi-analyte custom standard solutions traceable to the National Institute of Standards and Technology (NIST). Midrange calibration verification standards (MCVA and MCVB) were used to verify acceptance of the two-point calibration curves obtained for each analyte and for continuing calibration verification.

The controlling documents were procedures RPG-CMC-211, Rev. 4, *Determination of Elemental Composition by Inductively Coupled Argon Plasma Optical Emission Spectrometry (ICP-OES)*, and ASO-QAP-001, Rev. 11, *Analytical Support Operations (ASO) Quality Assurance Plan*. Instrument calibrations, QC checks and blanks (e.g., ICV/ICB, CCV/CCB, LLS, ICS), matrix spike, post-digestion spikes, duplicate, reagent spike, blank spike, and serial dilution were conducted during the analysis run.

Preparation Blank (PB):

A preparation blank (reagents only) was prepared for the extraction process. All analytes were within the acceptance criteria of <EQL (estimated quantitation level), \leq 50% regulatory decision level, or \leq 10% of the concentration in the samples. The lab diluent passed these criteria.

Blank Spike (BS)/Laboratory Control Sample (LCS):

A reagent spike (BS) sample (reagents and spikes) was prepared for the extraction process. Recovery values are listed for all analytes included in the RS that were measured at or above the EQL. Recovery values for the AOI meeting this requirement ranged from 83% to 109% and were within the acceptance criterion of 80% to 120%.

Duplicate/Replicate Relative Percent Difference (RPD)/Relative Standard Deviation (RSD): A duplicate of sample 19-1774 was prepared and analyzed. RPDs are listed for all analytes that were measured at or above the EQL. RPDs for the AOI meeting this requirement

ranged from 0.8% to 4.4% and were within the acceptance criterion of $\leq 20\%$ for liquid samples.

Matrix-Spike (MS) Sample:

A matrix spike (MS) of sample 19-1774 was prepared for the extraction process. Recovery values are listed for all analytes included in the MS that were measured at or above the EQL. Recovery values for the AOI meeting this requirement ranged from 50% to 122% and were within the acceptance criterion of 75% to 125%, with the exception of Ag (50%).

Initial/Continuing Calibration Verification (ICV/CCV):

MCVA and MCVB solutions were analyzed immediately after calibration, after each group of not more than ten samples, and at the end of the analytical run. The concentrations of all AOI were within the acceptance criteria of 90% to 110% recovery.

Initial/Continuing Calibration Blank (ICB/CCB):

The ICB/CCB solution (5% v/v HNO₃) was analyzed immediately after the ICV solutions and after the CCV solutions (after each group of not more than ten samples and at the end of the analytical run). The concentration of all AOI were within the acceptance criteria of $\langle EQL \rangle$, with the exception of sodium in the final CCB solution (0.372ppm). Associated samples are greater than 10x the EQL.

Low-Level Standard (LLS):

The LLS solution was analyzed immediately after the first CCB solution. The concentrations of all AOI were within the acceptance criteria of 70% to 130% recovery.

Interference Check Standard (ICS/SST):

The ICS solution was analyzed immediately after the first LLS solution and immediately prior to analyzing the final CCV solutions. The concentrations of all AOI were within the acceptance criteria of 80% to 120% recovery.

Serial Dilution (SD):

Five-fold serial dilution was conducted on sample 19-1773. Percent differences (%Ds) are listed for all analytes that had a concentration at or above the EQL in the diluted sample. The %Ds for the AOI meeting this requirement ranged from 0.3% to 24.3% and were within the acceptance criterion of \leq 10% except for potassium which had a %D of 24.3%.

Post-Digestion Spike (PS-A) - Sample (A Component):

In addition to the BS sample, a post-digestion spike (A Component) was conducted on sample 19-1775. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration $\geq 25\%$ of that in the sample. Recovery values for the AOI meeting this requirement ranged from 96% to 117%, and were within the acceptance criterion of 80% to 120%.

Post-Digestion Spike (PS-B) - Sample (B Component):

In addition to the MS sample, a post-digestion spike (B Component) was conducted on sample 19-1775. Recovery values are listed for all analytes in the spike that were measured at or above the EQL, and that had a spike concentration $\geq 25\%$ of that in the sample. Recovery values for the AOI meeting this requirement was 99% and within the acceptance criterion of 80% to 120%.

Other QC:

All other instrument-related QC tests for the AOI passed within their respective acceptance criteria.

Comments:

- 1) The "Final Results" have been corrected for all laboratory dilutions performed on the samples during processing and analysis, unless specifically noted.
- 2) Instrument detection limits (IDL) and estimated quantitation limits (EQL) shown are for acidified water and/or fusion flux matrices as applicable. Method detection limits (MDL) for individual samples can be estimated by multiplying the IDL by the "Process Factor" for that individual sample. The estimated quantitation limit (EQL) for each concentration value can be obtained by multiplying the EQL by the "Process Factor".
- 3) Routine precision and bias is typically $\pm 15\%$ or better for samples in dilute, acidified water (e.g. 5% v/v HNO₃ or less) at analyte concentrations > EQL up to the upper calibration level. This also presumes that the total dissolved solids concentration in the sample is less than 5000 µg/mL (0.5 per cent by weight). Note that bracketed values listed in the data report are within the MDL and the EQL, and have potential uncertainties greater than 15%. Concentration values < MDL are listed as "--". Note, that calibration and QC standard samples are validated to a precision of $\pm 10\%$.
- 4) Analytes included in the spike A component (for the AS/PS) are; Ag, Al, As, B, Ba, Be, Bi, Ca, Cd, Co, Cr, Cu, Fe, K, Li, Mg, Mn, Mo, Na, Ni, P, Pb, Sb, Se, Si, Sn, Sr, Ta, Ti, Tl, V, W, Y, Zn, and Zr. Analytes included in the spike B component are; Ce, Dy, Eu, La, Nd, Pd, Rh, Ru, S, Te, Th, and U.

PNNL-28958, Rev. 0 RPT-DFTP-013, Rev. 0 Page 1 of 3

Battelle PNNL/RPG/Inorganic Analysis ... ICPOES Data Report

		Run Date >	7/3/2019	7/3/2019	7/3/2019	7/3/2019	7/3/2019	7/3/2019	7/3/2019	7/3/2019
		Process	4.0	4.0	242.0			247.0	246.6	248.0
		Factor >	1.0	1.0	248.9	249.6		247.0	246.6	240.9
					19-1773 @ 5x	19-1774 @ 5x	DUP-1774 @ 5x	19-1775 @ 5x	19-1776 @ 5x	19-1777 @ 5x
Instr. Det. Limit (IDL)	Est. Quant. Limit (EQL)	Client ID >	Lab Diluent	<u>Reagent</u> <u>Blank</u>	TI061-COMP- FEED	<u>TI061-CC</u>	OMP-EFF	<u>TI061-L-F1-A</u>	<u>TI061-L-F4-A</u>	<u>TI061-L-F8-A</u>
(µg/mL)	(µg/mL)	(Analyte)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)
0.0019	0.019	Ag								
0.0122	0.122	AI		[0.35]	10014.22	9829.47	9676.03	8519.36	9134.23	8959.56
0.0529	0.529	As					-			
0.0003	0.003	Ba			[0.41]		-			
0.0177	0.177	Ca		[2.13]	[34.48]	[23.09]	[12.60]	[29.03]	[24.36]	[23.05]
0.0022	0.022	Cd			6.51	[5.11]	5.43	[2.38]	5.51	[5.45]
0.0025	0.025	Cr			526.82	525.56	503.84	486.89	494.73	488.30
0.0052	0.052	Fe		[0.34]	14.77	[12.48]	[11.79]	15.26	26.42	[12.14]
0.0240	0.240	к			4684.88	4518.87	4484.80	4143.80	4359.04	4187.15
0.0246	0.246	Na		s 	137330.45	133410.93	129259.79	122715.68	128843.96	126427.90
0.0049	0.049	Ni			24.58	23.60	22.72	20.61	21.64	20.96
0.0554	0.554	Р			873.09	845.57	808.90	774.90	804.26	774.71
0.0162	0.162	Pb			[8.12]					
0.1186	1.186	S		S. 	1932.77	1862.85	1826.32	1761.04	1831.36	1743.59
0.0005	0.005	Sr			[0.13]			-	-	
0.0004	0.004	Ti				[0.19]	[0.18]	[0.30]	[0.26]	[0.23]
0.0043	0.043	Zn		[0.52]		[1.53]	[1.08]	[1.11]	-	
0.0019	0.019	Zr				[2.09]	[2.09]	[1.37]	[1.52]	[1.66]
Other Analyte	es									
0.0067	0.067	В	[0.04]	[1.01]	31.35	23.21	20.59	107.27	99.41	87.96
0.0001	0.001	Be			[0.14]	[0.11]	[0.12]	-	[0.09]	[0.10]
0.0494	0.494	Bi							-	
0.0069	0.069	Ce						-	-	
0.0035	0.035	Co					[1.40]		[1.07]	
0.0016	0.016	Cu		[0.07]	[1.14]	[0.55]				[0.62]
0.0029	0.029	Dy		-						
0.0013	0.013	Eu								
0.0022	0.022	La								
0.0009	0.009	Li			[0.23]					
0.0068	0.068	Mg		[0.24]						
0.0002	0.002	Mn								
0.0045	0.045	Мо			44.28	42.27	40.76	39.66	40.59	40.54
0.0086	0.086	Nd							:. 	
0.0077	0.077	Pd		-						
0.0165	0.165	Rh		-				-		
0.0097	0.097	Ru			[7.06]	[7.54]	[5.67]	[7.92]	[7.29]	[6.03]
0.0518	0.518	Sb								
0.1595	1.595	Se								
0.0237	0.237	Si		[0.86]	70.96	[41.84]	[40.66]	122.87	105.73	108.96
0.0313	0.313	Sn		[0.80]	[13.68]	[8.24]	[13.99]	[16.39]	[16.96]	
0.0217	0.217	Та								
0.0311	0.311	Те			[8.45]					
0.0189	0.189	Th			-			-		
0.0530	0.530	TI								
0.0360	0.360	U			[15.24]		[10.10]			[9.00]
0.0021	0.021	V			[0.78]					
0.0216	0.216	W			70.89	72.72	62.83	65.68	62.41	60.32
0.0012	0.012	Y								

1) "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the "multiplier"

near the top of each column. The estimated sample quantitation limit = EQL (in Column 2)

times the "multiplier". Overall error for values \geq EQL is estimated to be within ±15%.

2) Values in brackets [] are ≥ MDL but < EQL, with errors likely to exceed 15%.

ASR-0787 Results from C0804 ASR-0787 Fiskum (128)

PNNL-28958, Rev. 0 RPT-DFTP-013, Rev. 0 Page 2 of 3

Battelle PNNL/RPG/Inorganic Analysis ... ICPOES Data Report

	1								
		Run Date >	7/3/2019	7/3/2019	7/3/2019	7/3/2019	7/3/2019	7/3/2019	7/3/2019
		Process							
		Factor >	250.2	250.7	247.0	248.5	251.9	248.3	246.8
			19-1778 @ 5x	19-1779 @ 5x	19-1780 @ 5x	19-1781 @ 5x	19-1782 @ 5x	19-1783 @ 5x	19-1784 @ 5x
Instr. Det. Limit (IDL)	Est. Quant. Limit (EQL)	Client ID >	<u>TI061-L-F11-A</u>	<u>TI061-L-F14-A</u>	<u>TI061-L-F18-A</u>	TI061-L-F22-A	<u>TI061-L-F25-A</u>	<u>TI061-L-F27-A</u>	<u>TI061-L-F29-A</u>
(µg/mL)	(µg/mL)	(Analyte)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)	(µg/mL)
0.0019	0.019	Ag							
0.0122	0.122	AI	9032.73	8772.12	9431.78	9526.53	9598.06	9576.94	9636.19
0.0529	0.529	As				-	-		
0.0003	0.003	Ba		-					
0.0177	0.177	Ca	[20.47]	[21.08]	[23.83]	[21.94]	[19.51]	[21.19]	[21.77]
0.0022	0.022	Cd	6.00	[5.47]	5.95	6.39	5.96	6.50	5.98
0.0025	0.025	Cr	494.77	477.32	497.64	518.84	504.67	508.87	515.21
0.0052	0.052	Fe	14.13	[11.75]	13.81	13.19	14.18	13.37	33.90
0.0240	0.240	к	4328.35	4159.06	4512.37	4514.00	4492.38	4459.46	4554.30
0.0246	0.246	Na	127623.19	125402.26	128473.07	131485.59	132018.56	130987.29	133132.27
0.0049	0.049	Ni	20.89	21.20	21.51	21.58	22.75	23.01	22.25
0.0554	0.554	Р	803.51	764.11	791.91	823.87	812.99	819.64	821.55
0.0162	0.162	Pb					[4.23]		[5.99]
0.1186	1.186	S	1851.86	1786.69	1846.62	1906.33	1831.28	1847.77	1855.36
0.0005	0.005	Sr							
0.0004	0.004	Ti	[0.29]	[0.21]	[0.18]	[0.24]	[0.14]	[0.21]	[0.21]
0.0043	0.043	Zn	[1.34]		[1.16]	[1.37]			
0.0019	0.019	Zr	[1.51]	[1.13]	[1.38]	[1.68]	[1.28]	[1.39]	[1.10]
Other Analyte	s								
0.0067	0.067	В	106.90	96.55	83.88	83.78	74.02	62.54	77.17
0.0001	0.001	Be	[0.12]	[0.11]	[0.12]	[0.12]	[0.13]	[0.12]	[0.12]
0.0494	0.494	Bi							
0.0069	0.069	Ce							
0.0035	0.035	Co					[1.12]		[0.98]
0.0016	0.016	Cu	[1.30]	[0.78]	[0.86]	[1.25]	[1.23]	[1.05]	[1.05]
0.0029	0.029	Dy							
0.0013	0.013	Eu							
0.0022	0.022	La							
0.0009	0.009	Li							
0.0068	0.068	Mg							
0.0002	0.002	Mn							
0.0045	0.045	Мо	42.41	40.52	41.87	43.87	41.49	42.31	42.89
0.0086	0.086	Nd							
0.0077	0.077	Pd		[2.39]		[2.80]			
0.0165	0.165	Rh			[4.79]				
0.0097	0.097	Ru	[7.40]	[6.07]	[8.41]	[5.29]	[6.17]	[7.37]	[5.61]
0.0518	0.518	Sb			[15.65]	[17.18]	[13.10]		
0.1595	1.595	Se							
0.0237	0.237	Si	135.56	137.68	176.52	166.97	150.31	162.13	146.35
0.0313	0.313	Sn	[11.49]	[16.61]	[16.51]	[9.26]	[14.51]	[14.84]	[9.56]
0.0217	0.217	Та							
0.0311	0.311	Те	[8.77]						
0.0189	0.189	Th							
0.0530	0.530	TI							
0.0360	0.360	U			[9.40]		[10.20]		
0.0021	0.021	v			-				
0.0216	0.216	w	67.88	64.92	67.48	74.50	70.61	73.16	67.19
0.0012	0.012	Y							

1) "--" indicates the value is < MDL. The method detection limit (MDL) = IDL times the "multiplier"

near the top of each column. The estimated sample quantitation limit = EQL (in Column 2)

2) Values in brackets [] are ≥ MDL but < EQL, with errors likely to exceed 15%.

ASR-0787 Results from C0804 ASR-0787 Fiskum (128)

times the "multiplier". Overall error for values \geq EQL is estimated to be within ±15%.

Battelle PNNL/RPG/Inorganic Analysis ... ICPOES Data Report

Criteria >	≤ 20%	80%-120%	75%-125%	80%-120%	80%-120%	≤ 10%
QC ID >	19-1774 Dup	BS-1720	19-1774 MS	19-1775 + AS-A	19-1775 + AS-В	19-1773 5-fold Serial Dil
Analytes	RPD (%)	%Rec	%Rec	%Rec	%Rec	%Diff
Ag		83	50	96		
AI	1.6	97	89	105		0.3
As				102		
Ba		100	99	104		
Ca		104	111	113		
Cd		97	96	101		
Cr	4.2	100	nr	102		1.7
Fe		100	98	105		
к	0.8	92	122	117		24.3
Na	3.2	109	nr	nr		1.6
Ni	3.8	99	98	104		
P	4.4	98	89	98		1.2
Ph		100	99	100		
S	2.0	96	94	100	99	39
Sr	2.0	101	07	101		5.8
Ti		100	00	102		
7-		100	99	103		
Zn		97	95	102		
Zr		99	97	104		
Other Analy	tes					
В	11.9	101	96	101		
Be		99	100	105		
Bi		83		96		
Ce		91	89		93	
Co				101		
Cu		100	103	107		
Dy					100	
Eu					99	
La		98	97		98	
Li		111	123	120		
Mg		101	103	107	1	
Mn		98	96	101		
Mo	3.6	98	94	99		
Nd		98	97		100	
Pd					96	
Rh					102	
Ru					96	
Sb				101		
Se				103		
Si		90	101	106		
Sn				99		
Ta				100		
Te					100	
Th		00	99		00	
TI		33	30	01	33	
		100	07	31	400	
U V		100	9/	00	103	
V I		95	93	99		
14/	410	00	00	464		

QC Performance 7/3/2019

nr = spike concentration less than 25% of sample concentration. Matrix effects can be assessed from the serial dilution.

Battelle PNNL/RPL/Inorganic Analysis ... ICP-MS Analysis Report PO Box 999, Richland, Washington 99352

Project / WP#:	73312 / NC4189
ASR#:	0787
Client:	S. Fiskum
Total Samples:	12 (Aqueous)
Sample Description	AP-107 Tank Waste

ASO Sample ID	Client Sample ID	Client Sample Description	Sample Weight (g)
19-1773	TI061-COMP-FEED	AP-107 Tank Waste	NA
19-1774	TI061-COMP-EFF	AP-107 Tank Waste Cs Removed	NA
19-1775	TI061-L-F1-A		NA
19-1776	TI061-L-F4-A		NA
19-1777	TI061-L-F8-A		NA
19-1778	TI061-L-F11-A		NA
19-1779	TI061-L-F14-A	A.D. 107 Tenk Wests Co. Removed	NA
19-1780	TI061-L-F18-A	AP-107 Tank waste Cs Removed	NA
19-1781	TI061-L-F22-A		NA
19-1782	TI061-L-F25-A		NA
19-1783	TI061-L-F27-A		NA
19-1784	TI061-L-F29-A		NA

Sample Preparation: RPG-CMC-128, Rev. 1. "HNO3-HCl Acid Extraction of Liquids for Metals Analysis Using a Dry-Block Heater" performed by L. Darnell on 07/02/19; ICP-MS bench dilution in 2% v/v HNO₃ performed by D. Cherkasov on 07/09/19.

Procedure:	ion b 5)."	у										
Analyst:	D.E.Cherkasov	Analysis Date:	07/09/2019 ICP			e:	M0181					
See Chemical Measurement Center 98620 file: ICP-325-405-3 (Calibration and Maintenance Records)												
M&TE:	PerkinElmer Nex	ION [™] 350X ICP·	MS	SN: 85VN4	070702	RPL 405						
	Sartorius R200D	Balance		SN: 390800	42	RPL 405						
	Mettler AT400 B	alance		SN: M19445			L 405 FH					
	Sartorius ME414	S Balance		SN: 21308482 RPL 420			L 420					
	Ohaus EX324 Ba	alance		SN: 8033311209 SAL Cell			L Cell 2					

Denis erKasa

Report Preparer

Review and Concurrence

07/15 Date

07/18/2019 Date

Appendix E

Twelve aqueous samples submitted under Analytical Service Request (ASR) 0787 were analyzed by ICP-MS. Sample 19-1774 was prepared in duplicate following RPL procedure RPG-CMC-128 by diluting 1 mL to approximately 25 mL. Prior to analysis all samples were further diluted in 2% HNO₃. None of the samples were filtered.

All results are reported on a mass per unit volume basis (ng/mL) for each detected analyte. The data have been adjusted for instrument dilutions.

Analyte of interest (AOI) was specified in the ASR and is listed in the upper section of the attached ICP-MS Data Report. The U-238 was the only analyte of interest for samples 19-1773 to 19-17844. The quality control (QC) results for the AOI have been evaluated and are presented below.

Calibration of the ICP-MS was done following the manufacturer's recommended calibration procedure using multi-analyte custom standard solutions traceable to the National Institute of Standards and Technology (NIST). Midrange calibration verification standards were used to verify acceptance of the six-point calibration curve and for initial and continuing calibration verification (ICV/CCV).

The controlling document was procedure RPG-CMC-292, Rev 0, *Determination of Elemental Composition by Inductively Coupled Argon Plasma Mass Spectrometry (ICP-MS)* and ASO-QAP-001, Rev. 11, *Analytical Support Operations (ASO) Quality Assurance Plan.* Instrument calibrations, QC checks and blanks (e.g., ICV/ICB, CCV/CCB, LLS, ICS), post-digestion spike, duplicate, blank spike, and serial dilution were conducted during the analysis run.

Internal Standard (IS):

All solutions (blanks, standards, QC checks, and samples) were mixed in-line with a solution containing 10 ppb each of Tb-159, and Bi-209 as the internal standard (IS). The AOI data were normalized using the IS mass (e.g., Bi-209). The Bi-209 IS recoveries ranged from 95.3% to 113.7%, which were within the acceptance criterion of 30% to 120%.

Preparation Blank (PB):

A preparation blank (reagents only) was prepared for the dissolution process. The concentration of U was within the acceptance criteria of \leq EQL (estimated quantitation level), \leq 50% regulatory decision level, or less than \leq 10% of the concentration in the samples.

Blank Spike (BS)/Laboratory Control Sample (LCS):

A blank spike (BS) sample was prepared by spiking a 2% HNO3 blank with an equivalent volume of the CCV-71A-2ppb standard (1:1 ratio). The recovery for the AOI was 99.7%, which is within the acceptance criteria of 80% to 120% recovery.

Duplicate/Replicate Relative Percent Difference (RPD)/Relative Standard Deviation (RSD):

A duplicate of sample 19-1774 was prepared. The RPD% for Uranium was 2.3% which is within the acceptance criterion of $\leq 20\%$ RPD for liquid samples.

Matrix-Spike (MS) Sample:

Post digestion spike sample (PS-A) was run on the samples 19-1775 as an alternative and the results are discussed later in this report.

Initial/Continuing Calibration Verification (ICV/CCV):

The ICV/CCV solution was analyzed immediately after calibration, after each group of not more than ten samples, and at the end of the analytical run. The concentrations of all AOI were within the acceptance criteria of 90% to 110% recovery.

Initial/Continuing Calibration Blank (ICB/CCB):

The ICB/CCB solution (2% HNO₃) was analyzed immediately after the ICV solutions and after the CCV solutions (after each group of not more than ten samples and at the end of the analytical run). The concentration of all AOI were within the acceptance criteria of $\langle EQL \rangle$.

Low-Level Standard (LLS):

The LLS solution was analyzed immediately after the first CCB solution. The concentrations of all AOI were within the acceptance criteria of 70% to 130% recovery.

Interference Check Standard (ICS):

The ICS solution was analyzed immediately after the first LLS solution and immediately prior to analyzing the final CCV solutions. The concentrations of all AOI were within the acceptance criteria of 80% to 120% recovery.

Serial Dilution (SD):

Five-fold serial dilution was conducted on sample 19-1775. Percent differences (%D) are listed for all analytes that had a concentration at or above the EQL in the diluted sample. The Uranium result was below the EQL and was not evaluated.

Post-Digestion Spike (PS-71A)/Analytical Spike):

A post spike sample was prepared for sample 19-1775 with the addition of CCV-71A. The recovery of U was 96.8, which met the acceptance criterion of 80% to 120%.

Other QC:

All other instrument-related QC tests for the AOI passed within their respective acceptance criteria.

Comments:

- 1) The "Final Results" have been corrected for all laboratory dilutions performed on the samples during processing and analysis, unless specifically noted.
- 2) Instrument detection limits (IDL) and estimated quantitation limits (EQL) shown are for acidified water and/or fusion flux matrices as applicable. Method detection limits (MDL) for individual samples can be estimated by multiplying the IDL by the "Process Factor" for that individual sample. The estimated quantitation limit (EQL) for each concentration value can be obtained by multiplying the EQL by the "Process Factor".
- 3) Routine precision and bias is typically $\pm 15\%$ or better for samples in dilute, acidified water (e.g. 2% v/v HNO₃ or less) at analyte concentrations > EQL up to the upper calibration level. This also presumes that the total dissolved solids concentration in the sample is less than 50 µg/mL (0.005 per cent by weight). Note that bracketed values listed in the data report are within the MDL and the EQL, and have potential uncertainties greater than 15%. Concentration values < MDL are listed as "--". Note, that calibration and QC standard samples are validated to a precision of $\pm 10\%$.
- 4) Analytes included in the spike 71A component (for the AS/PS) are; Ag, Al, As, B, Ba, Be, Ca, Cd, Ce, Co, Cr, Cs, Cu, Dy, Er, Eu, Fe, Ga, Gd, Ho, K, La, Lu, Mg, Mn, Na, Nd, Ni, P, Pb, Pr, Rb, S, Se, Sm, Sr, Th, Tl, Tm, U, V, Yb, and Zn. Analytes included in the spike 71B component are; Ge, Hf, Mo, Nb, Sb, Si, Sn, Ta, Te, Ti, W, and Zr. Analytes included in the spike 71C component are; Ir, Os, Pd, Pt, Re, Rh, and Ru. Analytes included in the spike 71D component are; Bi, In, Li, Sc, Tb, and Y. Analytes included in the spike Hg component are; Hg.

PNNL-28958, Rev. 0 RPT-DFTP-@163,1Brev. 0

Battelle PNNL/RPG/Inorganic Analysis ... ICP-MS Data Report

		Run Date >	7/9/2019	7/9/2019	7/9/2019	7/9/2019	7/9/2019	7/9/2019	7/9/2019	7/9/2019	7/9/2019	7/9/2019	7/9/2019	7/9/2019	7/9/2019	7/9/2019	7/9/2019	7/9/2019
		Process																
		Factor >	1.00	43488.1	44525.6	44743.1	43460.4	44148.5	44065.0	44036.6	44492.4	44710.5	44802.6	44123.6	44423.3	44958.0	44266.7	44022.9
			2% HNO3	(900x) BLK-	(900x) 19-	(900x) 19-	(900x) DUP-	(900x) 19-	(900x) 19-	(900x) 19-	(900x) 19-	(900x) 19-	(900x) 19-	(900x) 19-	(900x) 19-	(900x) 19-	(900x) 19-	(900x) 19-
		RPL/LAB >	Lab Blank	1720	1773	1774	1774	1775	1775 Rep	1776	1777	1778	1779	1780	1781	1782	1783	1784
Instr. Det.	Est. Quant.		2% HNO3 Lab Blank	Process Blank	TI061-COMP-	<u>T1061-C0</u>	OMP-EFF	<u>TI061-</u>	<u>L-F1-A</u>	TI061-L-F4-A	<u>TI061-L-F8-A</u>	TI061-L-F11-	TI061-L-F14-	TI061-L-F18-	<u>TI061-L-F22-</u>	<u>TI061-L-F25-</u>	<u>TI061-L-F27-</u>	TI061-L-F29-
Limit (IDL)	Limit (EQL)	Client ID >										-	-		-	-	4	-
(ng/mL)	(ng/mL)	(Analyte)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)	(ng/mL)
0.00004	0.00039	U-238		[3.3E+00]	17878	10819	10752	617	628	11580	14884	15432	15576	15685	16738	17130	16959	17045

nternal Standa	ard % Recover	ry														
Bi 209 (IS)	104%	106%	107%	108%	106%	102%	105%	103%	109%	102%	103%	106%	105%	103%	101%	104%

 "--" indicates the value is < MDL, The method detection limit (MDL) = IDL times the "multiplier" near the top of each column. The estimated sample quantitation limit = EQL (in Column 2) times the "multiplier". Overall error for values ≿ EQL is estimated to be within ±15%.
 Values in brackets [] are ≥ MDL but < EQL, with errors likely to exceed 15%.

IS = Internal Standard. The concentration of certain elements cannot be determined due to the presence of the IS in all solutions.

QC Performance7/18/2018

Criteria >	≤ 20%	≤ 20%	80%-120%	75%-125%	≤ 10%
QC ID >	19-1774 DUP	19-1775 Rep	BS/LCS	19-1775 + PSA	19-1775 5-fold Serial Dil
Analytes	RPD (%)	RPD (%)	%Rec	%Rec	%Diff
U-238	2.3%	2.0%	99.7%	96.8%	

Internal Standard % Recovery

	Bi 209 (IS)	106%	105%	105%	110%	102%
--	-------------	------	------	------	------	------

Shaded results are outside the acceptance criteria.

nr = spike concentration less than 25% of sample concentration. Matrix effects can be assessed from the serial dilution. na = not applicable; KOH flux and Ni crucible or Na2O2 flux and Zr crucible for fusion preparations, or Si for HF assisted digests. IS = Internal Standard. The concentration of certain elements cannot be determined due to the presence of the IS in all solutions. NM = Not measured. The isotope was not measure due to method or molecular interference limitations. Battelle PNNL/RPL/ASO Hydroxide Analysis Report P.O. Box 999, 902 Battelle Blvd., Richland, Washington 99352

Hvdroxide	Analysis	bv	Titration

Project / WP#:	73325/NC2504
-	73312/NC4189
ASR#:	0731.00 & 0787.01
Client:	SK Fiskum
Total # of Samples:	2

RPL ID	Client Sample ID
19-1107	TCT008-COMP-FEED
19-1773	TI061-COMP-FEED

Analysis Type:	Hydroxide
Sample Processing Prior to Radiochemical Processing/Analysis	 None Digested as per RPG-CMC-128, Rev.1, HNO₃-HCL Acid extraction of Liquids for Metals Analysis Using a Dry-Block Heater Fusion as per RPG-CMC-115, Solubilization of Metals from Solids Using a KOH-KNO₃ Fusion Other:
Pre-dilution Prior to Radiochemical Processing?	No Yes
Analysis Procedure:	RPG-CMC-228, Rev. 0.1, Determination of Hydroxyl (OH) and Alkalinity of Aqueous Solutions, Leachates & Supernates
Analysis Date or Date Range:	5/30/19
Technician/Analyst:	KN Pool & AM Carney
Electronic Data File:	ASR 0731 & 0787.01_Fiskum.xls
ASO Project 98620 File:	File Plan 5872: Sample preparation and analysis records; LSC 3100 TR calibration, daily checks, and maintenance records; and standard certificates and preparation. Also, balance calibration and performance check records.
M&TE Number(s):	Beckman Coulter pH Meter, SN: 110650046

M. Pul , 6/2/19 Preparer Date

Andrew Cary 1_6/4/19 Reviewer Date

Battelle PNNL/RPL/ASO Hydroxide Analysis Report

SAMPLE RESULTS

See attached data report, Sample Results for ASR's 0731.00 and 0787.01. All sample data are molarity of hydroxide at each end point.

ASO Project Files, ASR 0731.00 and 0787.01 have been created for this report including all appropriate supporting records which may include the Pipette Performance Check Worksheet form and laboratory bench records, and records associated with hydroxide analysis. Standard certificates and balance calibration records can be found in the ASO Records.

Sample preparation and analysis

Hydroxide analysis was performed for diluted sample aliquots of 2 aqueous samples (19-1107 and 19-1773). Samples were analyzed by manual titration for the base constituents content following procedure RPG-CMC-228, Rev. 0.1, *Determination of Hydroxyl (OH) and Alkalinity of Aqueous Solutions, Leachates & Supernates.*

A sample size of 0.100 mL was added to 2.5 mL of DI water. The titrant used was 0.0887 M HCl. Standardized HCl was prepared on 5/6/19 and documented on Chem_Rec 238. Titration pH measurements were obtained using a Beckman Coulter 560 pH meter with serial number (SN) 110650046. The pH meter was calibrated using 3 buffers, pH 4, 7 and 10 and the calibration verified using an independent pH 7 buffer.

The customer requested target MDL of 0.1M was achieved for all samples. The titrant used for the sample analyses is 0.0887M HCl along with a pH meter accurate to within 0.05 pH units. The low concentration of the titrant along with the sensitivity of the pH meter provides the detection capability to meet this requirement.

The initial diluted pH is reported on the attached Report Summary along with the free hydroxide molarity. pH measured at each inflection point are provided below.

Sample ID	1st Inflection point	2nd Inflection point	3rd Inflection point
19-1107	10.9	8.3	4.8
19-1773	10.9	7.9	4.8

Included in the data package is a report summary and the sample results calculated from the raw data. A copy of the titration curve data for each sample is also included with this report.

Battelle PNNL/RPL/ASO Hydroxide Analysis Report

QUALITY CONTROL RESULTS

This analysis has limited quality control (QC) samples that are prepared. There are no laboratory blanks (LB), matrix spikes or reagent spikes analyzed.

Instrument Calibration Control

The pH meter was calibrated using 3 buffers, pH 4, 7 and 10 and the calibration verified using an independent pH 7 buffer.

Assumption and Limitations of the Data

Generally ASO analysis methods are associated with reported uncertainty. Titrations are not amenable to calculations of uncertainty.

Comments

1. The results have been corrected for all dilution factors resulting from sample processing.

Attachment: Data Report -- Sample Results for ASR 0731.00 and 0787.01

Battelle Pacific Northwest Laboratory Nuclear Chemistry and Engineering Group Chemical Measurements Center

ASR # 731.00 & 0787.01

WP# NC2504 & NC4189

Hydroxide and Alkalinity Determination Procedure: RPG-CMC-228, Rev 0.1

Equip # Beckman Coulter 560, SN#110650046

	Repo	rt Summary fo	r ASR #	731.00 & 0787.01 Rev-00 & 01	Report Date Analysis Date	6/2/2019 5/30/2019	
				Co	ncentration, moles / Lit	er	
		Diluted	OH conc.	First Point	Second Point	Third Point	
RPG #	Client ID	pH	ug/mL	Molarity	Molarity	Molarity	
19-1107	TCT008-COMP-FEED	12.39	1.66E+04	0.98	1.24	0.71	
19-1773	TIGGL COMP FFFD	12 42	1.51E+04	0.89	1.15	0.71	

OH conc (mg/L) = M (g/L) * 17,000

Note: Results are presented for the first, second, and third inflection points on the titration curves for the hydroxide titration, as applicable. The first inflection point is generally associated with the free hydroxide concentration. The second inflection point generally represents total hydroxide, or carbonate or a combination of aluminate and carbonate. The third inflection point is usually indicative of bicarbonate or other weak acids or possibly the continued protonation of alumina.

Analyst: and an and 6/4/19 Reviewer: Terl N. Poel 6/2/19

PNNL-28958, Rev. 0 RPT-DFTP-013, Rev. 0

Battelle Pacific Northwest Laboratory Nuclear Chemistry and Engineering Group

ASR #	and Rev #	731.00 & 07	787.0 Rev-00 & 01
	Client:	S. Fiskum	
Project:	71274	WP#	NC2504
			NC4189

 Report Date:
 6/2/2019

 Analysis Date:
 5/30/2019

Procedure: **RPG-CMC-228, Rev 0.1** Determination of Hydroxyl (OH-) and Alkalinity of Aqueous Solutions, Leachates and Supernates

Chem

Spreadsheet: OH-TemplateLocked07.xls 4/1/2007

Analyst: andrew any

Titrant	Molarity	Rec#	_						OH			
HCI	0.0887	238						Diluted	1st Equivale	ence		
 , , ,			-				Titrator	Initial	Point		Found	
			Dilution	Sample	Sample	Density	Routine	pН	Titrant		millimoles	Molarity
RPG #	Sample ID		Factor	Vol. (mL)	Wt. (g)	g/mL	#	reading	Vol. (mL)	pН	base	base
19-1107	TCT008-COMP-FEED		NA	0.105	NA	NA	NA	12.39	1.15	10.86	0.102	0.976
19-1773	TI061-COMP-FEED		NA	0.105	NA	NΛ	NA	12.42	1.05	10.89	0.093	0.887

Slope NA

Equip # 110650046 Lab Loc. 420

Instrument Calibration

By Manual Titration

Buffer	Vendor	Lot Number	Expire Date
4	Inorganic Ventures	N2-WCS673492	30-May-20
7	Inorganic Ventures	N2-WCS674708	30-May-20
10	Inorganic Ventures	P2-WCS675599	30-May-19
2 111.10	X.7	T () 1	T'D'

2-nd Verif	Vendor	Lot Number	Expire Date
7	Spex	1-53MJX	23-May-20

.

ASR # and Rev # 31.00 & 0787.0 Rev-00 & 01

NC2504 NC4189

Battelle Pacific Northwest Laboratory Nuclear Chemistry and Engineering Group

WP#

Procedure:

Alkalinity of Aqueous Solutions, Leachates and Supernates

By Manual Titration

Equip # Beckman Coulter 560, SN#110650046

Titrant	Molarity					
HCl	0.0887	2nd Equivalence				
			Point		Found	
		Sample	Titrant		millimoles	Molarity
RPG #		Vol. (mL)	Vol. (mL)	pH	base	base
19-1107		0.105	2.618	8.310	0.130	1.242
19-1773		0.105	2.408	7.940	0.121	1.153

3rd Equiva	lence		
Point		Found	
Titrant		millimoles	Molarity
Vol. (mL)	pН	base	base
3.455	4.790	0.074	0.710
3.246	4.840	0.074	0.710

PNNL-28958, Rev. 0 RPT-DFTP-013, Rev. 0


PNNL-28958, Rev. 0 RPT-DFTP-013, Rev. 0



Project Number: Charge Code: ASR Number: Client:

Total Samples:

73312 NC4189 0787 S. Fiskum A. Rovira 1 liquid

	RPL Numbers	Client IDs
Samples	19-1773	TI061-COMP-FEED

Analysis Procedure	RPG-CMC-386 Rev. 1, "Carbon Measured in Solids,
	Sludge, and Liquid Matrices"
Prep Procedure	None
Analyst	A. Carney
Analysis Date	June 10, 2019
CCV Standards	TIC/TOC CMS # 543144 and 543143
BS/LCS/MS Standards	TIC/TOC CMS # 542973 and 543145
Excel Data File	ASR-0787-Fiskum.xlsx
M&TE Numbers	Carbon System (WD36639, RPL/701)
	Balance : Sartorius R200D, S/N 30809774
All Analysis Records	5015_06-10-2019-131336.CSV

Andraw Car Prepared By

Cal a

<u>b/20/19</u> Date <u>6/21/19</u> Date

Reviewed By

Table 1: TOC/TIC Results for ASR 0787

TIC in Sample 19-1773 (mg C/L):	7776.69
MDL (mg C/L):	194
EQL (mg C/L):	970
TOC in Sample 19-1773 (mg C/L):	2154.96
MDL (mg C/L):	194
EQL (mg C/L):	970
TIC in Sample 19-1773-Dup:	7720.46
MDL (mg C/L):	194
EQL (mg C/L):	970
TOC in Sample 19-1773-Dup :	2184.46
MDL (mg C/L):	194
EQL (mg C/L):	970
19-1773 TIC RPD:	0.7%
19-1773 TOC RPD:	1.4%

Sample Analysis/Results Discussion

One liquid sample was submitted under Analytical Service Request (ASR) 0787 for total inorganic and total organic carbon analysis. The analysis was performed by the hot persulfate wet oxidation method, with the results summarized in Table 1. The TIC is measured first with additions of heated sulfuric acid followed by the addition of a silver catalyzed acidic potassium persulfate solution for oxidation at 92-95 °C for TOC. The analyses were performed following procedure RPG-CMC-386, Rev. 1, *Carbon Analyses in Solids, Sludge and Liquid Matrices*.

The sample was analyzed with one duplicate for each TIC and TOC. An analytical spike was also run for TIC and TOC on the sample. The sample results are corrected for the contribution from the system blank, as per procedure RPG-CMC-386, Rev. 1. All data are reported as mg C/L of sample.

Data Limitations

None

Quality Control Discussion

The calibration and QC sample standards for the TOC initial/continuing calibration verification check (ICV/CCV) sample is a 1000 μ g/mL solution of total organic carbon standard. The calibration and QC sample standards for the TIC initial/continuing calibration verification check (ICV/CCV) sample is a 1000 μ g/mL total inorganic standard. The identification of the standards and their Chemical Management System (CMS) numbers are included on the raw data bench sheets for traceability.

The QC samples analyzed as part of the method include initial and continuing calibration verification samples (ICV/CCV), initial and continuing calibration blanks (ICB/CCB), laboratory duplicate for the sample, a laboratory control sample/blank spike (LCS/BS), and an analytical spike (AS). The work was performed in one batch.

Two blanks are run at the beginning of each batch and a blank is run after ICV/CCV. The blanks must be <EQL. The blanks run in the batch are <EQL.

Initial Calibration Check and Continuing Calibration Verification Standards:

The calibration of the coulometer analysis system was checked by calibration verification standards analyzed at the beginning and end of the analysis run. TOC results for the two ICVs were 100.2% and 97.5% recovery, and for the two TIC ICVs the results were 98.0% and 99.4% recovery, within the acceptance criterion of 90% to 110%. The TOC result for the CCV was 98.6% recovery and the TIC CCV was 99.0% recovery, within the acceptance criterion of 85% to 115%.

- Laboratory Control Sample/Blank Spike: One TIC and TOC LCS/BS was analyzed. The TIC LCS/BS result was 100.9% recovery and the TOC LCS/BS result was 99.9% recovery, meeting the acceptance criteria of 75% to 125%.
- <u>Duplicate/Replicate</u>: Precision of the carbon measurements is demonstrated by the relative percent difference (RPD) between sample and duplicate/replicate. Sample 19-1773 TIC RPD was 0.7% and TOC was 1.4%. Both samples meet the acceptance criteria of ≤20%.
- <u>Analytical Spike (AS)</u>: The accuracy of the carbon measurements can be estimated by the recovery from the AS. The results for the analytical spike for TIC is 99.4% recovery and for the TOC, 93.3% recovery. The AS recovery for the TOC and TIC results meets the acceptance criterion of 75% to 125%.

Deviation from Procedure:

None

General Comments

- 1) Routine precision and bias are typically $\pm 15\%$ or better for non-complex samples that are free of interferences.
- 2) For the TIC/TOC, the analysis MDL is calculated by dividing the batch IDL by the sample mass and is therefore dependent on sample size. The estimated quantitation limit (EQL) is defined as 5x the MDL. Results <5x MDL have higher uncertainties and RPDs are not calculated if the results are <5x MDL.</p>
- 3) Where applicable, the reported "Final Results" have been corrected for any dilution performed on the sample prior to analysis.

Gamma Energy Analysis (GEA)

Project / WP#:	73312/NC4189
ASR#:	0787.00
Client:	SK Fiskum
Total Samples:	2

	RPL ID	-	Client Sample ID	
	19-1773		TI061-COMP-FEED	
	19-1774		TI061-COMP-EFF	
Analysis Type:		GEA- f	or all positively measured or non-detected	isotopes
Sample Processing Prior to Radiochemical Processing/Analysis None □ Digested as per RPG-CMC-129, Rev. 0 HNO3-HCl Solids Using a Dry Block Heater □ Fusion as per RPG-CMC-115, Solubilization of Me Using a KOH-KNO3 Fusion □ Other: Preparation may also involve attaining a GEA geometry with the calibration geometry.		-HCl Acid Extraction of of Metals from Solids metry that is compatible		
Analysis Procedure	:	RPG-CMC-450, Rev. 3 Gamma Energy Analysis (GEA) and Low-Er Photon Spectrometry (LEPS)		(GEA) and Low-Energy
Reference Date:		None		
Analysis Date or Da	te Range:	June 10	, 2019	
Technician/Analyst	:	M Cantaloub		
Rad Chem Electron	ic Data File:	19-1773 Fiskum.xls		
ASO Project 98620	File:	File Plan 5872, Technical (Radiochemistry), Gamma Calibration, daily checks, and maintenance records; and standard certificates and preparation. Also, balance calibration and performance check records.		ma Calibration, daily ertificates and mance check records.
M&TE Number(s):		Detector T,Z		

Irang le Prepare

6/19/19 Date

19 Jun 2019 Date Reviewer

SAMPLE RESULTS

The client requested measurement of Cs-137, Co-60 and Eu-154 and all detected isotopes in these samples.

ASO Project File, ASR 0787.00 has been created for this report including all appropriate supporting records which may include the Pipette Performance Check Worksheet form, standard certificates, laboratory bench records, Sample Preparation Laboratory Bench Sheets, and Gamma Energy Analysis printouts. Detector calibration records, control charts and balance calibration records can be found in the ASO Records.

Sample Preparation, Separation, Mounting and Counting Methods

The quality control (QC) steps for direct GEA are discussed below.

QUALITY CONTROL RESULTS

Tracer:

Tracers are not used for ASO GEA methods.

Process Blank (PB):

No process blank was prepared for gamma counting.

Required Detection Limits

There are required detection limits for these samples. MDL for Eu-154 of 1.0E-5 uCi/mL not achieved on COMP FEED sample due to high Cs-137 activity.

Blank Spike (BS)/Laboratory Control Sample (LCS)/ Matrix Spike (MS):

There are no BS, LCS or MS samples analyzed for ASO GEA analyses. Instrument performance is assessed by the analyses of daily control counts and weekly background counts, as discussed below.

Duplicate Relative Percent Difference (RPD):

No duplicate samples were provided for gamma counting.

Instrument Calibration and Quality Control

Gamma detectors are calibrated using multi-isotope standards that are NIST-traceable and prepared in the identical counting geometry to all samples and detectors, if possible. Counter control sources containing Am-241, Cs-137 and Co-60 are analyzed daily before the use of each detector. Procedure RPG-CMC-450 requires that a counter control source is checked daily and

must be within ± 3 sigma or $\pm 3\%$ of the control value, whichever is greater. Gamma counting was not performed unless the control counts were within the required limits. Background counts are performed on all gamma detectors at least weekly for either an overnight or weekend count. The most recent background is subtracted from all sample counts.

Assumptions and Limitations of the Data

For these samples, the sample geometry was not the same as our calibrated geometries. Consequently, we used calibration geometries that were as close as possible to the sample geometries. This introduces some additional uncertainty above what is listed in our report. However, this additional uncertainty is thought to be small since all samples were counted at distances of about 10 cm from the detector face.

Interferences/Resolution

None.

Uncertainty

For gamma counting, the uncertainty in the counting data, photon abundance and the nuclear half-life, and efficiency are included in the calculation of the total uncertainty along with a systematic uncertainty for sample prep. The Canberra Genie software includes both random and systematic uncertainties in the calculation of the total uncertainties which are listed on the report. We conservatively estimate that 2% is the lowest uncertainty possible for our GEA measurements taking into account systematic uncertainties in gamma calibration standards.

Comments

None.

Attachment: Data Report Sample Results for ASR 0787.00.

8/22/19

Pacific Northwest Na PO Box 999, Richlan Radiochemical Scien	ational Laboratory d, WA ces and Engineering Group		filename	19-1773 8/22/2019 Revision 1
Client: Fiskum	Project: 73312 NC4189	Prepared by: Midul Call	8/22	2/19

Irang-le

 Procedures:
 RPG-CMC-450, Rev. 3 Gamma Energy Analysis (GEA) and Low-Energy Photon Spectrometry (LEPS)

 Spectrometry
 M&TE:

 Gamma detectors T, Z
 10-Jun-19

Technical Reviewer:

	Measured Activity, µCi/mL ± 1s			
RPL ID:	19-1	773	19-1	774
Sample ID:	TI061-CON	MP-FEED	TI061-COMP-EFF	
Isotope				
Co-60	4.96E-04	$\pm 34\%$	5.29E-04	± 2%
Sb-126	< 2E-3		1.37E-04	$\pm 2\%$
Sn-126	1.36E-01	± 6%	3.19E-04	± 3%
Cs-137	1.54E+02	±1%	5.29E-04	± 2%
Eu-154	< 2E-3		4.65E-05	± 2%
Eu-155	< 3E-2		1.04E-05	±15%
Am-241	<1E-1		2.45E-04	± 3%

Notes: MDL for Eu-154 of 1.0E-5 uCi/mL not achieved on COMP FEED sample due to high Cs-137 activity

Revision 1 COMP-FEED Sn-126 result impacted by gamma-induced Pb x-ry fluencence interference with the primary Sn-126 line.

Revision 1 COMP-FEED Sn-126 results should be considered an upper limit. MDA is 2E-2.

Revision 1 Sb-126 is decay daughter of Sn-126

Closest matching geometry used for COMP FEED sample analysis - PSV10mlw at 20 inches from detector.

or yo by Liquid	Semimation Spectrometry
Project / WP#:	73312 / NC4189
ASR#:	0787.02
Client:	S. Fiskum
Total # of Samples:	12

Sr-90 by Liquid Scintillation Spectrometry

RPL ID	Client Sample ID	RPL ID	Client Sample ID
19-1773	TI061-COMP-FEED	19-1779	TI061-L-F14-A
19-1774	TI061-COMP-EFF	19-1780	TI061-L-F18-A
19-1775	TI061-L-F1-A	19-1781	TI061-L-F22-A
19-1776	TI061-L-F4-A	19-1782	TI061-L-F25-A
19-1777	TI061-L-F8-A	19-1783	TI061-L-F27-A
19-1778	TI061-L-F11-A	19-1784	TI061-L-F29-A

Analysis Type:	Sr-90	
Sample Processing Prior to Radiochemical Processing/Analysis	 None Digested as per RPG-CMC-128, Rev.1, HNO₃-HCL Acid extraction of Liquids for Metals Analysis Using a Dry-Block Heater Fusion as per RPG-CMC-115, Solubilization of Metals from Solids Using a KOH-KNO₃ Fusion Other: 	
Pre-dilution Prior to Radiochemical	□ No	
Processing?	Yes	
Separation Procedure:	RPG-CMC-476, Rev. 0, Strontium Separation Using Eichrom Strontium Resin	
Spike Standard ID:	R-693-b-10 (Sr-90)	
Separation Date:	08/06/2019 @ 11:00	
Technician/Analyst:	L. Darnell	
Analysis Procedure:	RPG-CMC-474, Rev. 1, Measurement of Alpha and Beta Activity by Liquid Scintillation Spectrometry	
Reference Date:	08/6/19	
Analysis Date or Date Range:	08/07/19 (first count), 08/11/19 (second count)	
Technician/Analyst:	LP Darnell & T. Trang-Le	
Rad Chem Electronic Data File:	RPG-RC\PNL\Projects\Backup files\Backup 19\19-1773 Fiskum.xls	
ASO Project 98620 File:	File Plan 5872: T 73312: Sample preparation and analysis records; T-4.4 LSC 3100 calibration, daily checks, and maintenance records; and T3 standard certificates and preparation. Also, balance calibration and performance check records.	
M&TE Number(s):	Perkin Elmer Tri-Carb 3100, Serial # DG08061340, RPL 425, Tri-Carb 2700TR software version 1.04 dated 9/99	

l 8/12

Preparer

Date

Trang-le 1 8/12/19 Reviewer Date

SAMPLE RESULTS

See attached data report, Sample Results for ASR 0787.02. All sample data are reported in μ Ci/mL with a 1- σ uncertainty (see Comments).

Sample preparation, separation, mounting, and counting

All twelve samples submitted under Analytical Service Request (ASR) 0787.02 were analyzed for Sr-90 by chemical separation and liquid scintillation counting. All the samples were prepared in RPL/420. Aliquots of the as received samples were used for radioanalytical analyses; only Sr-90 data are presented in this report.

QUALITY CONTROL RESULTS

Quality control (QC) samples include a laboratory preparation blank (PB) and sample duplicate. Additional laboratory QC samples were prepared prior to separations; these include a reagent blank spike (RS), and a matrix spike (MS), each made with the addition of Sr-90 standard to an aliquot of the samples.

Laboratory Preparation Blank (PB):

The activity level of Sr-90 present in the laboratory separation blank (1.55E-5 μ Ci/mL) is both below the activity present in the samples (1.19 E-3 μ Ci/mL or higher) and less than sample MDC (2.2 E-5 μ Ci/mL) meeting the acceptance criteria of less than 5% of the sample activity or less than the sample MDC.

Blank Spike (BS) - Reagent Spike (RS):

The RS recovery of 97% meets the procedure acceptance criteria of 80% to 120% recovery.

Matrix Spike (MS):

The MS recovery of 102% meets the acceptance criterion of 75% to 125% recovery. Note: the MS sample was prepared "after" digestion (see comments), by adding a known Sr-90 standard quantity to an aliquot of 19-1774 (TI061-COMP-EFF).

Duplicate -- Relative Percent Difference (RPD):

Duplicate results are required to agree within $\leq 20\%$ RPD. The ASO QAP further specifies that the two results need to be >5 times the MDA or have individual uncertainties $\leq 20\%$. The sample and duplicate result were both less than 5 times MDA. The RPD is not calculated when one or both of the results used to calculate the RPD is less than 5 times MDA.

Instrument Quality Control

The liquid scintillation counter is calibrated for tritium and C-14 using quenched standard sets that are purchased from the vendor. Daily control counts are then performed using a tritium, C-14, and a background count sample. The instrument software assesses the performance of the control counts and provides control charts to ensure the continuing calibration of the instrument. If the daily performance check fails, then the instrument is not used. Preventative maintenance and repairs are performed by the vendor under our service contract. The counting efficiency for Sr-90 is assumed to be 100%; therefore, no specific Sr-90 calibration is performed. The LSC system calibration and performance is verified by

assessing the recovery of a reagent spike and a matrix spike that are included in every batch of samples. A preparation blank (i.e., digestion blank) and a laboratory separations blank are also included with every batch of samples; the instrument background is subtracted from all results and the preparation and separation blanks are used to assess sample contamination during sample processing steps.

Assumption and Limitations of the Data

None

Comments

- 1. The results have been corrected for all dilution factors resulting from sample processing.
- The stated 1-σ uncertainty represents the total propagated error associated with processing and counting operations and includes weighing errors, volume uncertainties, and counting error. Generally, errors are set at 2% to provide a more realistic estimate of the uncertainty in view of the systematic uncertainties that are not fully accounted for in the uncertainty calculations.

Attachment: Data Report -- Sample Results for ASR 0787.02

Pacific Northwest National Laboratory PO Box 999, Richland, WA Radiochemical Sciences and Engineering Group

filename 19-1773 8/12/2019

Client: Fiskum ASR 0787.02	Project: 73312 NC4189	×	Prepared by: Trang-le 8/12/19 Technical Reviewer: Vala 22/19
Procedures:	RPG-CMC-474	, Rev 1, Mea	surement of Alpha and Beta Activity by Liquid Scintillation
METE	Perkin Elmer Ti	riCarh model	3100TR liquid scintillation spectrometer
Count dates:	August 7 & 11	2019	5100 TK liquid semimation spectrometer
Count dates.	August / & II,	2017	
	Lab	Sr-90 Cond	centration,
Sample	ID	μCi/mL ± 1	1s uncertainty
TIACL COMD FEED	10,1772	4.525.1	1.20/
11001-COMP-FEED	19-1//3	4.52E-1	± 2%
TI061-COMP-EFF	19-1774	<1.E-3	
	DUP-1774	<1.E-3	
	RPD	-	
TI061-L-F1-A	19-1775	<1.E-3	
T1061-L-F4-A	19-1776	1.24E-3	± 16%
T1061-L-F8-A	19-1777	<1.E-3	
T1061-L-F11-A	19-1778	1.51E-2	±2%
TI061-L-F14-A	19-1779	<1.E-3	
TI061-L-F18-A	19-1780	1.19E-3	$\pm 17\%$
TI061-L-F22-A	19-1781	<1.E-3	
TI061-L-F25-A	19-1782	1.68E-3	± 12%
TI061-L-F27-A	19-1783	2.14E-3	±10%
TI061-L-F29-A	19-1784	2.73E-3	± 8%
	Reagent Spike	97%	
	Matrix Spike	102%	
	Blank	1.55E-5	± 25%

Tc-99 Analysis

Project / WP#:	73312/NC4189
ASR#:	0787.02
Client:	SK Fiskum
Total # of Samples:	2

RPL ID	Client Sample ID
19-1773	TI061-COMP-FEED
19-1774	TI061-COMP-EFF

Analysis Type:	Tc-99
Sample Processing Prior to Radiochemical Processing/Analysis	None None
1 rocosing rinnigolo	Digested as per PNL-ALO-106, Rev. 1, Acid Digestion of Waters, Soils, and Sludges for Subsequent Radiochemical Sample Analyses
	□ Fusion as per RPG-CMC-115 Rev. 0, Solubilization of Metals from Solids Using a KOH-KNO ₃ Fusion
	Digested as per RPG-CMC-128. Rev.1, HNO ₃ -HCL Acid extraction of Liquids for Metals Analysis Using a Dry-Block Heater
Pre-dilution Prior to Radiochemical Processing?	□ No
	Yes
RadioChemical Preparation Procedure:	RPG-CMC-432, Rev. 0, Technicium-99 Analysis
Technician/Analyst:	LP Darnell, (8/7/19)
Spike and Tracer Standard ID's:	R-714-b-1 (10/25/19) (Tc-99)
Analysis Procedure:	RPG-CMC-474, Rev. 1, Measurement of Alpha and Beta Activity by Liquid Scintillation Spectrometry
Reference Date:	Not applicable
Analysis Date(s) or Date Range:	8/13/19
Technician/Analyst:	LP Darnell
Analysis Data (File):	RPG-RC\PNL\Projects\Backup files\Backup 19\19-1773 Fiskum.xls
CMC Project 98620 File:	File Plan 5872: T 73312: Sample preparation and analysis records; T-4.4 Alpha Detector calibration, calibration verification checks, and maintenance records; and T3 Standard certificates and preparation. Also balance calibration and instrument performance checks
M&TE Number(s):	Perkin Elmer 3100 TR Liquid scintillation spectrometer– See attached M&TE list

8/16/19 Date ad

Preparer

Reviewer Date

Sample Results

See attached data report, Sample Results for ASR 0787.02. All data are reported in units of μ Ci/mL with a 1- σ uncertainty (see comments).

Sample Preparation, Separation, Mounting and Counting Methods

Two samples submitted under ASR 0787.02 were analyzed for Tc-99. All the samples were prepared in Laboratory 420. A direct aliquot of each of the parent samples was processed using procedure RPG-CMC-432, Rev. 0, *Technicium-99 Analysis*.

The samples were counted on August 13, 2019; no decay corrections were made.

QUALITY CONTROL RESULTS

Quality control (QC) samples prepared in laboratory 420 include a laboratory blank and sample duplicate, matrix spike sample and reagent spike sample.

The QC sample results for Tc-99 have been evaluated and are discussed below. A summary of the Tc-99 analysis results, including QC sample performance, is given in the attached data report.

Tracer:

There is no tracer for Tc-99 analysis.

Laboratory Separation Blank (LB):

The activity level of Tc-99 present in the laboratory separation blank (1.3E-6 μ Ci/mL) is both below the activity present in the samples (9.79E-2 μ Ci/mL or higher) and less than sample MDC (4.0 E-5 μ Ci/mL) meeting the acceptance criteria of less than 5% of the sample activity or less than the sample MDC.

Reagent Spike (RS):

The RS recovery of 101% (Tc-99) meets the acceptance criteria of 80% to 120% recovery.

Matrix Spike (MS):

A matrix spike sample was prepared using sample 19-1774 (TI061-COMP-EFF). The sample activity was significantly higher (~200x) than the activity of spike added for the matrix spike. The matrix spike is not calculated in this case.

Laboratory Duplicate - Relative Percent Difference (RPD):

Duplicate results for 19-1774 (TI061-COMP-EFF) are required to agree within $\leq 20\%$ RPD. The ASO QAP further specifies that the two results need to be > 5 times the MDC or have individual uncertainties < 20%. Duplicate results were 1% RPD, thus meeting the $\leq 20\%$ requirement.

Instrument Quality Control

LSC counters receive initial calibration with NIST traceable sources to determine the counter efficiency. When internal tracers are not used, the counter efficiency is used in calculation of final results.

The liquid scintillation counter (LSC) is calibrated for tritium and C-14 using quenched standard sets that are purchased from the vendor. Daily control counts are then performed using a tritium, C-14, and a background count sample. The instrument software assesses the performance of the control counts and provides control charts to ensure the continuing calibration of the instrument. If the daily performance check fails, then the instrument is not used. Preventative maintenance and repairs are performed by the vendor under our service contract. The counting efficiency for Tc-99 is assumed to be 100%; therefore no specific Tc-99 calibration is performed. The LSC system calibration and performance is verified by assessing the recovery of a reagent spike and a matrix spike that are included in every batch of samples. A laboratory separation blank is also included with every batch of samples; the instrument background is subtracted from all results and separation blanks are used to assess sample contamination during sample processing steps.

Assumption and Limitations of the Data

None

Attachment: Data Report -- Sample Results for ASR 0787.02.

Neptunium 237 Analysis

Project / WP#:	73312/NC4189
ASR#:	0787.02
Client:	SK Fiskum
Total # of Samples:	12

RPL ID	Client Sample ID	RPL ID	Client Sample ID
19-1773	TI061-COMP-FEED	19-1779	TI061-L-F14-A
19-1774	TI061-COMP-EFF	19-1780	TI061-L-F18-A
19-1775	TI061-L-F1-A	19-1781	TI061-L-F22-A
19-1776	TI061-L-F4-A	19-1782	TI061-L-F25-A
19-1777	TI061-L-F8-A	19-1783	TI061-L-F27-A
19-1778	TI061-L-F11-A	19-1784	TI061-L-F29-A

Analysis Type:	AEA – Np-237				
Sample Processing Prior to Radiochemical Processing/Analysis	 None Digested as per PNL-ALO-106, Rev. 1, Acid Digestion of Waters, Soils, and Sludges for Subsequent Radiochemical Sample Analyses Fusion as per RPG-CMC-115, Rev0, Solubilization of Metals from Solids Using a KOH-KNO3 Fusion Digested as per RPG-CMC-128, Rev.1, HNO3-HCL Acid extraction of Liquids for Metals Analysis Using a Dry-Block Heater 				
Pre-dilution Prior to Radiochemical Processing?	No Yes example 2 mL to 100 mL; 50x dilution				
Neptunium Separation Procedure:	RPG-CMC-4017, Rev. 0, Analysis of Environmental Water Samples for Actinides and Strontium- 90				
Technician/Analyst:	LP Darnell, (08/01/2019)				
Spike Standard ID's:	R-686-a-2 (Np-237)				
Co-Precipitation Procedure:	RPG-CMC-496, Rev. 1, Coprecipitation Mounting of Actinides for Alpha Spectroscopy				
Technician/Analyst:	LP Darnell, (08/01/2019)				
Analysis Procedure:	RPG-CMC-422, Rev. 2, Solutions Analysis: Alpha Spectrometry				
Reference Date:	Same as counting dates				
Analysis Date or Date Range:	August 5-8, 2019				
Technician/Analyst:	T. Trang-Le				
Analysis Data (File):	RPG-RC\PNL\Projects\Backup files\Backup 19\19-1773 Fiskum.xls				
CMC Project 98620 File:	File Plan 5871: T 73312: Sample preparation and analysis records; T-4.4 Alpha Detector calibration, calibration verification checks, and maintenance records; and T3 Standard certificates and preparation. Also balance calibration and instrument performance checks.				
M&TE Number(s):	Ortec AEA counters – 32 counters – See attached M&TE list				

<u>8/16/19</u> Date

Preparer

T Irang-le , 8/16/19 Reviewer Date

Sample Results

See attached data report, Sample Results for ASR 0787.02. All data are reported in units of μ Ci per mL with a 1- σ uncertainty unless noted otherwise (see Comments).

Sample Preparation, Separation, Mounting and Counting Methods

All twelve samples submitted under Analytical Service Request (ASR) 0787.02 were analyzed for neptunium by Alpha Spectrometry. All the samples were prepared in RPL/420, only Np-AEA is presented in this report.

Following the digestion process of samples, the Np was separated from the aqueous samples using anion exchange chromatography using procedure RPG-CMC-4017. The separated Np fraction was mounted for alpha spectrometry by co-precipitation using procedure RPG-CMC-496, and then counted by alpha spectrometry using procedure RPC-CMC-422. The samples were counted on August 5-8, 2019; no decay corrections were made.

Alpha and beta analyses were performed on each sample to determine appropriate aliquot sizes for the chemical separations needed for the alpha and beta emitters of interest. Gross alpha and gross beta activity were measured by evaporating small aliquots of leachate onto counting planchets per procedure RPG-CMC-4001 and counting per procedure RPG-CMC-408.

QUALITY CONTROL RESULTS

Quality control (QC) samples include a laboratory preparation blank and sample duplicate. Additional QC samples were prepared prior to separations; these include a reagent spike (RS), and a matrix spike (MS) made by adding Np-237 standard to a diluted sample.

Tracer:

Tracer is not used for analyses of Np.

Laboratory Preparation Blank (PB)):

The activity level of Np-237 present in the laboratory preparation blank (-6.1E-9 μ Ci/mL) is both below the activity present in the samples (7.43E-6 μ Ci/mL or higher) and less than lowest sample MDC (3.4 E-8 μ Ci/mL) meeting the acceptance criteria of less than 5% of the sample activity or less than the sample MDC.

Blank Spike (BS) - Reagent Spike (RS):

The RS recovery of 93% meets the acceptance criteria of 80% to 120% recovery.

Matrix Spike (MS):

The MS recovery of 87% meets the acceptance criterion of 75% to 125% recovery. Note: the MS sample was prepared "after" digestion, by adding a known Np-237 standard quantity to an aliquot of the leachate. Sample number 19-1774 (TI061-COMP-EFF) was selected as the matrix spike sample.

Duplicate -- Relative Percent Difference (RPD):

Duplicate results are required to agree within $\leq 20\%$ RPD. The ASO QAP further specifies that the two results need to be >5 times the MDA or have individual uncertainties <20%. The duplicate result was 25% RPD; exceeding the $\leq 20\%$ requirement.

Instrument Quality Control

Alpha counters undergo calibration annually to determine the counter's efficiency over the normal calibration range of 3 to 6 MeV. The vendor software determines a constant detector efficiency for this energy range. Np samples are counted and results calculated using the established detector efficiency.

Detector backgrounds are determined every 4 weeks or after the last analytical run, whichever is longer. Detector background counts are subtracted from all subsequent sample counts. A process blank is analyzed with each analytical batch to evaluate for contamination in the sample preparation process.

Assumption and Limitations of the Data

None.

Comments

- 1. The results have been corrected for all dilution factors resulting from sample processing.
- 2. The 1-sigma uncertainty represents the total propagated error associated with processing and counting operations and include; weighing errors, volume uncertainties, and counting error.

Attachment: Data Report -- Sample Results for ASR 0787.02.

Project / WP#:	73312/NC4189
ASR#:	0787.02
Client:	SK Fiskum
Total # of Samples:	12

Plutonium 238, 239+240 Analysis

RPL ID	Client Sample ID	RPL ID	Client Sample ID
19-1773	TI061-COMP-FEED	19-1779	TI061-L-F14-A
19-1774	TI061-COMP-EFF	19-1780	TI061-L-F18-A
19-1775	TI061-L-F1-A	19-1781	TI061-L-F22-A
19-1776	TI061-L-F4-A	19-1782	TI061-L-F25-A
19-1777	TI061-L-F8-A	19-1783	TI061-L-F27-A
19-1778	TI061-L-F11-A	19-1784	TI061-L-F29-A

Analysis Type:	AEA – Pu-238, Pu-239+240				
Sample Processing Prior to Radiochemical Processing/Analysis Pre-dilution Prior to Radiochemical Processing?	 None Digested as per PNL-ALO-106, Rev. 1, Acid Digestion of Waters, Soils, and Sludges for Subsequent Radiochemical Sample Analyses Fusion as per RPG-CMC-115, Rev0, Solubilization of Metals from Solids Using a KOH-KNO3 Fusion Digested as per RPG-CMC-128, Rev.1, HNO3-HCL Acid extraction of Liquids for Metals Analysis Using a Dry-Block Heater No 				
	\boxtimes Yes example 2 mL to 100 mL; 50x dilution				
Plutonium Separation Procedure:	RPG-CMC-4017, Rev. 0, Analysis of Environmental Water Samples for Actinides and Strontium-90				
Technician/Analyst:	LP Darnell, (07/25/2019)				
Co-Precipitation Procedure:	RPG-CMC-496, Rev. 1, Coprecipitation Mounting of Actinides for Alpha Spectroscopy				
Technician/Analyst:	LP Darnell, (07/25/2019)				
Spike and Tracer Standard ID's:	R-687-a-7 (Pu-239), R-719-1 (Pu-242 tracer)				
Analysis Procedure:	RPG-CMC-422, Rev. 2, Solutions Analysis: Alpha Spectrometry				
Reference Date:	Same as counting dates				
Analysis Date or Date Range:	July 29 - August 1, 2019				
Technician/Analyst:	T. Trang-Le				
Analysis Data (File):	RPG-RC\PNL\Projects\Backup files\Backup 19\19-1773 Fiskum.xls				
CMC Project 98620 File:	File Plan 5871: T 73312: Sample preparation and analysis records; T-4.4 Alpha Detector calibration, calibration verification checks, and maintenance records; and T3 Standard certificates and preparation. Also balance calibration and instrument performance checks.				
M&TE Number(s):	Ortec AEA counters – 32 counters – See attached M&TE list				
Hall. Park 1 8/16	119 _ T TRang-le 18/16/19				

Preparer Preparer

Date

-le _/_<u>&//6/19</u> Date IRang Reviewer U

Sample Results

See attached data report, Sample Results for ASR 0787.02. All data are reported in units of μ Ci per mL with a 1- σ uncertainty unless noted otherwise (see Comments).

Sample Preparation, Separation, Mounting and Counting Methods

All twelve samples submitted under Analytical Service Request (ASR) 0787.02 were analyzed for plutonium isotopes by Alpha Spectrometry. All the samples were prepared in RPL/420, only Pu-AEA is presented in this report.

Following the digestion process of the samples, the Pu was separated from the aqueous samples by anion exchange chromatography using procedure RPG-CMC-4017. The separated Pu fraction was mounted for alpha spectrometry by co-precipitation using procedure RPG-CMC-496, and then counted by alpha spectrometry using procedure RPC-CMC-422. The samples were counted on July 29-August 1, 2019; no decay corrections were made.

QUALITY CONTROL RESULTS

Quality control (QC) samples include a laboratory preparation blank and sample duplicate. Additional QC samples were prepared prior to separations; these included a reagent spike (RS) and a matrix spike (MS) made by adding Pu-239 standard to a diluted sample.

Tracer:

The Pu-242 tracer is added to every sample after appropriate dilution and prior to plutonium separations. The use of a Pu-242 tracer corrects for radiochemical yield and mathematically removes the detector counting efficiency from the results calculations. Tracer recovery is required to be high enough to provide acceptable counting statistics. The Pu-242 tracer counting statistics were acceptable for all samples. The tracer recoveries ranged from 37% to 106%.

Laboratory Preparation Blank (PB):

The activity levels of Pu-239/240 and Pu-238 present in the laboratory preparation blank (9.59E-6 μ Ci/mL, Pu-239/240 and 1.08E-6 μ Ci/mL, Pu-238) are both below the activities present in the samples (1.49E-4 μ Ci/mL Pu-239/240 and 2.09E-5 μ Ci/mL Pu-238 or higher). The blank result for the Pu-239/240 is 6.4% of the activity present in the lowest activity sample, thus not meeting the less than 5% of the activity present in the samples criteria. The blank result for the Pu-238 is 5.1% of the activity present in the lowest activity sample, thus not meeting the less than 5% of the activity present in the samples criteria. Sample TI061-L-F1-A is the only sample in the group of 12 where the activity present in the blank is greater than 5% or the activity present in the samples.

Blank Spike (BS) – Reagent Spike (RS):

The RS recovery of 98% meets the acceptance criteria of 80% to 120% recovery.

Matrix Spike (MS):

The MS recovery of 102% meets the acceptance criterion of 75% to 125% recovery. Note: the MS sample was prepared "after" digestion, by adding a known Pu-239 standard quantity to an aliquot of the leachate. Sample number 19-1774 (TI061-COMP-EFF) was selected as the matrix spike sample.

Duplicate -- Relative Percent Difference (RPD):

Duplicate results are required to agree within $\leq 20\%$ RPD. The ASO QAP further specifies that the two results need to be >5 times the MDA or have individual uncertainties $\leq 20\%$. The duplicate results were $\leq 2\%$ RPD for Pu-238 and Pu-239/240; thus meeting the $\leq 20\%$ requirement.

Instrument Quality Control

Alpha counters receive initial calibration with NIST traceable sources to determine the counter efficiency. When internal tracers are not used, the counter efficiency is used in calculation of final results.

Detector backgrounds are determined every 4 weeks or after the last analytical run, whichever is longer. Detector background counts are subtracted from all subsequent sample counts. A process blank is analyzed with each analytical batch to evaluate for contamination in the sample preparation process.

Assumption and Limitations of the Data

None

Comments

- 1. The results have been corrected for all dilution factors resulting from sample processing.
- 2. The 1-sigma uncertainty represents the total propagated error associated with processing and counting operations and include; weighing errors, volume uncertainties, and counting error.

Attachment: Data Report -- Sample Results for ASR 0787.02.

Pacific Northwest National Laboratory PO Box 999, Richland, WA				filename 19-1773 Fiskum 8/15/2019							
Radiochemical Scien	ces and Engine	ering Gro	սթ		c (u)	65.60	a lal a	11 10			
Client: Fiskum	Project: 73312	2		Prepared by:	Chuc	Foder	94194 3-	-16-19			
ASR 0787.02	NC4189		Technical Reviewer: 1 IRCORD - 10 8/16/19								
			rechnic	al Keviewer.		ang-	u spe	2/1			
Procedures:	RPG-CMC-474, Rev 1, Measurement of Alpha and Beta Activity by Liquid Scintillation										
	Spectrometry										
	RPG-CMC-432, Rev 0, Technetium-99 Analysis										
	RPG-CMC-4017, Rev 0, Analysis of Environmental Water Samples for Actinides and Strontium-90										
	RPG-CMC-422, Rev 2, Solutions Analysis: Alpha Spectrometry (Np, Pu, Am, Cm)										
	RPG-CMC-4	96, Rev 1, 0	Coprecipita	tion Mounting	g of Actinic	les for Alph	a Spectrome	try			
M&TE:	Perkin Elmer	TriCarb mo	odel 31001	R liquid scint	mation spe	ctrometer					
Count datas:	Alpha spectro	metry cour	iting system	1							
Count dates.	July 51-Augu	St 11, 2019									
	Lab			M	easured Co	oncentration	, μCi/mL ± 1	s uncertainty	/		
Sample	ID	Sr	-90	Tc-	99	Np	-237	Pu-2	238	Pu-239)+240
TH061-COMP-FEED	19-1773	4.52E-1	+ 2%	1.00E-1	± 5%	4.76E-5	± 2%	7.54E-5	± 4%	5.60E-4	± 2%
TI061-COMP-EFF	19-1774	<1.E-3	/ 0	9.84E-2	± 5%	1.77E-5	$\pm 3\%$	2.78E-5	$\pm 5\%$	2.01E-4	± 2%
	DUP-1774	<1.E-3		9.79E-2	± 5%	2.28E-5	± 3%	2.71E-5	± 5%	2.01E-4	± 2%
	RPD	-		1%		25%		2%		0.3%	
TI061-L-F1-A	19-1775	<1.E-3				7.43E-6	± 4%	2.09E-5	± 6%	1.49E-4	± 3%
TI061-L-F4-A	19-1776	1.24E-3	± 16%			2.52E-5	± 3%	4.05E-5	± 4%	2.72E-4	± 2%
TI061-L-F8-A	19-1777	<1.E-3				3.98E-5	± 2%	4.10E-5	± 4%	3.07E-4	± 2%
TI061-L-F11-A	19-1778	1.51E-2	± 2%			4.51E-5	± 2%	4.98E-5	± 4%	3.42E-4	± 2%
TI061-L-F14-A	19-1779	<1.E-3				4.61E-5	± 2%	4.56E-5	± 4%	3.42E-4	± 2%
TI061-L-F18-A	19-1780	1.19E-3	± 17%			4.58E-5	± 2%	5.00E-5	± 4%	3.77E-4	± 2%
TI061-L-F22-A	19-1781	<1.E-3				4.60E-5	± 2%	4.89E-5	± 4%	3.68E-4	± 2%
TI061-L-F25-A	19-1782	1.68E-3	± 12%			4.35E-5	± 2%	5.10E-5	± 4%	3.73E-4	± 2%
TI061-L-F27-A	19-1783	2.14E-3	± 10%			4.16E-5	± 2%	5.44E-5	± 4%	3.96E-4	± 2%
TI061-L-F29-A	19-1784	2.73E-3	± 8%			4.22E-5	± 2%	5.07E-5	± 6%	3.95E-4	± 3%
	Reagent Spike	97%		101%		93%				98%	
	Matrix Spike	102%				87%				102%	
	Lab Blank	1.55E-5	± 25%	1.3E-6	± 952%	-6.1E-9	± 173%	1.08E-6	± 31%	9.59E-6	± 8%

[Tc-99 spike was too small for the sample activity]

Pacific Northwest National Laboratory

902 Battelle Boulevard P.O. Box 999 Richland, WA 99354 1-888-375-PNNL (7665)

www.pnnl.gov